

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070824\
 Data File : BG061941.D
 Acq On : 8 Jul 2024 14:05
 Operator : MA/JU
 Sample : P3059-07
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 E1GM7

Quant Time: Jul 08 15:28:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070224.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 03 02:38:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.094	152	42068	20.000	ng/u1	0.00
20) Naphthalene-d8	10.908	136	205209	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.715	164	157501	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.459	188	376556	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.719	240	333236	20.000	ng/u1	0.00
88) Perylene-d12	24.956	264	403334	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.481	96	5062	4.522	ng/uL	0.00
4) Pyridine-d5	3.899	84	33331	9.683	ng/u1	0.00
7) Phenol-d5	7.242	99	47211	10.255	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.418	67	80134	27.486	ng/u1	0.00
11) 2-Chlorophenol-d4	7.624	132	88933	28.162	ng/u1	0.00
15) 4-Methylphenol-d8	8.793	113	80679	22.258	ng/u1	0.00
21) Nitrobenzene-d5	9.263	128	52182	33.435	ng/u1	0.00
24) 2-Nitrophenol-d4	9.986	143	61023	34.239	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.520	165	114624	33.343	ng/u1	-0.01
31) 4-Chloroaniline-d4	11.043	131	135712	27.192	ng/u1	0.00
46) Dimethylphthalate-d6	14.116	166	408690	31.562	ng/u1	0.00
49) Acenaphthylene-d8	14.416	160	445914	32.934	ng/u1	0.00
54) 4-Nitrophenol-d4	14.903	143	23336	11.285	ng/u1	-0.01
60) Fluorene-d10	15.708	176	359486	32.978	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.820	200	70153	35.474	ng/u1	0.00
73) Anthracene-d10	17.559	188	610193	34.708	ng/u1	0.00
81) Pyrene-d10	19.839	212	686128	35.005	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.733	264	716554	35.822	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070824\
Data File : BG061941.D
Acq On : 8 Jul 2024 14:05
Operator : MA/JU
Sample : P3059-07
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E1GM7

Quant Time: Jul 08 15:28:51 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070224.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jul 03 02:38:46 2024
Response via : Initial Calibration

