

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070824\
 Data File : BG061956.D
 Acq On : 9 Jul 2024 00:57
 Operator : MA/JU
 Sample : P3062-02MS 10X
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 E1GN8MS

Quant Time: Jul 09 01:45:01 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070224.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 03 02:38:46 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut
 Upadhyay

07/09/2024
 Supervised By :mohammad
 ahmed

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.094	152	41993	20.000	ng/u1	0.00
20) Naphthalene-d8	10.908	136	223610	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.715	164	159809	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.459	188	370111	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.719	240	330219	20.000	ng/u1	# 0.00
88) Perylene-d12	24.956	264	395949	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.481	96	3611	3.231	ng/uL	0.00
4) Pyridine-d5	3.904	84	19370	5.637	ng/u1	0.00
7) Phenol-d5	7.242	99	54698	11.902	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.418	67	113488	38.996	ng/u1	0.00
11) 2-Chlorophenol-d4	7.624	132	124650	39.542	ng/u1	0.00
15) 4-Methylphenol-d8	8.805	113	102113	28.221	ng/u1	0.00
21) Nitrobenzene-d5	9.263	128	79968	47.022	ng/u1	0.00
24) 2-Nitrophenol-d4	9.986	143	90650	46.677	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.526	165	178426	47.632	ng/u1	0.00
31) 4-Chloroaniline-d4	11.037	131	107463	19.760	ng/u1	0.00
46) Dimethylphthalate-d6	14.122	166	635360	48.359	ng/u1	0.00
49) Acenaphthylene-d8	14.416	160	654630	47.650	ng/u1	0.00
54) 4-Nitrophenol-d4	14.909	143	29497	14.059	ng/u1	0.00
60) Fluorene-d10	15.708	176	524241	47.398	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.826	200	114699	59.009	ng/u1	0.00
73) Anthracene-d10	17.559	188	811935	46.987	ng/u1	0.00
81) Pyrene-d10	19.839	212	927982	47.777	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.739	264	1026355	52.266	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.517	88	8440	6.843	ng/uL	94
5) Pyridine	3.928	79	20017	5.680	ng/u1	87
6) Benzaldehyde	7.230	77	65965	27.121	ng/u1	92
8) Phenol	7.271	94	59042	12.549	ng/u1#	80
10) Bis(2-Chloroethyl)ether	7.512	93	143006	39.565	ng/u1	99
12) 2-Chlorophenol	7.659	128	127477	39.693	ng/u1	96
13) 2-Methylphenol	8.534	108	109407	32.579	ng/u1	93
14) 2,2'-oxybis(1-Chloropr...	8.617	45	286623	36.297	ng/u1	100
16) Acetophenone	8.916	105	268476	44.919	ng/u1	94
17) N-Nitroso-di-n-propyla...	8.899	70	137565	41.112	ng/u1	95
18) 4-Methylphenol	8.863	108	109821	29.202	ng/u1	99
19) Hexachloroethane	9.181	117	57957	41.600	ng/u1#	77
22) Nitrobenzene	9.304	77	214801	44.496	ng/u1	98
23) Isophorone	9.827	82	424912	38.961	ng/u1#	96
25) 2-Nitrophenol	10.015	139	88644	44.757	ng/u1#	92
26) 2,4-Dimethylphenol	10.074	107	174220	36.755	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.309	93	220146	36.655	ng/u1	95
29) 2,4-Dichlorophenol	10.550	162	160934	45.736	ng/u1	88
30) Naphthalene	10.961	128	506478	41.234	ng/u1	96
32) 4-Chloroaniline	11.067	127	91702	17.184	ng/u1	100
33) Hexachlorobutadiene	11.231	225	112862	42.623	ng/u1#	86
34) Caprolactam	11.830	113	9771m	7.088	ng/u1	
35) 4-Chloro-3-methylphenol	12.177	107	208177	44.422	ng/u1	94
36) 2-Methylnaphthalene	12.553	142	388058	43.609	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070824\
 Data File : BG061956.D
 Acq On : 9 Jul 2024 00:57
 Operator : MA/JU
 Sample : P3062-02MS 10X
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :

BNA_G

ClientSampleId :

E1GN8MS

Manual Integrations

APPROVED

Reviewed By :Jagrut
Upadhyay


07/09/2024

Supervised By :mohammad
ahmed

Quant Time: Jul 09 01:45:01 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070224.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 03 02:38:46 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.771	142	410923	44.436	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.917	216	235703	49.114	ng/ul	98
40) Hexachlorocyclopentadiene	12.888	237	75854	23.744	ng/ul#	98
41) 2,4,6-Trichlorophenol	13.152	196	157482	49.405	ng/ul	99
42) 2,4,5-Trichlorophenol	13.229	196	167917	47.955	ng/ul	89
43) 1,1'-Biphenyl	13.552	154	561445	45.597	ng/ul	99
44) 2-Chloronaphthalene	13.599	162	425160	46.007	ng/ul	91
45) 2-Nitroaniline	13.805	65	192536	52.027	ng/ul	99
47) Dimethylphthalate	14.169	163	569186	45.040	ng/ul	99
48) 2,6-Dinitrotoluene	14.292	165	127992	53.389	ng/ul	100
50) Acenaphthylene	14.445	152	663874	44.029	ng/ul	99
51) 3-Nitroaniline	14.621	138	99688	42.718	ng/ul	99
52) Acenaphthene	14.780	153	468825	45.029	ng/ul	94
53) 2,4-Dinitrophenol	14.827	184	74545	59.517	ng/ul	95
55) 4-Nitrophenol	14.921	109	37011	15.455	ng/ul	91
56) Dibenzofuran	15.115	168	683983	46.482	ng/ul	93
57) 2,4-Dinitrotoluene	15.080	165	182830	52.002	ng/ul#	93
58) 2,3,4,6-Tetrachlorophenol	15.338	232	151276	47.844	ng/ul	96
59) Diethylphthalate	15.526	149	604302	44.375	ng/ul	96
61) Fluorene	15.761	166	565121	45.159	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.749	204	289799	46.103	ng/ul	91
63) 4-Nitroaniline	15.785	138	119242	50.358	ng/ul#	78
66) 4,6-Dinitro-2-methylph...	15.838	198	116881	55.873	ng/ul#	88
67) N-Nitrosodiphenylamine	15.967	169	470101	46.510	ng/ul	99
68) 4-Bromophenyl-phenylether	16.642	248	206540	49.190	ng/ul	95
69) Hexachlorobenzene	16.760	284	239539	49.613	ng/ul	98
70) Atrazine	16.907	200	85441	21.414	ng/ul	95
71) Pentachlorophenol	17.107	266	123521	46.544	ng/ul	96
72) Phenanthrene	17.500	178	896892	45.959	ng/ul	99
74) Anthracene	17.594	178	882301	43.748	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.523	216	254235	55.307	ng/ul	96
76) Pentachlorobenzene	15.033	250	247430	47.194	ng/ul	96
77) Carbazole	17.859	167	796547	46.931	ng/ul	99
78) Di-n-butylphthalate	18.405	149	971819	44.597	ng/ul	98
80) Fluoranthene	19.504	202	1046990	45.549	ng/ul#	93
82) Pyrene	19.868	202	1089944	45.904	ng/ul#	95
83) Butylbenzylphthalate	20.738	149	445560	44.713	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.613	252	273899	34.865	ng/ul	96
85) Benzo(a)anthracene	21.701	228	1141138	47.728	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.596	149	655560	45.846	ng/ul#	95
87) Chrysene	21.766	228	1074333	47.999	ng/ul	100
89) Di-n-octyl phthalate	22.818	149	1126681	46.577	ng/ul	100
90) Benzo(b)fluoranthene	23.928	252	1239716	49.696	ng/ul	97
91) Benzo(k)fluoranthene	23.999	252	1220959	49.081	ng/ul#	93
93) Benzo(a)pyrene	24.815	252	1034665	43.800	ng/ul#	96
94) Indeno(1,2,3-cd)pyrene	28.670	276	1542850m	51.109	ng/ul	
95) Dibenzo(a,h)anthracene	28.734	278	1276727m	51.209	ng/ul	
96) Benzo(g,h,i)perylene	29.827	276	1143004m	47.594	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 6.843 ng/uL
RT: 3.517 min Scan# 39
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS





Tgt Ion: 88 Resp: 8440
Ion Ratio Lower Upper
88 100
43 55.4 39.4 59.2
58 102.1 78.1 117.1

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#5
Pyridine
Concen: 5.680 ng/uL
RT: 3.928 min Scan# 109
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion: 79 Resp: 20017
Ion Ratio Lower Upper
79 100
52 127.1 86.0 129.0
51 57.2 46.4 69.6





#6
Benzaldehyde
Concen: 27.121 ng/ul
RT: 7.230 min Scan# 671
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion: 77 Resp: 6596
Ion Ratio Lower Upper
77 100
105 87.6 79.6 119.4
106 76.6 62.6 93.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#8
Phenol
Concen: 12.549 ng/ul
RT: 7.271 min Scan# 678
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion: 94 Resp: 59042
Ion Ratio Lower Upper
94 100
65 44.5 29.0 43.4#
66 62.9 37.4 56.2#





#10
Bis(2-Chloroethyl)ether
Concen: 39.565 ng/ul
RT: 7.512 min Scan# 719
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion: 93 Resp: 143000
Ion Ratio Lower Upper
93 100
63 89.2 72.2 108.2
95 34.2 26.7 40.1

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#12
2-Chlorophenol
Concen: 39.693 ng/ul
RT: 7.659 min Scan# 744
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:128 Resp: 127477
Ion Ratio Lower Upper
128 100
64 56.2 42.6 63.8
130 26.9 23.3 34.9





#13
2-Methylphenol
Concen: 32.579 ng/ul
RT: 8.534 min Scan# 892
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS





Tgt Ion:108 Resp: 10940
Ion Ratio Lower Upper
108 100
107 87.2 74.9 112.3

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#14
2,2'-oxybis(1-Chloropropane)
Concen: 36.297 ng/ul
RT: 8.617 min Scan# 907
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion: 45 Resp: 286623
Ion Ratio Lower Upper
45 100
77 9.1 7.4 11.0
79 6.2 4.8 7.2





#16
Acetophenone
Concen: 44.919 ng/ul
RT: 8.916 min Scan# 916
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:105 Resp: 268470
Ion Ratio Lower Upper
105 100
77 86.3 64.1 96.1
51 44.2 33.8 50.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#17
N-Nitroso-di-n-propylamine
Concen: 41.112 ng/ul
RT: 8.899 min Scan# 955
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion: 70 Resp: 137565
Ion Ratio Lower Upper
70 100
42 85.3 64.7 97.1
101 8.2 7.3 10.9
130 17.2 17.0 25.6





#18
4-Methylphenol
Concen: 29.202 ng/ul
RT: 8.863 min Scan# 948
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:108 Resp: 10982
Ion Ratio Lower Upper
108 100
107 109.0 88.4 132.6

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#19
Hexachloroethane
Concen: 41.600 ng/ul
RT: 9.181 min Scan# 1003
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:117 Resp: 57957
Ion Ratio Lower Upper
117 100
201 112.4 69.3 103.9#
199 72.6 48.9 73.3





#22
Nitrobenzene
Concen: 44.496 ng/ul
RT: 9.304 min Scan# 1024
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion: 77 Resp: 21480
Ion Ratio Lower Upper
77 100
123 39.7 31.0 46.6
65 16.6 12.7 19.1

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#23
Isophorone
Concen: 38.961 ng/ul
RT: 9.827 min Scan# 1113
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion: 82 Resp: 424912
Ion Ratio Lower Upper
82 100
95 9.3 6.0 9.0#
138 13.6 11.9 17.9





#25
2-Nitrophenol
Concen: 44.757 ng/ul
RT: 10.015 min Scan# 1145
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:139 Resp: 88644
Ion Ratio Lower Upper
139 100
65 78.5 61.6 92.4
109 56.7 35.4 53.2

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#26
2,4-Dimethylphenol
Concen: 36.755 ng/ul
RT: 10.074 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:107 Resp: 174220
Ion Ratio Lower Upper
107 100
121 46.5 35.6 53.4
122 75.8 62.2 93.2





#27
 Bis(2-Chloroethoxy)methane
 Concen: 36.655 ng/ul
 RT: 10.309 min Scan# 1195
 Delta R.T. -0.000 min
 Lab File: BG061956.D
 Acq: 9 Jul 2024 00:57

Instrument :
 BNA_G
 ClientSampleId :
 E1GN8MS






Tgt Ion: 93 Resp: 220140
 Ion Ratio Lower Upper
 93 100
 95 29.7 26.2 39.2
 123 8.1 7.5 11.3

Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay




07/09/2024
 Supervised By :mohammad
 ahmed


07/10/2024



#29
 2,4-Dichlorophenol
 Concen: 45.736 ng/ul
 RT: 10.550 min Scan# 1236
 Delta R.T. -0.006 min
 Lab File: BG061956.D
 Acq: 9 Jul 2024 00:57


Tgt Ion:162 Resp: 160934
 Ion Ratio Lower Upper
 162 100
 164 69.4 46.9 70.3
 98 46.8 33.4 50.2





#30
Naphthalene
Concen: 41.234 ng/ul
RT: 10.961 min Scan# 1306
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:128 Resp: 506478
Ion Ratio Lower Upper
128 100
129 10.3 7.7 11.5
127 13.9 9.6 14.4

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#32
4-Chloroaniline
Concen: 17.184 ng/ul
RT: 11.067 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:127 Resp: 91702
Ion Ratio Lower Upper
127 100
129 30.2 24.2 36.4





#33
Hexachlorobutadiene
Concen: 42.623 ng/ul
RT: 11.231 min Scan# 11286
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:225 Resp: 11286
Ion Ratio Lower Upper
225 100
223 63.4 45.4 68.2
227 68.5 44.0 66.0

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#34
Caprolactam
Concen: 7.088 ng/ul m
RT: 11.830 min Scan# 1454
Delta R.T. -0.012 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Tgt Ion:113 Resp: 9771
Ion Ratio Lower Upper
113 100
55 284.4 184.6 276.8#
56 178.4 135.0 202.6





#35
4-Chloro-3-methylphenol
Concen: 44.422 ng/ul
RT: 12.177 min Scan# 1512
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :

BNA_G

ClientSampleId :

E1GN8MS





Tgt Ion:107 Resp: 20817
Ion Ratio Lower Upper
107 100
144 21.7 19.6 29.4
142 70.6 60.3 90.5

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#36
2-Methylnaphthalene
Concen: 43.609 ng/ul
RT: 12.553 min Scan# 1577
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:142 Resp: 388058
Ion Ratio Lower Upper
142 100
141 87.8 69.4 104.2





#37
1-Methylnaphthalene
Concen: 44.436 ng/ul
RT: 12.771 min Scan# 1614
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:142 Resp: 41092
Ion Ratio Lower Upper
142 100
141 89.9 71.4 107.2
116 5.3 3.8 5.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#39
1,2,4,5-Tetrachlorobenzene
Concen: 49.114 ng/ul
RT: 12.917 min Scan# 1639
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:216 Resp: 235703
Ion Ratio Lower Upper
216 100
214 76.6 62.5 93.7
179 20.9 18.6 27.8
108 16.0 13.7 20.5





#40
Hexachlorocyclopentadiene
Concen: 23.744 ng/ul
RT: 12.888 min Scan# 1634
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:237 Resp: 75854
Ion Ratio Lower Upper
237 100
235 65.5 51.7 77.5
272 12.1 12.4 18.6

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#41
2,4,6-Trichlorophenol
Concen: 49.405 ng/ul
RT: 13.152 min Scan# 1679
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:196 Resp: 157482
Ion Ratio Lower Upper
196 100
198 101.2 79.9 119.9
200 32.5 26.7 40.1





#42
2,4,5-Trichlorophenol
Concen: 47.955 ng/ul
RT: 13.229 min Scan# 1692
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:196 Resp: 16791
Ion Ratio Lower Upper
196 100
198 102.8 72.5 108.7
200 31.4 23.7 35.5

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#43
1,1'-Biphenyl
Concen: 45.597 ng/ul
RT: 13.552 min Scan# 1747
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:154 Resp: 561445
Ion Ratio Lower Upper
154 100
153 39.2 31.7 47.5
76 11.0 10.0 15.0





#44
2-Chloronaphthalene
Concen: 46.007 ng/ul
RT: 13.599 min Scan# 1755
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:162 Resp: 425160
Ion Ratio Lower Upper
162 100
164 34.0 23.3 34.9
127 47.6 33.2 49.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#45
2-Nitroaniline
Concen: 52.027 ng/ul
RT: 13.805 min Scan# 1790
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion: 65 Resp: 192536
Ion Ratio Lower Upper
65 100
92 60.3 48.2 72.4
138 72.8 56.8 85.2





#47
Dimethylphthalate
Concen: 45.040 ng/ul
RT: 14.169 min Scan# 1852
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:163 Resp: 569180
Ion Ratio Lower Upper
163 100
194 3.2 2.6 4.0
164 10.6 8.2 12.4

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#48
2,6-Dinitrotoluene
Concen: 53.389 ng/ul
RT: 14.292 min Scan# 1873
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:165 Resp: 127992
Ion Ratio Lower Upper
165 100
89 78.0 62.6 93.8
121 23.6 18.4 27.6





#50
Acenaphthylene
Concen: 44.029 ng/ul
RT: 14.445 min Scan# 1898
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS





Tgt Ion:152 Resp: 663874
Ion Ratio Lower Upper
152 100
151 20.0 16.5 24.7
153 13.8 11.4 17.0

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#51
3-Nitroaniline
Concen: 42.718 ng/ul
RT: 14.621 min Scan# 1929
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:138 Resp: 99688
Ion Ratio Lower Upper
138 100
108 12.4 12.3 18.5
92 145.2 116.3 174.5





#52
Acenaphthene
Concen: 45.029 ng/ul
RT: 14.780 min Scan# 1956
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS





Tgt Ion:153 Resp: 468829
Ion Ratio Lower Upper
153 100
152 44.9 37.2 55.8
154 86.0 75.0 112.6

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#53
2,4-Dinitrophenol
Concen: 59.517 ng/ul
RT: 14.827 min Scan# 1964
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:184 Resp: 74545
Ion Ratio Lower Upper
184 100
63 104.6 83.6 125.4
154 93.4 66.8 100.2





#55
4-Nitrophenol
Concen: 15.455 ng/ul
RT: 14.921 min Scan# 1980
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:109 Resp: 3701
Ion Ratio Lower Upper
109 100
139 79.9 67.4 101.2
65 115.0 102.1 153.1

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#56
Dibenzofuran
Concen: 46.482 ng/ul
RT: 15.115 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:168 Resp: 683983
Ion Ratio Lower Upper
168 100
139 42.9 31.0 46.4





#57
2,4-Dinitrotoluene
Concen: 52.002 ng/ul
RT: 15.080 min Scan# 2007
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:165 Resp: 182830
Ion Ratio Lower Upper
165 100
63 61.0 44.9 67.3
182 2.1 2.6 3.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#58
2,3,4,6-Tetrachlorophenol
Concen: 47.844 ng/ul
RT: 15.338 min Scan# 2051
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Tgt Ion:232 Resp: 151276
Ion Ratio Lower Upper
232 100
131 44.0 35.8 53.8
130 2.9 2.0 3.0
166 30.2 28.0 42.0





#59
Diethylphthalate
Concen: 44.375 ng/ul
RT: 15.526 min Scan# 2082
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS





Tgt Ion:149 Resp: 604302
Ion Ratio Lower Upper
149 100
177 23.2 16.3 24.5
150 12.8 9.6 14.4

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#61
Fluorene
Concen: 45.159 ng/ul
RT: 15.761 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:166 Resp: 565121
Ion Ratio Lower Upper
166 100
165 102.6 78.4 117.6
167 13.6 10.4 15.6





#62
4-Chlorophenyl-phenylether
Concen: 46.103 ng/ul
RT: 15.749 min Scan# 2121
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:204 Resp: 289799
Ion Ratio Lower Upper
204 100
206 35.2 24.5 36.7
141 64.6 46.6 69.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#63
4-Nitroaniline
Concen: 50.358 ng/ul
RT: 15.785 min Scan# 2127
Delta R.T. 0.005 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:138 Resp: 119242
Ion Ratio Lower Upper
138 100
92 71.6 52.3 78.5
108 150.9 94.1 141.1#





#66
4,6-Dinitro-2-methylphenol
Concen: 55.873 ng/ul
RT: 15.838 min Scan# 2135
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:198 Resp: 11688
Ion Ratio Lower Upper
198 100
182 4.0 3.6 5.4
77 34.6 21.8 32.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#67
N-Nitrosodiphenylamine
Concen: 46.510 ng/ul
RT: 15.967 min Scan# 2158
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:169 Resp: 470101
Ion Ratio Lower Upper
169 100
168 69.2 55.6 83.4
167 36.1 29.8 44.8





#68
4-Bromophenyl-phenylether
Concen: 49.190 ng/ul
RT: 16.642 min Scan# 21
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:248 Resp: 206540
Ion Ratio Lower Upper
248 100
250 93.2 77.2 115.8
141 77.8 57.8 86.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#69
Hexachlorobenzene
Concen: 49.613 ng/ul
RT: 16.760 min Scan# 2293
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:284 Resp: 239539
Ion Ratio Lower Upper
284 100
142 33.5 26.4 39.6
249 31.7 27.2 40.8





#70
Atrazine
Concen: 21.414 ng/ul
RT: 16.907 min Scan# 21
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS





Tgt Ion:200 Resp: 8544
Ion Ratio Lower Upper
200 100
173 20.3 18.7 28.1
215 43.9 37.8 56.6

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#71
Pentachlorophenol
Concen: 46.544 ng/ul
RT: 17.107 min Scan# 2352
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Tgt Ion:266 Resp: 123521
Ion Ratio Lower Upper
266 100
264 60.9 45.0 67.6
268 62.2 50.5 75.7





#72

Phenanthrene

Concen: 45.959 ng/ul

RT: 17.500 min Scan# 2419

Delta R.T. -0.006 min

Lab File: BG061956.D

Acq: 9 Jul 2024 00:57

Instrument :

BNA_G

ClientSampleId :

E1GN8MS

Tgt Ion:178 Resp: 896892

Ion Ratio Lower Upper

178 100

179 15.6 11.9 17.9

176 19.9 15.5 23.3

Manual Integrations

APPROVED





Reviewed By :Jagrut

Upadhyay

07/09/2024

Supervised By :mohammad
ahmed

07/10/2024



#74

Anthracene

Concen: 43.748 ng/ul

RT: 17.594 min Scan# 2435

Delta R.T. -0.000 min

Lab File: BG061956.D

Acq: 9 Jul 2024 00:57




Tgt Ion:178 Resp: 882301


Ion Ratio Lower Upper

178 100

179 15.8 11.9 17.9


176 18.2 14.7 22.1





#75
1,2,3,4-Tetrachlorobenzene
Concen: 55.307 ng/ul
RT: 13.523 min Scan# 1742
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS





Tgt Ion:216 Resp: 254239
Ion Ratio Lower Upper
216 100
214 80.6 61.7 92.5
218 47.9 35.7 53.5

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#76
Pentachlorobenzene
Concen: 47.194 ng/ul
RT: 15.033 min Scan# 1999
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:250 Resp: 247430
Ion Ratio Lower Upper
250 100
248 61.8 47.5 71.3
252 65.4 50.0 75.0





#77
 Carbazole
 Concen: 46.931 ng/ul
 RT: 17.859 min Scan# 2480
 Delta R.T. -0.006 min
 Lab File: BG061956.D
 Acq: 9 Jul 2024 00:57

Instrument :
 BNA_G
 ClientSampleId :
 E1GN8MS






Tgt Ion:167 Resp: 79654
 Ion Ratio Lower Upper
 167 100
 166 22.3 18.2 27.2
 139 13.0 9.6 14.4

Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay




07/09/2024
 Supervised By :mohammad
 ahmed


07/10/2024



#78
 Di-n-butylphthalate
 Concen: 44.597 ng/ul
 RT: 18.405 min Scan# 2573
 Delta R.T. -0.006 min
 Lab File: BG061956.D
 Acq: 9 Jul 2024 00:57


Tgt Ion:149 Resp: 971819
 Ion Ratio Lower Upper
 149 100
 150 8.8 7.5 11.3
 104 6.8 5.0 7.4





#80
Fluoranthene
Concen: 45.549 ng/ul
RT: 19.504 min Scan# 21
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:202 Resp: 1046990
Ion Ratio Lower Upper
202 100
101 6.3 7.4 11.2
100 6.5 6.6 10.0

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#82
Pyrene
Concen: 45.904 ng/ul
RT: 19.868 min Scan# 2822
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Tgt Ion:202 Resp: 1089944
Ion Ratio Lower Upper
202 100
101 8.6 8.8 13.2
100 8.5 7.6 11.4





#83

Butylbenzylphthalate

Concen: 44.713 ng/ul

RT: 20.738 min Scan# 2970

Delta R.T. -0.006 min

Lab File: BG061956.D

Acq: 9 Jul 2024 00:57

Instrument :

BNA_G

ClientSampleId :

E1GN8MS

Tgt Ion:	149	Resp:	445560
Ion Ratio	Lower	Upper	
149	100		
91	83.7	60.6	91.0
206	20.1	14.6	21.8

Manual Integrations





APPROVED

Reviewed By :Jagrut
Upadhyay

07/09/2024

Supervised By :mohammad
ahmed

07/10/2024



#84

3,3'-Dichlorobenzidine

Concen: 34.865 ng/ul




RT: 21.613 min Scan# 3119


Delta R.T. -0.006 min

Lab File: BG061956.D

Acq: 9 Jul 2024 00:57


Tgt Ion:	252	Resp:	273899
Ion Ratio	Lower	Upper	
252	100		
254	62.5	52.8	79.2
126	9.1	6.6	10.0





#85
Benzo(a)anthracene
Concen: 47.728 ng/ul
RT: 21.701 min Scan# 3113
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS





Tgt Ion:228 Resp: 1141138
Ion Ratio Lower Upper
228 100
229 19.4 15.7 23.5
226 26.2 21.7 32.5

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#86
Bis(2-ethylhexyl)phthalate
Concen: 45.846 ng/ul
RT: 21.596 min Scan# 3116
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:149 Resp: 655560
Ion Ratio Lower Upper
149 100
167 24.7 21.7 32.5
279 4.6 2.8 4.2#





#87
Chrysene
Concen: 47.999 ng/ul
RT: 21.766 min Scan# 3145
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:228 Resp: 107433
Ion Ratio Lower Upper
228 100
226 29.2 23.3 34.9
229 19.6 15.5 23.3

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#89
Di-n-octyl phthalate
Concen: 46.577 ng/ul
RT: 22.818 min Scan# 3324
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57


Tgt Ion:149 Resp: 1126681





#90
Benzo(b)fluoranthene
Concen: 49.696 ng/ul
RT: 23.928 min Scan# 3512
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS





Tgt Ion:252 Resp: 1239710
Ion Ratio Lower Upper
252 100
253 22.8 17.2 25.8
125 6.1 6.1 9.1

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed


07/10/2024



#91
Benzo(k)fluoranthene
Concen: 49.081 ng/ul
RT: 23.999 min Scan# 3525
Delta R.T. -0.006 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Tgt Ion:252 Resp: 1220959
Ion Ratio Lower Upper
252 100
253 20.8 19.3 28.9
125 5.7 6.3 9.5#






#93
Benzo(a)pyrene
Concen: 43.800 ng/ul
RT: 24.815 min Scan# 30
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :

BNA_G

ClientSampleId :

E1GN8MS



Tgt Ion:252 Resp: 103466
Ion Ratio Lower Upper
252 100
253 21.3 16.5 24.7
125 5.8 7.4 11.2




Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay


07/09/2024

Supervised By :mohammad
ahmed



07/10/2024




#94
Indeno(1,2,3-cd)pyrene
Concen: 51.109 ng/ul m
RT: 28.670 min Scan# 4320
Delta R.T. 0.005 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57




Tgt Ion:276 Resp: 1542850
Ion Ratio Lower Upper
276 100
138 12.7 11.9 17.9
277 26.2 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 51.209 ng/ul m
RT: 28.734 min Scan# 41
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Instrument :
BNA_G
ClientSampleId :
E1GN8MS






Tgt Ion:278 Resp: 127672
Ion Ratio Lower Upper
278 100
139 7.4 7.6 11.4#
279 22.6 18.8 28.2

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

07/09/2024
Supervised By :mohammad
ahmed

07/10/2024



#96
Benzo(g,h,i)perylene
Concen: 47.594 ng/ul m
RT: 29.827 min Scan# 4517
Delta R.T. -0.000 min
Lab File: BG061956.D
Acq: 9 Jul 2024 00:57

Tgt Ion:276 Resp: 1143004
Ion Ratio Lower Upper
276 100
138 10.3 11.1 16.7#
277 24.5 18.8 28.2

