

Data Path : V:\HPCHEM1\BNA G\DATA\BG070916\
 Data File : BG022951.D
 Acq On : 9 Jul 2016 4:46
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 11 04:06:46 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG070816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 07 16:51:07 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2	1,4-Dioxane	0.475	0.480	-1.1	126	0.00
3	Pyridine	1.627	1.547	4.9	115	0.00
4	n-Nitrosodimethylamine	0.698	0.723	-3.6	123	0.00
5 S	2-Fluorophenol	1.223	1.207	1.3	121	0.00
6	Aniline	2.654	2.652	0.1	119	0.00
7 S	Phenol-d6	1.915	1.923	-0.4	118	0.00
8	2-Chlorophenol	1.301	1.324	-1.8	122	0.00
9	Benzaldehyde	1.280	1.275	0.4	122	0.00
10 C	Phenol	1.971	1.989	-0.9	120	0.00
11	bis(2-Chloroethyl)ether	1.562	1.556	0.4	120	0.00
12	1,3-Dichlorobenzene	1.593	1.551	2.6	121	0.00
13 C	1,4-Dichlorobenzene	1.595	1.630	-2.2	124	0.00
14	1,2-Dichlorobenzene	1.585	1.578	0.4	125	0.00
15	Benzyl Alcohol	1.754	1.759	-0.3	117	0.00
16	2,2'-oxybis(1-Chloropropane	1.908	1.940	-1.7	123	0.00
17	2-Methylphenol	1.283	1.277	0.5	119	0.00
18	Hexachloroethane	0.673	0.655	2.7	126	0.00
19 P	n-Nitroso-di-n-propylamine	1.727	1.689	2.2	114	0.00
20	3+4-Methylphenols	1.789	1.820	-1.7	115	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
22	Acetophenone	0.600	0.575	4.2	113	0.00
23 S	Nitrobenzene-d5	0.467	0.460	1.5	116	0.00
24	Nitrobenzene	0.496	0.485	2.2	117	0.00
25	Isophorone	0.939	0.853	9.2	104	0.00
26 C	2-Nitrophenol	0.195	0.194	0.5	110	0.00
27	2,4-Dimethylphenol	0.337	0.320	5.0	112	0.00
28	bis(2-Chloroethoxy)methane	0.524	0.495	5.5	109	0.00
29 C	2,4-Dichlorophenol	0.359	0.357	0.6	114	0.00
30	1,2,4-Trichlorobenzene	0.411	0.399	2.9	115	0.00
31	Naphthalene	1.018	0.992	2.6	115	0.00
32	Benzoic acid	0.306	0.286	6.5	103	0.02
33	4-Chloroaniline	0.498	0.482	3.2	113	0.00
34 C	Hexachlorobutadiene	0.334	0.329	1.5	119	0.00
35	Caprolactam	0.148	0.121	18.2	96	0.02
36 C	4-Chloro-3-methylphenol	0.491	0.453	7.7	106	0.00
37	2-Methylnaphthalene	0.805	0.767	4.7	109	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
39	1,2,4,5-Tetrachlorobenzene	0.862	0.918	-6.5	109	0.00
40 P	Hexachlorocyclopentadiene	0.496	0.482	2.8	99	0.00
41 S	2,4,6-Tribromophenol	0.359	0.329	8.4	95	0.00
42 C	2,4,6-Trichlorophenol	0.495	0.508	-2.6	103	0.00
43	2,4,5-Trichlorophenol	0.509	0.519	-2.0	105	0.00
44 S	2-Fluorobiphenyl	1.270	1.295	-2.0	106	0.00

Data Path : V:\HPCHEM1\BNA G\DATA\BG070916\
 Data File : BG022951.D
 Acq On : 9 Jul 2016 4:46
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 11 04:06:46 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG070816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 07 16:51:07 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.501	1.569	-4.5	106	0.00
46	2-Chloronaphthalene	1.217	1.267	-4.1	106	0.00
47	2-Nitroaniline	0.519	0.525	-1.2	103	0.00
48	Acenaphthylene	1.826	1.840	-0.8	103	0.00
49	Dimethylphthalate	1.634	1.545	5.4	98	0.00
50	2,6-Dinitrotoluene	0.367	0.358	2.5	100	0.00
51 C	Acenaphthene	1.193	1.200	-0.6	103	0.00
52	3-Nitroaniline	0.366	0.370	-1.1	102	0.00
53 P	2,4-Dinitrophenol	0.252	0.189	25.0	73	0.00
54	Dibenzofuran	1.786	1.741	2.5	101	0.00
55 P	4-Nitrophenol	0.307	0.272	11.4	90	0.00
56	2,4-Dinitrotoluene	0.535	0.518	3.2	100	0.00
57	Fluorene	1.537	1.478	3.8	100	0.00
58	2,3,4,6-Tetrachlorophenol	0.509	0.478	6.1	96	0.00
59	Diethylphthalate	1.662	1.539	7.4	96	0.00
60	4-Chlorophenyl-phenylether	0.936	0.901	3.7	98	0.00
61	4-Nitroaniline	0.396	0.372	6.1	97	0.00
62	Azobenzene	1.590	1.571	1.2	102	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
64	4,6-Dinitro-2-methylphenol	0.149	0.140	6.0	88	0.00
65 c	n-Nitrosodiphenylamine	0.576	0.583	-1.2	97	0.00
66	4-Bromophenyl-phenylether	0.260	0.255	1.9	94	0.00
67	Hexachlorobenzene	0.283	0.278	1.8	95	0.00
68	Atrazine	0.256	0.249	2.7	96	0.00
69 C	Pentachlorophenol	0.183	0.173	5.5	87	0.00
70	Phenanthrene	0.980	0.960	2.0	97	0.00
71	Anthracene	0.992	0.984	0.8	98	0.00
72	Carbazole	0.858	0.845	1.5	97	0.00
73	Di-n-butylphthalate	0.954	0.976	-2.3	97	0.00
74 C	Fluoranthene	1.203	1.141	5.2	97	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
76	Benzidine	0.573	0.542	5.4	91	0.00
77	Pyrene	1.124	1.049	6.7	98	0.00
78 S	Terphenyl-d14	0.646	0.640	0.9	98	0.00
79	Butylbenzylphthalate	0.445	0.449	-0.9	101	0.00
80	Benzo(a)anthracene	1.119	1.098	1.9	99	0.00
81	3,3'-Dichlorobenzidine	0.491	0.481	2.0	96	0.00
82	Chrysene	1.050	1.047	0.3	101	0.00
83	Bis(2-ethylhexyl)phthalate	0.618	0.604	2.3	99	0.00
84 c	Di-n-octyl phthalate	1.031	1.027	0.4	99	0.02
85	Indeno(1,2,3-cd)pyrene	1.338	1.298	3.0	99	0.04
86 I	Perylene-d12	1.000	1.000	0.0	96	0.02
87	Benzo(b)fluoranthene	1.154	1.154	0.0	99	0.02

Data Path : V:\HPCHEM1\BNA G\DATA\BG070916\
Data File : BG022951.D
Acq On : 9 Jul 2016 4:46
Operator : UM/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jul 11 04:06:46 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG070816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 07 16:51:07 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.095	1.129	-3.1	98	0.03
89 C	Benzo(a)pyrene	1.106	1.115	-0.8	98	0.02
90	Dibenzo(a,h)anthracene	1.098	1.108	-0.9	99	0.05
91	Benzo(a,h,i)perylene	1.099	1.112	-1.2	98	0.06

(#) = Out of Range

SPCC's out = 0 CCC's out = 0