

Data Path : Z:\HPCHEM1\BNA_G\Data\BG071015\
 Data File : BG017813.D
 Acq On : 10 Jul 2015 16:43
 Operator : TP/UM
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02045

Quant Time: Jul 10 23:34:00 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG070915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 10 23:32:40 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.524	0.550	-5.0	106	0.00
3 S	1,4-Dioxane-d8	0.496	0.559	-12.7	115	0.00
4	Benzaldehyde	1.158	1.401	-21.0#	105	0.00
5 S	Phenol-d5	1.913	1.937	-1.3	100	0.00
6	Phenol	2.055	2.109	-2.6	100	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.126	1.211	-7.5	104	0.00
8	Bis(2-Chloroethyl)ether	1.506	1.538	-2.1	100	0.00
9 S	2-Chlorophenol-d4	1.544	1.496	3.1	94	0.00
10	2-Chlorophenol	1.531	1.522	0.6	98	0.00
11	2-Methylphenol	1.562	1.564	-0.1	99	0.00
12	2,2'-oxybis(1-Chloropropane	2.255	2.503	-11.0	106	0.00
13 S	4-Methylphenol-d8	1.599	1.539	3.8	95	0.00
14	Acetophenone	2.348	2.475	-5.4	101	0.00
15 P	N-Nitroso-di-n-propylamine	1.456	1.532	-5.2	101	0.00
16	4-Methylphenol	1.763	1.742	1.2	95	0.00
17	Hexachloroethane	0.633	0.642	-1.4	101	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
19 S	Nitrobenzene-d5	0.168	0.157	6.5	90	0.00
20	Nitrobenzene	0.401	0.415	-3.5	97	0.00
21	Isophorone	0.794	0.847	-6.7	99	0.00
22 S	2-Nitrophenol-d4	0.169	0.153	9.5	85	0.00
23 C	2-Nitrophenol	0.184	0.169	8.2	84	0.00
24	2,4-Dimethylphenol	0.388	0.408	-5.2	98	0.00
25	Bis(2-Chloroethoxy)methane	0.424	0.447	-5.4	98	0.00
26 S	2,4-Dichlorophenol-d3	0.296	0.291	1.7	94	0.00
27 C	2,4-Dichlorophenol	0.292	0.292	0.0	93	0.00
28	Naphthalene	1.049	1.085	-3.4	98	0.00
29 S	4-Chloroaniline-d4	0.411	0.478	-16.3	98	0.00
30	4-Chloroaniline	0.406	0.474	-16.7	96	0.00
31 C	Hexachlorobutadiene	0.158	0.151	4.4	91	0.00
32	Caprolactam	0.233	0.244	-4.7	96	0.00
33 C	4-Chloro-3-methylphenol	0.374	0.385	-2.9	94	0.00
34	2-Methylnaphthalene	0.757	0.772	-2.0	96	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
36	1,2,4,5-Tetrachlorobenzene	0.517	0.522	-1.0	97	0.00
37	Hexachlorocyclopentadiene	0.294	0.270	8.2	91	0.00
38 C	2,4,6-Trichlorophenol	0.358	0.331	7.5	85	0.00
39	2,4,5-Trichlorophenol	0.391	0.379	3.1	91	0.00
40	1,1'-Biphenyl	1.538	1.539	-0.1	94	0.00
41	2-Chloronaphthalene	1.166	1.174	-0.7	96	0.00
42	2-Nitroaniline	0.429	0.429	0.0	93	0.00
43 S	Dimethylphthalate-d6	1.521	1.524	-0.2	93	0.00
44	Dimethylphthalate	1.562	1.591	-1.9	96	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG071015\
 Data File : BG017813.D
 Acq On : 10 Jul 2015 16:43
 Operator : TP/UM
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02045

Quant Time: Jul 10 23:34:00 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG070915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 10 23:32:40 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.325	0.293	9.8	83	0.00
46 S	Acenaphthylene-d8	1.837	1.866	-1.6	94	0.00
47	Acenaphthylene	1.999	2.061	-3.1	96	0.00
48	3-Nitroaniline	0.356	0.379	-6.5	89	0.00
49 C	Acenaphthene	1.334	1.374	-3.0	97	0.00
50	2,4-Dinitrophenol	0.170	0.144	15.3	92	0.00
51 S	4-Nitrophenol-d4	0.324	0.320	1.2	93	0.00
52	4-Nitrophenol	0.262	0.275	-5.0	96	0.00
53	Dibenzofuran	1.816	1.831	-0.8	95	0.00
54	2,4-Dinitrotoluene	0.458	0.446	2.6	90	0.00
55	2,3,4,6-Tetrachlorophenol	0.329	0.303	7.9	87	0.00
56	Diethylphthalate	1.676	1.749	-4.4	97	0.00
57 S	Fluorene-d10	1.355	1.364	-0.7	95	0.00
58	Fluorene	1.613	1.635	-1.4	94	0.00
59	4-Chlorophenyl-phenylether	0.742	0.733	1.2	93	0.00
60	4-Nitroaniline	0.418	0.441	-5.5	93	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.121	0.102	15.7	87	0.00
63	4,6-Dinitro-2-methylphenol	0.128	0.114	10.9	91	0.00
64	N-Nitrosodiphenylamine	0.625	0.630	-0.8	93	0.00
65	4-Bromophenyl-phenylether	0.200	0.198	1.0	95	0.00
66	Hexachlorobenzene	0.215	0.209	2.8	91	0.00
67	Atrazine	0.231	0.245	-6.1	95	0.00
68 C	Pentachlorophenol	0.133	0.114	14.3	89	0.00
69	Phenanthrene	1.119	1.157	-3.4	96	0.00
70 S	Anthracene-d10	0.935	0.951	-1.7	95	0.00
71	Anthracene	1.149	1.199	-4.4	96	0.00
72	1,2,3,4-Tetrachlorobenzene	0.241	0.235	2.5	93	0.00
73	Pentachlorobenzene	0.248	0.243	2.0	95	0.00
74	Carbazole	1.019	1.156	-13.4	97	0.00
75	Di-n-butylphthalate	1.397	1.515	-8.4	100	0.00
76 C	Fluoranthene	1.146	1.328	-15.9	97	0.00
77 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
78 S	Pyrene-d10	0.964	0.992	-2.9	96	0.00
79	Pyrene	1.245	1.289	-3.5	97	0.00
80	Butylbenzylphthalate	0.585	0.591	-1.0	95	0.00
81	3,3'-Dichlorobenzidine	0.356	0.391	-9.8	94	0.00
82	Benzo(a)anthracene	1.130	1.160	-2.7	97	0.00
83	Bis(2-ethylhexyl)phthalate	0.802	0.816	-1.7	97	0.00
84	Chrysene	1.060	1.083	-2.2	97	0.00
85 I	Perylene-d12	1.000	1.000	0.0	96	0.00
86	Di-n-octyl phthalate	1.280	1.476	-15.3	96	0.00
87	Benzo(b)fluoranthene	1.157	1.162	-0.4	95	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG071015\
Data File : BG017813.D
Acq On : 10 Jul 2015 16:43
Operator : TP/UM
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02045

Quant Time: Jul 10 23:34:00 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG070915.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jul 10 23:32:40 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.101	1.180	-7.2	99	0.00
89 S	Benzo(a)pyrene-d12	0.998	1.028	-3.0	97	0.00
90 C	Benzo(a)pyrene	1.112	1.162	-4.5	98	0.00
91	Indeno(1,2,3-cd)pyrene	1.251	1.267	-1.3	97	0.00
92	Dibenzo(a,h)anthracene	1.073	1.086	-1.2	96	0.00
93	Benzo(g,h,i)perylene	1.036	1.056	-1.9	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0