


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG071015\
Data File : BG017809.D
Acq On : 10 Jul 2015 10:40
Operator : TP/UM
Sample : DFTPP11
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP11

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG070915.M
Title : SVOA CALIBRATION
Last Update : Fri Jul 10 23:32:40 2015



AutoFind: Scans 2514, 2515, 2516; Background Corrected with Scan 2509

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	59.6	27333	PASS
68	69	0.00	2	0.8	201	PASS
69	198	0.00	100	58.5	26800	PASS
70	69	0.00	2	0.7	189	PASS
127	198	10	80	60.9	27917	PASS
197	198	0.00	2	0.5	218	PASS
198	198	100	100	100.0	45829	PASS
199	198	5	9	7.2	3299	PASS
275	198	10	60	25.8	11814	PASS
365	198	1	100	4.4	2019	PASS
441	442	0.01	24	15.2	4581	PASS
442	198	50	100	65.8	30144	PASS
443	442	15	24	19.1	5763	PASS

Scan 2509 (17.429 min): BG017809.D
DFTPP11

m/z	Abundance
40.00	1214.0
41.00	196.0
44.00	905.0
49.00	326.0
55.00	165.0
70.90	194.0
83.80	159.0
202.10	155.0
206.90	385.0
207.90	158.0
426.00	152.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

m/z	Abundance
37.00	215.0
38.00	788.0
39.00	2750.0
39.90	1080.0
44.00	489.0
49.00	311.0
50.00	9699.0
51.00	32224.0
52.00	1794.0
55.00	286.0
56.00	769.0
57.00	2507.0
58.00	195.0
60.90	339.0
62.00	723.0
63.00	1182.0
64.00	173.0
65.10	434.0
69.00	31912.0
69.90	320.0
72.90	606.0
74.00	3587.0
75.00	5928.0
76.10	2061.0
77.00	35520.0
78.00	2526.0
79.00	2591.0
80.00	2271.0
81.00	2451.0
81.90	525.0
82.90	441.0
85.00	439.0
85.90	755.0
86.90	371.0
91.00	671.0
91.90	918.0
93.00	4322.0
93.90	466.0
95.90	153.0
98.00	3305.0
99.00	2455.0
100.00	218.0
101.00	1294.0
102.10	207.0
102.90	591.0
103.90	1092.0
105.00	826.0
106.00	385.0
106.90	10689.0
108.00	1782.0
108.90	566.0
110.00	19128.0
111.00	3857.0
111.80	546.0
113.00	151.0
116.00	831.0
117.00	10222.0
118.00	811.0
118.90	222.0
121.90	547.0
123.00	699.0
123.90	672.0
125.00	549.0
127.00	33384.0
128.00	2353.0
128.90	13357.0
130.00	1219.0
131.00	221.0
132.90	154.0
133.90	556.0
134.90	1232.0
136.10	405.0
137.10	582.0
139.00	180.0
139.80	156.0
141.00	1800.0
141.90	520.0
142.90	654.0
145.80	276.0
146.90	1113.0
147.90	2456.0
148.90	601.0
151.00	294.0
151.80	249.0
153.00	498.0
153.90	284.0
155.00	1080.0
156.00	1411.0
156.90	310.0
158.10	397.0
158.90	361.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

159.90	439.0
161.00	1292.0
162.80	287.0
164.90	642.0
166.00	835.0
167.00	3628.0
168.00	2408.0
169.00	323.0
171.00	179.0
171.90	240.0
172.90	432.0
174.00	671.0
174.90	1261.0
176.00	382.0
176.90	649.0
178.00	269.0
179.00	2322.0
179.90	1419.0
181.00	864.0
182.20	153.0
185.00	1196.0
186.00	9713.0
187.00	2586.0
188.00	273.0
188.90	645.0
191.00	283.0
191.90	945.0
192.90	941.0
194.90	198.0
196.00	2238.0
196.70	654.0
197.90	52464.0
198.90	4251.0
200.00	336.0
201.50	492.0
202.30	175.0
202.90	705.0
203.90	2454.0
205.00	3515.0
206.00	16496.0
207.00	2311.0
208.00	898.0
208.90	620.0
209.80	207.0
210.90	900.0
214.90	378.0
216.00	336.0
216.90	4406.0
217.90	581.0
221.00	2298.0
221.70	1030.0
223.00	924.0
224.00	9131.0
225.00	2536.0
226.10	260.0
226.90	4483.0
227.90	319.0
228.90	1312.0
233.80	177.0
234.90	409.0
235.80	211.0
237.00	258.0
238.00	162.0
240.00	199.0
240.80	214.0
242.00	616.0
243.20	474.0
244.00	7767.0
245.00	827.0
245.90	1526.0
247.00	536.0
248.20	173.0
248.90	234.0
253.00	180.0
253.20	178.0
254.90	34504.0
256.00	4998.0
256.90	746.0
257.90	2240.0
258.90	407.0
265.00	1136.0
271.40	181.0
272.00	351.0
273.00	1080.0
274.00	3683.0
275.00	13978.0
275.90	1473.0
276.90	1144.0
278.10	180.0
283.00	431.0
284.90	210.0
295.10	219.0
295.90	5549.0
296.90	733.0
300.90	341.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

303.10	552.0
314.00	179.0
315.00	492.0
315.60	163.0
316.10	204.0
321.10	259.0
323.00	1459.0
323.90	164.0
326.90	519.0
327.70	239.0
334.00	1016.0
334.90	324.0
336.10	163.0
341.00	185.0
346.00	551.0
352.00	667.0
352.80	163.0
353.80	509.0
365.00	2188.0
365.70	176.0
371.00	249.0
371.90	571.0
372.80	185.0
402.00	183.0
403.00	235.0
404.00	155.0
420.90	308.0
421.80	362.0
423.00	2051.0
424.00	402.0
439.40	177.0
441.00	4899.0
442.00	34576.0
443.00	7164.0
443.90	447.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

m/z	Abundance
37.90	176.0
39.00	2356.0
39.90	1046.0
43.00	262.0
44.00	455.0
49.10	556.0
50.00	7056.0
51.00	26728.0
52.00	1427.0
56.00	785.0
57.00	1618.0
59.80	199.0
60.80	386.0
62.00	354.0
62.90	1115.0
63.90	165.0
65.00	637.0
68.90	25176.0
69.90	249.0
73.10	385.0
74.00	2750.0
75.00	4247.0
76.00	1014.0
77.00	27968.0
78.00	1998.0
78.90	2242.0
80.00	1586.0
81.00	2620.0
81.90	578.0
83.10	332.0
83.90	369.0
84.90	410.0
86.00	1016.0
87.20	203.0
88.10	199.0
91.00	479.0
92.00	815.0
93.00	3347.0
96.00	407.0
98.00	2977.0
99.00	2170.0
100.90	1407.0
102.80	409.0
104.00	512.0
104.90	728.0
106.00	413.0
107.00	9496.0
107.90	1304.0
110.00	16141.0
110.90	2857.0
111.90	290.0
117.00	8569.0
118.00	934.0
122.00	658.0
123.00	1409.0
124.00	330.0
125.00	353.0
125.70	152.0
127.00	27280.0
128.00	2460.0
128.90	11791.0
129.90	887.0
130.90	167.0
134.00	467.0
135.00	901.0
135.90	426.0
136.90	371.0
138.00	263.0
140.00	308.0
140.90	1212.0
141.90	371.0
142.90	234.0
144.00	191.0
146.10	450.0
147.00	1049.0
147.90	2285.0
148.80	533.0
150.00	157.0
151.60	230.0
152.90	288.0
153.90	407.0
155.00	915.0
156.00	1124.0
156.90	375.0
158.00	298.0
158.90	404.0
160.00	495.0
161.00	774.0
161.90	477.0
162.80	223.0
163.60	154.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

165.00	426.0
166.00	531.0
167.00	3540.0
168.00	1843.0
168.90	315.0
169.70	301.0
170.10	178.0
170.90	193.0
172.10	285.0
172.90	420.0
173.90	621.0
174.90	1143.0
176.00	473.0
176.90	634.0
177.90	174.0
179.00	2122.0
179.90	1238.0
181.00	718.0
181.80	329.0
185.00	1372.0
186.00	8464.0
187.00	2094.0
188.00	334.0
188.80	359.0
191.20	211.0
191.90	905.0
193.00	813.0
195.90	1384.0
197.90	51048.0
198.90	2986.0
200.00	401.0
201.30	333.0
203.00	420.0
204.00	2463.0
205.00	3882.0
206.00	16392.0
206.90	1746.0
208.00	1029.0
208.90	438.0
209.90	307.0
210.50	573.0
211.00	671.0
212.00	234.0
215.00	282.0
215.90	576.0
216.90	4159.0
218.00	528.0
220.90	2236.0
221.80	517.0
223.00	928.0
224.00	8814.0
224.90	2261.0
225.90	413.0
227.00	3465.0
227.90	571.0
228.90	946.0
230.90	597.0
232.90	164.0
233.80	281.0
234.90	285.0
235.90	259.0
236.70	301.0
240.00	182.0
241.00	539.0
241.90	506.0
242.90	398.0
244.00	7583.0
245.00	716.0
245.90	1739.0
247.00	302.0
249.00	241.0
250.80	266.0
252.10	203.0
252.80	396.0
255.00	31648.0
255.90	4427.0
256.80	304.0
257.90	1972.0
258.90	632.0
264.90	1586.0
265.90	281.0
273.00	1161.0
274.00	2895.0
274.90	13513.0
275.90	1989.0
276.90	1755.0
278.00	159.0
280.80	184.0
282.90	216.0
285.00	255.0
292.00	161.0
293.00	398.0
296.00	4649.0
296.90	640.0
303.20	371.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

313.90	189.0
314.90	527.0
316.00	432.0
323.00	1378.0
324.10	330.0
326.80	274.0
332.90	200.0
334.00	1162.0
334.80	242.0
340.90	339.0
346.00	364.0
351.90	437.0
353.00	343.0
353.90	406.0
365.00	2459.0
366.00	389.0
366.90	154.0
370.00	173.0
372.00	734.0
373.00	196.0
382.90	205.0
383.90	240.0
402.10	208.0
403.10	503.0
421.10	349.0
421.80	291.0
422.90	1732.0
424.10	373.0
436.50	183.0
438.00	168.0
439.30	259.0
440.00	194.0
441.00	6423.0
442.00	39064.0
443.00	7427.0
444.00	750.0

Instrument :

BNA_G

ClientSampleId :

DFTPP11

m/z	Abundance
38.00	506.0
39.00	2118.0
40.00	1202.0
41.10	248.0
44.00	567.0
44.80	178.0
47.90	178.0
49.00	458.0
50.00	6933.0
51.00	23048.0
52.00	1411.0
54.90	204.0
55.90	608.0
57.00	1608.0
61.10	235.0
62.10	425.0
63.00	863.0
65.00	456.0
67.10	227.0
68.00	603.0
69.00	23312.0
70.20	209.0
72.90	225.0
74.00	3116.0
75.00	4049.0
76.00	1502.0
77.00	24304.0
78.00	1941.0
78.90	1891.0
80.00	1058.0
81.00	2126.0
82.10	708.0
82.90	532.0
83.90	152.0
85.00	338.0
85.90	576.0
87.00	306.0
90.90	440.0
92.00	467.0
93.00	2865.0
93.90	222.0
95.90	158.0
97.90	2348.0
99.00	1757.0
99.90	213.0
101.00	1037.0
102.00	212.0
102.90	215.0
104.00	696.0
105.00	502.0
107.00	6804.0
108.00	1264.0
108.90	266.0
110.00	11468.0
111.00	2517.0
112.00	230.0
113.00	206.0
115.00	178.0
116.10	605.0
117.00	5961.0
117.90	635.0
120.70	229.0
122.00	572.0
123.00	954.0
124.00	462.0
125.00	503.0
127.00	23088.0
128.00	1971.0
128.90	8638.0
130.00	899.0
131.00	395.0
133.00	203.0
134.00	326.0
134.90	859.0
135.90	205.0
136.90	468.0
140.90	1261.0
142.00	521.0
142.90	406.0
144.90	151.0
145.90	284.0
147.00	864.0
148.00	1430.0
149.00	286.0
151.00	163.0
152.90	429.0
153.80	234.0
154.90	521.0
156.00	662.0
157.00	315.0
157.90	446.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

159.00	210.0
159.90	364.0
161.00	445.0
162.10	168.0
164.90	559.0
165.90	439.0
167.00	2297.0
168.00	1424.0
169.10	231.0
171.00	163.0
172.00	179.0
173.00	267.0
174.00	525.0
175.00	836.0
176.90	353.0
178.00	175.0
179.00	1850.0
179.90	1052.0
181.00	618.0
183.10	179.0
184.00	318.0
185.10	861.0
186.00	5661.0
187.00	1762.0
188.00	290.0
188.90	437.0
190.90	293.0
191.90	422.0
193.00	587.0
195.90	1179.0
197.90	33976.0
198.90	2660.0
199.90	155.0
201.30	178.0
202.90	422.0
203.90	1971.0
204.90	2113.0
206.00	9573.0
207.00	1821.0
208.00	402.0
208.90	183.0
210.10	317.0
211.00	182.0
214.80	151.0
215.80	379.0
216.90	2540.0
218.00	239.0
220.90	1665.0
221.60	367.0
222.00	375.0
222.90	734.0
224.00	5111.0
225.00	1217.0
226.00	174.0
227.00	3310.0
228.10	210.0
228.90	547.0
231.00	257.0
234.00	209.0
234.90	161.0
236.80	168.0
238.90	211.0
240.90	313.0
242.00	510.0
242.90	354.0
244.00	3908.0
245.00	385.0
245.90	641.0
250.10	154.0
253.10	247.0
254.90	19728.0
256.00	3192.0
256.80	192.0
257.00	182.0
257.90	1359.0
258.80	203.0
264.90	756.0
272.90	1146.0
273.90	1459.0
275.00	7952.0
276.00	870.0
277.00	816.0
284.70	170.0
285.10	258.0
295.90	2727.0
297.00	485.0
301.10	196.0
303.00	409.0
314.80	317.0
323.10	623.0
326.90	257.0
334.00	601.0
334.90	189.0
352.00	198.0
354.10	289.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

364.90	1410.0
365.80	197.0
372.00	546.0
402.00	167.0
402.80	230.0
420.90	170.0
422.00	270.0
423.00	1047.0
441.00	2422.0
442.00	16792.0
443.00	2699.0
444.00	179.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11