



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071024\
Data File : BG061990.D
Acq On : 10 Jul 2024 8:59
Operator : MA/JU
Sample : DFTPP594
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP594

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070224.MA.M
Title : SVOA CALIBRATION
Last Update : Wed Jul 03 02:38:46 2024



AutoFind: Scans 2481, 2482, 2483; Background Corrected with Scan 2476

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	50.7	70515	PASS
68	69	0.00	2	1.4	835	PASS
69	198	0.00	100	41.5	57757	PASS
70	69	0.00	2	0.4	251	PASS
127	198	10	80	44.9	62408	PASS
197	198	0.00	2	0.5	688	PASS
198	198	100	100	100.0	139061	PASS
199	198	5	9	6.5	9013	PASS
275	198	10	60	30.4	42208	PASS
365	198	1	100	5.6	7828	PASS
441	198	0.01	100	15.9	22166	PASS
442	198	50	100	91.3	126899	PASS
443	442	15	24	19.8	25120	PASS

m/z	Abundance
40.10	197.0
41.10	248.0
43.10	197.0
44.00	2835.0
49.00	211.0
61.20	185.0
84.90	177.0
109.80	264.0
165.70	193.0
166.20	335.0
168.20	285.0
184.80	456.0
220.10	251.0

Instrument :
BNA_G
ClientSampleId :
DFTPP594

m/z	Abundance
36.20	204.0
37.00	375.0
37.70	544.0
38.10	614.0
39.00	5233.0
40.00	571.0
41.10	246.0
44.00	1746.0
44.90	501.0
49.20	893.0
50.00	18312.0
51.00	68168.0
52.00	3154.0
53.00	217.0
55.20	388.0
56.00	2292.0
57.00	4716.0
58.00	354.0
60.00	180.0
61.00	956.0
62.00	919.0
63.00	2759.0
64.00	290.0
64.90	952.0
69.00	51888.0
69.80	565.0
71.20	152.0
72.80	340.0
74.00	5870.0
75.00	7791.0
76.00	2604.0
77.00	62792.0
78.10	5064.0
79.00	4614.0
80.00	2995.0
81.00	3512.0
81.90	1419.0
83.00	624.0
84.90	696.0
85.90	962.0
87.00	728.0
88.00	284.0
89.10	341.0
91.10	828.0
93.00	7910.0
94.00	455.0
95.80	308.0
96.40	202.0
98.00	5837.0
99.00	4641.0
99.90	461.0
101.00	2914.0
101.80	402.0
103.10	560.0
104.00	1533.0
104.80	1711.0
105.80	486.0
107.00	19488.0
108.00	2859.0
110.00	28552.0
111.00	5323.0
112.00	834.0
114.80	207.0
115.80	1067.0
117.00	15357.0
118.00	1376.0
120.00	590.0
121.90	1022.0
122.90	1963.0
123.90	1006.0
124.90	912.0
127.00	53944.0
128.00	3885.0
129.00	24560.0
130.00	1606.0
131.00	381.0
131.80	203.0
132.60	325.0
133.90	950.0
134.80	1245.0
135.90	1028.0
137.00	704.0
138.00	188.0
138.80	271.0
140.00	642.0
141.00	2420.0
142.10	693.0
143.00	825.0
146.00	689.0
147.00	1856.0
148.10	3654.0
149.10	1581.0

Instrument :
BNA_G
ClientSampleId :
DFTPP594

150.10	286.0
151.10	446.0
153.10	1739.0
154.00	478.0
155.00	2076.0
156.00	2457.0
156.80	474.0
158.00	637.0
158.90	492.0
160.00	418.0
161.00	1384.0
162.20	160.0
162.70	194.0
165.00	1419.0
165.90	1040.0
167.00	6211.0
168.00	3097.0
169.00	308.0
170.50	340.0
171.10	617.0
171.90	765.0
173.00	763.0
174.10	1167.0
175.00	2131.0
175.90	775.0
176.90	1317.0
178.00	332.0
178.90	3479.0
180.00	3263.0
181.00	1411.0
182.90	397.0
183.10	422.0
184.00	491.0
185.00	2493.0
186.00	17688.0
186.90	5369.0
188.10	825.0
189.00	1737.0
189.90	337.0
191.10	496.0
191.80	1268.0
193.00	1597.0
193.80	437.0
194.90	415.0
195.90	3919.0
196.60	451.0
198.00	117408.0
199.00	7557.0
200.00	938.0
200.90	210.0
201.50	799.0
203.00	1487.0
204.10	3844.0
205.10	7268.0
206.00	33232.0
207.00	4150.0
207.90	1472.0
209.00	356.0
209.90	386.0
210.90	1289.0
211.60	379.0
214.70	711.0
216.00	967.0
217.00	9570.0
218.00	1511.0
219.20	253.0
219.70	231.0
221.00	3509.0
221.80	1419.0
223.00	2593.0
224.00	20432.0
225.10	5675.0
226.10	705.0
227.00	8776.0
228.10	1945.0
229.00	1953.0
229.90	256.0
230.80	827.0
231.90	262.0
233.10	318.0
233.90	571.0
235.00	600.0
236.00	529.0
237.00	1049.0
238.80	278.0
239.10	373.0
239.90	373.0
241.10	258.0
242.00	1211.0
242.90	752.0
244.00	13926.0
245.10	2332.0
246.00	3822.0
247.00	1087.0
247.70	262.0
249.10	350.0

Instrument :
BNA_G
ClientSampleId :
DFTPP594

250.90	381.0
252.00	239.0
253.20	622.0
253.80	595.0
255.00	78680.0
256.00	13015.0
256.90	1077.0
258.00	5414.0
259.00	1334.0
261.10	175.0
262.90	173.0
263.80	241.0
264.90	2082.0
271.00	401.0
272.10	649.0
272.90	2031.0
274.00	6107.0
275.00	31072.0
276.00	4860.0
277.00	2633.0
278.10	1044.0
282.00	286.0
282.90	634.0
283.90	506.0
285.00	544.0
292.10	513.0
293.00	841.0
296.00	12911.0
296.90	2339.0
298.00	175.0
303.10	1220.0
304.10	462.0
305.20	186.0
310.00	565.0
312.70	218.0
314.00	732.0
314.90	1411.0
316.00	902.0
317.10	468.0
319.80	191.0
321.00	372.0
321.90	440.0
323.00	3796.0
324.10	959.0
326.30	262.0
326.90	525.0
327.90	384.0
331.90	317.0
332.90	385.0
334.00	3360.0
334.90	1022.0
340.10	245.0
341.10	1332.0
341.90	350.0
346.00	716.0
351.90	1264.0
353.00	769.0
354.20	1518.0
354.90	472.0
365.00	6817.0
366.00	1429.0
369.90	381.0
371.10	335.0
371.90	1132.0
372.90	468.0
382.80	536.0
390.10	394.0
390.80	176.0
402.00	876.0
403.00	1366.0
403.70	276.0
404.20	343.0
405.00	172.0
414.70	189.0
420.90	1350.0
421.90	1080.0
422.90	5541.0
423.90	1394.0
431.20	191.0
432.90	172.0
433.50	318.0
441.00	16185.0
442.00	81368.0
443.00	16568.0
443.80	1867.0

Instrument :
BNA_G
ClientSampleId :
DFTPP594

m/z	Abundance
36.90	353.0
38.00	943.0
39.00	5939.0
40.00	709.0
41.00	279.0
44.00	1737.0
45.00	305.0
50.00	21816.0
51.00	81880.0
52.10	4243.0
53.10	300.0
54.90	269.0
55.90	2532.0
57.00	5171.0
58.00	171.0
61.00	1320.0
62.00	853.0
62.90	2728.0
64.00	882.0
65.00	1145.0
66.20	187.0
67.90	1651.0
69.00	68280.0
70.00	474.0
71.00	202.0
72.90	670.0
74.00	7422.0
75.00	9688.0
76.00	4671.0
77.00	71112.0
78.00	4946.0
79.00	6055.0
80.00	4384.0
81.00	5978.0
82.00	1118.0
83.00	1003.0
84.10	186.0
85.00	1435.0
86.10	1938.0
87.10	685.0
87.90	316.0
89.00	242.0
91.00	2038.0
91.90	1634.0
93.00	9900.0
94.00	755.0
95.00	164.0
95.90	276.0
96.90	261.0
98.00	8515.0
99.00	6727.0
99.90	927.0
101.00	3244.0
102.10	216.0
103.00	772.0
104.00	1949.0
105.00	1453.0
106.00	948.0
107.00	26232.0
108.00	4171.0
110.00	36584.0
111.00	7159.0
112.00	1185.0
113.00	158.0
114.10	163.0
116.00	1457.0
117.00	23936.0
118.00	1256.0
120.00	530.0
120.90	344.0
122.00	1628.0
123.00	2794.0
124.00	1773.0
124.80	843.0
127.00	75224.0
128.00	6656.0
129.00	32464.0
130.00	2160.0
131.00	434.0
132.10	239.0
132.50	326.0
134.00	1230.0
135.00	1672.0
136.10	1272.0
137.00	1231.0
137.70	178.0
138.10	328.0
140.10	343.0
140.90	3785.0
142.00	1693.0
143.00	850.0
144.20	370.0

Instrument :
BNA_G
ClientSampleId :
DFTPP594

144.80	261.0
145.90	699.0
147.00	1802.0
148.00	4535.0
149.00	1567.0
151.00	434.0
152.00	395.0
153.00	1693.0
154.10	1111.0
155.00	2686.0
156.00	3041.0
157.00	765.0
157.70	600.0
159.10	936.0
159.90	1439.0
161.00	2074.0
162.00	728.0
163.00	153.0
163.90	201.0
165.00	1373.0
165.90	1113.0
167.00	7529.0
168.10	3606.0
168.90	830.0
169.50	197.0
170.90	523.0
171.80	706.0
173.00	1231.0
174.00	1680.0
175.10	3758.0
176.00	894.0
177.00	1498.0
177.90	753.0
179.00	6590.0
180.00	4505.0
180.90	2445.0
181.90	568.0
183.00	155.0
184.00	839.0
185.10	3830.0
186.00	27024.0
187.00	8272.0
188.10	676.0
189.00	1666.0
191.00	1466.0
191.90	2062.0
193.10	2830.0
193.80	652.0
194.80	605.0
196.00	5174.0
196.80	2065.0
198.00	165760.0
199.00	10477.0
199.90	869.0
201.80	929.0
202.90	2482.0
204.10	6775.0
205.00	12773.0
206.00	47352.0
207.10	7724.0
208.00	1744.0
208.90	732.0
210.10	1500.0
211.00	2293.0
212.10	414.0
213.20	180.0
213.90	223.0
215.00	827.0
216.00	1091.0
217.00	14558.0
218.00	2415.0
221.00	4588.0
221.90	2471.0
223.00	3743.0
224.00	28872.0
225.00	8946.0
226.00	632.0
227.00	15617.0
227.90	1442.0
229.00	2735.0
230.10	359.0
231.00	698.0
232.00	332.0
233.10	415.0
234.00	600.0
235.00	858.0
235.80	606.0
237.00	875.0
238.20	293.0
238.90	390.0
239.10	396.0
239.90	531.0
241.00	944.0
241.90	1414.0
243.00	1302.0
244.00	18680.0

Instrument :
BNA_G
ClientSampleId :
DFTPP594

244.90	1975.0
246.00	5338.0
247.10	1204.0
248.10	229.0
248.30	216.0
249.10	856.0
251.30	473.0
251.90	822.0
253.10	696.0
255.00	120240.0
256.00	17680.0
257.00	1502.0
258.00	8310.0
258.90	1352.0
259.90	220.0
261.10	222.0
263.10	187.0
264.10	512.0
265.00	3125.0
266.00	422.0
266.80	166.0
269.00	162.0
270.00	333.0
270.50	227.0
271.00	277.0
271.70	444.0
272.10	507.0
273.00	3110.0
274.00	8625.0
275.00	49392.0
276.00	7997.0
277.00	5301.0
278.10	932.0
282.10	204.0
283.10	884.0
284.00	450.0
284.90	530.0
286.20	192.0
287.10	161.0
288.80	482.0
290.00	340.0
291.10	336.0
292.00	324.0
293.20	1039.0
294.90	633.0
296.00	18520.0
296.90	3065.0
301.00	428.0
301.90	298.0
303.10	2498.0
304.10	540.0
305.60	190.0
308.10	498.0
309.30	210.0
310.10	236.0
310.80	265.0
313.10	594.0
314.10	875.0
315.00	1869.0
315.90	1049.0
319.00	174.0
320.70	490.0
321.80	375.0
323.00	7199.0
324.00	1498.0
324.70	172.0
326.20	410.0
326.90	1269.0
328.00	676.0
329.10	194.0
332.20	452.0
332.90	792.0
334.10	3781.0
335.10	985.0
340.40	191.0
340.90	527.0
341.80	240.0
342.10	203.0
346.00	745.0
346.90	345.0
352.10	2134.0
353.00	1209.0
354.00	1351.0
355.10	565.0
359.00	231.0
365.00	8683.0
365.90	1201.0
366.90	201.0
369.80	170.0
371.20	646.0
371.90	3963.0
373.00	576.0
377.60	207.0
382.90	837.0
383.90	280.0
386.10	168.0

Instrument :
BNA_G
ClientSampleId :
DFTPP594

389.80	312.0
390.90	303.0
400.80	172.0
401.90	1169.0
403.00	2448.0
404.00	947.0
408.20	198.0
420.30	178.0
421.00	1824.0
423.00	10484.0
423.90	1746.0
425.00	536.0
426.80	200.0
427.70	202.0
428.50	222.0
428.90	215.0
430.10	202.0
431.20	298.0
435.50	246.0
441.00	24632.0
442.00	144256.0
443.00	29800.0
443.90	3792.0
450.10	157.0
483.10	215.0
525.50	157.0

Instrument :
BNA_G
ClientSampleId :
DFTPP594

m/z	Abundance
36.90	219.0
38.00	635.0
39.00	5921.0
40.00	305.0
44.00	2634.0
44.90	208.0
48.50	166.0
50.00	15818.0
51.00	61496.0
52.00	2526.0
55.10	602.0
56.00	2081.0
57.00	4026.0
57.90	190.0
61.00	477.0
62.00	974.0
63.00	2288.0
64.00	485.0
65.00	1272.0
67.90	853.0
69.00	53104.0
70.00	280.0
71.00	166.0
73.00	866.0
74.00	6068.0
75.00	8849.0
76.20	2109.0
77.00	52312.0
78.00	3436.0
79.00	4432.0
80.00	3569.0
80.90	5595.0
81.90	1219.0
83.00	357.0
85.00	992.0
86.00	924.0
86.90	420.0
87.90	277.0
90.90	1594.0
92.00	1484.0
93.00	7944.0
94.10	523.0
96.20	306.0
98.00	6711.0
99.00	5191.0
99.80	848.0
101.00	2110.0
102.00	301.0
103.00	996.0
103.90	1820.0
105.00	1651.0
105.90	811.0
107.00	19816.0
108.00	3806.0
108.90	453.0
110.00	33880.0
111.00	4375.0
112.00	581.0
113.10	187.0
115.90	981.0
117.00	18272.0
117.90	1150.0
120.10	303.0
122.00	1039.0
123.10	1537.0
124.00	1421.0
124.90	799.0
127.00	58056.0
128.00	4136.0
129.00	27408.0
130.00	2536.0
130.90	295.0
131.70	159.0
132.60	198.0
133.00	212.0
133.90	677.0
135.00	2166.0
136.10	598.0
137.00	1020.0
137.40	638.0
139.80	483.0
141.00	4135.0
141.90	984.0
143.00	903.0
144.00	369.0
146.10	606.0
147.10	1980.0
148.00	3330.0
149.10	559.0
150.20	271.0
151.00	639.0
151.60	241.0

Instrument :
BNA_G
ClientSampleId :
DFTPP594

151.90	322.0
152.90	1143.0
154.00	870.0
155.10	1965.0
156.00	2954.0
156.80	659.0
157.40	544.0
157.70	535.0
158.10	891.0
159.00	576.0
160.10	1287.0
161.10	1517.0
162.00	483.0
163.00	470.0
165.00	1442.0
166.10	1313.0
167.00	7825.0
168.10	3086.0
169.10	608.0
171.00	413.0
171.90	987.0
173.00	529.0
173.90	1740.0
175.00	2530.0
176.00	880.0
176.90	1954.0
178.00	685.0
179.00	4983.0
180.00	4953.0
181.00	2717.0
182.10	390.0
182.70	165.0
183.50	213.0
184.00	707.0
185.00	2531.0
186.00	20792.0
187.00	8441.0
188.10	781.0
189.00	1716.0
190.00	861.0
191.00	1007.0
192.00	1699.0
193.00	2024.0
194.10	344.0
194.70	377.0
196.00	3894.0
198.00	134016.0
198.90	9006.0
199.90	535.0
201.40	985.0
202.00	597.0
203.00	1462.0
204.10	7068.0
205.00	9537.0
206.00	38568.0
207.00	5532.0
207.90	1671.0
209.00	475.0
210.10	505.0
211.00	1747.0
212.00	539.0
213.60	181.0
215.00	774.0
215.50	469.0
216.10	878.0
217.00	12571.0
217.90	1356.0
218.70	229.0
220.10	301.0
220.90	3126.0
221.60	1661.0
221.90	1671.0
223.00	3010.0
224.00	25000.0
225.00	7342.0
226.00	415.0
227.00	12872.0
228.00	2209.0
229.00	2538.0
230.00	474.0
231.00	1160.0
233.10	353.0
234.00	737.0
235.00	1001.0
236.00	601.0
236.90	773.0
238.10	196.0
239.00	524.0
239.80	462.0
241.20	603.0
241.90	1015.0
244.00	18168.0
245.10	2624.0
246.00	3918.0
247.10	1135.0
247.80	374.0

Instrument :
BNA_G
ClientSampleId :
DFTPP594

248.90	734.0
249.80	181.0
250.80	262.0
251.90	614.0
253.40	676.0
255.00	106744.0
256.00	14871.0
256.80	1677.0
258.00	7046.0
258.90	1362.0
259.80	188.0
260.90	367.0
262.00	469.0
265.00	2254.0
266.20	268.0
267.60	180.0
270.00	228.0
270.80	220.0
271.40	196.0
272.10	394.0
272.90	3136.0
273.90	7161.0
275.00	46160.0
276.00	5706.0
277.10	4454.0
278.00	672.0
279.00	277.0
281.00	262.0
282.00	453.0
282.80	353.0
283.00	372.0
283.40	403.0
284.20	569.0
285.10	707.0
289.10	496.0
290.00	158.0
291.20	363.0
292.20	348.0
293.00	605.0
293.90	283.0
296.00	15999.0
296.90	1833.0
298.00	524.0
299.00	368.0
299.80	242.0
301.00	509.0
302.00	267.0
303.00	1334.0
303.90	469.0
305.00	223.0
307.90	437.0
308.80	167.0
310.00	204.0
313.20	395.0
314.00	619.0
315.00	1523.0
315.90	1082.0
317.10	272.0
320.90	664.0
321.80	184.0
323.10	5277.0
324.00	843.0
327.10	947.0
328.00	872.0
332.00	359.0
332.90	700.0
334.10	3695.0
335.00	918.0
336.00	202.0
338.80	319.0
341.10	761.0
341.90	301.0
345.90	1453.0
346.90	344.0
351.00	332.0
352.00	2099.0
352.80	894.0
354.10	1964.0
355.00	584.0
359.00	290.0
365.00	7984.0
365.90	928.0
366.90	291.0
368.20	186.0
370.10	296.0
370.90	390.0
372.00	3163.0
373.10	728.0
383.00	1209.0
383.90	194.0
389.70	699.0
390.80	520.0
401.90	1439.0
403.10	1903.0
403.90	1346.0
411.30	172.0

Instrument :
BNA_G
ClientSampleId :
DFTPP594

419.00	181.0
420.90	1124.0
422.00	1363.0
423.00	9749.0
424.10	2142.0
424.80	250.0
428.40	231.0
431.30	191.0
434.30	168.0
435.70	400.0
441.00	25680.0
442.00	155072.0
443.00	28992.0
443.90	2341.0

Instrument :
BNA_G
ClientSampleId :
DFTPP594