

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG071315\
 Data File : BG017846.D
 Acq On : 14 Jul 2015 6:16
 Operator : TP/UM
 Sample : G2938-04
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0203

Quant Time: Jul 14 07:30:21 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG070915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 14 07:26:36 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	72037	20.00	ng/ul	0.00
18) Naphthalene-d8	10.37	136	326697	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.24	164	183289	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.00	188	385781	20.00	ng/ul	0.00
77) Chrysene-d12	21.20	240	405860	20.00	ng/ul	0.00
85) Perylene-d12	23.42	264	382595	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.13	96	2924	1.64	ng/uL	0.00
5) Phenol-d5	6.77	99	43282	6.28	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.93	67	131336	32.39	ng/ul	0.00
9) 2-Chlorophenol-d4	7.12	132	125432	22.55	ng/ul	0.00
13) 4-Methylphenol-d8	8.30	113	76649	13.31	ng/ul	0.00
19) Nitrobenzene-d5	8.74	128	82772	30.16	ng/ul	0.00
22) 2-Nitrophenol-d4	9.46	143	71019	25.73	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.99	165	115964	23.98	ng/ul	0.00
29) 4-Chloroaniline-d4	10.51	131	15197	2.26	ng/ul	0.00
43) Dimethylphthalate-d6	13.66	166	419145	30.06	ng/ul	0.00
46) Acenaphthylene-d8	13.93	160	544774	32.35	ng/ul	0.00
51) 4-Nitrophenol-d4	14.45	143	11973	4.03	ng/ul	0.00
57) Fluorene-d10	15.24	176	384547	30.98	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.36	200	48304	20.73	ng/ul	0.00
70) Anthracene-d10	17.10	188	574032	31.82	ng/ul	0.00
78) Pyrene-d10	19.40	212	645730	33.00	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.28	264	543784	28.49	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.16	88	16560	8.77	ng/uL	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG071315\
Data File : BG017846.D
Acq On : 14 Jul 2015 6:16
Operator : TP/UM
Sample : G2938-04
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0203

Quant Time: Jul 14 07:30:21 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG070915.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 14 07:26:36 2015
Response via : Initial Calibration

