


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG071315\
Data File : BG017823.D
Acq On : 13 Jul 2015 10:19
Operator : TP/UM
Sample : DFTPP12
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP12

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG070915.M
Title : SVOA CALIBRATION
Last Update : Tue Jul 14 01:25:57 2015



AutoFind: Scans 2514, 2515, 2516; Background Corrected with Scan 2509

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	52.1	24966	PASS
68	69	0.00	2	1.2	306	PASS
69	198	0.00	100	53.9	25794	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	61.2	29322	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	47893	PASS
199	198	5	9	7.2	3467	PASS
275	198	10	60	27.9	13345	PASS
365	198	1	100	4.6	2201	PASS
441	442	0.01	24	14.6	4957	PASS
442	198	50	100	70.8	33906	PASS
443	442	15	24	19.3	6530	PASS

m/z	Abundance
36.80	331.0
38.00	556.0
39.10	2615.0
40.00	1245.0
42.90	195.0
44.00	1370.0
48.90	592.0
50.00	9104.0
51.00	29784.0
52.00	1714.0
53.10	172.0
55.00	305.0
55.90	1059.0
57.00	1938.0
57.90	181.0
61.00	420.0
62.00	547.0
63.00	1459.0
64.00	219.0
65.00	498.0
69.00	29616.0
73.00	469.0
74.00	3316.0
75.00	5267.0
76.00	1674.0
77.00	34112.0
78.10	2383.0
79.00	1981.0
79.90	1853.0
81.00	2440.0
82.00	663.0
83.00	612.0
83.90	231.0
85.10	333.0
85.90	641.0
87.00	691.0
87.90	170.0
90.90	718.0
91.90	799.0
93.00	4021.0
94.00	197.0
95.90	363.0
98.00	3599.0
99.00	2317.0
99.90	375.0
101.00	1633.0
102.70	333.0
103.90	773.0
105.00	843.0
106.00	295.0
107.00	10833.0
108.00	1389.0
109.00	453.0
110.00	17472.0
110.90	2533.0
111.90	286.0
112.90	250.0
117.00	11119.0
117.90	847.0
119.90	171.0
122.00	980.0
123.00	1167.0
123.90	343.0
124.90	537.0
127.00	33144.0
127.90	2846.0
129.00	14290.0
130.00	1130.0
131.00	205.0
131.60	272.0
133.90	196.0
135.00	1076.0
135.90	483.0
136.90	795.0
138.00	241.0
140.90	1593.0
141.90	701.0
142.90	239.0
145.90	359.0
147.00	954.0
147.90	2048.0
149.00	429.0
151.00	168.0
151.90	166.0
152.90	554.0
153.90	281.0
154.90	1192.0
156.00	1292.0
156.90	315.0
157.90	268.0
159.20	297.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

159.90	415.0
161.00	756.0
162.00	171.0
165.00	717.0
166.10	507.0
167.00	4318.0
167.90	2187.0
168.80	385.0
171.10	305.0
172.00	394.0
173.10	304.0
173.90	630.0
175.00	1529.0
176.00	328.0
176.90	983.0
177.80	230.0
178.90	2522.0
179.90	1511.0
180.90	493.0
184.90	1347.0
186.00	8512.0
187.00	2521.0
188.00	220.0
189.00	746.0
189.80	258.0
190.60	243.0
191.90	447.0
192.90	690.0
193.90	280.0
194.90	253.0
196.00	2019.0
197.90	48720.0
198.90	3216.0
200.00	453.0
201.50	244.0
202.80	450.0
203.10	424.0
203.90	2564.0
205.00	3969.0
206.00	15468.0
207.00	2101.0
208.00	479.0
208.90	362.0
209.80	315.0
210.90	713.0
211.50	355.0
214.80	294.0
216.00	637.0
216.90	4908.0
218.00	710.0
220.90	2525.0
221.90	643.0
223.00	768.0
224.00	9665.0
225.00	2218.0
226.90	4622.0
227.90	466.0
228.90	709.0
229.90	252.0
231.00	181.0
234.10	246.0
234.80	348.0
236.00	183.0
236.90	340.0
238.90	259.0
242.00	780.0
243.00	497.0
244.00	6751.0
244.90	555.0
246.00	1401.0
248.10	158.0
248.90	244.0
252.00	263.0
252.90	373.0
253.40	162.0
255.00	31496.0
255.90	4290.0
256.90	234.0
258.00	1723.0
258.90	728.0
264.80	1361.0
272.90	725.0
273.90	2113.0
275.00	13581.0
276.00	1373.0
277.00	1744.0
278.00	156.0
282.90	184.0
283.90	230.0
284.90	448.0
291.90	160.0
292.80	325.0
294.60	191.0
296.00	4565.0
296.90	647.0

Instrument :

BNA_G

ClientSampleId :

DFTPP12

302.00	316.0
303.00	550.0
304.10	193.0
313.90	166.0
315.20	335.0
316.00	363.0
322.00	193.0
323.00	860.0
323.90	308.0
327.00	358.0
334.00	536.0
335.10	329.0
340.90	281.0
345.80	167.0
351.70	244.0
352.00	265.0
352.80	256.0
353.30	213.0
354.00	472.0
365.00	1788.0
366.00	251.0
372.00	538.0
373.00	299.0
382.70	161.0
401.90	214.0
403.00	417.0
421.00	310.0
421.90	361.0
422.90	1183.0
424.00	256.0
439.30	218.0
439.90	276.0
441.00	4183.0
442.00	27384.0
443.00	5913.0
443.90	514.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

m/z	Abundance
38.00	533.0
39.10	2558.0
40.00	1002.0
43.00	224.0
44.00	1241.0
44.80	153.0
49.10	531.0
50.00	7384.0
51.00	29792.0
52.00	1197.0
54.90	184.0
55.80	667.0
57.00	1897.0
58.00	187.0
61.00	385.0
62.00	546.0
63.00	1007.0
65.00	595.0
67.90	918.0
69.00	31136.0
73.10	439.0
74.00	3299.0
75.00	5011.0
76.10	2140.0
77.00	33720.0
78.00	2441.0
79.00	3107.0
80.00	1867.0
81.00	2728.0
81.90	522.0
83.00	584.0
84.00	201.0
85.00	354.0
86.00	961.0
87.00	197.0
88.90	192.0
91.00	994.0
92.00	630.0
93.00	4362.0
94.00	336.0
96.00	243.0
97.90	3072.0
99.00	2803.0
100.00	244.0
101.00	1139.0
102.90	503.0
104.00	963.0
104.90	817.0
107.00	10308.0
107.90	1776.0
110.00	16464.0
110.90	3120.0
111.80	312.0
113.00	329.0
116.00	794.0
117.00	11219.0
118.00	721.0
119.00	269.0
122.00	861.0
122.90	1140.0
123.90	730.0
124.90	658.0
127.00	35520.0
128.00	2977.0
129.00	14255.0
130.00	1180.0
131.00	244.0
132.20	177.0
133.90	445.0
135.00	988.0
136.00	384.0
136.90	708.0
139.90	294.0
140.90	2064.0
141.90	655.0
142.90	526.0
145.90	343.0
147.00	1129.0
148.00	2243.0
148.90	735.0
149.90	283.0
151.10	263.0
152.90	1075.0
153.90	513.0
154.90	1143.0
156.00	1366.0
157.90	335.0
159.00	249.0
159.80	603.0
161.10	1151.0
161.90	340.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

163.80	271.0
164.90	911.0
166.00	847.0
167.00	4076.0
168.00	2433.0
168.80	270.0
169.90	163.0
170.90	183.0
172.00	400.0
173.10	452.0
174.00	764.0
175.00	1476.0
175.90	336.0
177.00	686.0
178.90	2634.0
180.00	1784.0
180.90	815.0
182.10	202.0
184.00	369.0
184.90	1880.0
186.00	9350.0
186.90	2431.0
188.00	241.0
189.00	640.0
190.90	326.0
191.90	1089.0
193.00	1384.0
193.90	243.0
194.80	195.0
195.90	1735.0
197.90	59048.0
198.90	4283.0
199.80	587.0
200.90	271.0
201.30	273.0
202.90	603.0
204.00	2579.0
205.00	4162.0
206.00	16896.0
207.00	2205.0
208.00	950.0
209.00	530.0
209.90	706.0
211.00	866.0
215.00	201.0
216.90	5099.0
218.00	865.0
220.90	2942.0
222.00	396.0
223.00	1556.0
224.00	11687.0
225.00	2966.0
226.00	301.0
227.00	4936.0
228.00	657.0
229.00	726.0
231.00	469.0
233.90	379.0
234.90	256.0
235.90	163.0
237.00	355.0
237.90	153.0
238.80	225.0
240.10	270.0
240.80	288.0
242.00	346.0
243.00	777.0
244.00	6973.0
244.90	1059.0
246.00	1794.0
247.00	440.0
248.10	201.0
248.80	246.0
252.00	187.0
252.70	184.0
254.90	38072.0
256.00	4967.0
256.90	374.0
258.00	2382.0
258.90	357.0
260.00	195.0
264.90	1169.0
270.90	175.0
273.00	1074.0
273.90	3007.0
275.00	15874.0
276.00	2029.0
276.90	1565.0
277.90	166.0
281.50	172.0
283.00	416.0
283.90	262.0
285.00	275.0
285.80	156.0
293.00	189.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

294.80	243.0
295.90	5087.0
296.90	1318.0
302.90	672.0
307.90	214.0
313.90	316.0
314.90	760.0
316.10	304.0
320.70	287.0
323.00	2035.0
324.10	245.0
326.90	423.0
327.90	188.0
332.00	165.0
334.00	1136.0
335.10	241.0
338.80	151.0
341.00	251.0
345.90	339.0
351.90	437.0
353.90	434.0
364.90	2427.0
365.90	284.0
368.80	289.0
370.60	157.0
371.10	153.0
372.00	934.0
372.90	341.0
383.00	268.0
402.00	407.0
403.00	441.0
403.90	167.0
421.00	417.0
422.00	496.0
423.00	2306.0
424.10	910.0
435.30	270.0
438.60	181.0
439.10	234.0
441.00	6300.0
442.00	40696.0
443.00	7794.0
444.10	715.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

m/z	Abundance
37.00	160.0
38.10	252.0
39.00	1220.0
40.00	982.0
41.00	163.0
44.00	1141.0
48.90	246.0
49.20	265.0
50.00	4579.0
51.00	15323.0
51.90	396.0
56.00	408.0
57.00	1688.0
63.00	789.0
64.00	174.0
69.00	16632.0
73.00	296.0
74.00	1806.0
75.00	2960.0
76.00	1070.0
77.00	19680.0
78.00	862.0
79.00	1118.0
80.00	941.0
81.00	1192.0
82.00	375.0
83.00	442.0
84.90	202.0
85.90	238.0
86.90	255.0
90.90	446.0
91.90	381.0
93.00	2526.0
98.00	2099.0
98.90	1617.0
101.00	705.0
103.10	203.0
103.90	474.0
104.90	616.0
107.00	5792.0
107.90	1333.0
110.00	11610.0
110.90	1303.0
115.90	317.0
117.00	6645.0
117.90	619.0
119.90	152.0
122.00	364.0
123.00	496.0
123.80	210.0
125.00	293.0
127.00	19304.0
128.00	1620.0
129.00	8074.0
129.90	503.0
131.00	169.0
134.00	229.0
134.90	944.0
135.90	278.0
137.20	174.0
140.90	1557.0
142.00	489.0
142.80	235.0
146.90	726.0
147.90	1400.0
148.90	261.0
152.90	303.0
153.90	163.0
154.10	169.0
155.00	559.0
156.00	660.0
156.80	222.0
159.90	398.0
160.80	450.0
165.00	489.0
165.90	663.0
166.90	2483.0
168.00	1114.0
169.00	396.0
172.00	212.0
172.80	266.0
173.90	339.0
175.00	1170.0
176.10	187.0
177.00	726.0
178.90	1757.0
179.90	1203.0
180.90	615.0
184.00	172.0
184.90	806.0
186.00	6296.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

186.90	1683.0
189.00	424.0
189.70	156.0
190.80	261.0
191.90	620.0
192.90	796.0
195.90	1008.0
196.50	574.0
197.90	35912.0
198.90	2902.0
200.00	215.0
201.40	421.0
202.80	420.0
204.00	2034.0
205.00	2857.0
206.00	10117.0
207.00	2033.0
208.00	842.0
208.70	167.0
210.00	192.0
210.70	245.0
216.00	299.0
216.90	2857.0
217.80	407.0
221.00	1846.0
222.90	565.0
223.90	6084.0
224.90	1395.0
226.00	168.0
226.90	2284.0
228.00	422.0
228.80	469.0
233.90	160.0
235.90	218.0
236.90	287.0
242.00	400.0
243.00	681.0
244.00	5589.0
245.00	800.0
246.00	927.0
252.90	190.0
253.70	174.0
255.00	24880.0
256.00	3697.0
258.00	2223.0
259.00	521.0
259.90	162.0
264.90	761.0
272.90	510.0
273.90	2086.0
275.00	10581.0
275.90	1209.0
276.90	874.0
282.80	177.0
284.90	184.0
288.80	164.0
293.00	277.0
295.90	4291.0
296.90	561.0
302.00	170.0
303.10	354.0
304.10	346.0
314.90	534.0
315.80	188.0
321.00	293.0
322.80	975.0
327.00	275.0
327.80	156.0
333.00	190.0
334.10	771.0
340.90	274.0
345.90	255.0
352.00	219.0
353.00	158.0
354.10	255.0
365.00	2388.0
365.90	336.0
372.00	269.0
402.00	319.0
403.00	408.0
421.00	193.0
422.10	265.0
423.00	1764.0
424.10	348.0
436.10	288.0
441.00	4390.0
442.00	33640.0
443.00	5883.0
444.00	669.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

m/z	Abundance
36.20	202.0
39.90	878.0
43.90	1481.0
84.00	202.0
85.60	160.0
132.90	180.0
165.00	185.0
207.00	336.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12