

Data Path : Z:\HPCHEM1\BNA G\DATA\BG071316\
 Data File : BG023080.D
 Acq On : 13 Jul 2016 21:51
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 14 04:36:48 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG070816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 12 18:30:58 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.475	0.402	15.4	88	0.00
3	Pyridine	1.627	1.469	9.7	91	0.00
4	n-Nitrosodimethylamine	0.698	0.641	8.2	91	0.00
5 S	2-Fluorophenol	1.223	1.161	5.1	97	0.00
6	Aniline	2.654	2.475	6.7	93	0.00
7 S	Phenol-d6	1.915	1.798	6.1	92	0.00
8	2-Chlorophenol	1.301	1.217	6.5	94	0.00
9	Benzaldehyde	1.280	1.163	9.1	93	0.00
10 C	Phenol	1.971	1.878	4.7	95	0.00
11	bis(2-Chloroethyl)ether	1.562	1.453	7.0	94	0.00
12	1,3-Dichlorobenzene	1.593	1.488	6.6	97	0.00
13 C	1,4-Dichlorobenzene	1.595	1.529	4.1	97	0.00
14	1,2-Dichlorobenzene	1.585	1.441	9.1	95	0.00
15	Benzyl Alcohol	1.754	1.637	6.7	91	0.00
16	2,2'-oxybis(1-Chloropropane	1.908	1.851	3.0	98	0.00
17	2-Methylphenol	1.283	1.198	6.6	93	0.00
18	Hexachloroethane	0.673	0.640	4.9	103	0.00
19 P	n-Nitroso-di-n-propylamine	1.727	1.564	9.4	88	0.00
20	3+4-Methylphenols	1.789	1.653	7.6	87	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
22	Acetophenone	0.600	0.542	9.7	86	0.00
23 S	Nitrobenzene-d5	0.467	0.440	5.8	90	0.00
24	Nitrobenzene	0.496	0.474	4.4	93	0.00
25	Isophorone	0.939	0.840	10.5	83	0.00
26 C	2-Nitrophenol	0.195	0.186	4.6	86	0.00
27	2,4-Dimethylphenol	0.337	0.293	13.1	83	0.00
28	bis(2-Chloroethoxy)methane	0.524	0.494	5.7	89	0.00
29 C	2,4-Dichlorophenol	0.359	0.333	7.2	86	0.00
30	1,2,4-Trichlorobenzene	0.411	0.380	7.5	89	0.00
31	Naphthalene	1.018	0.940	7.7	89	0.00
32	Benzoic acid	0.306	0.273	10.8	80	0.02
33	4-Chloroaniline	0.498	0.459	7.8	88	0.00
34 C	Hexachlorobutadiene	0.334	0.302	9.6	88	0.00
35	Caprolactam	0.148	0.123	16.9	79	0.01
36 C	4-Chloro-3-methylphenol	0.491	0.443	9.8	84	0.01
37	2-Methylnaphthalene	0.805	0.743	7.7	86	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
39	1,2,4,5-Tetrachlorobenzene	0.862	0.787	8.7	81	0.00
40 P	Hexachlorocyclopentadiene	0.496	0.526	-6.0	93	0.00
41 S	2,4,6-Tribromophenol	0.359	0.296	17.5	74	0.00
42 C	2,4,6-Trichlorophenol	0.495	0.457	7.7	80	0.01
43	2,4,5-Trichlorophenol	0.509	0.469	7.9	83	0.01
44 S	2-Fluorobiphenyl	1.270	1.198	5.7	85	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG071316\
 Data File : BG023080.D
 Acq On : 13 Jul 2016 21:51
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 14 04:36:48 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG070816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 12 18:30:58 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.501	1.436	4.3	84	0.00
46	2-Chloronaphthalene	1.217	1.177	3.3	85	0.00
47	2-Nitroaniline	0.519	0.511	1.5	87	0.01
48	Acenaphthylene	1.826	1.718	5.9	83	0.00
49	Dimethylphthalate	1.634	1.463	10.5	81	0.00
50	2,6-Dinitrotoluene	0.367	0.334	9.0	81	0.01
51 C	Acenaphthene	1.193	1.130	5.3	85	0.00
52	3-Nitroaniline	0.366	0.344	6.0	83	0.01
53 P	2,4-Dinitrophenol	0.252	0.273	-8.3	92	0.00
54	Dibenzofuran	1.786	1.645	7.9	83	0.00
55 P	4-Nitrophenol	0.307	0.259	15.6	74	0.02
56	2,4-Dinitrotoluene	0.535	0.492	8.0	82	0.00
57	Fluorene	1.537	1.419	7.7	84	0.00
58	2,3,4,6-Tetrachlorophenol	0.509	0.440	13.6	76	0.00
59	Diethylphthalate	1.662	1.505	9.4	81	0.00
60	4-Chlorophenyl-phenylether	0.936	0.816	12.8	77	0.00
61	4-Nitroaniline	0.396	0.370	6.6	84	0.00
62	Azobenzene	1.590	1.554	2.3	88	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
64	4,6-Dinitro-2-methylphenol	0.149	0.130	12.8	73	0.00
65 c	n-Nitrosodiphenylamine	0.576	0.547	5.0	82	0.00
66	4-Bromophenyl-phenylether	0.260	0.225	13.5	75	0.00
67	Hexachlorobenzene	0.283	0.242	14.5	74	0.00
68	Atrazine	0.256	0.237	7.4	82	0.00
69 C	Pentachlorophenol	0.183	0.154	15.8	69	0.00
70	Phenanthrene	0.980	0.914	6.7	83	0.00
71	Anthracene	0.992	0.928	6.5	83	0.00
72	Carbazole	0.858	0.809	5.7	84	0.00
73	Di-n-butylphthalate	0.954	0.958	-0.4	86	0.00
74 C	Fluoranthene	1.203	1.067	11.3	82	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	82	0.02
76	Benzidine	0.573	0.365	36.3#	50	0.00
77	Pyrene	1.124	1.080	3.9	83	0.00
78 S	Terphenyl-d14	0.646	0.660	-2.2	84	0.00
79	Butylbenzylphthalate	0.445	0.453	-1.8	84	0.01
80	Benzo(a)anthracene	1.119	1.055	5.7	79	0.02
81	3,3'-Dichlorobenzidine	0.491	0.445	9.4	73	0.02
82	Chrysene	1.050	0.997	5.0	80	0.03
83	Bis(2-ethylhexyl)phthalate	0.618	0.614	0.6	83	0.03
84 c	Di-n-octyl phthalate	1.031	1.032	-0.1	82	0.04
85	Indeno(1,2,3-cd)pyrene	1.338	1.127	15.8	71	0.11
86 I	Perylene-d12	1.000	1.000	0.0	79	0.06
87	Benzo(b)fluoranthene	1.154	1.083	6.2	76	0.05

Data Path : Z:\HPCHEM1\BNA G\DATA\BG071316\
Data File : BG023080.D
Acq On : 13 Jul 2016 21:51
Operator : UM/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jul 14 04:36:48 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG070816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 12 18:30:58 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.095	1.043	4.7	75	0.06
89 C	Benzo(a)pyrene	1.106	1.029	7.0	75	0.06
90	Dibenzo(a,h)anthracene	1.098	0.971	11.6	71	0.10
91	Benzo(a,h,i)perylene	1.099	0.948	13.7	69	0.13

(#) = Out of Range

SPCC's out = 0 CCC's out = 0