


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG071318\
 Data File : BG035624.D
 Acq On : 13 Jul 2018 10:49
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG071218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 31 12:30:06 2018



AutoFind: Scans 2477, 2478, 2479; Background Corrected with Scan 2471

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.9	15507	PASS
68	69	0.00	2	1.8	323	PASS
69	198	0.00	100	31.2	17948	PASS
70	69	0.00	2	0.3	55	PASS
127	198	10	80	35.9	20690	PASS
197	198	0.00	2	0.3	174	PASS
198	198	100	100	100.0	57560	PASS
199	198	5	9	8.5	4898	PASS
275	198	10	60	28.8	16559	PASS
365	198	1	100	3.5	2003	PASS
441	198	0.01	100	9.3	5367	PASS
442	198	50	100	61.7	35517	PASS
443	442	15	24	19.8	7024	PASS