

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071421\
 Data File : BG049341.D
 Acq On : 14 Jul 2021 23:37
 Operator : CG/JU
 Sample : PB137619BL
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK619

Quant Time: Jul 15 01:27:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG070921.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 09 03:15:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.061	152	28815	20.000	ng/ul	-0.02
20) Naphthalene-d8	10.881	136	118745	20.000	ng/ul	-0.02
38) Acenaphthene-d10	14.712	164	87243	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.461	188	198478	20.000	ng/ul	0.00
79) Chrysene-d12	21.745	240	197799	20.000	ng/ul	0.00
88) Perylene-d12	25.029	264	203060	20.000	ng/ul	-0.02

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.454	96	4918	8.028	ng/uL	-0.02
4) Pyridine-d5	3.877	84	70384	39.076	ng/ul	0.00
7) Phenol-d5	7.221	99	99177	39.568	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.385	67	56083	38.578	ng/ul	0.00
11) 2-Chlorophenol-d4	7.597	132	75559	40.788	ng/ul	0.00
15) 4-Methylphenol-d8	8.772	113	80991	40.329	ng/ul	0.00
21) Nitrobenzene-d5	9.230	128	37831	41.518	ng/ul	0.00
24) 2-Nitrophenol-d4	9.958	143	44182	42.239	ng/ul	-0.02
28) 2,4-Dichlorophenol-d3	10.505	165	88725	43.470	ng/ul	0.00
31) 4-Chloroaniline-d4	11.016	131	108185	40.674	ng/ul	0.00
46) Dimethylphthalate-d6	14.112	166	290540	43.303	ng/ul	0.00
49) Acenaphthylene-d8	14.406	160	337055	42.815	ng/ul	0.00
54) 4-Nitrophenol-d4	14.906	143	41785	37.626	ng/ul	0.00
60) Fluorene-d10	15.705	176	238082	41.911	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.816	200	37833	29.989	ng/ul	0.00
73) Anthracene-d10	17.561	188	399474	44.207	ng/ul	0.00
81) Pyrene-d10	19.847	212	457847	45.162	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.806	264	467410	44.280	ng/ul	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071421\
Data File : BG049341.D
Acq On : 14 Jul 2021 23:37
Operator : CG/JU
Sample : PB137619BL
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK619

Quant Time: Jul 15 01:27:46 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG070921.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jul 09 03:15:38 2021
Response via : Initial Calibration

