

Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG071516\  
 Data File : BG023112.D  
 Acq On : 15 Jul 2016 14:05  
 Operator : UM/SJ  
 Sample : PB92146BL  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 PB92146BL

Quant Time: Jul 16 00:48:22 2016  
 Quant Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG070816.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Jul 12 18:30:58 2016  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 8.16  | 152  | 98851    | 20.00  | ng    | 0.01     |
| 21) Naphthalene-d8          | 10.99 | 136  | 456657   | 20.00  | ng    | 0.02     |
| 38) Acenaphthene-d10        | 14.81 | 164  | 372761   | 20.00  | ng    | 0.01     |
| 63) Phenanthrene-d10        | 17.56 | 188  | 915371   | 20.00  | ng    | 0.00     |
| 75) Chrysene-d12            | 21.86 | 240  | 985775   | 20.00  | ng    | 0.01     |
| 86) Perylene-d12            | 25.24 | 264  | 936383   | 20.00  | ng    | 0.05     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.71  | 112  | 858107   | 141.98 | ng    | 0.01     |
| 7) Phenol-d6                | 7.32  | 99   | 1311313  | 138.52 | ng    | 0.02     |
| 23) Nitrobenzene-d5         | 9.34  | 82   | 936739   | 87.84  | ng    | 0.01     |
| 41) 2,4,6-Tribromophenol    | 16.30 | 330  | 685825   | 102.37 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl        | 13.43 | 172  | 1958438  | 82.76  | ng    | 0.01     |
| 78) Terphenyl-d14           | 20.16 | 244  | 2856486  | 96.95  | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG071516\  
Data File : BG023112.D  
Acq On : 15 Jul 2016 14:05  
Operator : UM/SJ  
Sample : PB92146BL  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
PB92146BL

Quant Time: Jul 16 00:48:22 2016  
Quant Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG070816.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Jul 12 18:30:58 2016  
Response via : Initial Calibration

