




Data Path : Z:\HPCHEM1\BNA_G\DATA\BG071516\
Data File : BG023110.D
Acq On : 15 Jul 2016 12:29
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 16 06:59:10 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG070816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 12 18:30:58 2016
Response via : Initial Calibration



TIC: BG023110.D

(76) Benzidine




19.789min (+0.009) 15983.40ng

response 1381081

Ion	Exp%	Act%
184.00	100	100
185.00	15.80	16.47
183.00	13.40	14.37
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG071516\
Data File : BG023110.D
Acq On : 15 Jul 2016 12:29
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 16 06:59:10 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG070816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 12 18:30:58 2016
Response via : Initial Calibration



TIC: BG023110.D

(69) Pentachlorophenol (C)

17.209min (+0.009) 11573.58ng

response 285313

Ion	Exp%	Act%
265.70	100	100
268.00	64.90	65.94
264.00	63.20	61.86
0.00	0.00	0.00