

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG071616\
 Data File : BG023139.D
 Acq On : 16 Jul 2016 9:38
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 18 01:16:12 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG070816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 12 18:30:58 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
2	1,4-Dioxane	0.475	0.475	0.0	91	0.00
3	Pyridine	1.627	1.591	2.2	86	0.00
4	n-Nitrosodimethylamine	0.698	0.728	-4.3	90	0.00
5 S	2-Fluorophenol	1.223	1.215	0.7	89	0.00
6	Aniline	2.654	2.627	1.0	86	0.00
7 S	Phenol-d6	1.915	1.939	-1.3	87	0.00
8	2-Chlorophenol	1.301	1.320	-1.5	89	0.00
9	Benzaldehyde	1.280	1.257	1.8	88	0.00
10 C	Phenol	1.971	1.999	-1.4	88	0.00
11	bis(2-Chloroethyl)ether	1.562	1.561	0.1	88	0.00
12	1,3-Dichlorobenzene	1.593	1.573	1.3	89	0.00
13 C	1,4-Dichlorobenzene	1.595	1.618	-1.4	90	0.00
14	1,2-Dichlorobenzene	1.585	1.555	1.9	89	0.00
15	Benzyl Alcohol	1.754	1.753	0.1	85	0.00
16	2,2'-oxybis(1-Chloropropane	1.908	1.989	-4.2	92	0.00
17	2-Methylphenol	1.283	1.274	0.7	86	0.00
18	Hexachloroethane	0.673	0.677	-0.6	95	0.00
19 P	n-Nitroso-di-n-propylamine	1.727	1.673	3.1	82	0.00
20	3+4-Methylphenols	1.789	1.844	-3.1	85	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
22	Acetophenone	0.600	0.566	5.7	81	0.00
23 S	Nitrobenzene-d5	0.467	0.471	-0.9	87	0.00
24	Nitrobenzene	0.496	0.490	1.2	87	0.00
25	Isophorone	0.939	0.863	8.1	77	0.00
26 C	2-Nitrophenol	0.195	0.187	4.1	78	0.00
27	2,4-Dimethylphenol	0.337	0.317	5.9	81	0.00
28	bis(2-Chloroethoxy)methane	0.524	0.503	4.0	81	0.00
29 C	2,4-Dichlorophenol	0.359	0.347	3.3	81	0.00
30	1,2,4-Trichlorobenzene	0.411	0.399	2.9	84	0.00
31	Naphthalene	1.018	1.005	1.3	86	0.00
32	Benzoic acid	0.306	0.238	22.2	63	0.00
33	4-Chloroaniline	0.498	0.474	4.8	82	0.00
34 C	Hexachlorobutadiene	0.334	0.312	6.6	82	0.00
35	Caprolactam	0.148	0.120	18.9	70	0.00
36 C	4-Chloro-3-methylphenol	0.491	0.462	5.9	79	0.00
37	2-Methylnaphthalene	0.805	0.778	3.4	81	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	77	0.00
39	1,2,4,5-Tetrachlorobenzene	0.862	0.838	2.8	76	0.00
40 P	Hexachlorocyclopentadiene	0.496	0.681	-37.3#	107	0.00
41 S	2,4,6-Tribromophenol	0.359	0.292	18.7	65	0.00
42 C	2,4,6-Trichlorophenol	0.495	0.455	8.1	70	0.00
43	2,4,5-Trichlorophenol	0.509	0.463	9.0	72	0.00
44 S	2-Fluorobiphenyl	1.270	1.256	1.1	79	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG071616\
 Data File : BG023139.D
 Acq On : 16 Jul 2016 9:38
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 18 01:16:12 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG070816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 12 18:30:58 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.501	1.520	-1.3	79	0.00
46	2-Chloronaphthalene	1.217	1.242	-2.1	79	0.00
47	2-Nitroaniline	0.519	0.526	-1.3	79	0.01
48	Acenaphthylene	1.826	1.840	-0.8	79	0.00
49	Dimethylphthalate	1.634	1.482	9.3	72	0.00
50	2,6-Dinitrotoluene	0.367	0.340	7.4	73	0.00
51 C	Acenaphthene	1.193	1.177	1.3	78	0.00
52	3-Nitroaniline	0.366	0.356	2.7	75	0.00
53 P	2,4-Dinitrophenol	0.252	0.237	6.0	71	0.00
54	Dibenzofuran	1.786	1.734	2.9	77	0.00
55 P	4-Nitrophenol	0.307	0.310	-1.0	79	0.01
56	2,4-Dinitrotoluene	0.535	0.492	8.0	73	0.00
57	Fluorene	1.537	1.502	2.3	78	0.00
58	2,3,4,6-Tetrachlorophenol	0.509	0.439	13.8	67	0.00
59	Diethylphthalate	1.662	1.544	7.1	74	0.00
60	4-Chlorophenyl-phenylether	0.936	0.861	8.0	71	0.00
61	4-Nitroaniline	0.396	0.385	2.8	77	0.00
62	Azobenzene	1.590	1.659	-4.3	83	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	74	0.00
64	4,6-Dinitro-2-methylphenol	0.149	0.125	16.1	59	0.00
65 c	n-Nitrosodiphenylamine	0.576	0.585	-1.6	74	0.00
66	4-Bromophenyl-phenylether	0.260	0.239	8.1	67	0.00
67	Hexachlorobenzene	0.283	0.264	6.7	69	0.00
68	Atrazine	0.256	0.245	4.3	72	0.00
69 C	Pentachlorophenol	0.183	0.185	-1.1	71	0.00
70	Phenanthrene	0.980	0.998	-1.8	76	0.00
71	Anthracene	0.992	1.023	-3.1	78	0.00
72	Carbazole	0.858	0.910	-6.1	80	0.00
73	Di-n-butylphthalate	0.954	1.075	-12.7	82	0.00
74 C	Fluoranthene	1.203	1.223	-1.7	80	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	76	0.01
76	Benzidine	0.573	0.520	9.2	67	0.00
77	Pyrene	1.124	1.129	-0.4	81	0.00
78 S	Terphenyl-d14	0.646	0.691	-7.0	81	0.00
79	Butylbenzylphthalate	0.445	0.477	-7.2	82	0.00
80	Benzo(a)anthracene	1.119	1.141	-2.0	79	0.02
81	3,3'-Dichlorobenzidine	0.491	0.470	4.3	72	0.01
82	Chrysene	1.050	1.072	-2.1	80	0.01
83	Bis(2-ethylhexyl)phthalate	0.618	0.660	-6.8	83	0.02
84 c	Di-n-octyl phthalate	1.031	1.121	-8.7	83	0.03
85	Indeno(1,2,3-cd)pyrene	1.338	1.169	12.6	68	0.07
86 I	Perylene-d12	1.000	1.000	0.0	74	0.05
87	Benzo(b)fluoranthene	1.154	1.145	0.8	75	0.04

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG071616\
Data File : BG023139.D
Acq On : 16 Jul 2016 9:38
Operator : UM/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jul 18 01:16:12 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG070816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 12 18:30:58 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.095	1.132	-3.4	76	0.05
89 C	Benzo(a)pyrene	1.106	1.111	-0.5	75	0.04
90	Dibenzo(a,h)anthracene	1.098	1.020	7.1	70	0.07
91	Benzo(g,h,i)perylene	1.099	0.948	13.7	64	0.08

(#) = Out of Range

SPCC's out = 0 CCC's out = 0