


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG071717\
Data File : BG027942.D
Acq On : 17 Jul 2017 10:26
Operator : SJ/JU
Sample : DFTPP51
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP51

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG071717.M
Title : SVOA CALIBRATION
Last Update : Mon Jul 17 16:40:57 2017



AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2468

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.6	25512	PASS
68	69	0.00	2	0.8	180	PASS
69	198	0.00	100	30.1	22858	PASS
70	69	0.00	2	0.3	66	PASS
127	198	10	80	38.6	29314	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	76018	PASS
199	198	5	9	6.5	4916	PASS
275	198	10	60	27.7	21072	PASS
365	198	1	100	3.6	2704	PASS
441	442	0.01	100	14.9	7609	PASS
442	198	50	100	67.4	51202	PASS
443	442	15	24	19.5	9974	PASS

m/z	Abundance
37.90	272.0
39.00	1972.0
40.00	228.0
42.90	186.0
44.00	1030.0
49.00	348.0
50.00	6405.0
51.00	24840.0
52.00	1275.0
55.00	284.0
55.90	836.0
57.00	1765.0
60.90	241.0
62.00	434.0
63.00	1368.0
64.10	151.0
65.00	600.0
68.90	22144.0
72.90	466.0
74.00	2508.0
75.00	2866.0
76.00	1410.0
77.00	23792.0
78.00	1738.0
78.90	1653.0
79.90	1262.0
81.00	1806.0
82.00	206.0
83.00	490.0
85.00	363.0
86.00	651.0
87.00	187.0
90.90	824.0
92.00	461.0
93.00	2885.0
98.00	2287.0
99.00	1938.0
101.00	1089.0
102.90	248.0
103.80	548.0
105.00	768.0
107.00	7477.0
108.00	1040.0
109.90	12406.0
111.00	1996.0
111.70	240.0
116.00	240.0
117.00	5984.0
118.00	564.0
121.90	502.0
122.90	759.0
123.90	264.0
124.90	357.0
127.00	26256.0
127.90	1929.0
129.00	10894.0
129.90	727.0
131.00	258.0
134.00	346.0
135.00	1190.0
136.10	482.0
137.00	848.0
138.20	219.0
140.00	173.0
140.90	1452.0
141.90	518.0
142.90	517.0
145.90	162.0
147.00	653.0
147.90	1878.0
149.00	497.0
151.10	233.0
153.00	394.0
153.90	392.0
155.00	655.0
156.00	985.0
157.00	290.0
157.90	259.0
158.90	331.0
159.90	497.0
160.90	837.0
161.90	200.0
164.00	156.0
165.00	358.0
166.00	479.0
167.00	3001.0
167.90	1516.0
168.90	304.0
170.10	153.0
171.00	263.0
172.00	289.0

Instrument :
BNA_G
ClientSampleId :
DFTPP51

172.80	293.0
174.00	840.0
175.00	1452.0
175.90	395.0
176.90	630.0
178.00	190.0
178.90	2485.0
180.00	2091.0
181.00	727.0
181.90	174.0
183.90	229.0
185.00	1283.0
186.00	9190.0
187.00	1772.0
187.90	254.0
188.90	528.0
190.90	331.0
191.90	411.0
193.00	1042.0
194.00	188.0
194.80	186.0
196.00	2220.0
197.90	61048.0
198.90	3880.0
200.00	295.0
201.30	240.0
203.00	449.0
204.00	2895.0
205.00	3992.0
206.00	18472.0
207.00	2667.0
207.90	640.0
210.10	226.0
210.80	875.0
211.70	221.0
214.80	157.0
215.90	395.0
216.90	4926.0
217.90	556.0
221.00	2205.0
222.00	573.0
222.90	787.0
224.00	9188.0
225.00	2292.0
226.10	422.0
227.00	3622.0
227.90	632.0
229.00	658.0
229.90	166.0
230.90	337.0
234.00	328.0
234.90	580.0
235.90	164.0
237.00	626.0
239.10	164.0
239.90	308.0
241.00	449.0
242.00	598.0
242.90	526.0
244.00	7390.0
245.00	888.0
245.90	1282.0
246.90	494.0
249.00	394.0
250.70	175.0
252.90	318.0
255.00	37536.0
256.00	5635.0
257.00	722.0
258.00	1902.0
258.90	352.0
260.00	160.0
264.90	969.0
266.00	180.0
272.00	158.0
273.00	1252.0
273.90	2354.0
275.00	16824.0
276.00	1856.0
277.00	1457.0
278.00	312.0
282.90	237.0
285.00	269.0
293.00	342.0
294.90	177.0
296.00	4582.0
297.00	841.0
298.00	160.0
303.00	785.0
314.00	322.0
315.00	653.0
316.00	174.0
321.10	274.0
323.00	1686.0
324.10	305.0

Instrument :
BNA_G
ClientSampleId :
DFTPP51

327.00	307.0
327.70	163.0
334.00	797.0
335.00	342.0
341.00	153.0
345.90	178.0
351.90	424.0
353.00	379.0
354.00	379.0
365.00	1694.0
366.00	408.0
372.00	849.0
372.90	179.0
383.00	293.0
402.00	364.0
403.00	379.0
420.90	205.0
422.00	405.0
423.00	1818.0
423.90	683.0
441.00	4711.0
442.00	31976.0
443.10	6495.0
444.00	337.0

Instrument :
BNA_G
ClientSampleId :
DFTPP51

m/z	Abundance
38.00	455.0
39.00	2284.0
41.00	198.0
43.00	183.0
44.00	1044.0
48.90	231.0
50.00	7603.0
51.00	30256.0
52.00	1094.0
55.10	373.0
55.90	946.0
56.90	2569.0
61.00	331.0
62.00	342.0
63.00	1054.0
63.80	179.0
65.00	575.0
68.00	540.0
69.00	26704.0
69.80	199.0
73.00	431.0
74.00	2941.0
74.90	4472.0
76.00	1461.0
77.00	28336.0
78.00	1897.0
79.00	2041.0
80.00	1521.0
81.00	1916.0
82.00	500.0
83.00	656.0
84.90	528.0
86.00	1064.0
86.90	433.0
87.90	189.0
91.00	567.0
92.00	730.0
93.00	2990.0
93.90	615.0
95.90	177.0
98.00	2848.0
98.90	2271.0
99.90	270.0
101.00	1275.0
103.00	411.0
104.00	777.0
104.90	946.0
107.00	10241.0
108.00	1878.0
109.00	580.0
109.90	16408.0
110.90	2878.0
112.00	502.0
116.00	646.0
117.00	7625.0
117.90	575.0
122.00	820.0
123.00	1112.0
124.00	567.0
125.00	533.0
127.00	35312.0
128.00	2872.0
129.00	12637.0
129.90	1042.0
131.80	180.0
133.90	398.0
135.00	1165.0
135.90	445.0
137.00	911.0
140.90	2018.0
142.10	718.0
143.00	443.0
145.10	197.0
145.90	429.0
146.90	950.0
148.00	2254.0
148.90	441.0
153.00	627.0
153.90	533.0
154.90	970.0
156.00	1145.0
157.00	317.0
157.90	534.0
158.90	265.0
159.90	757.0
160.90	998.0
162.00	319.0
165.00	795.0
165.90	785.0
167.00	4442.0
168.00	2115.0

Instrument :
BNA_G
ClientSampleId :
DFTPP51

169.00	516.0
170.00	156.0
171.00	380.0
172.00	410.0
173.00	714.0
174.00	778.0
174.90	1409.0
175.80	279.0
177.00	1023.0
178.00	305.0
178.90	2987.0
180.00	2463.0
181.00	1084.0
182.00	354.0
182.90	164.0
184.00	309.0
185.00	1595.0
186.00	11198.0
187.00	3918.0
187.90	459.0
189.00	804.0
190.10	216.0
190.90	474.0
191.90	1148.0
192.90	1268.0
193.80	252.0
196.00	2553.0
197.90	91232.0
198.90	5868.0
199.90	329.0
201.90	244.0
202.90	761.0
204.00	3477.0
205.00	6171.0
206.00	25504.0
207.00	3272.0
208.00	1002.0
209.20	244.0
209.80	259.0
211.00	1165.0
211.90	226.0
215.00	353.0
216.00	474.0
216.90	6602.0
218.00	636.0
221.00	3962.0
222.90	1384.0
224.00	14407.0
225.00	3798.0
226.00	509.0
226.90	5944.0
228.00	896.0
229.00	1608.0
231.10	372.0
233.90	446.0
234.90	258.0
235.90	207.0
237.00	655.0
238.80	190.0
240.90	581.0
241.90	841.0
243.00	972.0
244.00	11346.0
245.00	1719.0
246.00	2066.0
247.00	484.0
248.90	494.0
252.10	207.0
253.20	321.0
255.00	55184.0
256.00	8323.0
257.00	668.0
258.00	3393.0
258.90	734.0
265.00	1641.0
265.90	216.0
271.00	241.0
272.00	176.0
273.00	1890.0
274.00	4345.0
275.00	24320.0
276.00	2843.0
276.90	2390.0
278.00	290.0
283.00	309.0
285.00	543.0
285.90	169.0
292.10	229.0
293.00	425.0
295.90	6245.0
296.90	651.0
297.90	208.0
303.00	688.0
304.00	268.0
313.90	220.0

Instrument :
BNA_G
ClientSampleId :
DFTPP51

314.90	821.0
315.90	572.0
316.80	172.0
320.90	156.0
323.10	2152.0
324.10	643.0
327.00	472.0
327.90	189.0
331.90	197.0
332.90	232.0
334.10	1203.0
335.10	313.0
340.90	369.0
346.00	480.0
352.00	639.0
353.00	513.0
354.00	493.0
364.90	3076.0
365.80	383.0
370.90	167.0
372.00	835.0
373.10	255.0
382.90	473.0
391.00	242.0
402.00	361.0
402.90	608.0
404.10	253.0
421.00	536.0
422.00	396.0
423.00	3874.0
423.90	1046.0
425.00	189.0
441.00	9032.0
442.00	57312.0
443.00	10728.0
444.00	1068.0

Instrument :
BNA_G
ClientSampleId :
DFTPP51

m/z	Abundance
37.90	237.0
39.00	1280.0
39.90	181.0
44.00	1290.0
49.10	167.0
50.00	5547.0
51.00	21440.0
52.00	1040.0
55.00	159.0
55.90	830.0
56.90	1612.0
60.90	293.0
61.90	438.0
63.00	928.0
65.00	598.0
66.90	170.0
69.00	19728.0
71.00	185.0
73.00	320.0
74.00	2359.0
75.00	2909.0
76.00	1092.0
77.00	24536.0
78.00	1463.0
79.00	1973.0
79.90	1227.0
81.00	1751.0
82.00	431.0
83.00	406.0
85.00	466.0
85.90	555.0
86.90	344.0
90.80	197.0
92.00	464.0
92.90	3084.0
93.90	165.0
98.00	1769.0
99.00	1813.0
101.00	692.0
103.00	248.0
103.90	630.0
105.00	742.0
107.00	7014.0
108.00	1035.0
110.00	13099.0
111.00	1977.0
111.90	439.0
113.00	177.0
115.10	202.0
116.00	499.0
117.00	5391.0
118.00	305.0
118.90	150.0
120.70	201.0
122.00	684.0
122.90	984.0
124.10	317.0
125.00	240.0
125.80	193.0
127.00	26376.0
127.90	1979.0
129.00	11918.0
130.00	761.0
130.80	208.0
131.00	211.0
134.00	380.0
134.90	1227.0
135.90	477.0
136.90	595.0
140.90	1796.0
142.00	350.0
143.00	437.0
147.10	839.0
148.00	1892.0
149.00	388.0
149.90	261.0
151.00	339.0
153.00	601.0
154.00	328.0
155.00	857.0
156.00	1225.0
157.10	289.0
157.90	398.0
158.80	175.0
159.90	428.0
161.00	888.0
161.90	373.0
162.90	244.0
164.00	204.0
165.00	471.0
166.00	736.0

Instrument :
BNA_G
ClientSampleId :
DFTPP51

166.90	3216.0
167.90	1858.0
169.10	287.0
170.90	260.0
173.00	425.0
174.00	897.0
175.00	1289.0
175.90	479.0
176.90	726.0
178.00	229.0
179.00	3091.0
180.00	1706.0
180.90	657.0
184.00	222.0
185.00	1604.0
186.00	10635.0
187.00	3238.0
188.10	399.0
188.90	458.0
190.90	478.0
192.00	784.0
192.90	908.0
193.90	260.0
196.00	2466.0
197.90	75776.0
198.90	5001.0
199.90	257.0
200.90	243.0
201.40	287.0
202.90	413.0
204.00	2900.0
205.00	5485.0
206.00	22152.0
207.00	2994.0
207.90	753.0
209.00	241.0
210.20	444.0
211.00	1223.0
213.10	171.0
214.90	194.0
215.90	542.0
217.00	5267.0
218.00	691.0
221.00	2821.0
222.00	1037.0
223.00	1347.0
224.00	11824.0
225.00	2777.0
226.00	561.0
227.00	5054.0
227.90	930.0
228.90	855.0
229.90	195.0
231.00	653.0
234.00	274.0
234.90	317.0
235.90	214.0
237.00	672.0
238.90	159.0
241.00	351.0
241.90	761.0
243.00	575.0
244.00	8041.0
245.00	1391.0
245.90	1652.0
247.00	405.0
249.00	342.0
249.90	208.0
252.90	214.0
255.00	49608.0
256.00	6936.0
256.90	664.0
258.00	2690.0
259.00	450.0
265.00	1191.0
272.10	163.0
273.00	1786.0
274.00	4416.0
275.00	22072.0
276.00	2866.0
277.00	1987.0
278.00	272.0
282.90	366.0
285.00	416.0
293.00	544.0
294.00	223.0
295.10	234.0
296.00	7111.0
297.00	990.0
302.00	222.0
303.00	1037.0
304.20	163.0
314.00	614.0
314.90	646.0
315.90	312.0

Instrument :
BNA_G
ClientSampleId :
DFTPP51

321.90	180.0
323.00	1871.0
324.10	358.0
327.00	492.0
328.00	181.0
331.80	213.0
332.90	213.0
334.00	1237.0
334.90	511.0
341.00	316.0
346.00	363.0
351.80	408.0
352.90	448.0
354.00	771.0
365.00	3344.0
365.90	500.0
372.00	691.0
373.00	314.0
383.10	433.0
383.90	157.0
392.10	189.0
401.90	642.0
403.00	436.0
403.90	335.0
421.00	701.0
421.90	809.0
423.00	3429.0
424.00	894.0
441.00	9086.0
442.00	64320.0
443.00	12700.0
444.00	879.0

Instrument :
BNA_G
ClientSampleId :
DFTPP51

m/z	Abundance
44.00	1282.0
207.00	229.0

Instrument :
BNA_G
ClientSampleId :
DFTPP51