

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG071719\
 Data File : BG041699.D
 Acq On : 17 Jul 2019 22:17
 Operator : HP/JU
 Sample : K3835-08
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SB-4(7-9)

Quant Time: Jul 18 07:45:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG071519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 15 16:16:21 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.05	152	106986	20.00	ng	0.00
21) Naphthalene-d8	10.85	136	424884	20.00	ng	0.00
39) Acenaphthene-d10	14.66	164	255374	20.00	ng	0.00
64) Phenanthrene-d10	17.40	188	647866	20.00	ng	0.00
76) Chrysene-d12	21.65	240	686403	20.00	ng	-0.01
87) Perylene-d12	24.84	264	785333	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.61	112	789567	120.65	ng	0.00
7) Phenol-d6	7.20	99	1092531	124.11	ng	0.00
23) Nitrobenzene-d5	9.20	82	655955	93.89	ng	0.00
42) 2,4,6-Tribromophenol	16.14	330	446498	155.14	ng	0.00
45) 2-Fluorobiphenyl	13.29	172	1388149	84.17	ng	0.00
79) Terphenyl-d14	20.00	244	2166710	73.77	ng	0.00
Target Compounds						
50) Dimethylphthalate	14.11	163	221577	11.047	ng	Qvalue 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG071719\
Data File : BG041699.D
Acq On : 17 Jul 2019 22:17
Operator : HP/JU
Sample : K3835-08
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SB-4(7-9)

Quant Time: Jul 18 07:45:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG071519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 15 16:16:21 2019
Response via : Initial Calibration






#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.05 min Scan# 844
Delta R.T. -0.01 min
Lab File: BG041699.D
Acq: 17 Jul 2019 22:17

Instrument :
BNA_G
ClientSampleId :
SB-4(7-9)

Tot Ion	Ratio	Lower	Upper
152	100		
150	151.3	126.9	190.3
115	58.6	45.0	67.6






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol
Concen: 120.646 ng
RT: 5.61 min Scan# 429
Delta R.T. -0.00 min
Lab File: BG041699.D
Acq: 17 Jul 2019 22:17

Tgt Ion	Ratio	Lower	Upper
112	100		
64	70.6	55.4	83.2
63	37.3	29.1	43.7






Abundance

Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BG

Ion 63.00 (62.70 to 63.70): BG





#7

Phenol-d6

Concen: 124.114 ng

RT: 7.20 min Scan# 700

Delta R.T. -0.00 min

Lab File: BG041699.D

Acq: 17 Jul 2019 22:17

Instrument :

BNA_G

ClientSampleId :

SB-4(7-9)


Tot Ion: 99 Resp: 1092531

Ion Ratio Lower Upper




99 100

42 21.3 17.6 26.4

71 35.2 27.8 41.6



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.85 min Scan# 1322

Delta R.T. -0.00 min

Lab File: BG041699.D

Acq: 17 Jul 2019 22:17

Tgt Ion: 136 Resp: 424884


Ion Ratio Lower Upper

136 100



137 10.5 9.2 13.8


54 10.2 8.7 13.1

68 6.2 5.5 8.3



Abundance Ion 136.00 (135.70 to 136.70): BGC
Ion 137.00 (136.70 to 137.70): BGC
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC









#23
 Nitrobenzene-d5
 Concen: 93.889 ng
 RT: 9.20 min Scan# 1041
 Delta R.T. -0.01 min
 Lab File: BG041699.D
 Acq: 17 Jul 2019 22:17




Instrument :
 BNA_G
 ClientSampled :
 SB-4(7-9)


Tot Ion	82.00	Resp	655955
Ion Ratio	Lower	Upper	
82	100		
128	44.2	35.1	52.7
54	58.4	48.7	73.1



#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.66 min Scan# 1970
 Delta R.T. -0.01 min
 Lab File: BG041699.D
 Acq: 17 Jul 2019 22:17

Tgt Ion	164.00	Resp	255374
Ion Ratio	Lower	Upper	
164	100		
162	101.4	79.6	119.4
160	44.3	35.4	53.0






#42
 2,4,6-Tribromophenol
 Concen: 155.140 ng
 RT: 16.14 min Scan# 2222
 Delta R.T. -0.01 min
 Lab File: BG041699.D
 Acq: 17 Jul 2019 22:17

Instrument :
 BNA_G
 ClientSampleId :
 SB-4(7-9)

Tot Ion	Ratio	Lower	Upper
330	100		
332	96.5	77.4	116.0
141	39.6	31.9	47.9






Abundance

Ion 329.80 (329.50 to 330.50): E


Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#45
 2-Fluorobiphenyl
 Concen: 84.172 ng
 RT: 13.29 min Scan# 1737
 Delta R.T. -0.01 min
 Lab File: BG041699.D
 Acq: 17 Jul 2019 22:17

Tgt Ion	Ratio	Lower	Upper
172	100		
171	39.8	35.0	52.4
170	27.1	23.5	35.3






Abundance

Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E





#50

Dimethylphthalate

Concen: 11.047 ng

RT: 14.11 min Scan# 1876

Delta R.T. -0.02 min

Lab File: BG041699.D

Acq: 17 Jul 2019 22:17

Instrument :

BNA_G

ClientSampleId :

SB-4(7-9)


Tot Ion:163 Resp: 221577

Ion Ratio Lower Upper

163 100

194 4.5 4.2 6.2

164 11.5 9.4 14.2






Abundance

Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.40 min Scan# 2436

Delta R.T. -0.01 min

Lab File: BG041699.D

Acq: 17 Jul 2019 22:17


Tgt Ion:188 Resp: 647866

Ion Ratio Lower Upper

188 100

94 8.0 6.9 10.3

80 9.9 8.0 12.0






Abundance

Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BG

Ion 80.00 (79.70 to 80.70): BG









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.65 min Scan# 3159
Delta R.T. -0.01 min
Lab File: BG041699.D
Acq: 17 Jul 2019 22:17




Instrument :
BNA_G
ClientSampled :
SB-4(7-9)


Tot Ion:240 Resp: 686403
Ion Ratio Lower Upper
240 100
120 9.4 8.0 12.0
236 26.9 22.6 33.8



#79
Terphenyl-d14
Concen: 73.766 ng
RT: 20.00 min Scan# 2879
Delta R.T. -0.01 min
Lab File: BG041699.D
Acq: 17 Jul 2019 22:17

Tgt Ion:244 Resp: 2166710
Ion Ratio Lower Upper
244 100
212 10.2 9.7 14.5
122 13.2 12.5 18.7





#87
Pervlene-d12
Concen: 20.000 ng
RT: 24.84 min Scan# 3703
Delta R.T. -0.02 min
Lab File: BG041699.D
Acq: 17 Jul 2019 22:17

Instrument :
BNA_G
ClientSampleId :
SB-4(7-9)

Tot Ion	Ratio	Lower	Upper
264	100		
260	24.0	19.5	29.3
265	22.4	18.0	27.0

