



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071724\
Data File : BG062105.D
Acq On : 17 Jul 2024 7:51
Operator : MA/JU
Sample : DFTPP599
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP599

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070224.MA.M
Title : SVOA CALIBRATION
Last Update : Tue Jul 16 12:23:16 2024



AutoFind: Scans 2475, 2476, 2477; Background Corrected with Scan 2470

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	54.6	57869	PASS
68	69	0.00	2	1.6	748	PASS
69	198	0.00	100	42.8	45371	PASS
70	69	0.00	2	0.6	272	PASS
127	198	10	80	46.5	49227	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	105941	PASS
199	198	5	9	7.5	7998	PASS
275	198	10	60	32.1	33987	PASS
365	198	1	100	6.4	6807	PASS
441	198	0.01	100	15.5	16401	PASS
442	198	50	100	96.3	102008	PASS
443	442	15	24	18.9	19248	PASS

m/z	Abundance
43.10	192.0
43.90	1095.0
54.90	364.0
55.20	375.0
55.90	182.0
71.00	204.0
72.80	320.0
93.10	157.0
111.00	241.0
114.90	240.0
119.10	174.0
133.00	164.0
143.00	196.0
170.10	169.0
174.20	185.0
193.00	181.0
194.70	155.0
196.10	216.0
207.00	326.0
208.30	212.0
259.00	215.0
265.10	186.0
323.20	197.0
350.00	151.0
362.00	227.0
363.00	190.0
389.40	161.0
390.00	172.0
465.60	171.0
532.10	196.0

Instrument :
BNA_G
ClientSampleId :
DFTPP599

m/z	Abundance
37.10	174.0
38.20	406.0
38.90	2597.0
40.00	570.0
41.00	337.0
44.00	1601.0
44.90	196.0
48.80	598.0
50.00	13674.0
51.00	50128.0
52.00	2662.0
54.80	617.0
55.30	439.0
56.00	1422.0
57.00	3292.0
58.00	358.0
61.00	453.0
61.90	641.0
63.10	1856.0
64.00	426.0
65.00	829.0
68.90	37688.0
70.40	151.0
70.60	153.0
73.00	522.0
74.00	3452.0
75.00	5931.0
76.00	2205.0
77.00	37712.0
78.00	2855.0
79.00	3895.0
79.90	2385.0
80.90	4252.0
82.00	585.0
83.00	1356.0
85.10	531.0
85.90	1062.0
87.10	531.0
88.00	192.0
90.90	1289.0
92.00	1163.0
93.00	4771.0
93.80	488.0
94.90	419.0
96.00	406.0
96.90	243.0
98.10	4668.0
99.00	3464.0
99.90	353.0
101.00	1613.0
101.90	195.0
103.10	446.0
103.90	1439.0
105.00	943.0
105.70	248.0
107.00	12435.0
108.00	2743.0
109.00	635.0
110.00	22888.0
110.90	3417.0
112.00	538.0
114.90	169.0
115.90	546.0
117.00	12936.0
118.00	1177.0
118.90	270.0
120.00	196.0
122.00	1327.0
123.00	1461.0
124.00	349.0
125.10	850.0
125.80	193.0
127.00	41680.0
128.00	4186.0
129.00	18520.0
130.00	1186.0
131.10	512.0
133.20	265.0
134.00	1203.0
135.00	1938.0
136.00	677.0
136.90	542.0
138.00	370.0
140.20	250.0
141.00	2243.0
141.90	645.0
143.10	525.0
145.80	340.0
147.10	2047.0
148.00	2931.0
149.00	643.0
151.00	267.0

Instrument :
BNA_G
ClientSampleId :
DFTPP599

152.10	190.0
153.00	812.0
154.00	800.0
154.90	1447.0
156.00	2111.0
157.70	510.0
159.00	783.0
160.00	1323.0
161.00	1149.0
161.90	230.0
162.10	227.0
164.90	1243.0
165.80	563.0
167.00	5003.0
168.00	1824.0
171.10	186.0
171.90	301.0
173.10	280.0
174.00	1063.0
175.00	1827.0
175.90	930.0
177.00	1199.0
177.90	818.0
178.90	2851.0
180.00	2802.0
181.20	855.0
182.00	394.0
183.00	415.0
183.80	215.0
184.20	264.0
184.90	1720.0
186.00	11835.0
187.00	3145.0
188.00	480.0
189.00	829.0
191.00	669.0
192.00	1404.0
193.00	1513.0
194.10	512.0
195.10	156.0
196.00	3444.0
198.00	79384.0
199.00	7050.0
199.90	602.0
200.70	274.0
201.60	507.0
203.00	605.0
204.00	3717.0
205.00	5903.0
206.00	23728.0
207.00	3174.0
208.00	1162.0
209.10	680.0
210.10	590.0
211.00	1569.0
213.40	193.0
215.10	273.0
216.00	821.0
216.90	5825.0
217.90	1016.0
219.10	222.0
220.00	188.0
221.00	2291.0
221.80	996.0
223.00	1409.0
224.00	14840.0
225.00	3896.0
226.20	739.0
226.90	7391.0
228.10	1444.0
228.90	1078.0
229.80	223.0
231.00	463.0
232.80	374.0
233.90	242.0
235.00	876.0
235.80	456.0
237.00	787.0
238.90	355.0
240.10	262.0
241.00	402.0
242.10	610.0
244.10	9388.0
245.10	1199.0
246.00	2661.0
246.90	667.0
248.90	578.0
250.00	169.0
250.80	324.0
251.70	285.0
253.00	449.0
255.00	59256.0
256.00	8757.0
256.90	718.0
258.00	4169.0
259.00	767.0

Instrument :
BNA_G
ClientSampleId :
DFTPP599

261.00	396.0
261.90	197.0
263.10	201.0
264.90	2206.0
266.00	424.0
266.70	210.0
270.70	177.0
271.30	225.0
272.00	211.0
273.00	1968.0
274.00	3947.0
275.00	22928.0
276.00	3630.0
277.00	2838.0
277.80	643.0
280.80	152.0
283.00	411.0
285.00	389.0
289.90	178.0
292.10	169.0
292.90	649.0
294.10	266.0
296.00	7228.0
297.10	1047.0
300.90	360.0
302.10	165.0
303.00	689.0
313.30	290.0
313.60	293.0
315.00	840.0
315.90	674.0
320.90	239.0
323.10	1731.0
324.10	375.0
326.80	785.0
328.00	540.0
331.90	189.0
332.70	198.0
334.00	1682.0
335.10	703.0
346.00	627.0
352.00	1127.0
353.10	328.0
353.90	748.0
358.90	217.0
365.00	4310.0
365.90	626.0
370.80	198.0
372.10	1408.0
373.00	840.0
383.70	160.0
384.10	187.0
389.80	205.0
390.10	346.0
402.00	493.0
403.10	804.0
404.10	318.0
420.80	724.0
421.90	639.0
423.00	2446.0
424.00	666.0
428.60	171.0
438.50	231.0
441.00	9476.0
442.00	58224.0
443.00	12536.0
443.90	932.0
549.30	158.0

Instrument :
BNA_G
ClientSampleId :
DFTPP599

m/z	Abundance
37.90	554.0
39.00	6216.0
39.90	243.0
41.00	282.0
42.70	171.0
44.00	911.0
44.90	194.0
50.00	19496.0
51.00	69512.0
52.00	3351.0
52.80	473.0
54.80	404.0
55.10	490.0
55.90	1995.0
57.00	3871.0
57.80	416.0
61.00	520.0
62.00	863.0
63.00	2324.0
63.90	441.0
65.00	1306.0
67.90	1295.0
69.00	52784.0
70.00	370.0
70.90	207.0
73.10	787.0
74.00	4963.0
75.00	8011.0
76.10	4186.0
77.00	53128.0
78.00	3490.0
79.00	3967.0
80.00	3569.0
81.00	5535.0
81.90	1405.0
83.00	703.0
84.00	260.0
85.00	1435.0
86.00	1660.0
87.00	601.0
87.90	185.0
89.00	194.0
90.10	282.0
91.00	980.0
91.90	1179.0
93.00	6951.0
94.10	421.0
95.10	239.0
96.00	584.0
96.70	280.0
96.90	278.0
97.90	5243.0
99.00	5040.0
100.10	385.0
101.00	2594.0
102.10	350.0
103.00	700.0
104.00	1669.0
105.00	1708.0
107.00	20672.0
108.00	2912.0
108.90	603.0
110.00	29016.0
111.00	5673.0
112.00	472.0
112.60	165.0
113.20	203.0
113.80	172.0
115.00	170.0
116.10	1052.0
117.00	19904.0
118.00	1607.0
118.90	159.0
120.10	265.0
122.10	1416.0
123.10	2058.0
124.00	628.0
124.80	806.0
127.00	56528.0
128.00	4571.0
129.00	28008.0
130.00	3474.0
130.80	417.0
131.90	216.0
132.90	367.0
134.00	792.0
135.00	2285.0
135.90	594.0
137.00	855.0
137.80	497.0
140.00	305.0
141.00	3877.0

Instrument :
BNA_G
ClientSampleId :
DFTPP599

142.10	1252.0
143.00	856.0
143.90	310.0
144.70	157.0
145.90	354.0
147.00	2327.0
148.00	4036.0
149.00	580.0
149.90	226.0
151.00	759.0
151.60	300.0
153.00	911.0
154.00	689.0
155.00	2084.0
156.10	2504.0
157.10	427.0
158.00	675.0
159.00	741.0
159.80	816.0
160.90	1854.0
162.10	509.0
165.00	1301.0
166.30	1087.0
167.00	6715.0
168.00	2567.0
169.00	530.0
170.40	546.0
170.90	330.0
171.80	576.0
173.10	1520.0
174.10	2496.0
175.00	2467.0
176.00	885.0
176.90	1791.0
177.80	970.0
179.00	4990.0
180.00	3725.0
180.90	1991.0
182.00	263.0
183.00	577.0
183.90	514.0
185.00	2933.0
186.00	19320.0
187.00	6066.0
188.00	506.0
189.00	1882.0
191.10	594.0
191.90	1638.0
193.00	2180.0
193.90	461.0
196.00	2656.0
198.00	122168.0
199.00	8939.0
199.90	1170.0
201.40	414.0
202.10	266.0
203.00	1580.0
203.90	5696.0
205.00	10113.0
206.00	36552.0
207.10	5532.0
208.00	1131.0
209.00	748.0
209.80	553.0
210.20	536.0
210.50	610.0
211.10	1387.0
212.10	345.0
212.90	161.0
214.00	162.0
215.00	416.0
216.10	852.0
217.00	10788.0
218.00	1536.0
221.00	3628.0
222.00	1952.0
223.10	3081.0
224.00	21552.0
225.00	5678.0
227.00	11579.0
228.00	1559.0
229.00	1907.0
230.00	529.0
230.90	701.0
232.10	224.0
232.90	186.0
233.90	521.0
234.90	600.0
236.10	543.0
237.00	1268.0
238.90	435.0
239.90	662.0
241.00	557.0
241.90	881.0
242.90	1248.0
244.00	15248.0

Instrument :
BNA_G
ClientSampleId :
DFTPP599

245.00	2354.0
246.00	3629.0
246.80	659.0
248.00	207.0
249.00	530.0
249.60	267.0
250.90	324.0
251.50	271.0
252.10	515.0
252.90	565.0
254.00	1101.0
255.00	96016.0
256.00	16145.0
257.20	848.0
258.00	5863.0
258.90	933.0
260.00	226.0
261.00	168.0
265.00	2924.0
266.00	583.0
270.00	184.0
271.00	229.0
273.00	3271.0
274.00	8257.0
275.00	40648.0
276.10	6256.0
277.00	3510.0
278.00	976.0
279.70	388.0
283.10	486.0
283.80	299.0
284.80	589.0
286.00	231.0
290.70	151.0
293.10	925.0
296.00	13726.0
296.80	1514.0
301.90	226.0
303.10	1382.0
304.00	433.0
310.00	289.0
313.80	417.0
314.10	405.0
315.00	2390.0
316.00	866.0
317.00	302.0
320.90	545.0
321.80	160.0
323.10	5018.0
324.00	1314.0
327.10	825.0
327.90	422.0
331.90	758.0
332.40	199.0
333.00	331.0
334.10	4468.0
335.00	1242.0
338.80	185.0
341.10	544.0
341.90	215.0
342.90	258.0
344.00	235.0
345.90	794.0
351.90	1535.0
353.00	1577.0
354.00	1491.0
355.20	189.0
356.00	175.0
359.10	296.0
364.90	7123.0
365.90	640.0
370.80	460.0
372.00	3234.0
372.80	872.0
378.90	219.0
382.80	937.0
390.00	156.0
391.10	416.0
392.00	192.0
400.80	408.0
401.90	1154.0
402.90	1377.0
404.20	897.0
405.10	187.0
414.40	165.0
417.70	195.0
421.00	1080.0
423.00	6411.0
424.00	617.0
424.80	226.0
427.70	201.0
434.00	183.0
441.00	17376.0
441.90	115512.0
443.00	22288.0
444.00	2796.0

Instrument :
BNA_G
ClientSampleId :
DFTPP599

445.00316.0

537.70257.0

Instrument :
BNA_G
ClientSampleId :
DFTPP599

m/z	Abundance
37.90	507.0
39.00	4121.0
40.00	152.0
41.10	297.0
44.00	749.0
49.00	371.0
50.00	13647.0
51.00	53968.0
52.00	3271.0
53.00	282.0
55.20	415.0
56.00	1975.0
57.00	3469.0
57.90	285.0
60.90	943.0
62.10	675.0
62.90	3390.0
64.00	456.0
64.90	873.0
66.00	377.0
67.10	244.0
68.00	948.0
69.00	45640.0
69.90	447.0
73.00	755.0
74.00	5462.0
75.00	6987.0
76.10	2771.0
77.00	47928.0
77.90	3640.0
79.00	3542.0
80.00	2694.0
81.00	4177.0
81.90	737.0
82.90	826.0
84.90	935.0
86.10	1228.0
87.00	731.0
90.90	745.0
92.00	754.0
93.00	6247.0
93.90	569.0
95.00	316.0
96.10	280.0
97.10	417.0
98.00	5286.0
99.00	4343.0
100.20	194.0
101.00	1342.0
103.00	400.0
103.90	1606.0
105.00	1196.0
107.00	18912.0
108.00	3508.0
108.80	432.0
110.00	28072.0
111.00	4837.0
112.10	671.0
112.90	464.0
116.10	888.0
117.00	14457.0
118.00	1315.0
118.80	244.0
119.30	183.0
122.10	1362.0
123.00	1367.0
124.00	1128.0
124.90	522.0
127.00	49472.0
128.00	4185.0
129.00	20784.0
130.00	2310.0
131.10	309.0
132.00	239.0
133.80	776.0
135.00	1753.0
136.00	722.0
137.00	1322.0
138.90	261.0
141.00	3097.0
142.00	814.0
142.80	818.0
143.80	426.0
146.00	803.0
147.00	1561.0
148.00	3603.0
148.90	688.0
149.80	220.0
151.10	373.0
151.90	328.0
153.00	888.0
153.90	966.0

Instrument :
BNA_G
ClientSampleId :
DFTPP599

155.00	1666.0
156.00	2000.0
157.10	886.0
157.90	399.0
159.10	314.0
159.80	1469.0
161.10	1988.0
161.80	544.0
164.90	927.0
166.10	1160.0
167.00	7085.0
168.00	2904.0
169.00	679.0
170.10	627.0
171.00	394.0
172.00	587.0
173.20	930.0
174.10	1410.0
175.00	3606.0
176.00	619.0
177.00	1896.0
178.00	493.0
179.00	5131.0
180.00	3254.0
181.10	1719.0
181.90	498.0
183.50	181.0
184.00	323.0
185.00	3359.0
186.00	20200.0
187.00	5339.0
188.00	1023.0
188.90	1431.0
190.00	358.0
191.00	837.0
192.00	1183.0
192.90	1638.0
193.90	386.0
196.00	2961.0
198.00	116272.0
199.00	8006.0
199.80	524.0
200.30	235.0
200.60	239.0
200.90	174.0
201.60	692.0
202.10	481.0
203.00	1458.0
204.00	5323.0
205.00	10007.0
206.00	36408.0
207.00	4777.0
208.00	1476.0
208.90	592.0
210.00	840.0
211.00	2366.0
214.90	1029.0
216.10	1315.0
217.00	10240.0
218.00	1310.0
219.80	227.0
221.00	5005.0
221.50	2014.0
223.00	2062.0
224.00	21680.0
225.00	5806.0
226.10	666.0
227.00	11238.0
228.00	1569.0
229.00	2719.0
231.00	641.0
232.10	188.0
232.90	629.0
234.10	485.0
235.10	988.0
236.00	411.0
236.80	743.0
237.80	215.0
238.90	421.0
239.90	491.0
240.90	383.0
241.40	535.0
242.30	1262.0
243.00	1527.0
244.00	16213.0
245.00	1427.0
246.00	4354.0
246.90	719.0
249.10	322.0
249.90	160.0
250.90	644.0
252.30	546.0
252.90	789.0
255.00	96840.0
256.00	13631.0
257.00	1654.0

Instrument :
BNA_G
ClientSampleId :
DFTPP599

258.00	6192.0
258.70	865.0
261.20	265.0
263.10	166.0
265.00	2457.0
265.90	340.0
266.20	350.0
270.00	193.0
271.20	398.0
273.00	2957.0
274.00	7226.0
275.00	38384.0
276.00	5235.0
277.00	4521.0
278.00	937.0
281.80	336.0
283.10	333.0
283.90	291.0
285.00	1081.0
286.00	465.0
289.00	356.0
292.00	192.0
292.90	1155.0
294.00	256.0
294.80	321.0
296.00	14655.0
297.00	2019.0
298.00	461.0
301.10	219.0
302.00	451.0
302.90	1474.0
304.10	720.0
304.90	343.0
307.90	307.0
313.90	1111.0
315.00	2512.0
316.00	990.0
317.10	412.0
319.90	369.0
321.10	268.0
321.90	236.0
323.00	4805.0
323.90	610.0
326.90	808.0
327.90	478.0
328.90	199.0
332.00	551.0
333.00	455.0
334.10	3718.0
334.90	612.0
341.10	1218.0
342.00	172.0
345.90	915.0
352.00	1265.0
353.00	1928.0
354.00	2055.0
355.00	370.0
356.50	160.0
358.80	289.0
360.80	168.0
365.00	8988.0
365.80	1049.0
369.40	293.0
369.70	252.0
370.80	350.0
372.00	3192.0
373.20	567.0
374.30	224.0
377.20	170.0
383.00	762.0
384.10	622.0
389.80	437.0
390.80	570.0
392.20	212.0
399.40	170.0
401.90	1351.0
403.10	1835.0
404.00	655.0
420.90	1329.0
421.80	1853.0
423.00	8347.0
424.20	1354.0
424.80	429.0
428.30	228.0
429.80	198.0
431.20	227.0
440.90	22352.0
442.00	132288.0
443.00	22920.0
444.00	2923.0
445.10	275.0
473.20	225.0
485.80	189.0

Instrument :
BNA_G
ClientSampleId :
DFTPP599