

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071822\
 Data File : BG054308.D
 Acq On : 18 Jul 2022 10:29
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 18 11:03:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 15 15:54:11 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.381	0.395	-3.7	97	0.00
3	Pyridine	0.967	1.049	-8.5	92	0.00
4	n-Nitrosodimethylamine	0.541	0.600	-10.9	95	0.00
5 S	2-Fluorophenol	0.958	0.989	-3.2	92	0.00
6	Aniline	1.515	1.519	-0.3	89	0.00
7 S	Phenol-d6	1.313	1.330	-1.3	89	0.00
8	2-Chlorophenol	1.101	1.099	0.2	89	0.00
9	Benzaldehyde	0.737	0.709	3.8	97	0.00
10 C	Phenol	1.304	1.328	-1.8	91	0.00
11	bis(2-Chloroethyl)ether	0.951	0.902	5.2	86	0.00
12	1,3-Dichlorobenzene	1.341	1.348	-0.5	90	0.00
13 C	1,4-Dichlorobenzene	1.385	1.428	-3.1	92	0.00
14	1,2-Dichlorobenzene	1.306	1.330	-1.8	91	0.00
15	Benzyl Alcohol	1.080	1.046	3.1	83	0.00
16	2,2'-oxybis(1-Chloropropane	1.198	1.126	6.0	82	0.00
17	2-Methylphenol	0.940	0.971	-3.3	91	0.00
18	Hexachloroethane	0.459	0.467	-1.7	91	0.00
19 P	n-Nitroso-di-n-propylamine	0.877	0.829	5.5	82	0.00
20	3+4-Methylphenols	1.324	1.307	1.3	87	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
22	Acetophenone	0.481	0.495	-2.9	85	0.00
23 S	Nitrobenzene-d5	0.367	0.388	-5.7	87	0.00
24	Nitrobenzene	0.359	0.379	-5.6	89	0.00
25	Isophorone	0.639	0.652	-2.0	84	0.00
26 C	2-Nitrophenol	0.181	0.186	-2.8	84	0.00
27	2,4-Dimethylphenol	0.250	0.266	-6.4	89	0.00
28	bis(2-Chloroethoxy)methane	0.322	0.326	-1.2	83	0.00
29 C	2,4-Dichlorophenol	0.377	0.397	-5.3	87	0.00
30	1,2,4-Trichlorobenzene	0.467	0.502	-7.5	89	0.00
31	Naphthalene	1.004	1.040	-3.6	86	0.00
32	Benzoic acid	0.080	0.084	-5.0	77	0.00
33	4-Chloroaniline	0.408	0.434	-6.4	89	0.00
34 C	Hexachlorobutadiene	0.413	0.436	-5.6	87	0.00
35	Caprolactam	0.095	0.105	-10.5	97	0.00
36 C	4-Chloro-3-methylphenol	0.343	0.360	-5.0	89	0.00
37	2-Methylnaphthalene	0.771	0.810	-5.1	89	0.00
38	1-Methylnaphthalene	0.751	0.785	-4.5	89	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
40	1,2,4,5-Tetrachlorobenzene	0.876	0.909	-3.8	88	0.00
41 P	Hexachlorocyclopentadiene	0.206	0.227	-10.2	86	0.00
42 S	2,4,6-Tribromophenol	0.420	0.454	-8.1	94	0.00
43 C	2,4,6-Trichlorophenol	0.477	0.500	-4.8	88	0.00
44	2,4,5-Trichlorophenol	0.539	0.575	-6.7	94	0.00
45 S	2-Fluorobiphenyl	1.418	1.446	-2.0	88	0.00
46	1,1'-Biphenyl	1.368	1.408	-2.9	90	0.00
47	2-Chloronaphthalene	1.145	1.164	-1.7	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071822\
 Data File : BG054308.D
 Acq On : 18 Jul 2022 10:29
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 18 11:03:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 15 15:54:11 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.299	0.318	-6.4	90	0.00
49	Acenaphthylene	1.701	1.756	-3.2	89	0.00
50	Dimethylphthalate	1.562	1.619	-3.6	91	0.00
51	2,6-Dinitrotoluene	0.343	0.364	-6.1	93	0.00
52 C	Acenaphthene	1.030	1.052	-2.1	89	0.00
53	3-Nitroaniline	0.287	0.296	-3.1	91	0.00
54 P	2,4-Dinitrophenol	0.166	0.163	1.8	85	0.00
55	Dibenzofuran	1.833	1.848	-0.8	89	0.00
56 P	4-Nitrophenol	0.191	0.205	-7.3	92	0.00
57	2,4-Dinitrotoluene	0.491	0.522	-6.3	92	0.00
58	Fluorene	1.507	1.541	-2.3	90	0.00
59	2,3,4,6-Tetrachlorophenol	0.525	0.554	-5.5	90	0.00
60	Diethylphthalate	1.499	1.568	-4.6	92	0.00
61	4-Chlorophenyl-phenylether	0.989	1.046	-5.8	92	0.00
62	4-Nitroaniline	0.301	0.335	-11.3	97	0.00
63	Azobenzene	1.083	1.108	-2.3	89	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
65	4,6-Dinitro-2-methylphenol	0.124	0.126	-1.6	90	0.00
66 c	n-Nitrosodiphenylamine	0.499	0.512	-2.6	93	0.00
67	4-Bromophenyl-phenylether	0.281	0.290	-3.2	93	0.00
68	Hexachlorobenzene	0.321	0.333	-3.7	93	0.00
69	Atrazine	0.222	0.236	-6.3	97	0.00
70 C	Pentachlorophenol	0.126	0.127	-0.8	89	0.00
71	Phenanthrene	0.996	1.034	-3.8	95	0.00
72	Anthracene	0.977	1.010	-3.4	94	0.00
73	Carbazole	0.805	0.856	-6.3	98	0.00
74	Di-n-butylphthalate	0.950	0.995	-4.7	96	0.00
75 C	Fluoranthene	1.358	1.464	-7.8	99	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
77	Benzydine	0.362	0.411	-13.5	117	0.00
78	Pyrene	1.148	1.150	-0.2	99	0.00
79 S	Terphenyl-d14	1.008	1.033	-2.5	101	0.00
80	Butylbenzylphthalate	0.358	0.367	-2.5	103	0.00
81	Benzo(a)anthracene	1.270	1.310	-3.1	105	0.00
82	3,3'-Dichlorobenzidine	0.395	0.415	-5.1	104	0.00
83	Chrysene	1.199	1.227	-2.3	103	0.00
84	Bis(2-ethylhexyl)phthalate	0.506	0.516	-2.0	103	0.00
85 c	Di-n-octyl phthalate	0.863	0.886	-2.7	102	0.00
86 I	Perylene-d12	1.000	1.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	1.521	1.566	-3.0	103	0.00
88	Benzo(b)fluoranthene	1.183	1.214	-2.6	102	0.00
89	Benzo(k)fluoranthene	1.177	1.219	-3.6	105	0.00
90 C	Benzo(a)pyrene	1.010	1.040	-3.0	103	0.00
91	Dibenzo(a,h)anthracene	1.251	1.285	-2.7	103	0.00
92	Benzo(g,h,i)perylene	1.235	1.268	-2.7	103	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG071822\
Data File : BG054308.D
Acq On : 18 Jul 2022 10:29
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jul 18 11:03:57 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 15 15:54:11 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0