

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG071915\
 Data File : BG017971.D
 Acq On : 19 Jul 2015 23:30
 Operator : TP/UM
 Sample : G2966-08
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TEPMW-PCB2

Quant Time: Jul 20 03:47:25 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG071715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 18 01:50:30 2015
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	88254	20.00	ng	0.00
21) Naphthalene-d8	10.36	136	397509	20.00	ng	0.00
38) Acenaphthene-d10	14.24	164	250341	20.00	ng	0.00
63) Phenanthrene-d10	16.99	188	569596	20.00	ng	0.00
75) Chrysene-d12	21.20	240	594491	20.00	ng	0.00
86) Perylene-d12	23.41	264	589354	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.19	112	214742	42.45	ng	0.00
7) Phenol-d6	6.77	99	183057	27.62	ng	0.00
23) Nitrobenzene-d5	8.73	82	383012	62.41	ng	0.00
41) 2,4,6-Tribromophenol	15.74	330	272307	107.36	ng	0.00
44) 2-Fluorobiphenyl	12.86	172	957993	66.31	ng	0.00
78) Terphenyl-d14	19.64	244	1615091	68.43	ng	0.00
Target Compounds						
32) Benzoic acid	9.58	122	53	Below Cal	#	58
49) Dimethylphthalate	13.70	163	74544	4.29	ng	# 97
53) 2,4-Dinitrophenol	14.29	184	66	Below Cal	#	1
64) 4,6-Dinitro-2-methylphenol	15.35	198	70	Below Cal	#	1
69) Pentachlorophenol	16.71	266	75	Below Cal	#	18


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG071915\
Data File : BG017971.D
Acq On : 19 Jul 2015 23:30
Operator : TP/UM
Sample : G2966-08
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TEPMW-PCB2

Quant Time: Jul 20 03:47:25 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG071715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 18 01:50:30 2015
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.58 min Scan# 833
Delta R.T. -0.01 min
Lab File: BG017971.D
Acq: 19 Jul 2015 23:30

Instrument :
BNA_G
ClientSampleId :
TEPMW-PCB2

Tgt Ion:152 Resp: 88254
Ion Ratio Lower Upper
152 100
150 163.8 119.1 178.7
115 58.9 47.0 70.4




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 42.45 ng
RT: 5.19 min Scan# 426
Delta R.T. 0.00 min
Lab File: BG017971.D
Acq: 19 Jul 2015 23:30

Tgt Ion:112 Resp: 214742
Ion Ratio Lower Upper
112 100
64 49.3 45.4 68.0
63 26.6 25.3 37.9



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BGC
Ion 63.00 (62.70 to 63.70): BGC





#7

Phenol-d6

Concen: 27.62 ng

RT: 6.77 min Scan# 694

Delta R.T. 0.00 min

Lab File: BG017971.D

Acq: 19 Jul 2015 23:30

Instrument :

BNA_G

ClientSampleId :

TEPMW-PCB2


Tgt Ion: 99 Resp: 183057

Ion Ratio Lower Upper




99 100

42 15.4 16.2 24.4#

71 32.1 29.5 44.3



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 10.36 min Scan# 1306

Delta R.T. -0.01 min

Lab File: BG017971.D

Acq: 19 Jul 2015 23:30

Tgt Ion: 136 Resp: 397509


Ion Ratio Lower Upper

136 100



137 11.7 8.8 13.2


54 6.5 6.6 9.8#

68 4.9 5.1 7.7#



Abundance Ion 136.00 (135.70 to 136.70): BGC
Ion 137.00 (136.70 to 137.70): BGC
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC









#23
 Nitrobenzene-d5
 Concen: 62.41 ng
 RT: 8.73 min Scan# 1029
 Delta R.T. -0.01 min
 Lab File: BG017971.D
 Acq: 19 Jul 2015 23:30




Instrument :
 BNA_G
 ClientSampleId :
 TEPMW-PCB2


Tgt Ion	Ratio	Lower	Upper
82	100		
128	49.8	37.8	56.6
54	44.0	39.9	59.9



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 14.24 min Scan# 1966
 Delta R.T. -0.01 min
 Lab File: BG017971.D
 Acq: 19 Jul 2015 23:30

Tgt Ion	Ratio	Lower	Upper
164	100		
162	98.4	76.3	114.5
160	44.3	33.7	50.5









#41
 2,4,6-Tribromophenol
 Concen: 107.36 ng
 RT: 15.74 min Scan# 2221
 Delta R.T. -0.01 min
 Lab File: BG017971.D
 Acq: 19 Jul 2015 23:30

Instrument :
 BNA_G
 ClientSampleId :
 TEPMW-PCB2

Tgt Ion: 330 Resp: 272307
 Ion Ratio Lower Upper
 330 100
 332 95.4 76.8 115.2
 141 40.1 38.0 57.0




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E



#44
 2-Fluorobiphenyl
 Concen: 66.31 ng
 RT: 12.86 min Scan# 1731
 Delta R.T. 0.00 min
 Lab File: BG017971.D
 Acq: 19 Jul 2015 23:30

Tgt Ion: 172 Resp: 957993
 Ion Ratio Lower Upper
 172 100
 171 37.1 30.7 46.1
 170 25.2 19.7 29.5



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E

