

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG071915\
 Data File : BG017965.D
 Acq On : 19 Jul 2015 19:23
 Operator : TP/UM
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 20 03:43:49 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG071715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 18 01:50:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	127	0.00
2	1,4-Dioxane	0.504	0.439	12.9	118	0.00
3	Pyridine	1.347	1.225	9.1	115	0.00
4	n-Nitrosodimethylamine	0.511	0.456	10.8	118	0.00
5 S	2-Fluorophenol	1.146	1.089	5.0	123	0.00
6	Aniline	2.003	1.873	6.5	123	0.00
7 S	Phenol-d6	1.502	1.421	5.4	124	0.00
8	2-Chlorophenol	1.352	1.308	3.3	128	0.00
9	Benzaldehyde	0.920	0.846	8.0	121	0.00
10 C	Phenol	1.603	1.516	5.4	125	0.00
11	bis(2-Chloroethyl)ether	1.254	1.185	5.5	125	0.00
12	1,3-Dichlorobenzene	1.484	1.430	3.6	128	0.00
13 C	1,4-Dichlorobenzene	1.521	1.462	3.9	128	0.00
14	1,2-Dichlorobenzene	1.454	1.381	5.0	128	0.00
15	Benzyl Alcohol	1.111	1.061	4.5	126	0.00
16	2,2'-oxybis(1-Chloropropane	1.511	1.327	12.2	120	0.00
17	2-Methylphenol	1.118	1.068	4.5	126	0.00
18	Hexachloroethane	0.575	0.545	5.2	126	0.00
19 P	n-Nitroso-di-n-propylamine	1.097	0.999	8.9	120	0.00
20	3+4-Methylphenols	1.513	1.465	3.2	126	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	129	0.00
22	Acetophenone	0.432	0.401	7.2	125	0.00
23 S	Nitrobenzene-d5	0.309	0.289	6.5	124	0.00
24	Nitrobenzene	0.329	0.312	5.2	124	0.00
25	Isophorone	0.648	0.597	7.9	121	0.00
26 C	2-Nitrophenol	0.157	0.163	-3.8	141	0.00
27	2,4-Dimethylphenol	0.286	0.277	3.1	130	0.00
28	bis(2-Chloroethoxy)methane	0.361	0.338	6.4	124	0.00
29 C	2,4-Dichlorophenol	0.254	0.268	-5.5	135	0.00
30	1,2,4-Trichlorobenzene	0.302	0.293	3.0	132	0.00
31	Naphthalene	0.979	0.920	6.0	126	0.00
32	Benzoic acid	0.110	0.151	-37.3#	205#	0.01
33	4-Chloroaniline	0.436	0.419	3.9	126	0.00
34 C	Hexachlorobutadiene	0.170	0.175	-2.9	135	0.00
35	Caprolactam	0.129	0.125	3.1	130	0.01
36 C	4-Chloro-3-methylphenol	0.289	0.292	-1.0	132	0.00
37	2-Methylnaphthalene	0.703	0.671	4.6	128	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	131	0.00
39	1,2,4,5-Tetrachlorobenzene	0.536	0.527	1.7	133	0.00
40 P	Hexachlorocyclopentadiene	0.292	0.297	-1.7	140	0.00
41 S	2,4,6-Tribromophenol	0.203	0.211	-3.9	143	0.00
42 C	2,4,6-Trichlorophenol	0.349	0.365	-4.6	143	0.00
43	2,4,5-Trichlorophenol	0.363	0.385	-6.1	142	0.00
44 S	2-Fluorobiphenyl	1.154	1.087	5.8	127	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG071915\
 Data File : BG017965.D
 Acq On : 19 Jul 2015 19:23
 Operator : TP/UM
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 20 03:43:49 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG071715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 18 01:50:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.400	1.326	5.3	128	0.00
46	2-Chloronaphthalene	1.135	1.082	4.7	130	0.00
47	2-Nitroaniline	0.305	0.295	3.3	132	0.00
48	Acenaphthylene	1.892	1.795	5.1	126	0.00
49	Dimethylphthalate	1.389	1.323	4.8	127	0.00
50	2,6-Dinitrotoluene	0.315	0.319	-1.3	137	0.00
51 C	Acenaphthene	1.146	1.070	6.6	127	0.00
52	3-Nitroaniline	0.354	0.358	-1.1	132	0.00
53 P	2,4-Dinitrophenol	0.121	0.130	-7.4	152#	0.00
54	Dibenzofuran	1.654	1.548	6.4	127	0.00
55 P	4-Nitrophenol	0.280	0.279	0.4	134	0.00
56	2,4-Dinitrotoluene	0.422	0.425	-0.7	136	0.00
57	Fluorene	1.421	1.347	5.2	129	0.00
58	2,3,4,6-Tetrachlorophenol	0.313	0.334	-6.7	144	0.00
59	Diethylphthalate	1.439	1.357	5.7	127	0.00
60	4-Chlorophenyl-phenylether	0.697	0.671	3.7	130	0.00
61	4-Nitroaniline	0.415	0.400	3.6	132	0.00
62	Azobenzene	1.330	1.196	10.1	121	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	130	0.00
64	4,6-Dinitro-2-methylphenol	0.103	0.115	-11.7	154#	0.00
65 c	n-Nitrosodiphenylamine	0.601	0.585	2.7	128	0.00
66	4-Bromophenyl-phenylether	0.200	0.201	-0.5	135	0.00
67	Hexachlorobenzene	0.222	0.220	0.9	136	0.00
68	Atrazine	0.198	0.196	1.0	130	0.00
69 C	Pentachlorophenol	0.115	0.127	-10.4	148	0.00
70	Phenanthrene	1.032	0.951	7.8	125	0.00
71	Anthracene	1.004	0.963	4.1	126	0.00
72	Carbazole	0.948	0.899	5.2	126	0.00
73	Di-n-butylphthalate	1.162	1.113	4.2	125	0.00
74 C	Fluoranthene	1.129	1.063	5.8	125	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	127	0.00
76	Benzidine	0.568	0.580	-2.1	132	0.00
77	Pyrene	1.115	1.054	5.5	124	0.00
78 S	Terphenyl-d14	0.794	0.731	7.9	122	0.00
79	Butylbenzylphthalate	0.500	0.512	-2.4	128	0.00
80	Benzo(a)anthracene	1.056	1.010	4.4	125	0.00
81	3,3'-Dichlorobenzidine	0.383	0.396	-3.4	132	0.00
82	Chrysene	1.006	0.952	5.4	125	0.00
83	Bis(2-ethylhexyl)phthalate	0.688	0.682	0.9	124	0.00
84 c	Di-n-octyl phthalate	1.084	1.130	-4.2	129	0.00
85	Indeno(1,2,3-cd)pyrene	1.217	1.209	0.7	130	0.00
86 I	Perylene-d12	1.000	1.000	0.0	132	0.00
87	Benzo(b)fluoranthene	1.094	1.067	2.5	131	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG071915\
Data File : BG017965.D
Acq On : 19 Jul 2015 19:23
Operator : TP/UM
Sample : SSTDCCC040
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jul 20 03:43:49 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG071715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 18 01:50:30 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.093	0.995	9.0	120	0.00
89 C	Benzo(a)pyrene	1.020	0.979	4.0	128	0.00
90	Dibenzo(a,h)anthracene	1.057	1.029	2.6	132	0.00
91	Benzo(g,h,i)perylene	1.019	0.991	2.7	132	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0