

m/z	Abundance
43.90	798.0
207.00	242.0

Instrument :
BNA_G
ClientSampleId :
DFTPP52

m/z	Abundance
37.90	361.0
39.00	1948.0
41.10	178.0
44.00	928.0
48.90	227.0
50.00	8065.0
51.00	31704.0
52.00	1656.0
55.00	170.0
56.00	574.0
56.90	2234.0
60.90	363.0
61.90	516.0
62.90	1437.0
65.00	497.0
68.00	901.0
68.90	28696.0
70.00	266.0
70.90	150.0
73.00	342.0
74.00	3144.0
75.00	5072.0
76.00	2021.0
77.00	30904.0
78.00	2367.0
79.00	1638.0
80.00	1814.0
81.00	2434.0
82.00	617.0
83.00	515.0
85.00	355.0
86.00	980.0
86.90	390.0
90.90	440.0
92.00	706.0
93.00	3863.0
94.00	189.0
95.90	268.0
96.90	197.0
97.90	2601.0
99.00	2913.0
100.00	259.0
101.00	1100.0
103.00	542.0
103.90	841.0
105.00	1255.0
107.00	9713.0
108.00	1432.0
108.90	398.0
110.00	16976.0
111.00	3000.0
111.90	407.0
116.00	758.0
117.00	8231.0
117.90	688.0
119.90	261.0
122.00	1019.0
123.00	1241.0
124.00	570.0
124.90	492.0
127.00	32760.0
127.90	2556.0
129.00	15728.0
129.90	1036.0
131.10	195.0
133.90	413.0
134.90	1039.0
135.90	497.0
137.00	616.0
137.90	209.0
140.10	196.0
141.00	2139.0
142.00	841.0
143.00	484.0
146.10	399.0
147.00	1159.0
148.00	1799.0
149.00	629.0
151.10	482.0
153.00	779.0
154.00	646.0
155.00	1144.0
156.00	1784.0
156.90	387.0
157.90	384.0
159.00	315.0
160.00	954.0
160.90	1177.0
161.90	350.0
164.90	611.0
165.90	667.0

Instrument :
BNA_G
ClientSampleId :
DFTPP52

167.00	3848.0
168.00	2546.0
169.00	562.0
170.90	157.0
171.90	237.0
173.00	799.0
173.90	851.0
175.00	1662.0
176.00	529.0
176.90	903.0
177.90	251.0
178.90	3295.0
180.00	2550.0
181.00	847.0
181.90	276.0
184.00	403.0
184.90	1712.0
186.00	11521.0
187.00	3360.0
188.10	435.0
188.90	1073.0
189.90	379.0
190.90	340.0
192.00	908.0
193.00	1155.0
193.90	336.0
196.00	2493.0
197.90	88824.0
198.90	5647.0
199.90	368.0
201.30	266.0
202.90	521.0
204.00	3503.0
205.00	5557.0
206.00	24424.0
207.00	2904.0
207.90	758.0
209.00	475.0
210.90	1366.0
214.90	260.0
216.10	639.0
216.90	6948.0
217.90	980.0
220.90	3945.0
222.00	896.0
222.90	1450.0
224.00	14469.0
225.00	3389.0
226.00	685.0
227.00	6318.0
227.90	1013.0
228.90	1230.0
229.90	182.0
231.10	504.0
234.00	321.0
234.90	434.0
235.90	457.0
237.00	748.0
238.00	161.0
239.00	220.0
240.00	291.0
240.90	346.0
242.00	774.0
243.10	1019.0
244.00	10780.0
245.00	1440.0
245.90	2316.0
246.90	215.0
249.00	214.0
251.00	155.0
253.00	316.0
255.00	55472.0
256.00	8476.0
256.90	741.0
258.00	3379.0
258.90	439.0
260.80	157.0
265.00	1610.0
266.10	378.0
270.80	211.0
272.00	273.0
273.00	1648.0
273.90	4274.0
275.00	25120.0
276.00	3278.0
277.00	2108.0
278.00	444.0
281.00	202.0
283.00	282.0
285.00	356.0
292.90	365.0
296.00	6083.0
296.90	1077.0
298.00	164.0
300.90	235.0

Instrument :
BNA_G
ClientSampleId :
DFTPP52

302.20	159.0
303.00	748.0
304.10	431.0
307.80	182.0
313.90	237.0
315.00	584.0
316.00	565.0
322.10	189.0
323.00	2303.0
324.10	480.0
326.90	419.0
328.00	288.0
334.00	1110.0
335.00	255.0
340.90	309.0
345.90	550.0
352.00	911.0
353.00	476.0
354.10	761.0
355.10	203.0
365.00	2743.0
366.00	449.0
367.20	151.0
370.90	200.0
372.00	1051.0
373.10	400.0
383.00	286.0
390.90	226.0
402.00	461.0
403.10	791.0
404.00	551.0
421.00	470.0
422.10	442.0
423.00	3300.0
424.00	1092.0
441.00	8537.0
442.00	54104.0
443.00	10909.0
444.00	888.0

Instrument :
BNA_G
ClientSampleId :
DFTPP52

m/z	Abundance
38.00	382.0
39.00	2352.0
40.00	249.0
40.90	227.0
44.00	960.0
49.10	324.0
50.00	8112.0
51.00	31272.0
52.00	1732.0
55.90	908.0
57.00	2498.0
57.90	203.0
60.90	272.0
62.00	327.0
63.00	1538.0
63.90	154.0
65.00	509.0
69.00	28264.0
73.00	285.0
74.00	3556.0
75.00	5063.0
76.00	1518.0
77.00	34512.0
78.00	2237.0
79.00	1998.0
80.00	1423.0
81.00	2336.0
81.90	548.0
83.00	708.0
84.00	216.0
84.90	380.0
85.90	697.0
87.00	219.0
91.00	515.0
92.00	567.0
93.00	4044.0
94.10	351.0
95.90	252.0
98.00	2847.0
99.00	2391.0
99.80	291.0
101.00	1622.0
102.90	429.0
104.00	795.0
104.90	942.0
106.00	391.0
107.00	10841.0
107.90	1612.0
109.90	19096.0
111.00	3210.0
111.90	388.0
113.10	214.0
116.00	871.0
117.00	8886.0
117.90	738.0
119.70	224.0
122.00	935.0
122.90	1075.0
124.00	224.0
125.00	673.0
127.00	38368.0
128.00	3171.0
129.00	16600.0
129.90	1486.0
130.90	211.0
133.90	444.0
135.00	1290.0
136.00	310.0
137.00	947.0
139.70	259.0
140.90	2175.0
141.90	508.0
142.90	502.0
143.90	196.0
144.90	234.0
145.90	516.0
147.00	994.0
148.00	2461.0
149.00	646.0
150.00	182.0
150.80	395.0
151.40	202.0
151.90	241.0
153.00	753.0
153.90	517.0
154.90	1403.0
156.00	1765.0
157.00	374.0
157.90	369.0
159.00	246.0
159.90	638.0

Instrument :
BNA_G
ClientSampleId :
DFTPP52

160.90	1049.0
162.00	437.0
164.90	909.0
166.00	682.0
167.00	4441.0
168.00	2729.0
169.10	513.0
170.50	225.0
171.10	213.0
171.90	621.0
172.90	580.0
174.00	940.0
174.90	1557.0
176.00	652.0
176.90	560.0
177.90	407.0
178.90	3566.0
179.90	2982.0
181.00	1033.0
182.20	258.0
183.90	369.0
185.00	2037.0
186.00	15199.0
186.90	3888.0
188.10	237.0
189.00	890.0
191.20	407.0
191.90	1336.0
193.00	1373.0
194.00	298.0
196.00	2650.0
197.90	103368.0
198.90	6946.0
199.90	598.0
201.40	401.0
203.00	872.0
203.90	4355.0
205.00	6975.0
206.00	29496.0
207.00	4909.0
208.00	1036.0
208.90	347.0
210.00	491.0
211.00	1237.0
211.90	316.0
215.00	202.0
216.00	640.0
216.90	7889.0
217.90	1092.0
221.00	3978.0
222.00	1333.0
223.00	1775.0
224.00	17496.0
225.00	3891.0
226.10	842.0
227.00	7323.0
228.00	897.0
228.90	1311.0
229.90	161.0
231.00	333.0
233.90	413.0
235.00	523.0
236.00	255.0
236.90	782.0
239.00	401.0
240.00	436.0
240.90	411.0
241.90	978.0
243.00	977.0
244.00	12597.0
245.00	1676.0
246.00	2179.0
246.90	628.0
248.00	231.0
249.00	509.0
251.10	216.0
252.90	369.0
255.00	69856.0
256.00	11007.0
257.00	869.0
258.00	3390.0
258.90	858.0
263.90	189.0
265.00	1714.0
266.00	263.0
272.10	210.0
273.00	1922.0
273.90	5569.0
275.00	30688.0
276.00	4081.0
277.00	2630.0
277.90	442.0
281.80	154.0
283.00	301.0
284.10	273.0

Instrument :
BNA_G
ClientSampleId :
DFTPP52

284.90	673.0
293.00	504.0
294.00	280.0
294.80	179.0
295.90	9160.0
297.00	1207.0
302.10	179.0
303.00	1036.0
303.90	334.0
309.00	198.0
309.90	190.0
314.10	367.0
314.90	749.0
316.00	529.0
321.00	262.0
321.90	248.0
323.00	2641.0
324.00	558.0
326.90	457.0
327.80	274.0
332.20	175.0
333.10	343.0
334.00	1993.0
335.00	482.0
336.00	187.0
341.10	326.0
345.90	455.0
347.00	267.0
352.00	1037.0
353.00	567.0
354.00	880.0
358.90	180.0
365.00	3563.0
366.00	484.0
371.10	278.0
372.00	1310.0
373.00	256.0
383.10	437.0
383.90	153.0
389.90	216.0
392.00	169.0
401.90	562.0
402.90	647.0
404.00	269.0
421.00	654.0
421.90	401.0
423.00	5207.0
424.00	1165.0
425.00	224.0
441.00	12595.0
442.00	73072.0
443.00	14581.0
444.10	1098.0

Instrument :
BNA_G
ClientSampleId :
DFTPP52

m/z	Abundance
38.00	294.0
39.10	1030.0
40.10	170.0
40.90	208.0
43.90	744.0
49.00	211.0
50.00	5100.0
51.00	18096.0
52.00	1030.0
55.00	207.0
56.00	575.0
57.00	1403.0
61.00	306.0
61.90	265.0
63.00	930.0
64.90	349.0
68.90	14671.0
73.00	238.0
74.00	1840.0
75.00	2386.0
77.00	18704.0
78.00	1206.0
78.90	1392.0
80.00	1140.0
81.00	1137.0
81.90	390.0
82.90	231.0
85.00	300.0
85.90	360.0
87.00	200.0
90.90	577.0
91.90	327.0
93.00	2181.0
93.90	193.0
95.30	165.0
97.10	161.0
98.00	1703.0
99.00	1625.0
100.90	823.0
103.00	269.0
104.00	817.0
105.00	740.0
107.00	7116.0
108.00	924.0
109.90	10473.0
110.90	1963.0
111.90	390.0
115.90	399.0
117.00	5129.0
118.00	332.0
119.90	186.0
121.90	587.0
122.90	708.0
123.80	256.0
124.90	259.0
127.00	23600.0
128.00	1650.0
129.00	10291.0
130.00	695.0
134.90	699.0
135.70	354.0
137.00	546.0
139.90	211.0
141.00	1601.0
142.00	321.0
142.80	274.0
144.00	172.0
146.10	290.0
147.00	933.0
148.00	1460.0
148.90	356.0
151.10	155.0
152.90	447.0
154.00	414.0
155.00	966.0
156.00	840.0
157.20	152.0
157.90	247.0
158.90	273.0
160.00	382.0
160.90	614.0
164.90	572.0
165.90	298.0
167.00	2928.0
168.00	1409.0
168.90	339.0
170.00	195.0
171.00	205.0
171.80	356.0
172.90	518.0
174.00	686.0

Instrument :
BNA_G
ClientSampleId :
DFTPP52

175.00	1191.0
176.00	256.0
176.90	511.0
177.90	273.0
178.90	2502.0
180.00	1692.0
180.90	705.0
182.00	248.0
183.90	293.0
185.00	855.0
186.00	9267.0
187.00	2407.0
187.90	219.0
189.00	994.0
190.90	244.0
191.90	884.0
192.90	979.0
193.80	248.0
196.00	2627.0
197.90	64952.0
198.90	3775.0
200.10	197.0
203.00	678.0
203.90	2660.0
205.00	4414.0
206.00	18600.0
207.00	2679.0
208.00	667.0
209.00	208.0
210.30	461.0
210.90	1092.0
211.70	259.0
214.90	329.0
216.00	294.0
216.90	5116.0
217.90	443.0
219.10	187.0
220.90	2704.0
221.80	802.0
222.90	1417.0
224.00	10770.0
225.00	2554.0
225.90	296.0
227.00	4784.0
227.90	842.0
229.00	1011.0
230.00	160.0
231.00	374.0
234.00	289.0
234.90	351.0
235.80	246.0
236.90	404.0
239.00	298.0
240.00	249.0
241.00	202.0
242.00	776.0
243.00	751.0
244.00	8601.0
244.90	1032.0
246.00	1985.0
247.00	289.0
248.90	284.0
250.90	207.0
255.00	49088.0
256.00	6496.0
256.90	674.0
258.00	2538.0
259.00	410.0
263.90	277.0
265.00	1200.0
273.00	1999.0
273.90	3250.0
275.00	21032.0
275.90	2684.0
277.00	1742.0
278.00	349.0
283.00	290.0
285.00	261.0
291.90	157.0
292.90	446.0
294.00	177.0
296.00	5791.0
297.00	673.0
302.10	190.0
303.00	829.0
304.00	180.0
314.00	350.0
314.90	543.0
316.00	223.0
323.00	2735.0
324.00	466.0
326.90	298.0
332.00	212.0
334.00	1490.0
335.00	356.0

Instrument :
BNA_G
ClientSampleId :
DFTPP52

341.00	325.0
345.80	495.0
352.00	700.0
353.00	511.0
353.90	669.0
354.90	240.0
365.00	2935.0
365.80	294.0
371.10	238.0
372.00	1130.0
373.10	225.0
382.80	267.0
390.90	154.0
402.00	519.0
403.00	639.0
421.00	601.0
422.10	391.0
423.00	4183.0
424.10	865.0
437.10	206.0
441.00	9777.0
442.00	64384.0
443.00	12006.0
444.00	860.0


Instrument :
BNA_G
ClientSampleId :
DFTPP52

Data Path : Z:\HPCHEM1\BNA_G\Data\BG071917\
 Data File : BG027959.D
 Acq On : 19 Jul 2017 9:04
 Operator : SJ/JU
 Sample : DFTPP52
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP52

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG071717.M
 Title : SVOA CALIBRATION
 Last Update : Thu Jul 20 05:55:18 2017



AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.5	27024	PASS
68	69	0.00	2	1.3	300	PASS
69	198	0.00	100	27.9	23877	PASS
70	69	0.00	2	0.4	88	PASS
127	198	10	80	36.8	31576	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	85714	PASS
199	198	5	9	6.4	5456	PASS
275	198	10	60	29.9	25613	PASS
365	198	1	100	3.6	3080	PASS
441	442	0.01	100	16.1	10303	PASS
442	198	50	100	74.5	63853	PASS
443	442	15	24	19.6	12498	PASS