

Data Path : Z:\HPCHEM1\BNA_G\Data\BG072015\
 Data File : BG017998.D
 Acq On : 20 Jul 2015 20:37
 Operator : TP/UM
 Sample : G3024-01
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 UST-1(0-2)

Manual Integrations
 APPROVED

apatel
 7/22/2015 9:43:39 AM

Quant Time: Jul 21 01:47:56 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG071715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 18 01:50:30 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	111364	20.00	ng	0.00
21) Naphthalene-d8	10.37	136	465840	20.00	ng	0.00
38) Acenaphthene-d10	14.25	164	258236	20.00	ng	0.00
63) Phenanthrene-d10	17.00	188	493921	20.00	ng	0.00
75) Chrysene-d12	21.20	240	502749	20.00	ng	0.00
86) Perylene-d12	23.42	264	504132	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.20	112	702433	110.03	ng	0.01
7) Phenol-d6	6.78	99	902086	107.86	ng	0.01
23) Nitrobenzene-d5	8.74	82	536170	74.55	ng	0.00
41) 2,4,6-Tribromophenol	15.74	330	315713	120.67	ng	0.00
44) 2-Fluorobiphenyl	12.86	172	970928	65.16	ng	0.00
78) Terphenyl-d14	19.65	244	1174849	58.86	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	13.71	163	393318	21.93	ng	100
70) Phenanthrene	17.04	178	152895	6.00	ng	99
74) Fluoranthene	19.08	202	252204	9.04	ng	95
77) Pyrene	19.43	202	228397	8.15	ng	94
80) Benzo(a)anthracene	21.19	228	135491	5.10	ng	98
82) Chrysene	21.23	228	120781m	4.78	ng	
87) Benzo(b)fluoranthene	22.75	252	171810m	6.23	ng	
89) Benzo(a)pyrene	23.32	252	113840	4.43	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\Data\BG072015\
Data File : BG017998.D
Acq On : 20 Jul 2015 20:37
Operator : TP/UM
Sample : G3024-01
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
UST-1(0-2)

Manual Integrations
APPROVED

apatel
7/22/2015 9:43:39 AM

Quant Time: Jul 21 01:47:56 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG071715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 18 01:50:30 2015
Response via : Initial Calibration

