

Data Path : Z:\HPCHEM1\BNA G\Data\BG072017\
 Data File : BG027991.D
 Acq On : 20 Jul 2017 10:08
 Operator : SJ/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02065

Quant Time: Jul 20 18:48:59 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG071717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 20 05:55:18 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.440	0.419	4.8	85	0.00
3 S	1,4-Dioxane-d8	0.363	0.389	-7.2	88	0.00
4	Benzaldehyde	1.019	0.988	3.0	80	0.00
5 S	Phenol-d5	1.644	1.493	9.2	81	0.00
6	Phenol	1.595	1.454	8.8	79	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.948	0.883	6.9	79	0.00
8	Bis(2-Chloroethyl)ether	1.150	1.057	8.1	78	0.00
9 S	2-Chlorophenol-d4	1.291	1.253	2.9	81	0.00
10	2-Chlorophenol	1.200	1.121	6.6	80	0.00
11	2-Methylphenol	1.171	1.047	10.6	79	0.00
12	2,2'-oxybis(1-Chloropropane	2.168	1.911	11.9	74	0.00
13 S	4-Methylphenol-d8	1.360	1.219	10.4	78	0.00
14	Acetophenone	1.918	1.762	8.1	80	0.00
15 P	N-Nitroso-di-n-propylamine	0.983	0.871	11.4	77	0.00
16	4-Methylphenol	1.283	1.139	11.2	76	0.00
17	Hexachloroethane	0.483	0.484	-0.2	86	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	79	0.00
19 S	Nitrobenzene-d5	0.147	0.145	1.4	78	0.00
20	Nitrobenzene	0.340	0.343	-0.9	79	0.00
21	Isophorone	0.620	0.599	3.4	76	0.00
22 S	2-Nitrophenol-d4	0.177	0.181	-2.3	81	0.00
23 C	2-Nitrophenol	0.170	0.173	-1.8	81	0.00
24	2,4-Dimethylphenol	0.357	0.356	0.3	79	0.00
25	Bis(2-Chloroethoxy)methane	0.372	0.361	3.0	77	0.00
26 S	2,4-Dichlorophenol-d3	0.346	0.348	-0.6	80	0.00
27 C	2,4-Dichlorophenol	0.335	0.339	-1.2	79	0.00
28	Naphthalene	0.948	0.956	-0.8	79	0.00
29 S	4-Chloroaniline-d4	0.313	0.373	-19.2	103	0.00
30	4-Chloroaniline	0.290	0.342	-17.9	103	0.00
31 C	Hexachlorobutadiene	0.268	0.293	-9.3	87	0.00
32	Caprolactam	0.100	0.102	-2.0	80	0.00
33 C	4-Chloro-3-methylphenol	0.320	0.317	0.9	78	0.00
34	2-Methylnaphthalene	0.766	0.772	-0.8	78	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	78	0.00
36	1,2,4,5-Tetrachlorobenzene	0.752	0.800	-6.4	82	0.00
37	Hexachlorocyclopentadiene	0.445	0.406	8.8	80	0.00
38 C	2,4,6-Trichlorophenol	0.454	0.464	-2.2	78	0.00
39	2,4,5-Trichlorophenol	0.488	0.483	1.0	77	0.00
40	1,1'-Biphenyl	1.528	1.545	-1.1	76	0.00
41	2-Chloronaphthalene	1.208	1.221	-1.1	78	0.00
42	2-Nitroaniline	0.345	0.348	-0.9	76	0.00
43 S	Dimethylphthalate-d6	1.734	1.803	-4.0	80	0.00
44	Dimethylphthalate	1.579	1.645	-4.2	79	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG072017\
 Data File : BG027991.D
 Acq On : 20 Jul 2017 10:08
 Operator : SJ/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02065

Quant Time: Jul 20 18:48:59 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG071717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 20 05:55:18 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.319	0.336	-5.3	81	0.00
46 S	Acenaphthylene-d8	1.971	2.018	-2.4	79	0.00
47	Acenaphthylene	1.936	1.999	-3.3	78	0.00
48	3-Nitroaniline	0.278	0.310	-11.5	88	0.00
49 C	Acenaphthene	1.310	1.314	-0.3	78	0.00
50	2,4-Dinitrophenol	0.184	0.167	9.2	83	0.00
51 S	4-Nitrophenol-d4	0.251	0.280	-11.6	91	0.00
52	4-Nitrophenol	0.202	0.236	-16.8	90	0.00
53	Dibenzofuran	1.937	1.986	-2.5	78	0.00
54	2,4-Dinitrotoluene	0.466	0.515	-10.5	80	0.00
55	2,3,4,6-Tetrachlorophenol	0.465	0.520	-11.8	83	0.00
56	Diethylphthalate	1.646	1.714	-4.1	81	0.00
57 S	Fluorene-d10	1.634	1.718	-5.1	80	0.00
58	Fluorene	1.665	1.742	-4.6	79	0.00
59	4-Chlorophenyl-phenylether	0.923	0.963	-4.3	81	0.00
60	4-Nitroaniline	0.329	0.383	-16.4	87	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	85	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.139	0.130	6.5	86	0.00
63	4,6-Dinitro-2-methylphenol	0.133	0.131	1.5	94	0.00
64	N-Nitrosodiphenylamine	0.539	0.517	4.1	81	0.00
65	4-Bromophenyl-phenylether	0.244	0.238	2.5	85	0.00
66	Hexachlorobenzene	0.256	0.255	0.4	86	0.00
67	Atrazine	0.237	0.245	-3.4	89	0.00
68 C	Pentachlorophenol	0.139	0.128	7.9	87	0.00
69	Phenanthrene	1.025	1.042	-1.7	85	0.00
70 S	Anthracene-d10	0.964	0.981	-1.8	85	0.00
71	Anthracene	1.051	1.090	-3.7	86	0.00
72	Carbazole	0.866	0.942	-8.8	90	0.00
73	Di-n-butylphthalate	0.950	1.031	-8.5	88	0.00
74 C	Fluoranthene	1.255	1.442	-14.9	93	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
76 S	Pyrene-d10	0.899	0.902	-0.3	95	0.00
77	Pyrene	1.060	1.059	0.1	94	0.00
78	Butylbenzylphthalate	0.340	0.347	-2.1	98	0.00
79	3,3'-Dichlorobenzidine	0.325	0.349	-7.4	109	0.00
80	Benzo(a)anthracene	1.082	1.120	-3.5	97	0.00
81	Bis(2-ethylhexyl)phthalate	0.484	0.494	-2.1	97	0.00
82	Chrysene	0.985	1.007	-2.2	98	0.00
83 I	Perylene-d12	1.000	1.000	0.0	101	0.00
84	Di-n-octyl phthalate	0.928	0.889	4.2	97	0.00
85	Benzo(b)fluoranthene	1.141	1.149	-0.7	99	0.00
86	Benzo(k)fluoranthene	1.119	1.145	-2.3	102	0.00
87 S	Benzo(a)pyrene-d12	0.934	0.953	-2.0	100	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG072017\
Data File : BG027991.D
Acq On : 20 Jul 2017 10:08
Operator : SJ/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02065

Quant Time: Jul 20 18:48:59 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG071717.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jul 20 05:55:18 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.103	1.130	-2.4	101	0.00
89	Indeno(1,2,3-cd)pyrene	1.251	1.275	-1.9	100	0.00
90	Dibenzo(a,h)anthracene	1.050	1.072	-2.1	98	0.00
91	Benzo(a,h,i)perylene	1.032	1.061	-2.8	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0