

Data Path : Z:\HPCHEM1\BNA G\Data\BG072017\
 Data File : BG028015.D
 Acq On : 21 Jul 2017 2:03
 Operator : SJ/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02067

Quant Time: Jul 21 04:22:48 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG071717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 21 02:24:12 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.440	0.379	13.9	76	0.00
3 S	1,4-Dioxane-d8	0.363	0.329	9.4	74	0.00
4	Benzaldehyde	1.019	0.952	6.6	77	0.00
5 S	Phenol-d5	1.644	1.505	8.5	81	0.00
6	Phenol	1.595	1.474	7.6	80	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.948	0.857	9.6	76	0.00
8	Bis(2-Chloroethyl)ether	1.150	1.084	5.7	79	0.00
9 S	2-Chlorophenol-d4	1.291	1.303	-0.9	84	0.00
10	2-Chlorophenol	1.200	1.215	-1.3	86	0.00
11	2-Methylphenol	1.171	1.139	2.7	86	0.00
12	2,2'-oxybis(1-Chloropropane	2.168	1.764	18.6	68	0.00
13 S	4-Methylphenol-d8	1.360	1.341	1.4	85	0.00
14	Acetophenone	1.918	1.898	1.0	85	0.00
15 P	N-Nitroso-di-n-propylamine	0.983	0.948	3.6	83	0.00
16	4-Methylphenol	1.283	1.231	4.1	81	0.00
17	Hexachloroethane	0.483	0.487	-0.8	86	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
19 S	Nitrobenzene-d5	0.147	0.148	-0.7	88	0.00
20	Nitrobenzene	0.340	0.336	1.2	85	0.00
21	Isophorone	0.620	0.620	0.0	86	0.00
22 S	2-Nitrophenol-d4	0.177	0.192	-8.5	93	0.00
23 C	2-Nitrophenol	0.170	0.171	-0.6	87	0.00
24	2,4-Dimethylphenol	0.357	0.355	0.6	86	0.00
25	Bis(2-Chloroethoxy)methane	0.372	0.356	4.3	83	0.00
26 S	2,4-Dichlorophenol-d3	0.346	0.371	-7.2	93	0.00
27 C	2,4-Dichlorophenol	0.335	0.352	-5.1	90	0.00
28	Naphthalene	0.948	0.939	0.9	85	0.00
29 S	4-Chloroaniline-d4	0.313	0.392	-25.2#	118	0.00
30	4-Chloroaniline	0.290	0.359	-23.8#	118	0.00
31 C	Hexachlorobutadiene	0.268	0.288	-7.5	93	0.00
32	Caprolactam	0.100	0.097	3.0	84	0.00
33 C	4-Chloro-3-methylphenol	0.320	0.330	-3.1	89	0.00
34	2-Methylnaphthalene	0.766	0.787	-2.7	87	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	90	0.00
36	1,2,4,5-Tetrachlorobenzene	0.752	0.807	-7.3	96	0.00
37	Hexachlorocyclopentadiene	0.445	0.331	25.6#	76	0.00
38 C	2,4,6-Trichlorophenol	0.454	0.473	-4.2	92	0.00
39	2,4,5-Trichlorophenol	0.488	0.510	-4.5	94	0.00
40	1,1'-Biphenyl	1.528	1.532	-0.3	88	0.00
41	2-Chloronaphthalene	1.208	1.213	-0.4	90	0.00
42	2-Nitroaniline	0.345	0.329	4.6	83	0.00
43 S	Dimethylphthalate-d6	1.734	1.739	-0.3	89	0.00
44	Dimethylphthalate	1.579	1.602	-1.5	90	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG072017\
 Data File : BG028015.D
 Acq On : 21 Jul 2017 2:03
 Operator : SJ/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02067

Quant Time: Jul 21 04:22:48 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG071717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 21 02:24:12 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.319	0.323	-1.3	90	0.00
46 S	Acenaphthylene-d8	1.971	2.024	-2.7	92	0.00
47	Acenaphthylene	1.936	1.955	-1.0	89	0.00
48	3-Nitroaniline	0.278	0.283	-1.8	93	0.00
49 C	Acenaphthene	1.310	1.290	1.5	89	0.00
50	2,4-Dinitrophenol	0.184	0.158	14.1	91	0.00
51 S	4-Nitrophenol-d4	0.251	0.241	4.0	91	0.00
52	4-Nitrophenol	0.202	0.208	-3.0	92	0.00
53	Dibenzofuran	1.937	1.947	-0.5	89	0.00
54	2,4-Dinitrotoluene	0.466	0.479	-2.8	86	0.00
55	2,3,4,6-Tetrachlorophenol	0.465	0.484	-4.1	90	0.00
56	Diethylphthalate	1.646	1.574	4.4	86	0.00
57 S	Fluorene-d10	1.634	1.652	-1.1	90	0.00
58	Fluorene	1.665	1.664	0.1	88	0.00
59	4-Chlorophenyl-phenylether	0.923	0.953	-3.3	92	0.00
60	4-Nitroaniline	0.329	0.348	-5.8	92	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	90	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.139	0.128	7.9	90	0.00
63	4,6-Dinitro-2-methylphenol	0.133	0.126	5.3	95	0.00
64	N-Nitrosodiphenylamine	0.539	0.535	0.7	89	0.00
65	4-Bromophenyl-phenylether	0.244	0.253	-3.7	95	0.00
66	Hexachlorobenzene	0.256	0.260	-1.6	93	0.00
67	Atrazine	0.237	0.238	-0.4	92	0.00
68 C	Pentachlorophenol	0.139	0.130	6.5	93	0.00
69	Phenanthrene	1.025	1.023	0.2	88	0.00
70 S	Anthracene-d10	0.964	0.993	-3.0	91	0.00
71	Anthracene	1.051	1.065	-1.3	89	0.00
72	Carbazole	0.866	0.887	-2.4	90	0.00
73	Di-n-butylphthalate	0.950	0.995	-4.7	90	0.00
74 C	Fluoranthene	1.255	1.383	-10.2	95	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
76 S	Pyrene-d10	0.899	0.919	-2.2	96	0.00
77	Pyrene	1.060	1.077	-1.6	94	0.00
78	Butylbenzylphthalate	0.340	0.349	-2.6	98	0.00
79	3,3'-Dichlorobenzidine	0.325	0.350	-7.7	108	0.00
80	Benzo(a)anthracene	1.082	1.131	-4.5	98	0.00
81	Bis(2-ethylhexyl)phthalate	0.484	0.507	-4.8	99	0.00
82	Chrysene	0.985	1.001	-1.6	97	0.00
83 I	Perylene-d12	1.000	1.000	0.0	102	0.00
84	Di-n-octyl phthalate	0.928	0.886	4.5	98	0.00
85	Benzo(b)fluoranthene	1.141	1.173	-2.8	102	0.00
86	Benzo(k)fluoranthene	1.119	1.110	0.8	100	0.00
87 S	Benzo(a)pyrene-d12	0.934	0.965	-3.3	103	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG072017\
Data File : BG028015.D
Acq On : 21 Jul 2017 2:03
Operator : SJ/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02067

Quant Time: Jul 21 04:22:48 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG071717.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jul 21 02:24:12 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.103	1.122	-1.7	102	0.00
89	Indeno(1,2,3-cd)pyrene	1.251	1.277	-2.1	102	0.00
90	Dibenzo(a,h)anthracene	1.050	1.063	-1.2	98	0.00
91	Benzo(a,h,i)perylene	1.032	1.060	-2.7	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0