



Benzoic acid



Response = $2.193 \times 10^{-1} \times \text{Amt} - 5.969 \times 10^{-2}$
 Coef of Det (r^2) = 0.995526 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG072121.M
 Calibration Table Last Updated: Wed Jul 21 16:57:05 2021

Hexachlorocyclopentadiene

Response Ratio




$$\text{Response} = 3.132\text{e-}001 * \text{Amt} - 6.927\text{e-}002$$

Coef of Det (r^2) = 0.992880 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG072121.M


Calibration Table Last Updated: Wed Jul 21 16:57:05 2021

2,4-Dinitrophenol



Response = 1.945e-001 * Amt - 8.766e-002
 Coef of Det (r^2) = 0.995679 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG072121.M
 Calibration Table Last Updated: Wed Jul 21 16:57:05 2021

Pentachlorophenol



$\text{Response} = 1.492\text{e-}001 * \text{Amt} - 4.881\text{e-}002$
 Coef of Det (r^2) = 0.997524 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG072121.M
 Calibration Table Last Updated: Wed Jul 21 16:57:05 2021