

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072122\
 Data File : BG054341.D
 Acq On : 21 Jul 2022 11:33
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 22 02:36:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 15 15:54:11 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.381	0.365	4.2	102	0.00
3	Pyridine	0.967	1.065	-10.1	107	0.00
4	n-Nitrosodimethylamine	0.541	0.589	-8.9	107	0.00
5 S	2-Fluorophenol	0.958	1.002	-4.6	107	0.00
6	Aniline	1.515	1.609	-6.2	107	0.00
7 S	Phenol-d6	1.313	1.398	-6.5	108	0.00
8	2-Chlorophenol	1.101	1.143	-3.8	106	0.00
9	Benzaldehyde	0.737	0.734	0.4	115	0.00
10 C	Phenol	1.304	1.398	-7.2	110	0.00
11	bis(2-Chloroethyl)ether	0.951	1.007	-5.9	110	0.00
12	1,3-Dichlorobenzene	1.341	1.374	-2.5	105	0.00
13 C	1,4-Dichlorobenzene	1.385	1.404	-1.4	103	0.00
14	1,2-Dichlorobenzene	1.306	1.335	-2.2	105	0.00
15	Benzyl Alcohol	1.080	1.150	-6.5	105	0.00
16	2,2'-oxybis(1-Chloropropane	1.198	1.331	-11.1	112	0.00
17	2-Methylphenol	0.940	0.994	-5.7	106	0.00
18	Hexachloroethane	0.459	0.488	-6.3	108	0.00
19 P	n-Nitroso-di-n-propylamine	0.877	0.920	-4.9	104	0.00
20	3+4-Methylphenols	1.324	1.408	-6.3	107	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
22	Acetophenone	0.481	0.493	-2.5	105	0.00
23 S	Nitrobenzene-d5	0.367	0.380	-3.5	105	0.00
24	Nitrobenzene	0.359	0.367	-2.2	106	0.00
25	Isophorone	0.639	0.671	-5.0	107	0.00
26 C	2-Nitrophenol	0.181	0.188	-3.9	104	0.00
27	2,4-Dimethylphenol	0.250	0.267	-6.8	110	0.00
28	bis(2-Chloroethoxy)methane	0.322	0.337	-4.7	106	0.00
29 C	2,4-Dichlorophenol	0.377	0.396	-5.0	107	0.00
30	1,2,4-Trichlorobenzene	0.467	0.474	-1.5	104	0.00
31	Naphthalene	1.004	1.027	-2.3	105	0.00
32	Benzoic acid	0.080	0.141	-76.2#	160#	-0.01
33	4-Chloroaniline	0.408	0.426	-4.4	108	0.00
34 C	Hexachlorobutadiene	0.413	0.409	1.0	101	0.00
35	Caprolactam	0.095	0.105	-10.5	119	0.00
36 C	4-Chloro-3-methylphenol	0.343	0.353	-2.9	108	0.00
37	2-Methylnaphthalene	0.771	0.786	-1.9	106	0.00
38	1-Methylnaphthalene	0.751	0.756	-0.7	106	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
40	1,2,4,5-Tetrachlorobenzene	0.876	0.884	-0.9	102	0.00
41 P	Hexachlorocyclopentadiene	0.206	0.244	-18.4	109	0.00
42 S	2,4,6-Tribromophenol	0.420	0.443	-5.5	109	0.00
43 C	2,4,6-Trichlorophenol	0.477	0.514	-7.8	107	0.00
44	2,4,5-Trichlorophenol	0.539	0.578	-7.2	112	0.00
45 S	2-Fluorobiphenyl	1.418	1.454	-2.5	104	0.00
46	1,1'-Biphenyl	1.368	1.413	-3.3	107	0.00
47	2-Chloronaphthalene	1.145	1.187	-3.7	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072122\
 Data File : BG054341.D
 Acq On : 21 Jul 2022 11:33
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 22 02:36:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 15 15:54:11 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.299	0.327	-9.4	110	0.00
49	Acenaphthylene	1.701	1.781	-4.7	107	0.00
50	Dimethylphthalate	1.562	1.611	-3.1	107	0.00
51	2,6-Dinitrotoluene	0.343	0.361	-5.2	109	0.00
52 C	Acenaphthene	1.030	1.064	-3.3	107	0.00
53	3-Nitroaniline	0.287	0.313	-9.1	114	0.00
54 P	2,4-Dinitrophenol	0.166	0.184	-10.8	113	0.00
55	Dibenzofuran	1.833	1.885	-2.8	107	0.00
56 P	4-Nitrophenol	0.191	0.203	-6.3	108	0.02
57	2,4-Dinitrotoluene	0.491	0.518	-5.5	108	0.00
58	Fluorene	1.507	1.553	-3.1	107	0.00
59	2,3,4,6-Tetrachlorophenol	0.525	0.562	-7.0	108	0.00
60	Diethylphthalate	1.499	1.569	-4.7	109	0.00
61	4-Chlorophenyl-phenylether	0.989	1.002	-1.3	105	0.00
62	4-Nitroaniline	0.301	0.333	-10.6	114	0.00
63	Azobenzene	1.083	1.147	-5.9	109	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
65	4,6-Dinitro-2-methylphenol	0.124	0.129	-4.0	106	0.00
66 c	n-Nitrosodiphenylamine	0.499	0.514	-3.0	108	0.00
67	4-Bromophenyl-phenylether	0.281	0.286	-1.8	106	0.00
68	Hexachlorobenzene	0.321	0.320	0.3	104	0.00
69	Atrazine	0.222	0.226	-1.8	108	0.00
70 C	Pentachlorophenol	0.126	0.131	-4.0	106	0.00
71	Phenanthrene	0.996	1.020	-2.4	108	0.00
72	Anthracene	0.977	1.000	-2.4	108	0.00
73	Carbazole	0.805	0.843	-4.7	111	0.00
74	Di-n-butylphthalate	0.950	0.987	-3.9	110	0.00
75 C	Fluoranthene	1.358	1.344	1.0	105	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
77	Benzydine	0.362	0.398	-9.9	113	0.00
78	Pyrene	1.148	1.234	-7.5	105	0.00
79 S	Terphenyl-d14	1.008	1.074	-6.5	105	0.00
80	Butylbenzylphthalate	0.358	0.386	-7.8	108	0.00
81	Benzo(a)anthracene	1.270	1.312	-3.3	104	0.01
82	3,3'-Dichlorobenzidine	0.395	0.406	-2.8	101	0.00
83	Chrysene	1.199	1.242	-3.6	104	0.00
84	Bis(2-ethylhexyl)phthalate	0.506	0.538	-6.3	106	0.00
85 c	Di-n-octyl phthalate	0.863	0.906	-5.0	103	0.00
86 I	Perylene-d12	1.000	1.000	0.0	100	0.01
87	Indeno(1,2,3-cd)pyrene	1.521	1.557	-2.4	101	0.03
88	Benzo(b)fluoranthene	1.183	1.221	-3.2	100	0.01
89	Benzo(k)fluoranthene	1.177	1.215	-3.2	102	0.01
90 C	Benzo(a)pyrene	1.010	1.039	-2.9	100	0.02
91	Dibenzo(a,h)anthracene	1.251	1.287	-2.9	101	0.03
92	Benzo(g,h,i)perylene	1.235	1.269	-2.8	101	0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072122\
Data File : BG054341.D
Acq On : 21 Jul 2022 11:33
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jul 22 02:36:12 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 15 15:54:11 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0