



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072221\
 Data File : BG049403.D
 Acq On : 22 Jul 2021 7:36
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG072121.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jul 21 16:57:05 2021






AutoFind: Scans 2525, 2526, 2527; Background Corrected with Scan 2519

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.2	24357	PASS
68	69	0.00	2	0.8	227	PASS
69	198	0.00	100	48.4	26707	PASS
70	69	0.00	2	0.5	132	PASS
127	198	10	80	53.2	29328	PASS
197	198	0.00	2	0.7	397	PASS
198	198	100	100	100.0	55133	PASS
199	198	5	9	6.9	3812	PASS
275	198	10	60	31.3	17244	PASS
365	198	1	100	4.9	2715	PASS
441	198	0.01	100	13.2	7275	PASS
442	442	100	100	100.0	47480	PASS
443	442	15	24	20.1	9561	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072221\
Data File : BG049403.D
Acq On : 22 Jul 2021 7:36
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Jul 22 10:49:33 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG072121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 21 16:57:05 2021
Response via : Initial Calibration



TIC: BG049403.D\data.ms

(70) Pentachlorophenol (C)

17.102min (-0.000) 30467.23 ng




response 78407

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.20	57.83
264.00	61.30	63.19
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072221\
Data File : BG049403.D
Acq On : 22 Jul 2021 7:36
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Jul 22 10:49:33 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG072121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 21 16:57:05 2021
Response via : Initial Calibration



TIC: BG049403.D\data.ms

(77) Benzidine

19.680min (-0.001) 0.00 ng

response 407619

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.40	14.45
183.00	11.70	12.01
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
7/22/2021	BNA_G	BG049403.D
Compound Name	Response	Retention Time
DDT	198173	20.91
DDD	1330	20.47
DDE	100	19.96
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
1430	199603	0.72

Instrument :
BNA_G
ClientSampleId :
DFTPP