

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072222\
 Data File : BG054357.D
 Acq On : 22 Jul 2022 12:51
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 23 00:18:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 15 15:54:11 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	0.00
2	1,4-Dioxane	40.000	36.811	8.0	94	0.00
3	Pyridine	40.000	41.927	-4.8	98	0.00
4	n-Nitrosodimethylamine	40.000	44.757	-11.9	105	0.00
5 S	2-Fluorophenol	80.000	82.371	-3.0	101	0.00
6	Aniline	40.000	40.201	-0.5	98	0.00
7 S	Phenol-d6	80.000	82.588	-3.2	100	0.00
8	2-Chlorophenol	40.000	39.878	0.3	98	0.00
9	Benzaldehyde	40.000	41.792	-4.5	104	0.00
10 C	Phenol	40.000	40.727	-1.8	100	0.00
11	bis(2-Chloroethyl)ether	40.000	39.408	1.5	98	0.00
12	1,3-Dichlorobenzene	40.000	40.176	-0.4	99	0.00
13 C	1,4-Dichlorobenzene	40.000	39.696	0.8	97	0.00
14	1,2-Dichlorobenzene	40.000	40.249	-0.6	99	0.00
15	Benzyl Alcohol	40.000	40.251	-0.6	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.748	3.1	93	0.00
17	2-Methylphenol	40.000	40.465	-1.2	98	0.00
18	Hexachloroethane	40.000	40.904	-2.3	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.065	4.8	90	0.00
20	3+4-Methylphenols	40.000	39.047	2.4	95	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	95	0.00
22	Acetophenone	40.000	41.442	-3.6	95	0.00
23 S	Nitrobenzene-d5	80.000	84.048	-5.1	96	0.00
24	Nitrobenzene	40.000	42.578	-6.4	99	0.00
25	Isophorone	40.000	40.677	-1.7	93	0.00
26 C	2-Nitrophenol	40.000	41.660	-4.1	94	0.00
27	2,4-Dimethylphenol	40.000	42.251	-5.6	98	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.082	-0.2	91	0.00
29 C	2,4-Dichlorophenol	40.000	41.260	-3.1	95	0.00
30	1,2,4-Trichlorobenzene	40.000	41.796	-4.5	97	0.00
31	Naphthalene	40.000	41.249	-3.1	95	0.00
32	Benzoic acid	40.000	47.354	-18.4	120	0.00
33	4-Chloroaniline	40.000	42.258	-5.6	99	0.00
34 C	Hexachlorobutadiene	40.000	42.214	-5.5	97	0.00
35	Caprolactam	40.000	43.113	-7.8	105	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.490	-3.7	98	0.00
37	2-Methylnaphthalene	40.000	40.471	-1.2	95	0.00
38	1-Methylnaphthalene	40.000	40.709	-1.8	96	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	98	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.138	-2.8	95	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.368	6.6	94	0.00
42 S	2,4,6-Tribromophenol	80.000	88.468	-10.6	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.232	-8.1	98	0.00
44	2,4,5-Trichlorophenol	40.000	42.138	-5.3	100	0.00
45 S	2-Fluorobiphenyl	80.000	82.458	-3.1	96	0.00
46	1,1'-Biphenyl	40.000	40.251	-0.6	95	0.00
47	2-Chloronaphthalene	40.000	41.485	-3.7	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072222\
 Data File : BG054357.D
 Acq On : 22 Jul 2022 12:51
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 23 00:18:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 15 15:54:11 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.715	-9.3	100	0.00
49	Acenaphthylene	40.000	41.504	-3.8	97	0.00
50	Dimethylphthalate	40.000	41.686	-4.2	99	0.00
51	2,6-Dinitrotoluene	40.000	40.769	-1.9	97	0.00
52 C	Acenaphthene	40.000	41.774	-4.4	99	0.00
53	3-Nitroaniline	40.000	42.929	-7.3	102	0.00
54 P	2,4-Dinitrophenol	40.000	39.173	2.1	99	0.00
55	Dibenzofuran	40.000	41.254	-3.1	98	0.00
56 P	4-Nitrophenol	40.000	42.898	-7.2	100	0.02
57	2,4-Dinitrotoluene	40.000	42.625	-6.6	99	0.00
58	Fluorene	40.000	41.927	-4.8	100	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.485	-8.7	100	0.00
60	Diethylphthalate	40.000	42.699	-6.7	101	0.00
61	4-Chlorophenyl-phenylether	40.000	42.375	-5.9	100	0.00
62	4-Nitroaniline	40.000	44.037	-10.1	104	0.00
63	Azobenzene	40.000	41.671	-4.2	98	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.615	-4.0	101	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.915	-2.3	102	0.00
67	4-Bromophenyl-phenylether	40.000	40.052	-0.1	100	0.00
68	Hexachlorobenzene	40.000	40.246	-0.6	100	0.00
69	Atrazine	40.000	41.645	-4.1	105	0.00
70 C	Pentachlorophenol	40.000	38.925	2.7	99	0.00
71	Phenanthrene	40.000	40.584	-1.5	102	0.00
72	Anthracene	40.000	40.961	-2.4	103	0.00
73	Carbazole	40.000	41.985	-5.0	106	0.00
74	Di-n-butylphthalate	40.000	42.447	-6.1	108	0.00
75 C	Fluoranthene	40.000	41.842	-4.6	106	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	110	0.00
77	Benzydine	40.000	39.763	0.6	107	0.00
78	Pyrene	40.000	41.135	-2.8	106	0.00
79 S	Terphenyl-d14	80.000	83.612	-4.5	108	0.00
80	Butylbenzylphthalate	40.000	42.015	-5.0	110	0.00
81	Benzo(a)anthracene	40.000	40.752	-1.9	108	0.00
82	3,3'-Dichlorobenzidine	40.000	40.922	-2.3	106	0.00
83	Chrysene	40.000	41.092	-2.7	108	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.032	-2.6	108	0.00
85 c	Di-n-octyl phthalate	40.000	40.743	-1.9	105	0.00
86 I	Perylene-d12	20.000	20.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.930	-2.3	106	0.00
88	Benzo(b)fluoranthene	40.000	40.733	-1.8	104	0.00
89	Benzo(k)fluoranthene	40.000	41.252	-3.1	108	0.00
90 C	Benzo(a)pyrene	40.000	41.078	-2.7	105	0.00
91	Dibenzo(a,h)anthracene	40.000	40.680	-1.7	105	0.00
92	Benzo(g,h,i)perylene	40.000	40.921	-2.3	106	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072222\
Data File : BG054357.D
Acq On : 22 Jul 2022 12:51
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jul 23 00:18:55 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 15 15:54:11 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0