




Data Path : Z:\HPCHEM1\BNA_G\DATA\BG072316\
Data File : BG023236.D
Acq On : 23 Jul 2016 4:50
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 23 07:08:23 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG070816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 19 15:46:37 2016
Response via : Initial Calibration



TIC: BG023236.D

(76) Benzidine




19.789min (-19.789) 0.00ng

response 0

Ion	Exp%	Act%
184.00	100	0.00
185.00	15.80	0.00#
183.00	13.40	0.00#
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG072316\
Data File : BG023236.D
Acq On : 23 Jul 2016 4:50
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 23 07:08:23 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG070816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 19 15:46:37 2016
Response via : Initial Calibration



TIC: BG023236.D

(69) Pentachlorophenol (C)

17.221min (+0.017) 280206.06ng

response 159151

Ion	Exp%	Act%
265.70	100	100
268.00	64.90	69.95
264.00	63.20	65.66
0.00	0.00	0.00