

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072321\
 Data File : BG049427.D
 Acq On : 23 Jul 2021 14:33
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 23 16:11:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG072121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 21 16:57:05 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	91	0.00
2	1,4-Dioxane	40.000	40.381	-1.0	95	-0.01
3	Pyridine	40.000	43.090	-7.7	91	0.00
4	n-Nitrosodimethylamine	40.000	46.612	-16.5	101	0.00
5 S	2-Fluorophenol	80.000	85.141	-6.4	93	0.00
6	Aniline	40.000	40.970	-2.4	93	0.00
7 S	Phenol-d6	80.000	83.581	-4.5	93	0.00
8	2-Chlorophenol	40.000	42.374	-5.9	93	0.00
9	Benzaldehyde	40.000	42.642	-6.6	94	0.00
10 C	Phenol	40.000	43.666	-9.2	97	0.00
11	bis(2-Chloroethyl)ether	40.000	40.797	-2.0	93	0.00
12	1,3-Dichlorobenzene	40.000	40.661	-1.7	92	0.00
13 C	1,4-Dichlorobenzene	40.000	42.319	-5.8	94	0.00
14	1,2-Dichlorobenzene	40.000	41.164	-2.9	93	0.00
15	Benzyl Alcohol	40.000	42.888	-7.2	96	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.685	-4.2	96	0.00
17	2-Methylphenol	40.000	42.318	-5.8	94	0.00
18	Hexachloroethane	40.000	41.052	-2.6	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.077	-0.2	87	0.00
20	3+4-Methylphenols	40.000	41.472	-3.7	92	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	92	0.00
22	Acetophenone	40.000	40.655	-1.6	94	0.00
23 S	Nitrobenzene-d5	80.000	81.541	-1.9	94	0.00
24	Nitrobenzene	40.000	41.280	-3.2	94	0.00
25	Isophorone	40.000	41.381	-3.5	95	0.00
26 C	2-Nitrophenol	40.000	42.244	-5.6	94	0.00
27	2,4-Dimethylphenol	40.000	39.374	1.6	89	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.826	0.4	91	0.00
29 C	2,4-Dichlorophenol	40.000	40.899	-2.2	93	0.00
30	1,2,4-Trichlorobenzene	40.000	41.254	-3.1	96	0.00
31	Naphthalene	40.000	39.427	1.4	90	0.00
32	Benzoic acid	40.000	38.890	2.8	95	0.01
33	4-Chloroaniline	40.000	40.981	-2.5	92	0.00
34 C	Hexachlorobutadiene	40.000	42.099	-5.2	97	0.00
35	Caprolactam	40.000	41.172	-2.9	93	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.417	-1.0	93	0.00
37	2-Methylnaphthalene	40.000	40.581	-1.5	93	0.00
38	1-Methylnaphthalene	40.000	39.546	1.1	91	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	93	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.243	-0.6	95	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.628	-9.1	105	0.00
42 S	2,4,6-Tribromophenol	80.000	85.416	-6.8	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.920	-2.3	98	0.00
44	2,4,5-Trichlorophenol	40.000	40.806	-2.0	95	0.00
45 S	2-Fluorobiphenyl	80.000	77.787	2.8	93	0.00
46	1,1'-Biphenyl	40.000	38.887	2.8	94	0.00
47	2-Chloronaphthalene	40.000	39.465	1.3	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072321\
 Data File : BG049427.D
 Acq On : 23 Jul 2021 14:33
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 23 16:11:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG072121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 21 16:57:05 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.335	-5.8	100	0.00
49	Acenaphthylene	40.000	39.419	1.5	94	0.00
50	Dimethylphthalate	40.000	39.682	0.8	95	0.00
51	2,6-Dinitrotoluene	40.000	40.240	-0.6	98	0.00
52 C	Acenaphthene	40.000	39.729	0.7	95	0.00
53	3-Nitroaniline	40.000	40.485	-1.2	94	0.00
54 P	2,4-Dinitrophenol	40.000	41.413	-3.5	107	0.00
55	Dibenzofuran	40.000	39.200	2.0	95	0.00
56 P	4-Nitrophenol	40.000	39.114	2.2	98	0.00
57	2,4-Dinitrotoluene	40.000	41.427	-3.6	97	0.00
58	Fluorene	40.000	38.738	3.2	93	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.379	-0.9	94	0.00
60	Diethylphthalate	40.000	40.497	-1.2	96	0.00
61	4-Chlorophenyl-phenylether	40.000	39.794	0.5	94	0.00
62	4-Nitroaniline	40.000	39.131	2.2	91	0.00
63	Azobenzene	40.000	39.339	1.7	95	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	97	0.00
65	4,6-Dinitro-2-methylphenol	40.000	45.267	-13.2	108	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.974	5.1	95	0.00
67	4-Bromophenyl-phenylether	40.000	41.136	-2.8	100	0.00
68	Hexachlorobenzene	40.000	38.944	2.6	98	0.00
69	Atrazine	40.000	39.833	0.4	100	0.00
70 C	Pentachlorophenol	40.000	41.144	-2.9	105	0.00
71	Phenanthrene	40.000	38.501	3.7	95	0.00
72	Anthracene	40.000	38.614	3.5	95	0.00
73	Carbazole	40.000	39.348	1.6	96	0.00
74	Di-n-butylphthalate	40.000	39.963	0.1	98	0.00
75 C	Fluoranthene	40.000	39.608	1.0	97	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
77	Benzydine	40.000	40.303	-0.8	95	0.00
78	Pyrene	40.000	38.413	4.0	100	0.00
79 S	Terphenyl-d14	80.000	78.000	2.5	100	0.00
80	Butylbenzylphthalate	40.000	40.219	-0.5	100	0.00
81	Benzo(a)anthracene	40.000	38.643	3.4	100	0.00
82	3,3'-Dichlorobenzidine	40.000	40.193	-0.5	103	0.00
83	Chrysene	40.000	38.311	4.2	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.845	0.4	102	0.00
85 c	Di-n-octyl phthalate	40.000	39.047	2.4	100	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.308	1.7	100	0.00
88	Benzo(b)fluoranthene	40.000	40.127	-0.3	101	0.00
89	Benzo(k)fluoranthene	40.000	39.270	1.8	101	0.00
90 C	Benzo(a)pyrene	40.000	39.707	0.7	101	0.00
91	Dibenzo(a,h)anthracene	40.000	38.931	2.7	99	0.00
92	Benzo(g,h,i)perylene	40.000	38.849	2.9	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072321\
Data File : BG049427.D
Acq On : 23 Jul 2021 14:33
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jul 23 16:11:14 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG072121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 21 16:57:05 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0