

Data Path : Z:\HPCHEM1\BNA_G\Data\BG072415\
 Data File : BG018108.D
 Acq On : 25 Jul 2015 2:11
 Operator : TP/UM
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02034

Quant Time: Jul 25 03:36:35 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG072315.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 24 23:53:51 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.343	0.333	2.9	120	0.00
3 S	1,4-Dioxane-d8	0.320	0.322	-0.6	127	0.00
4	Benzaldehyde	0.793	0.717	9.6	106	0.00
5 S	Phenol-d5	1.543	1.347	12.7	108	0.00
6	Phenol	1.575	1.377	12.6	108	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.762	0.669	12.2	106	0.00
8	Bis(2-Chloroethyl)ether	1.175	1.076	8.4	109	0.00
9 S	2-Chlorophenol-d4	1.368	1.234	9.8	108	0.00
10	2-Chlorophenol	1.354	1.204	11.1	107	0.00
11	2-Methylphenol	1.301	1.138	12.5	109	0.00
12	2,2'-oxybis(1-Chloropropane	1.244	1.070	14.0	104	0.00
13 S	4-Methylphenol-d8	1.368	1.190	13.0	109	0.00
14	Acetophenone	1.929	1.730	10.3	109	0.00
15 P	N-Nitroso-di-n-propylamine	1.024	0.882	13.9	105	0.00
16	4-Methylphenol	1.440	1.283	10.9	108	0.00
17	Hexachloroethane	0.522	0.503	3.6	117	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
19 S	Nitrobenzene-d5	0.158	0.144	8.9	109	0.00
20	Nitrobenzene	0.291	0.270	7.2	105	0.00
21	Isophorone	0.606	0.579	4.5	111	0.00
22 S	2-Nitrophenol-d4	0.167	0.157	6.0	108	0.00
23 C	2-Nitrophenol	0.180	0.168	6.7	109	0.00
24	2,4-Dimethylphenol	0.316	0.300	5.1	111	0.00
25	Bis(2-Chloroethoxy)methane	0.341	0.317	7.0	108	0.00
26 S	2,4-Dichlorophenol-d3	0.291	0.274	5.8	110	0.00
27 C	2,4-Dichlorophenol	0.281	0.271	3.6	111	0.00
28	Naphthalene	0.937	0.879	6.2	109	0.00
29 S	4-Chloroaniline-d4	0.338	0.336	0.6	121	0.00
30	4-Chloroaniline	0.348	0.345	0.9	121	0.00
31 C	Hexachlorobutadiene	0.165	0.157	4.8	109	0.00
32	Caprolactam	0.134	0.134	0.0	114	0.00
33 C	4-Chloro-3-methylphenol	0.307	0.295	3.9	113	0.00
34	2-Methylnaphthalene	0.708	0.680	4.0	110	0.00
35	1-Methylnaphthalene	0.691	0.648	6.2	107	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
37	1,2,4,5-Tetrachlorobenzene	0.514	0.496	3.5	111	0.00
38	Hexachlorocyclopentadiene	0.301	0.264	12.3	108	0.00
39 C	2,4,6-Trichlorophenol	0.345	0.339	1.7	117	0.00
40	2,4,5-Trichlorophenol	0.375	0.368	1.9	116	0.00
41	1,1'-Biphenyl	1.369	1.324	3.3	109	0.00
42	2-Chloronaphthalene	1.078	1.033	4.2	109	0.00
43	2-Nitroaniline	0.283	0.252	11.0	102	0.00
44 S	Dimethylphthalate-d6	1.404	1.344	4.3	109	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG072415\
 Data File : BG018108.D
 Acq On : 25 Jul 2015 2:11
 Operator : TP/UM
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02034

Quant Time: Jul 25 03:36:35 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG072315.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 24 23:53:51 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.413	1.351	4.4	109	0.00
46	2,6-Dinitrotoluene	0.324	0.302	6.8	108	0.00
47 S	Acenaphthylene-d8	1.725	1.646	4.6	108	0.00
48	Acenaphthylene	1.812	1.733	4.4	108	0.00
49	3-Nitroaniline	0.291	0.266	8.6	108	0.00
50 C	Acenaphthene	1.186	1.105	6.8	107	0.00
51	2,4-Dinitrophenol	0.168	0.139	17.3	116	0.00
52 S	4-Nitrophenol-d4	0.295	0.261	11.5	104	0.00
53	4-Nitrophenol	0.194	0.178	8.2	104	0.00
54	Dibenzofuran	1.686	1.595	5.4	107	0.00
55	2,4-Dinitrotoluene	0.466	0.439	5.8	110	0.00
56	2,3,4,6-Tetrachlorophenol	0.335	0.318	5.1	111	0.00
57	Diethylphthalate	1.476	1.409	4.5	109	0.00
58 S	Fluorene-d10	1.298	1.235	4.9	109	0.00
59	Fluorene	1.451	1.352	6.8	106	0.00
60	4-Chlorophenyl-phenylether	0.727	0.684	5.9	109	0.00
61	4-Nitroaniline	0.360	0.325	9.7	105	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.128	0.113	11.7	107	0.00
64	4,6-Dinitro-2-methylphenol	0.135	0.119	11.9	106	0.00
65	N-Nitrosodiphenylamine	0.545	0.532	2.4	107	0.00
66	4-Bromophenyl-phenylether	0.197	0.192	2.5	111	0.00
67	Hexachlorobenzene	0.220	0.213	3.2	110	0.00
68	Atrazine	0.211	0.206	2.4	108	0.00
69 C	Pentachlorophenol	0.120	0.104	13.3	108	0.00
70	Phenanthrene	1.005	0.985	2.0	109	0.00
71 S	Anthracene-d10	0.866	0.833	3.8	107	0.00
72	Anthracene	1.020	0.987	3.2	107	0.00
73	1,2,3,4-Tetrachlorobenzene	0.225	0.223	0.9	111	0.00
74	Pentachlorobenzene	0.237	0.230	3.0	110	0.00
75	Carbazole	0.884	0.891	-0.8	105	0.00
76	Di-n-butylphthalate	1.164	1.133	2.7	106	0.00
77 C	Fluoranthene	1.070	1.132	-5.8	107	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
79 S	Pyrene-d10	0.916	0.860	6.1	108	0.00
80	Pyrene	1.145	1.069	6.6	106	0.00
81	Butylbenzylphthalate	0.469	0.446	4.9	111	0.00
82	3,3'-Dichlorobenzidine	0.317	0.289	8.8	108	0.00
83	Benzo(a)anthracene	1.058	1.024	3.2	112	0.00
84	Bis(2-ethylhexyl)phthalate	0.672	0.639	4.9	110	0.00
85	Chrysene	0.996	0.961	3.5	111	0.00
86 I	Perylene-d12	1.000	1.000	0.0	110	0.00
87	Di-n-octyl phthalate	1.072	1.090	-1.7	109	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG072415\
Data File : BG018108.D
Acq On : 25 Jul 2015 2:11
Operator : TP/UM
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02034

Quant Time: Jul 25 03:36:35 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG072315.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jul 24 23:53:51 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.096	1.039	5.2	114	0.00
89	Benzo(k)fluoranthene	1.061	1.013	4.5	110	0.00
90 S	Benzo(a)pyrene-d12	0.954	0.897	6.0	111	0.00
91 C	Benzo(a)pyrene	1.061	1.010	4.8	112	0.00
92	Indeno(1,2,3-cd)pyrene	1.224	1.142	6.7	112	-0.02
93	Dibenzo(a,h)anthracene	1.045	0.985	5.7	112	0.00
94	Benzo(g,h,i)perylene	1.012	0.946	6.5	112	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0