


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG072415\
Data File : BG018090.D
Acq On : 24 Jul 2015 14:03
Operator : TP/UM
Sample : DFTPP19
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP19

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG072315.M
Title : SVOA CALIBRATION
Last Update : Fri Jul 24 23:53:51 2015



AutoFind: Scans 2504, 2505, 2506; Background Corrected with Scan 2498

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.0	15309	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	31.9	18797	PASS
70	69	0.00	2	1.2	227	PASS
127	198	10	80	47.9	28194	PASS
197	198	0.00	2	0.3	179	PASS
198	198	100	100	100.0	58890	PASS
199	198	5	9	6.8	3992	PASS
275	198	10	60	28.6	16855	PASS
365	198	1	100	4.3	2559	PASS
441	442	0.01	24	14.6	8312	PASS
442	198	50	100	96.9	57066	PASS
443	442	15	24	19.3	11009	PASS

m/z	Abundance
38.10	262.0
39.00	1197.0
40.00	582.0
41.00	378.0
43.00	592.0
44.00	652.0
49.00	294.0
50.00	4575.0
51.00	16616.0
52.10	1085.0
54.90	511.0
56.00	667.0
57.00	1580.0
59.90	309.0
61.00	281.0
62.00	273.0
63.00	797.0
63.90	163.0
65.00	466.0
67.10	234.0
69.00	20496.0
69.90	203.0
71.10	244.0
73.00	715.0
74.00	2412.0
75.00	3531.0
76.10	1364.0
77.00	23288.0
78.00	1821.0
79.00	1589.0
80.00	1204.0
81.00	1902.0
82.00	485.0
82.90	516.0
84.00	299.0
85.10	413.0
85.90	756.0
87.00	250.0
91.00	433.0
91.90	350.0
93.00	2609.0
93.90	192.0
95.90	231.0
97.10	190.0
98.00	2197.0
99.00	1475.0
101.00	892.0
103.00	356.0
104.00	628.0
105.00	652.0
107.00	7888.0
108.00	1278.0
109.00	478.0
110.00	13762.0
111.00	2172.0
112.00	217.0
112.90	194.0
115.00	165.0
116.00	550.0
117.00	5753.0
118.00	470.0
121.90	740.0
123.00	761.0
123.90	313.0
125.00	422.0
127.00	27952.0
128.00	2028.0
129.00	10935.0
130.00	1130.0
131.00	210.0
133.90	306.0
135.00	730.0
136.00	342.0
137.10	409.0
138.80	176.0
141.00	1546.0
142.00	523.0
143.00	335.0
145.90	164.0
147.00	790.0
147.90	1500.0
149.00	394.0
151.10	227.0
151.90	214.0
152.90	597.0
153.90	349.0
155.00	1019.0
156.00	1361.0
157.00	215.0
158.10	218.0
158.90	223.0

Instrument :
BNA_G
ClientSampleId :
DFTPP19

159.90	417.0
160.90	628.0
162.00	283.0
165.00	644.0
167.00	2968.0
167.90	949.0
169.00	324.0
172.00	390.0
172.90	397.0
174.00	499.0
175.00	957.0
175.90	496.0
176.90	591.0
178.00	237.0
178.90	1953.0
180.00	1015.0
181.00	759.0
182.10	209.0
183.90	199.0
185.00	1094.0
186.00	7469.0
187.00	2391.0
188.00	235.0
189.00	436.0
190.90	195.0
192.00	722.0
192.90	1027.0
193.90	265.0
195.00	272.0
196.00	1230.0
197.90	55736.0
198.90	3591.0
200.00	406.0
201.40	330.0
203.10	396.0
204.00	2251.0
205.00	3998.0
206.00	14273.0
207.00	2341.0
208.00	675.0
209.00	191.0
210.90	516.0
215.00	183.0
216.00	409.0
216.90	4112.0
217.90	650.0
221.00	3107.0
222.00	582.0
223.00	1031.0
224.00	8722.0
225.00	2039.0
226.00	381.0
226.90	3475.0
227.90	490.0
228.90	682.0
230.10	150.0
231.00	268.0
233.90	277.0
235.00	306.0
235.90	187.0
240.90	283.0
241.90	408.0
243.00	784.0
244.00	6223.0
245.00	1218.0
246.00	1188.0
246.90	330.0
248.90	261.0
252.90	218.0
255.00	30248.0
256.00	4576.0
256.90	344.0
258.00	1978.0
258.90	259.0
264.90	773.0
273.00	851.0
273.90	2717.0
275.00	15335.0
276.00	2338.0
276.90	1543.0
278.00	311.0
280.90	168.0
283.00	230.0
285.00	171.0
293.00	352.0
296.00	4246.0
297.00	565.0
302.90	502.0
315.00	600.0
315.90	171.0
323.00	1348.0
324.00	383.0
326.80	184.0
334.10	609.0
335.00	153.0

Instrument :
BNA_G
ClientSampleId :
DFTPP19

341.00	256.0
345.90	309.0
352.00	527.0
353.00	404.0
354.00	396.0
355.00	152.0
364.90	1937.0
366.00	185.0
372.00	789.0
372.90	242.0
383.00	180.0
401.90	257.0
402.90	376.0
404.00	183.0
421.10	273.0
422.00	260.0
423.00	2077.0
424.00	608.0
441.00	6311.0
442.00	42368.0
443.00	8115.0
444.10	831.0

Instrument :
BNA_G
ClientSampleId :
DFTPP19

m/z	Abundance
37.00	211.0
38.10	192.0
39.00	1384.0
40.00	729.0
41.10	407.0
43.10	401.0
44.00	680.0
49.10	283.0
50.00	4759.0
51.00	18592.0
52.00	845.0
55.00	495.0
56.00	856.0
57.00	1847.0
60.00	350.0
61.00	345.0
62.00	219.0
63.00	756.0
63.90	153.0
65.00	458.0
69.00	22776.0
69.90	278.0
71.00	212.0
73.00	559.0
74.00	2593.0
75.00	3527.0
76.00	1608.0
77.00	26352.0
78.00	1745.0
79.00	1519.0
80.00	1375.0
81.00	2438.0
81.90	441.0
83.00	801.0
83.90	211.0
85.00	386.0
86.00	616.0
87.00	397.0
91.00	440.0
91.90	826.0
93.00	2647.0
94.00	366.0
96.10	296.0
97.20	371.0
98.00	2376.0
99.00	1811.0
101.00	1065.0
103.00	400.0
104.00	684.0
105.00	942.0
107.00	8675.0
108.00	1187.0
110.00	16141.0
111.00	2653.0
112.00	366.0
116.00	618.0
117.00	7500.0
117.90	281.0
121.90	883.0
122.90	1092.0
124.00	552.0
125.00	458.0
127.00	34296.0
128.00	2705.0
129.00	13032.0
129.90	1329.0
130.90	305.0
132.00	192.0
133.90	554.0
135.00	1160.0
135.90	468.0
137.00	401.0
139.90	385.0
140.90	2075.0
142.00	603.0
143.00	553.0
145.90	427.0
147.00	1023.0
148.00	2270.0
149.00	537.0
151.00	320.0
152.00	181.0
153.00	760.0
154.00	486.0
155.00	1238.0
156.00	1266.0
156.90	410.0
158.00	393.0
159.00	183.0
160.00	566.0
161.00	743.0

Instrument :
BNA_G
ClientSampleId :
DFTPP19

162.00	215.0
163.80	192.0
164.90	711.0
165.90	519.0
167.00	3588.0
168.00	1261.0
169.00	421.0
169.90	194.0
170.90	186.0
171.90	393.0
172.90	316.0
174.00	930.0
175.00	1330.0
176.00	662.0
176.90	873.0
177.90	249.0
178.90	2548.0
180.00	1692.0
181.00	781.0
184.10	206.0
184.90	1283.0
186.00	10245.0
187.00	2633.0
187.90	302.0
188.90	817.0
190.90	294.0
192.00	894.0
193.00	993.0
193.90	294.0
194.90	182.0
196.00	2597.0
197.90	70888.0
199.00	5039.0
200.00	291.0
201.30	388.0
202.90	647.0
204.00	2744.0
205.00	4405.0
206.00	19424.0
207.00	2907.0
208.00	716.0
208.90	264.0
209.90	297.0
210.10	308.0
211.00	900.0
213.00	238.0
215.00	258.0
216.00	307.0
216.90	5297.0
217.90	672.0
221.00	3800.0
221.80	898.0
222.90	1303.0
224.00	11167.0
225.00	2907.0
227.00	4644.0
227.90	770.0
228.90	1076.0
231.00	511.0
234.00	237.0
234.80	254.0
236.00	276.0
236.90	498.0
239.00	234.0
239.90	240.0
241.00	364.0
241.90	603.0
243.10	519.0
244.00	8502.0
245.00	1293.0
246.00	1997.0
246.80	336.0
248.90	354.0
252.00	176.0
253.00	162.0
255.00	44536.0
256.00	6480.0
257.00	514.0
257.90	3232.0
258.90	582.0
264.90	1130.0
271.90	353.0
272.90	991.0
274.00	3276.0
275.00	19056.0
276.00	2365.0
277.00	1789.0
277.90	224.0
281.10	182.0
282.70	237.0
284.10	241.0
285.00	297.0
292.90	676.0
294.10	208.0
296.00	6601.0

Instrument :
BNA_G
ClientSampleId :
DFTPP19

297.00	798.0
302.00	175.0
303.00	811.0
304.10	368.0
314.00	215.0
314.90	675.0
316.00	312.0
320.90	183.0
323.00	1845.0
324.00	298.0
327.00	434.0
327.90	160.0
331.90	226.0
333.10	264.0
334.00	1113.0
335.00	245.0
341.00	252.0
345.90	435.0
352.00	569.0
353.00	486.0
354.00	478.0
355.00	156.0
365.00	3048.0
366.00	512.0
372.00	1136.0
373.00	192.0
382.90	208.0
401.90	307.0
403.10	415.0
403.90	216.0
415.00	159.0
420.90	630.0
421.90	329.0
423.00	3892.0
424.00	662.0
441.00	9670.0
442.00	66872.0
443.00	12220.0
444.00	1232.0

Instrument :
BNA_G
ClientSampleId :
DFTPP19

m/z	Abundance
39.00	814.0
39.90	473.0
41.10	316.0
43.00	644.0
44.00	764.0
45.00	166.0
49.00	283.0
50.00	2635.0
51.00	10719.0
52.10	356.0
55.10	421.0
56.00	493.0
57.00	1262.0
60.00	445.0
63.00	565.0
64.90	266.0
67.00	175.0
69.00	13675.0
70.00	201.0
70.90	181.0
73.00	832.0
74.00	1609.0
75.00	2076.0
76.00	908.0
77.00	15632.0
78.10	1062.0
79.00	1221.0
80.00	1113.0
81.00	1373.0
81.90	439.0
83.00	663.0
83.90	241.0
85.00	380.0
85.90	370.0
86.90	181.0
90.90	449.0
91.90	277.0
93.00	1798.0
94.00	247.0
95.00	154.0
98.00	1683.0
99.00	1312.0
100.90	695.0
103.00	205.0
104.00	618.0
105.00	590.0
106.00	219.0
107.00	5857.0
108.00	1011.0
109.00	347.0
110.00	10210.0
111.00	1446.0
112.00	283.0
117.00	4182.0
118.00	554.0
120.90	166.0
122.00	463.0
122.90	909.0
123.90	425.0
124.90	575.0
127.00	22336.0
128.00	1722.0
129.00	8804.0
130.00	700.0
134.00	274.0
134.90	799.0
135.90	263.0
137.00	263.0
140.00	191.0
141.00	1233.0
141.90	449.0
142.90	292.0
146.00	191.0
147.00	651.0
148.00	928.0
148.90	553.0
151.70	150.0
153.00	385.0
154.00	449.0
155.00	950.0
156.00	971.0
157.10	205.0
157.90	304.0
159.90	263.0
161.00	508.0
161.90	195.0
164.90	163.0
166.00	435.0
167.00	2235.0
168.00	1029.0
169.10	196.0

Instrument :
BNA_G
ClientSampleId :
DFTPP19

170.10	162.0
172.00	205.0
173.00	264.0
174.00	560.0
175.00	851.0
176.30	165.0
176.90	472.0
177.90	343.0
179.00	1899.0
179.90	1235.0
181.00	503.0
181.90	213.0
183.90	283.0
185.00	788.0
186.00	7232.0
186.90	2118.0
187.90	190.0
188.90	350.0
190.90	233.0
192.00	579.0
193.10	619.0
193.80	185.0
196.00	1315.0
196.80	538.0
197.90	50048.0
198.90	3348.0
201.30	244.0
202.90	410.0
204.00	2038.0
205.00	3091.0
206.00	13113.0
207.00	2213.0
208.00	504.0
208.90	289.0
210.30	276.0
211.00	556.0
211.60	225.0
213.00	189.0
214.80	176.0
215.90	388.0
217.00	3672.0
218.00	578.0
221.00	2864.0
222.00	419.0
223.00	1018.0
224.00	7271.0
225.00	2084.0
226.10	290.0
226.90	3056.0
227.80	567.0
228.90	801.0
230.90	397.0
233.90	253.0
236.80	301.0
238.90	186.0
241.00	344.0
242.00	464.0
243.00	470.0
244.00	6287.0
245.00	869.0
246.00	1502.0
246.90	288.0
248.90	251.0
250.80	184.0
253.00	153.0
255.00	33624.0
256.00	4409.0
256.90	351.0
258.00	2153.0
258.90	333.0
265.00	857.0
271.90	263.0
273.00	724.0
274.00	2214.0
275.00	16174.0
276.00	1720.0
277.00	1355.0
278.00	220.0
281.00	192.0
283.00	216.0
284.10	255.0
285.10	241.0
292.90	167.0
295.90	4943.0
297.00	674.0
303.00	648.0
314.00	258.0
314.90	610.0
316.00	225.0
320.90	195.0
323.10	1731.0
323.90	284.0
326.90	225.0
331.80	157.0
332.90	196.0

Instrument :
BNA_G
ClientSampleId :
DFTPP19

334.00	1011.0
334.90	217.0
341.00	214.0
346.10	356.0
352.00	475.0
353.10	337.0
354.00	595.0
365.00	2692.0
365.90	297.0
371.00	154.0
372.00	1050.0
383.00	200.0
390.10	156.0
402.00	403.0
402.90	606.0
404.00	238.0
421.00	391.0
422.00	556.0
423.00	3505.0
424.00	495.0
441.00	8956.0
442.00	61960.0
443.00	12694.0
444.00	1225.0

Instrument :
BNA_G
ClientSampleId :
DFTPP19

m/z	Abundance
40.00	449.0
41.00	393.0
43.00	405.0
44.00	773.0
45.10	202.0
55.00	463.0
57.00	270.0
60.00	249.0
67.00	264.0
69.10	185.0
71.20	294.0
73.00	374.0
83.00	201.0
84.00	160.0
85.00	163.0
91.00	174.0
95.10	370.0
95.90	187.0
97.10	205.0
111.00	153.0
131.00	166.0
167.10	315.0
207.00	824.0
213.10	150.0
280.90	208.0

Instrument :
BNA_G
ClientSampleId :
DFTPP19