

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072521\
 Data File : BG049460.D
 Acq On : 24 Jul 2021 20:52
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020457

Quant Time: Jul 25 07:22:32 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG072421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 24 06:15:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.046	152	24286	20.000	ng/u1	0.00
20) Naphthalene-d8	10.866	136	97364	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.696	164	65110	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.445	188	147213	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.728	240	157848	20.000	ng/u1	0.00
88) Perylene-d12	25.005	264	166924	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.429	96	4569	8.607	ng/uL	0.00
4) Pyridine-d5	3.858	84	27805	18.236	ng/u1	0.00
7) Phenol-d5	7.200	99	41545	19.335	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.365	67	23534	19.338	ng/u1	0.00
11) 2-Chlorophenol-d4	7.576	132	31808	20.651	ng/u1	0.00
15) 4-Methylphenol-d8	8.751	113	30948	19.108	ng/u1	0.00
21) Nitrobenzene-d5	9.215	128	14980	20.012	ng/u1	0.00
24) 2-Nitrophenol-d4	9.938	143	17517	19.969	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.484	165	32358	20.282	ng/u1	0.00
31) 4-Chloroaniline-d4	11.001	131	42636	19.176	ng/u1	0.00
46) Dimethylphthalate-d6	14.097	166	102159	20.483	ng/u1	0.00
49) Acenaphthylene-d8	14.390	160	121864	21.037	ng/u1	0.00
54) 4-Nitrophenol-d4	14.890	143	15479	18.846	ng/u1	0.00
60) Fluorene-d10	15.689	176	85389	19.873	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.806	200	17244	18.903	ng/u1	0.00
73) Anthracene-d10	17.545	188	135333	19.780	ng/u1	0.00
81) Pyrene-d10	19.830	212	161991	20.070	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.776	264	172730	19.688	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.464	88	4801	6.617	ng/uL#	94
5) Pyridine	3.875	79	32633	20.009	ng/u1	94
6) Benzaldehyde	7.183	77	32287	25.182	ng/u1	98
8) Phenol	7.230	94	42057	20.037	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.464	93	29541	20.428	ng/u1	94
12) 2-Chlorophenol	7.605	128	30854	20.808	ng/u1	96
13) 2-Methylphenol	8.487	108	31293	19.561	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.569	45	32447	19.147	ng/u1	99
16) Acetophenone	8.868	105	52940	20.309	ng/u1#	96
17) N-Nitroso-di-n-propyla...	8.857	70	27779	20.666	ng/u1#	86
18) 4-Methylphenol	8.816	108	33618	19.975	ng/u1	98
19) Hexachloroethane	9.139	117	14577	21.821	ng/u1	89
22) Nitrobenzene	9.262	77	45393	20.310	ng/u1	98
23) Isophorone	9.779	82	77258	20.629	ng/u1	99
25) 2-Nitrophenol	9.973	139	17423	21.106	ng/u1#	78
26) 2,4-Dimethylphenol	10.026	107	41303	20.993	ng/u1	92
27) Bis(2-Chloroethoxy)met...	10.261	93	37883	20.280	ng/u1	97
29) 2,4-Dichlorophenol	10.513	162	32533	20.860	ng/u1	95
30) Naphthalene	10.919	128	99310	19.802	ng/u1	97
32) 4-Chloroaniline	11.024	127	40730	19.257	ng/u1#	82
33) Hexachlorobutadiene	11.201	225	26159	21.939	ng/u1	98
34) Caprolactam	11.776	113	3490	6.580	ng/u1	88
35) 4-Chloro-3-methylphenol	12.146	107	35472	20.214	ng/u1#	82
36) 2-Methylnaphthalene	12.528	142	74454	20.284	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072521\
 Data File : BG049460.D
 Acq On : 24 Jul 2021 20:52
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 16 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SSTD020457


Quant Time: Jul 25 07:22:32 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG072421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 24 06:15:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.746	142	73978	19.945	ng/ul	94
39) 1,2,4,5-Tetrachloroben...	12.892	216	39894	20.615	ng/ul	97
40) Hexachlorocyclopentadiene	12.869	237	22938	19.672	ng/ul	99
41) 2,4,6-Trichlorophenol	13.127	196	25271	21.250	ng/ul#	85
42) 2,4,5-Trichlorophenol	13.204	196	27464	21.625	ng/ul	94
43) 1,1'-Biphenyl	13.527	154	97538	21.037	ng/ul	98
44) 2-Chloronaphthalene	13.574	162	74780	20.726	ng/ul	96
45) 2-Nitroaniline	13.774	65	27426	21.420	ng/ul	90
47) Dimethylphthalate	14.144	163	99464	20.507	ng/ul	99
48) 2,6-Dinitrotoluene	14.267	165	19967	20.628	ng/ul#	86
50) Acenaphthylene	14.420	152	112682	20.369	ng/ul	96
51) 3-Nitroaniline	14.596	138	19262	19.952	ng/ul#	83
52) Acenaphthene	14.761	153	82008	20.669	ng/ul	99
53) 2,4-Dinitrophenol	14.802	184	11468	23.780	ng/ul#	80
55) 4-Nitrophenol	14.907	109	22303	20.026	ng/ul	89
56) Dibenzofuran	15.095	168	109542	20.446	ng/ul	99
57) 2,4-Dinitrotoluene	15.054	165	28488	20.763	ng/ul#	98
58) 2,3,4,6-Tetrachlorophenol	15.319	232	23852	20.696	ng/ul	99
59) Diethylphthalate	15.507	149	102759	20.471	ng/ul	97
61) Fluorene	15.747	166	94874	20.860	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.736	204	49209	20.784	ng/ul#	86
63) 4-Nitroaniline	15.759	138	21085	22.025	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.818	198	15699	18.695	ng/ul#	76
67) N-Nitrosodiphenylamine	15.947	169	80217	20.792	ng/ul	99
68) 4-Bromophenyl-phenylether	16.629	248	32772	19.993	ng/ul	94
69) Hexachlorobenzene	16.752	284	37046	19.975	ng/ul	95
70) Atrazine	16.893	200	34324	19.479	ng/ul	96
71) Pentachlorophenol	17.099	266	18822	21.919	ng/ul	85
72) Phenanthrene	17.492	178	151206	19.730	ng/ul	98
74) Anthracene	17.580	178	154053	19.836	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.498	216	40530	20.261	ng/ul	98
76) Pentachlorobenzene	15.013	250	44221	20.894	ng/ul	93
77) Carbazole	17.850	167	134671	19.396	ng/ul	98
78) Di-n-butylphthalate	18.403	149	171981	19.919	ng/ul	99
80) Fluoranthene	19.495	202	194709	19.868	ng/ul	98
82) Pyrene	19.859	202	197512	20.299	ng/ul	97
83) Butylbenzylphthalate	20.741	149	75601	19.957	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.622	252	74301	20.486	ng/ul	96
85) Benzo(a)anthracene	21.704	228	191928	19.887	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.604	149	112949	20.090	ng/ul#	98
87) Chrysene	21.775	228	180779	19.571	ng/ul	99
89) Di-n-octyl phthalate	22.832	149	185327	18.888	ng/ul	100
90) Benzo(b)fluoranthene	23.960	252	203632	19.463	ng/ul	98
91) Benzo(k)fluoranthene	23.960	252	203632	19.887	ng/ul	98
93) Benzo(a)pyrene	24.841	252	177423	19.548	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	28.742	276	234615	20.258	ng/ul#	92
95) Dibenzo(a,h)anthracene	28.800	278	193428	20.183	ng/ul	96
96) Benzo(g,h,i)perylene	29.911	276	194719	20.223	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 25 07:22:32 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG072421.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Jul 24 06:15:51 2021
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.617 ng/uL
RT: 3.464 min Scan# 72
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion: 88 Resp: 4801
Ion Ratio Lower Upper
88 100
43 28.4 29.5 44.3#
58 70.0 56.8 85.2



#5
Pyridine
Concen: 20.009 ng/uL
RT: 3.875 min Scan# 142
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion: 79 Resp: 32633
Ion Ratio Lower Upper
79 100
52 69.7 57.4 86.2
51 43.2 28.9 43.3









#6
Benzaldehyde
Concen: 25.182 ng/ul
RT: 7.183 min Scan# 705
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion: 77 Resp: 32287
Ion Ratio Lower Upper
77 100
105 90.2 70.0 105.0
106 88.9 70.1 105.1



#8
Phenol
Concen: 20.037 ng/ul
RT: 7.230 min Scan# 713
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion: 94 Resp: 42057
Ion Ratio Lower Upper
94 100
65 35.8 28.8 43.2
66 47.5 40.3 60.5









#10
Bis(2-Chloroethyl)ether
Concen: 20.428 ng/ul
RT: 7.464 min Scan# 753
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:	93	Resp:	29541
Ion	Ratio	Lower	Upper
93	100		
63	71.5	62.4	93.6
95	33.0	26.0	39.0



#12
2-Chlorophenol
Concen: 20.808 ng/ul
RT: 7.605 min Scan# 777
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:	128	Resp:	30854
Ion	Ratio	Lower	Upper
128	100		
64	58.6	44.2	66.4
130	31.2	24.1	36.1









#13
2-Methylphenol
Concen: 19.561 ng/ul
RT: 8.487 min Scan# 927
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:108 Resp: 31293
Ion Ratio Lower Upper
108 100
107 91.7 71.2 106.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.147 ng/ul
RT: 8.569 min Scan# 941
Delta R.T. -0.009 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion: 45 Resp: 32447
Ion Ratio Lower Upper
45 100
77 20.7 17.0 25.4
79 12.1 9.7 14.5









#16
Acetophenone
Concen: 20.309 ng/ul
RT: 8.868 min Scan# 992
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:	105	Resp:	52940
Ion Ratio	Lower	Upper	
105	100		
77	84.9	66.4	99.6
51	31.9	20.8	31.2#



#17
N-Nitroso-di-n-propylamine
Concen: 20.666 ng/ul
RT: 8.857 min Scan# 990
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:	70	Resp:	27779
Ion Ratio	Lower	Upper	
70	100		
42	68.8	44.2	66.4#
101	9.1	5.4	8.0#
130	15.9	12.4	18.6









#18
4-Methylphenol
Concen: 19.975 ng/ul
RT: 8.816 min Scan# 983
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampled :
SSTD020457


Tgt Ion:108 Resp: 33618
Ion Ratio Lower Upper
108 100
107 109.3 85.4 128.2



#19
Hexachloroethane
Concen: 21.821 ng/ul
RT: 9.139 min Scan# 1038
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:117 Resp: 14577
Ion Ratio Lower Upper
117 100
201 111.7 100.1 150.1
199 70.5 49.8 74.6









#22
Nitrobenzene
Concen: 20.310 ng/ul
RT: 9.262 min Scan# 1059
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion: 77	Resp: 45393
Ion Ratio	Lower Upper
77 100	
123 35.0	27.5 41.3
65 15.0	10.8 16.2



#23
Isophorone
Concen: 20.629 ng/ul
RT: 9.779 min Scan# 1147
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion: 82	Resp: 77258
Ion Ratio	Lower Upper
82 100	
95 9.9	7.4 11.0
138 13.3	10.7 16.1









#25
2-Nitrophenol
Concen: 21.106 ng/ul
RT: 9.973 min Scan# 1180
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:139 Resp: 17423
Ion Ratio Lower Upper
139 100
65 78.8 49.8 74.6#
109 54.0 31.9 47.9#



#26
2,4-Dimethylphenol
Concen: 20.993 ng/ul
RT: 10.026 min Scan# 1189
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:107 Resp: 41303
Ion Ratio Lower Upper
107 100
121 44.7 34.1 51.1
122 82.9 58.8 88.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.280 ng/ul
 RT: 10.261 min Scan# 1229
 Delta R.T. -0.003 min
 Lab File: BG049460.D
 Acq: 24 Jul 2021 20:52




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020457


Tgt Ion: 93	Resp:	37883
Ion Ratio	Lower	Upper
93	100	
95	31.7	26.8 40.2
123	9.3	8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 20.860 ng/ul
 RT: 10.513 min Scan# 1272
 Delta R.T. 0.003 min
 Lab File: BG049460.D
 Acq: 24 Jul 2021 20:52

Tgt Ion: 162	Resp:	32533
Ion Ratio	Lower	Upper
162	100	
164	61.4	52.2 78.2
98	29.2	25.4 38.2









#30
Naphthalene
Concen: 19.802 ng/ul
RT: 10.919 min Scan# 1341
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:	128	Resp:	99310
Ion Ratio	Lower	Upper	
128	100		
129	11.6	7.8	11.8
127	13.1	11.0	16.6



#32
4-Chloroaniline
Concen: 19.257 ng/ul
RT: 11.024 min Scan# 1359
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:	127	Resp:	40730
Ion Ratio	Lower	Upper	
127	100		
129	35.6	21.3	31.9#









#33
Hexachlorobutadiene
Concen: 21.939 ng/ul
RT: 11.201 min Scan# 1389
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:	225	Resp:	26159
Ion Ratio	Lower	Upper	
225	100		
223	66.6	52.1	78.1
227	67.1	52.4	78.6



#34
Caprolactam
Concen: 6.580 ng/ul
RT: 11.776 min Scan# 1487
Delta R.T. -0.009 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:	113	Resp:	3490
Ion Ratio	Lower	Upper	
113	100		
55	138.0	116.6	174.8
56	138.8	95.0	142.6









#35
4-Chloro-3-methylphenol
Concen: 20.214 ng/ul
RT: 12.146 min Scan# 1550
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:107 Resp: 35472
Ion Ratio Lower Upper
107 100
144 29.9 17.0 25.6#
142 80.5 53.3 79.9#



#36
2-Methylnaphthalene
Concen: 20.284 ng/ul
RT: 12.528 min Scan# 1615
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:142 Resp: 74454
Ion Ratio Lower Upper
142 100
141 87.2 69.9 104.9









#37
1-Methylnaphthalene
Concen: 19.945 ng/ul
RT: 12.746 min Scan# 1652
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:142 Resp: 73978
Ion Ratio Lower Upper
142 100
141 96.2 82.1 123.1
116 4.0 3.8 5.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.615 ng/ul
RT: 12.892 min Scan# 1677
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:216 Resp: 39894
Ion Ratio Lower Upper
216 100
214 75.8 58.6 88.0
179 20.8 17.0 25.4
108 19.9 14.4 21.6









#40
Hexachlorocyclopentadiene
Concen: 19.672 ng/ul
RT: 12.869 min Scan# 1673
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:	237	Resp:	22938
Ion Ratio	Lower	Upper	
237	100		
235	60.6	48.2	72.4
272	14.3	10.1	15.1



#41
2,4,6-Trichlorophenol
Concen: 21.250 ng/ul
RT: 13.127 min Scan# 1717
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:	196	Resp:	25271
Ion Ratio	Lower	Upper	
196	100		
198	90.7	61.6	92.4
200	32.5	21.3	31.9









#42
2,4,5-Trichlorophenol
Concen: 21.625 ng/ul
RT: 13.204 min Scan# 1730
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:196	Resp:	27464
Ion Ratio	Lower	Upper
196	100	
198	94.4	81.4 122.0
200	33.5	26.6 39.8



#43
1,1'-Biphenyl
Concen: 21.037 ng/ul
RT: 13.527 min Scan# 1785
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:154	Resp:	97538
Ion Ratio	Lower	Upper
154	100	
153	42.3	33.3 49.9
76	10.1	7.0 10.6









#44
2-Chloronaphthalene
Concen: 20.726 ng/ul
RT: 13.574 min Scan# 1793
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:	162	Resp:	74780
Ion Ratio	Lower	Upper	
162	100		
164	31.5	24.5	36.7
127	41.0	35.7	53.5



#45
2-Nitroaniline
Concen: 21.420 ng/ul
RT: 13.774 min Scan# 1827
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:	65	Resp:	27426
Ion Ratio	Lower	Upper	
65	100		
92	70.4	51.8	77.6
138	96.2	68.1	102.1









#47
Dimethylphthalate
Concen: 20.507 ng/ul
RT: 14.144 min Scan# 1890
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:163 Resp: 99464
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.6
164 10.0 8.5 12.7



#48
2,6-Dinitrotoluene
Concen: 20.628 ng/ul
RT: 14.267 min Scan# 1911
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52


Tgt Ion:165 Resp: 19967
Ion Ratio Lower Upper
165 100
89 73.6 67.4 101.2
121 19.0 24.6 36.8#








#50
Acenaphthylene
Concen: 20.369 ng/ul
RT: 14.420 min Scan# 1937
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52


Instrument :
BNA_G
ClientSampleId :
SSTD020457





Tgt Ion:152 Resp: 112682
Ion Ratio Lower Upper
152 100
151 20.3 14.4 21.6
153 12.8 9.5 14.3




#51
3-Nitroaniline
Concen: 19.952 ng/ul
RT: 14.596 min Scan# 1967
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52



Tgt Ion:138 Resp: 19262
Ion Ratio Lower Upper
138 100
108 14.7 8.2 12.2#
92 135.2 93.0 139.6









#52
Acenaphthene
Concen: 20.669 ng/ul
RT: 14.761 min Scan# 1995
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:153 Resp: 82008
Ion Ratio Lower Upper
153 100
152 47.1 37.4 56.0
154 82.6 66.7 100.1



#53
2,4-Dinitrophenol
Concen: 23.780 ng/ul
RT: 14.802 min Scan# 2002
Delta R.T. -0.002 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:184 Resp: 11468
Ion Ratio Lower Upper
184 100
63 88.1 85.5 128.3
154 73.2 75.8 113.6#









#55
4-Nitrophenol
Concen: 20.026 ng/ul
RT: 14.907 min Scan# 2020
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:109 Resp: 22303
Ion Ratio Lower Upper
109 100
139 66.6 57.0 85.4
65 89.7 83.1 124.7



#56
Dibenzofuran
Concen: 20.446 ng/ul
RT: 15.095 min Scan# 2052
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:168 Resp: 109542
Ion Ratio Lower Upper
168 100
139 45.2 35.5 53.3









#57
2,4-Dinitrotoluene
Concen: 20.763 ng/ul
RT: 15.054 min Scan# 2045
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:165 Resp: 28488
Ion Ratio Lower Upper
165 100
63 59.1 48.2 72.4
182 2.4 1.4 2.2#



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.696 ng/ul
RT: 15.319 min Scan# 2090
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:232 Resp: 23852
Ion Ratio Lower Upper
232 100
131 49.9 39.6 59.4
130 3.9 3.2 4.8
166 29.6 22.6 34.0









#59
Diethylphthalate
Concen: 20.471 ng/ul
RT: 15.507 min Scan# 2122
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:149 Resp: 102759
Ion Ratio Lower Upper
149 100
177 21.6 16.0 24.0
150 12.6 9.5 14.3



#61
Fluorene
Concen: 20.860 ng/ul
RT: 15.747 min Scan# 2163
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:166 Resp: 94874
Ion Ratio Lower Upper
166 100
165 99.2 81.8 122.8
167 12.9 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 20.784 ng/ul
RT: 15.736 min Scan# 2161
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:204 Resp: 49209
Ion Ratio Lower Upper
204 100
206 34.1 23.5 35.3
141 68.4 45.4 68.0#



#63
4-Nitroaniline
Concen: 22.025 ng/ul
RT: 15.759 min Scan# 2165
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:138 Resp: 21085
Ion Ratio Lower Upper
138 100
92 58.0 49.8 74.8
108 135.3 105.4 158.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.695 ng/ul
 RT: 15.818 min Scan# 2175
 Delta R.T. -0.003 min
 Lab File: BG049460.D
 Acq: 24 Jul 2021 20:52




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020457


Tgt Ion:198 Resp: 15699
 Ion Ratio Lower Upper
 198 100
 182 5.2 4.2 6.4
 77 41.0 20.8 31.2#



#67
 N-Nitrosodiphenylamine
 Concen: 20.792 ng/ul
 RT: 15.947 min Scan# 2197
 Delta R.T. 0.003 min
 Lab File: BG049460.D
 Acq: 24 Jul 2021 20:52

Tgt Ion:169 Resp: 80217
 Ion Ratio Lower Upper
 169 100
 168 70.1 57.3 85.9
 167 37.3 29.7 44.5









#68
4-Bromophenyl-phenylether
Concen: 19.993 ng/ul
RT: 16.629 min Scan# 2313
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:248 Resp: 32772
Ion Ratio Lower Upper
248 100
250 96.3 78.1 117.1
141 80.5 72.6 109.0



#69
Hexachlorobenzene
Concen: 19.975 ng/ul
RT: 16.752 min Scan# 2334
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:284 Resp: 37046
Ion Ratio Lower Upper
284 100
142 40.8 30.9 46.3
249 35.7 25.5 38.3









#70
Atrazine
Concen: 19.479 ng/ul
RT: 16.893 min Scan# 2358
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:200 Resp: 34324
Ion Ratio Lower Upper
200 100
173 22.3 17.6 26.4
215 46.0 33.5 50.3



#71
Pentachlorophenol
Concen: 21.919 ng/ul
RT: 17.099 min Scan# 2393
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:266 Resp: 18822
Ion Ratio Lower Upper
266 100
264 65.4 43.8 65.6
268 68.8 46.2 69.4









#72
Phenanthrene
Concen: 19.730 ng/ul
RT: 17.492 min Scan# 2460
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:178 Resp: 151206
Ion Ratio Lower Upper
178 100
179 15.2 13.2 19.8
176 19.3 15.0 22.6



#74
Anthracene
Concen: 19.836 ng/ul
RT: 17.580 min Scan# 2475
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:178 Resp: 154053
Ion Ratio Lower Upper
178 100
179 15.3 12.0 18.0
176 19.3 15.6 23.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.261 ng/uL
RT: 13.498 min Scan# 1780
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:216 Resp: 40530
Ion Ratio Lower Upper
216 100
214 76.6 60.7 91.1
218 48.5 40.2 60.4



#76
Pentachlorobenzene
Concen: 20.894 ng/uL
RT: 15.013 min Scan# 2038
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:250 Resp: 44221
Ion Ratio Lower Upper
250 100
248 67.9 48.3 72.5
252 61.7 52.6 78.8









#77
 Carbazole
 Concen: 19.396 ng/ul
 RT: 17.850 min Scan# 2521
 Delta R.T. 0.003 min
 Lab File: BG049460.D
 Acq: 24 Jul 2021 20:52




Instrument :
 BNA_G
 ClientSampleId :
 SST020457


Tgt Ion:167	Resp: 134671
Ion Ratio	Lower Upper
167 100	
166 23.2	17.5 26.3
139 14.0	11.4 17.2



#78
 Di-n-butylphthalate
 Concen: 19.919 ng/ul
 RT: 18.403 min Scan# 2615
 Delta R.T. 0.003 min
 Lab File: BG049460.D
 Acq: 24 Jul 2021 20:52

Tgt Ion:149	Resp: 171981
Ion Ratio	Lower Upper
149 100	
150 9.2	7.5 11.3
104 6.7	5.6 8.4









#80
Fluoranthene
Concen: 19.868 ng/ul
RT: 19.495 min Scan# 2801
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:202	Resp:	194709
Ion Ratio	Lower	Upper
202	100	
101	5.1	4.3 6.5
100	5.5	3.8 5.6



#82
Pyrene
Concen: 20.299 ng/ul
RT: 19.859 min Scan# 2863
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:202	Resp:	197512
Ion Ratio	Lower	Upper
202	100	
101	5.6	5.3 7.9
100	5.2	4.7 7.1









#83
Butylbenzylphthalate
Concen: 19.957 ng/ul
RT: 20.741 min Scan# 3013
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:	149	Resp:	75601
Ion Ratio	Lower	Upper	
149	100		
91	81.6	67.7	101.5
206	21.0	16.7	25.1



#84
3,3'-Dichlorobenzidine
Concen: 20.486 ng/ul
RT: 21.622 min Scan# 3163
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:	252	Resp:	74301
Ion Ratio	Lower	Upper	
252	100		
254	65.7	50.2	75.4
126	8.4	5.8	8.6









#85
Benzo(a)anthracene
Concen: 19.887 ng/ul
RT: 21.704 min Scan# 3177
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:228 Resp: 191928
Ion Ratio Lower Upper
228 100
229 19.1 15.4 23.2
226 26.3 21.1 31.7



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.090 ng/ul
RT: 21.604 min Scan# 3160
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:149 Resp: 112949
Ion Ratio Lower Upper
149 100
167 25.1 20.9 31.3
279 3.3 3.6 5.4#









#87
Chrysene
Concen: 19.571 ng/ul
RT: 21.775 min Scan# 3189
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:228 Resp: 180779
Ion Ratio Lower Upper
228 100
226 30.9 24.6 37.0
229 20.0 14.9 22.3



#89
Di-n-octyl phthalate
Concen: 18.888 ng/ul
RT: 22.832 min Scan# 3369
Delta R.T. -0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:149 Resp: 185327









#90
Benzo(b)fluoranthene
Concen: 19.463 ng/ul
RT: 23.960 min Scan# 3561
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:252 Resp: 203632
Ion Ratio Lower Upper
252 100
253 22.4 17.1 25.7
125 5.2 4.5 6.7



#91
Benzo(k)fluoranthene
Concen: 19.887 ng/ul
RT: 23.960 min Scan# 3561
Delta R.T. -0.062 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:252 Resp: 203632
Ion Ratio Lower Upper
252 100
253 22.4 17.3 25.9
125 5.2 4.6 7.0









#93
Benzo(a)pyrene
Concen: 19.548 ng/ul
RT: 24.841 min Scan# 3711
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52




Instrument :
BNA_G
ClientSampleId :
SSTD020457


Tgt Ion:252	Resp:	177423
Ion Ratio	Lower	Upper
252	100	
253	22.7	16.4 24.6
125	4.6	4.6 6.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 20.258 ng/ul
RT: 28.742 min Scan# 4375
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:276	Resp:	234615
Ion Ratio	Lower	Upper
276	100	
138	6.6	7.3 10.9#
277	20.1	19.2 28.8









#95
Dibenzo(a,h)anthracene
Concen: 20.183 ng/ul
RT: 28.800 min Scan# 4385
Delta R.T. 0.003 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Instrument :
BNA_G
ClientSampleId :
SSTD020457

Tgt Ion:278	Resp:	193428
Ion Ratio	Lower	Upper
278	100	
139	8.4	6.8 10.2
279	21.7	19.7 29.5



#96
Benzo(g,h,i)perylene
Concen: 20.223 ng/ul
RT: 29.911 min Scan# 4574
Delta R.T. 0.008 min
Lab File: BG049460.D
Acq: 24 Jul 2021 20:52

Tgt Ion:276	Resp:	194719
Ion Ratio	Lower	Upper
276	100	
138	10.8	10.8 16.2
277	23.7	20.6 31.0

