

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072521\
 Data File : BG049469.D
 Acq On : 25 Jul 2021 3:43
 Operator : CG/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020458

Quant Time: Jul 25 09:52:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG072421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 24 06:15:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.045	152	24221	20.000	ng/ul	0.00
20) Naphthalene-d8	10.871	136	96008	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.695	164	64111	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.444	188	148271	20.000	ng/ul	0.00
79) Chrysene-d12	21.727	240	156420	20.000	ng/ul	0.00
88) Perylene-d12	24.999	264	161814	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.428	96	4226	7.982	ng/uL	-0.01
4) Pyridine-d5	3.857	84	28969	19.051	ng/ul	0.00
7) Phenol-d5	7.199	99	41784	19.498	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.364	67	23825	19.630	ng/ul	0.00
11) 2-Chlorophenol-d4	7.575	132	31682	20.625	ng/ul	0.00
15) 4-Methylphenol-d8	8.756	113	31980	19.798	ng/ul	0.00
21) Nitrobenzene-d5	9.214	128	15582	21.110	ng/ul	0.00
24) 2-Nitrophenol-d4	9.943	143	17513	20.247	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.483	165	32006	20.345	ng/ul	0.00
31) 4-Chloroaniline-d4	11.000	131	40626	18.530	ng/ul	0.00
46) Dimethylphthalate-d6	14.096	166	100272	20.418	ng/ul	0.00
49) Acenaphthylene-d8	14.389	160	116612	20.444	ng/ul	0.00
54) 4-Nitrophenol-d4	14.889	143	16428	20.313	ng/ul	0.00
60) Fluorene-d10	15.688	176	86430	20.429	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.805	200	17849	19.427	ng/ul	0.00
73) Anthracene-d10	17.544	188	135774	19.703	ng/ul	0.00
81) Pyrene-d10	19.829	212	165017	20.632	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.770	264	174490	20.517	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.463	88	4257	5.883	ng/uL#	82
5) Pyridine	3.874	79	31553	19.399	ng/ul#	94
6) Benzaldehyde	7.182	77	30443	23.807	ng/ul	93
8) Phenol	7.223	94	40813	19.496	ng/ul	96
10) Bis(2-Chloroethyl)ether	7.458	93	27818	19.289	ng/ul	89
12) 2-Chlorophenol	7.605	128	30154	20.390	ng/ul	97
13) 2-Methylphenol	8.486	108	30925	19.383	ng/ul	83
14) 2,2'-oxybis(1-Chloropr...	8.574	45	32071	18.976	ng/ul#	95
16) Acetophenone	8.873	105	51754	19.908	ng/ul#	96
17) N-Nitroso-di-n-propyla...	8.856	70	27189	20.282	ng/ul#	97
18) 4-Methylphenol	8.815	108	33829	20.155	ng/ul	97
19) Hexachloroethane	9.138	117	13504	20.269	ng/ul	92
22) Nitrobenzene	9.255	77	45162	20.492	ng/ul	95
23) Isophorone	9.778	82	73615	19.933	ng/ul	99
25) 2-Nitrophenol	9.972	139	16870	20.724	ng/ul#	88
26) 2,4-Dimethylphenol	10.025	107	40619	20.937	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.266	93	37175	20.183	ng/ul#	98
29) 2,4-Dichlorophenol	10.512	162	31585	20.538	ng/ul	92
30) Naphthalene	10.918	128	98392	19.896	ng/ul	98
32) 4-Chloroaniline	11.023	127	41666	19.978	ng/ul#	88
33) Hexachlorobutadiene	11.206	225	24690	20.999	ng/ul#	86
34) Caprolactam	11.781	113	3715	7.103	ng/ul#	58
35) 4-Chloro-3-methylphenol	12.140	107	36450	21.065	ng/ul	99
36) 2-Methylnaphthalene	12.521	142	73029	20.176	ng/ul	90

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072521\
 Data File : BG049469.D
 Acq On : 25 Jul 2021 3:43
 Operator : CG/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020458

Quant Time: Jul 25 09:52:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG072421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 24 06:15:51 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.739	142	75357	20.603	ng/ul	88
39) 1,2,4,5-Tetrachloroben...	12.892	216	41236	21.641	ng/ul	98
40) Hexachlorocyclopentadiene	12.868	237	23063	20.087	ng/ul	97
41) 2,4,6-Trichlorophenol	13.126	196	24901	21.265	ng/ul#	78
42) 2,4,5-Trichlorophenol	13.209	196	26763	21.401	ng/ul	84
43) 1,1'-Biphenyl	13.526	154	95949	21.017	ng/ul#	96
44) 2-Chloronaphthalene	13.573	162	73268	20.624	ng/ul	96
45) 2-Nitroaniline	13.773	65	26141	20.735	ng/ul	95
47) Dimethylphthalate	14.143	163	97267	20.367	ng/ul	98
48) 2,6-Dinitrotoluene	14.266	165	19769	20.741	ng/ul#	92
50) Acenaphthylene	14.419	152	110711	20.324	ng/ul	96
51) 3-Nitroaniline	14.595	138	19492	20.505	ng/ul#	93
52) Acenaphthene	14.760	153	78705	20.145	ng/ul	94
53) 2,4-Dinitrophenol	14.801	184	11161	23.504	ng/ul	91
55) 4-Nitrophenol	14.901	109	22599	20.608	ng/ul	92
56) Dibenzofuran	15.094	168	111160	21.072	ng/ul	94
57) 2,4-Dinitrotoluene	15.047	165	28202	20.874	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.318	232	22900	20.179	ng/ul#	93
59) Diethylphthalate	15.506	149	99875	20.206	ng/ul	98
61) Fluorene	15.746	166	91448	20.420	ng/ul#	98
62) 4-Chlorophenyl-phenyle...	15.735	204	48900	20.976	ng/ul#	87
63) 4-Nitroaniline	15.758	138	20670	21.927	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.817	198	16108	19.045	ng/ul#	78
67) N-Nitrosodiphenylamine	15.946	169	79882	20.557	ng/ul	96
68) 4-Bromophenyl-phenylether	16.628	248	34006	20.598	ng/ul	95
69) Hexachlorobenzene	16.751	284	34896	18.682	ng/ul	95
70) Atrazine	16.892	200	34825	19.623	ng/ul	100
71) Pentachlorophenol	17.098	266	19041	22.016	ng/ul	96
72) Phenanthrene	17.485	178	151281	19.599	ng/ul	98
74) Anthracene	17.579	178	152899	19.547	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.497	216	41341	20.519	ng/ul	94
76) Pentachlorobenzene	15.012	250	42892	20.121	ng/ul	97
77) Carbazole	17.849	167	136438	19.510	ng/ul	100
78) Di-n-butylphthalate	18.402	149	168090	19.329	ng/ul	99
80) Fluoranthene	19.494	202	195817	20.163	ng/ul	99
82) Pyrene	19.859	202	194486	20.170	ng/ul	100
83) Butylbenzylphthalate	20.740	149	77979	20.772	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.615	252	74994	20.866	ng/ul	97
85) Benzo(a)anthracene	21.703	228	191382	20.012	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.603	149	112869	20.260	ng/ul	100
87) Chrysene	21.774	228	182612	19.950	ng/ul	96
89) Di-n-octyl phthalate	22.831	149	188031	19.768	ng/ul	100
90) Benzo(b)fluoranthene	24.018	252	197145	19.438	ng/ul	98
91) Benzo(k)fluoranthene	24.018	252	197145	19.862	ng/ul	98
93) Benzo(a)pyrene	24.840	252	180774	20.546	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	28.729	276	229706	20.461	ng/ul	96
95) Dibenzo(a,h)anthracene	28.799	278	189808	20.431	ng/ul	98
96) Benzo(g,h,i)perylene	29.904	276	187782	20.118	ng/ul#	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072521\
 Data File : BG049469.D
 Acq On : 25 Jul 2021 3:43
 Operator : CG/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020458

Quant Time: Jul 25 09:52:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG072421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 24 06:15:51 2021
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 5.883 ng/uL
RT: 3.463 min Scan# 72
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion: 88 Resp: 4257
Ion Ratio Lower Upper
88 100
43 19.4 29.5 44.3#
58 61.2 56.8 85.2



#5
Pyridine
Concen: 19.399 ng/uL
RT: 3.874 min Scan# 142
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion: 79 Resp: 31553
Ion Ratio Lower Upper
79 100
52 73.0 57.4 86.2
51 44.4 28.9 43.3#









#6
Benzaldehyde
Concen: 23.807 ng/ul
RT: 7.182 min Scan# 705
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion: 77 Resp: 30443
Ion Ratio Lower Upper
77 100
105 96.6 70.0 105.0
106 83.8 70.1 105.1



#8
Phenol
Concen: 19.496 ng/ul
RT: 7.223 min Scan# 712
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion: 94 Resp: 40813
Ion Ratio Lower Upper
94 100
65 34.2 28.8 43.2
66 47.1 40.3 60.5









#10
Bis(2-Chloroethyl)ether
Concen: 19.289 ng/ul
RT: 7.458 min Scan# 752
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion: 93 Resp: 27818
Ion Ratio Lower Upper
93 100
63 90.8 62.4 93.6
95 32.1 26.0 39.0



#12
2-Chlorophenol
Concen: 20.390 ng/ul
RT: 7.605 min Scan# 777
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion: 128 Resp: 30154
Ion Ratio Lower Upper
128 100
64 56.7 44.2 66.4
130 33.3 24.1 36.1





#13

2-Methylphenol

Concen: 19.383 ng/ul

RT: 8.486 min Scan# 927

Delta R.T. -0.004 min

Lab File: BG049469.D





Acq: 25 Jul 2021 3:43

Tgt Ion:108 Resp: 30925

Ion Ratio Lower Upper

108 100

107 73.1 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 18.976 ng/ul

RT: 8.574 min Scan# 942

Delta R.T. -0.004 min

Lab File: BG049469.D

Acq: 25 Jul 2021 3:43




Tgt Ion: 45 Resp: 32071


Ion Ratio Lower Upper

45 100

77 22.6 17.0 25.4

79 15.6 9.7 14.5#









#16
Acetophenone
Concen: 19.908 ng/ul
RT: 8.873 min Scan# 993
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:105 Resp: 51754
Ion Ratio Lower Upper
105 100
77 84.0 66.4 99.6
51 32.6 20.8 31.2#



#17
N-Nitroso-di-n-propylamine
Concen: 20.282 ng/ul
RT: 8.856 min Scan# 990
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion: 70 Resp: 27189
Ion Ratio Lower Upper
70 100
42 53.5 44.2 66.4
101 9.2 5.4 8.0#
130 17.6 12.4 18.6









#18
4-Methylphenol
Concen: 20.155 ng/ul
RT: 8.815 min Scan# 983
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:108 Resp: 33829
Ion Ratio Lower Upper
108 100
107 109.7 85.4 128.2



#19
Hexachloroethane
Concen: 20.269 ng/ul
RT: 9.138 min Scan# 1038
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:117 Resp: 13504
Ion Ratio Lower Upper
117 100
201 111.4 100.1 150.1
199 61.5 49.8 74.6









#22
Nitrobenzene
Concen: 20.492 ng/ul
RT: 9.255 min Scan# 1058
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:	77	Resp:	45162
Ion Ratio	Lower	Upper	
77	100		
123	30.4	27.5	41.3
65	13.3	10.8	16.2



#23
Isophorone
Concen: 19.933 ng/ul
RT: 9.778 min Scan# 1147
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:	82	Resp:	73615
Ion Ratio	Lower	Upper	
82	100		
95	8.0	7.4	11.0
138	13.4	10.7	16.1









#25
2-Nitrophenol
Concen: 20.724 ng/ul
RT: 9.972 min Scan# 1180
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:139 Resp: 16870
Ion Ratio Lower Upper
139 100
65 70.3 49.8 74.6
109 48.9 31.9 47.9#



#26
2,4-Dimethylphenol
Concen: 20.937 ng/ul
RT: 10.025 min Scan# 1189
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:107 Resp: 40619
Ion Ratio Lower Upper
107 100
121 42.9 34.1 51.1
122 74.7 58.8 88.2









#27
Bis(2-Chloroethoxy)methane
Concen: 20.183 ng/ul
RT: 10.266 min Scan# 1230
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion: 93	Resp:	37175
Ion Ratio	Lower	Upper
93	100	
95	33.4	26.8 40.2
123	7.1	8.3 12.5#



#29
2,4-Dichlorophenol
Concen: 20.538 ng/ul
RT: 10.512 min Scan# 1272
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion: 162	Resp:	31585
Ion Ratio	Lower	Upper
162	100	
164	59.5	52.2 78.2
98	37.0	25.4 38.2









#30
Naphthalene
Concen: 19.896 ng/ul
RT: 10.918 min Scan# 1341
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.1	7.8	11.8
127	13.7	11.0	16.6



#32
4-Chloroaniline
Concen: 19.978 ng/ul
RT: 11.023 min Scan# 1359
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion	Ratio	Lower	Upper
127	100		
129	32.6	21.3	31.9#









#33
Hexachlorobutadiene
Concen: 20.999 ng/ul
RT: 11.206 min Scan# 1390
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:225 Resp: 24690
Ion Ratio Lower Upper
225 100
223 48.6 52.1 78.1#
227 60.6 52.4 78.6



#34
Caprolactam
Concen: 7.103 ng/ul
RT: 11.781 min Scan# 1488
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:113 Resp: 3715
Ion Ratio Lower Upper
113 100
55 185.1 116.6 174.8#
56 179.3 95.0 142.6#









#35
 4-Chloro-3-methylphenol
 Concen: 21.065 ng/ul
 RT: 12.140 min Scan# 1549
 Delta R.T. -0.010 min
 Lab File: BG049469.D
 Acq: 25 Jul 2021 3:43




Instrument :
 BNA_G
 ClientSampleId :
 SST020458


Tgt Ion:107 Resp: 36450
 Ion Ratio Lower Upper
 107 100
 144 20.2 17.0 25.6
 142 66.6 53.3 79.9



#36
 2-Methylnaphthalene
 Concen: 20.176 ng/ul
 RT: 12.521 min Scan# 1614
 Delta R.T. -0.004 min
 Lab File: BG049469.D
 Acq: 25 Jul 2021 3:43

Tgt Ion:142 Resp: 73029
 Ion Ratio Lower Upper
 142 100
 141 96.9 69.9 104.9





#37

1-Methylnaphthalene

Concen: 20.603 ng/ul

RT: 12.739 min Scan# 1651

Delta R.T. -0.004 min

Lab File: BG049469.D

Acq: 25 Jul 2021 3:43





Instrument :

BNA_G

ClientSampleId :

SSTD020458

Tgt Ion:142	Resp:	75357
Ion Ratio	Lower	Upper
142	100	
141	90.2	82.1 123.1
116	4.5	3.8 5.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.641 ng/ul




RT: 12.892 min Scan# 1677


Delta R.T. 0.002 min

Lab File: BG049469.D

Acq: 25 Jul 2021 3:43

Tgt Ion:216	Resp:	41236
Ion Ratio	Lower	Upper
216	100	
214	74.7	58.6 88.0
179	21.3	17.0 25.4
108	19.3	14.4 21.6









#40
Hexachlorocyclopentadiene
Concen: 20.087 ng/ul
RT: 12.868 min Scan# 1673
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:	237	Resp:	23063
Ion Ratio	Lower	Upper	
237	100		
235	58.3	48.2	72.4
272	14.5	10.1	15.1



#41
2,4,6-Trichlorophenol
Concen: 21.265 ng/ul
RT: 13.126 min Scan# 1717
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:	196	Resp:	24901
Ion Ratio	Lower	Upper	
196	100		
198	98.9	61.6	92.4#
200	32.6	21.3	31.9#









#42
2,4,5-Trichlorophenol
Concen: 21.401 ng/ul
RT: 13.209 min Scan# 1731
Delta R.T. 0.008 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:196	Resp:	26763
Ion Ratio	Lower	Upper
196	100	
198	84.7	81.4 122.0
200	26.6	26.6 39.8



#43
1,1'-Biphenyl
Concen: 21.017 ng/ul
RT: 13.526 min Scan# 1785
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:154	Resp:	95949
Ion Ratio	Lower	Upper
154	100	
153	39.5	33.3 49.9
76	11.9	7.0 10.6#









#44
2-Chloronaphthalene
Concen: 20.624 ng/ul
RT: 13.573 min Scan# 1793
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:162 Resp: 73268
Ion Ratio Lower Upper
162 100
164 34.6 24.5 36.7
127 43.8 35.7 53.5



#45
2-Nitroaniline
Concen: 20.735 ng/ul
RT: 13.773 min Scan# 1827
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion: 65 Resp: 26141
Ion Ratio Lower Upper
65 100
92 59.5 51.8 77.6
138 87.9 68.1 102.1





#47

Dimethylphthalate

Concen: 20.367 ng/ul

RT: 14.143 min Scan# 1890

Delta R.T. 0.002 min

Lab File: BG049469.D

Acq: 25 Jul 2021 3:43





Instrument :

BNA_G

ClientSampleId :

SSTD020458

Tgt Ion:163	Resp:	97267
Ion Ratio	Lower	Upper
163	100	
194	4.0	3.0 4.6
164	11.4	8.5 12.7



#48

2,6-Dinitrotoluene

Concen: 20.741 ng/ul




RT: 14.266 min Scan# 1911


Delta R.T. 0.002 min

Lab File: BG049469.D

Acq: 25 Jul 2021 3:43

Tgt Ion:165	Resp:	19769
Ion Ratio	Lower	Upper
165	100	
89	79.8	67.4 101.2
121	22.7	24.6 36.8









#50
Acenaphthylene
Concen: 20.324 ng/ul
RT: 14.419 min Scan# 1937
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:152 Resp: 110711
Ion Ratio Lower Upper
152 100
151 20.3 14.4 21.6
153 12.9 9.5 14.3



#51
3-Nitroaniline
Concen: 20.505 ng/ul
RT: 14.595 min Scan# 1967
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:138 Resp: 19492
Ion Ratio Lower Upper
138 100
108 15.0 8.2 12.2#
92 123.4 93.0 139.6









#52
Acenaphthene
Concen: 20.145 ng/ul
RT: 14.760 min Scan# 1995
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:153 Resp: 78705
Ion Ratio Lower Upper
153 100
152 45.8 37.4 56.0
154 91.4 66.7 100.1



#53
2,4-Dinitrophenol
Concen: 23.504 ng/ul
RT: 14.801 min Scan# 2002
Delta R.T. -0.003 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:184 Resp: 11161
Ion Ratio Lower Upper
184 100
63 101.1 85.5 128.3
154 82.4 75.8 113.6









#55
4-Nitrophenol
Concen: 20.608 ng/ul
RT: 14.901 min Scan# 2019
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:109 Resp: 22599
Ion Ratio Lower Upper
109 100
139 66.2 57.0 85.4
65 94.5 83.1 124.7



#56
Dibenzofuran
Concen: 21.072 ng/ul
RT: 15.094 min Scan# 2052
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:168 Resp: 111160
Ion Ratio Lower Upper
168 100
139 40.7 35.5 53.3









#57
2,4-Dinitrotoluene
Concen: 20.874 ng/ul
RT: 15.047 min Scan# 2044
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:165 Resp: 28202
Ion Ratio Lower Upper
165 100
63 60.4 48.2 72.4
182 2.0 1.4 2.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.179 ng/ul
RT: 15.318 min Scan# 2090
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:232 Resp: 22900
Ion Ratio Lower Upper
232 100
131 43.7 39.6 59.4
130 1.7 3.2 4.8#
166 31.2 22.6 34.0









#59
Diethylphthalate
Concen: 20.206 ng/ul
RT: 15.506 min Scan# 2122
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:	149	Resp:	99875
Ion Ratio	Lower	Upper	
149	100		
177	20.6	16.0	24.0
150	13.0	9.5	14.3



#61
Fluorene
Concen: 20.420 ng/ul
RT: 15.746 min Scan# 2163
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:	166	Resp:	91448
Ion Ratio	Lower	Upper	
166	100		
165	103.4	81.8	122.8
167	10.8	11.0	16.4#









#62
4-Chlorophenyl-phenylether
Concen: 20.976 ng/ul
RT: 15.735 min Scan# 2161
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:204 Resp: 48900
Ion Ratio Lower Upper
204 100
206 35.3 23.5 35.3#
141 67.0 45.4 68.0



#63
4-Nitroaniline
Concen: 21.927 ng/ul
RT: 15.758 min Scan# 2165
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:138 Resp: 20670
Ion Ratio Lower Upper
138 100
92 54.4 49.8 74.8
108 137.5 105.4 158.2









#66
4,6-Dinitro-2-methylphenol
Concen: 19.045 ng/ul
RT: 15.817 min Scan# 2175
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:198 Resp: 16108
Ion Ratio Lower Upper
198 100
182 2.8 4.2 6.4#
77 38.9 20.8 31.2#



#67
N-Nitrosodiphenylamine
Concen: 20.557 ng/ul
RT: 15.946 min Scan# 2197
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:169 Resp: 79882
Ion Ratio Lower Upper
169 100
168 66.3 57.3 85.9
167 37.6 29.7 44.5









#68
4-Bromophenyl-phenylether
Concen: 20.598 ng/ul
RT: 16.628 min Scan# 2313
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:	248	Resp:	34006
Ion Ratio	Lower	Upper	
248	100		
250	97.8	78.1	117.1
141	81.1	72.6	109.0



#69
Hexachlorobenzene
Concen: 18.682 ng/ul
RT: 16.751 min Scan# 2334
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:	284	Resp:	34896
Ion Ratio	Lower	Upper	
284	100		
142	39.7	30.9	46.3
249	36.9	25.5	38.3









#70
Atrazine
Concen: 19.623 ng/ul
RT: 16.892 min Scan# 2358
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:200 Resp: 34825
Ion Ratio Lower Upper
200 100
173 21.8 17.6 26.4
215 42.0 33.5 50.3



#71
Pentachlorophenol
Concen: 22.016 ng/ul
RT: 17.098 min Scan# 2393
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:266 Resp: 19041
Ion Ratio Lower Upper
266 100
264 56.7 43.8 65.6
268 61.7 46.2 69.4









#72
 Phenanthrene
 Concen: 19.599 ng/ul
 RT: 17.485 min Scan# 2459
 Delta R.T. -0.004 min
 Lab File: BG049469.D
 Acq: 25 Jul 2021 3:43




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020458


Tgt Ion:178 Resp: 151281
 Ion Ratio Lower Upper
 178 100
 179 15.4 13.2 19.8
 176 19.5 15.0 22.6



#74
 Anthracene
 Concen: 19.547 ng/ul
 RT: 17.579 min Scan# 2475
 Delta R.T. -0.004 min
 Lab File: BG049469.D
 Acq: 25 Jul 2021 3:43

Tgt Ion:178 Resp: 152899
 Ion Ratio Lower Upper
 178 100
 179 15.4 12.0 18.0
 176 19.1 15.6 23.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.519 ng/uL
RT: 13.497 min Scan# 1780
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:216 Resp: 41341
Ion Ratio Lower Upper
216 100
214 82.0 60.7 91.1
218 47.5 40.2 60.4



#76
Pentachlorobenzene
Concen: 20.121 ng/uL
RT: 15.012 min Scan# 2038
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:250 Resp: 42892
Ion Ratio Lower Upper
250 100
248 60.6 48.3 72.5
252 61.5 52.6 78.8









#77
 Carbazole
 Concen: 19.510 ng/ul
 RT: 17.849 min Scan# 2521
 Delta R.T. 0.002 min
 Lab File: BG049469.D
 Acq: 25 Jul 2021 3:43




Instrument :
 BNA_G
 ClientSampleId :
 SST020458


Tgt Ion:167 Resp: 136438
 Ion Ratio Lower Upper
 167 100
 166 21.9 17.5 26.3
 139 14.3 11.4 17.2



#78
 Di-n-butylphthalate
 Concen: 19.329 ng/ul
 RT: 18.402 min Scan# 2615
 Delta R.T. 0.002 min
 Lab File: BG049469.D
 Acq: 25 Jul 2021 3:43

Tgt Ion:149 Resp: 168090
 Ion Ratio Lower Upper
 149 100
 150 8.6 7.5 11.3
 104 7.2 5.6 8.4









#80
Fluoranthene
Concen: 20.163 ng/ul
RT: 19.494 min Scan# 2801
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:202 Resp: 195817
Ion Ratio Lower Upper
202 100
101 5.9 4.3 6.5
100 5.0 3.8 5.6



#82
Pyrene
Concen: 20.170 ng/ul
RT: 19.859 min Scan# 2863
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:202 Resp: 194486
Ion Ratio Lower Upper
202 100
101 6.8 5.3 7.9
100 5.9 4.7 7.1









#83
Butylbenzylphthalate
Concen: 20.772 ng/ul
RT: 20.740 min Scan# 3013
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:	149	Resp:	77979
Ion Ratio	Lower	Upper	
149	100		
91	79.3	67.7	101.5
206	18.4	16.7	25.1



#84
3,3'-Dichlorobenzidine
Concen: 20.866 ng/ul
RT: 21.615 min Scan# 3162
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:	252	Resp:	74994
Ion Ratio	Lower	Upper	
252	100		
254	64.8	50.2	75.4
126	8.5	5.8	8.6









#85
Benzo(a)anthracene
Concen: 20.012 ng/ul
RT: 21.703 min Scan# 3177
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:228 Resp: 191382
Ion Ratio Lower Upper
228 100
229 19.6 15.4 23.2
226 28.0 21.1 31.7



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.260 ng/ul
RT: 21.603 min Scan# 3160
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:149 Resp: 112869
Ion Ratio Lower Upper
149 100
167 26.2 20.9 31.3
279 4.9 3.6 5.4









#87
Chrysene
Concen: 19.950 ng/ul
RT: 21.774 min Scan# 3189
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:228 Resp: 182612
Ion Ratio Lower Upper
228 100
226 28.6 24.6 37.0
229 20.0 14.9 22.3



#89
Di-n-octyl phthalate
Concen: 19.768 ng/ul
RT: 22.831 min Scan# 3369
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:149 Resp: 188031









#90
Benzo(b)fluoranthene
Concen: 19.438 ng/ul
RT: 24.018 min Scan# 3571
Delta R.T. 0.060 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:252 Resp: 197145
Ion Ratio Lower Upper
252 100
253 22.7 17.1 25.7
125 5.7 4.5 6.7



#91
Benzo(k)fluoranthene
Concen: 19.862 ng/ul
RT: 24.018 min Scan# 3571
Delta R.T. -0.004 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:252 Resp: 197145
Ion Ratio Lower Upper
252 100
253 22.7 17.3 25.9
125 5.7 4.6 7.0









#93
Benzo(a)pyrene
Concen: 20.546 ng/ul
RT: 24.840 min Scan# 3711
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43




Instrument :
BNA_G
ClientSampleId :
SSTD020458


Tgt Ion:252 Resp: 180774
Ion Ratio Lower Upper
252 100
253 21.2 16.4 24.6
125 5.5 4.6 6.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 20.461 ng/ul
RT: 28.729 min Scan# 4373
Delta R.T. -0.010 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:276 Resp: 229706
Ion Ratio Lower Upper
276 100
138 8.9 7.3 10.9
277 26.7 19.2 28.8









#95
Dibenzo(a,h)anthracene
Concen: 20.431 ng/ul
RT: 28.799 min Scan# 4385
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Instrument :
BNA_G
ClientSampleId :
SSTD020458

Tgt Ion:278 Resp: 189808
Ion Ratio Lower Upper
278 100
139 8.1 6.8 10.2
279 23.2 19.7 29.5



#96
Benzo(g,h,i)perylene
Concen: 20.118 ng/ul
RT: 29.904 min Scan# 4573
Delta R.T. 0.002 min
Lab File: BG049469.D
Acq: 25 Jul 2021 3:43

Tgt Ion:276 Resp: 187782
Ion Ratio Lower Upper
276 100
138 9.2 10.8 16.2#
277 23.3 20.6 31.0

