

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072522\
 Data File : BG054374.D
 Acq On : 25 Jul 2022 10:57
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 26 00:53:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 15 15:54:11 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	97	-0.01
2	1,4-Dioxane	40.000	41.054	-2.6	103	-0.02
3	Pyridine	40.000	42.309	-5.8	97	-0.01
4	n-Nitrosodimethylamine	40.000	44.969	-12.4	104	-0.01
5 S	2-Fluorophenol	80.000	84.941	-6.2	102	0.00
6	Aniline	40.000	41.680	-4.2	99	-0.01
7 S	Phenol-d6	80.000	83.361	-4.2	99	0.00
8	2-Chlorophenol	40.000	41.408	-3.5	100	0.00
9	Benzaldehyde	40.000	44.392	-11.0	107	-0.01
10 C	Phenol	40.000	41.354	-3.4	100	0.00
11	bis(2-Chloroethyl)ether	40.000	41.373	-3.4	101	-0.01
12	1,3-Dichlorobenzene	40.000	41.332	-3.3	99	-0.01
13 C	1,4-Dichlorobenzene	40.000	40.487	-1.2	97	-0.01
14	1,2-Dichlorobenzene	40.000	41.462	-3.7	100	-0.01
15	Benzyl Alcohol	40.000	41.158	-2.9	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.239	-3.1	97	-0.01
17	2-Methylphenol	40.000	41.641	-4.1	99	0.00
18	Hexachloroethane	40.000	43.000	-7.5	103	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	39.940	0.2	93	-0.01
20	3+4-Methylphenols	40.000	40.889	-2.2	97	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	95	-0.01
22	Acetophenone	40.000	42.506	-6.3	98	0.00
23 S	Nitrobenzene-d5	80.000	84.570	-5.7	97	0.00
24	Nitrobenzene	40.000	42.592	-6.5	100	0.00
25	Isophorone	40.000	42.054	-5.1	96	0.00
26 C	2-Nitrophenol	40.000	42.112	-5.3	95	-0.01
27	2,4-Dimethylphenol	40.000	41.623	-4.1	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.790	-2.0	93	-0.01
29 C	2,4-Dichlorophenol	40.000	42.247	-5.6	97	0.00
30	1,2,4-Trichlorobenzene	40.000	41.768	-4.4	97	-0.01
31	Naphthalene	40.000	41.933	-4.8	97	-0.01
32	Benzoic acid	40.000	48.144	-20.4	123	-0.02
33	4-Chloroaniline	40.000	42.088	-5.2	98	0.00
34 C	Hexachlorobutadiene	40.000	41.493	-3.7	95	0.00
35	Caprolactam	40.000	45.316	-13.3	110	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.009	-5.0	100	0.00
37	2-Methylnaphthalene	40.000	41.235	-3.1	97	-0.01
38	1-Methylnaphthalene	40.000	40.760	-1.9	96	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	99	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.465	-1.2	94	0.00
41 P	Hexachlorocyclopentadiene	40.000	34.721	13.2	86	0.00
42 S	2,4,6-Tribromophenol	80.000	86.732	-8.4	103	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.784	-7.0	99	0.00
44	2,4,5-Trichlorophenol	40.000	42.662	-6.7	103	0.00
45 S	2-Fluorobiphenyl	80.000	80.772	-1.0	95	0.00
46	1,1'-Biphenyl	40.000	40.511	-1.3	97	-0.01
47	2-Chloronaphthalene	40.000	40.592	-1.5	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072522\
 Data File : BG054374.D
 Acq On : 25 Jul 2022 10:57
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 26 00:53:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 15 15:54:11 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.702	-9.3	102	0.00
49	Acenaphthylene	40.000	41.352	-3.4	98	0.00
50	Dimethylphthalate	40.000	41.364	-3.4	99	0.00
51	2,6-Dinitrotoluene	40.000	41.861	-4.7	100	0.00
52 C	Acenaphthene	40.000	41.317	-3.3	99	0.00
53	3-Nitroaniline	40.000	43.253	-8.1	104	0.00
54 P	2,4-Dinitrophenol	40.000	40.221	-0.6	104	0.00
55	Dibenzofuran	40.000	41.065	-2.7	99	0.00
56 P	4-Nitrophenol	40.000	41.169	-2.9	97	0.02
57	2,4-Dinitrotoluene	40.000	42.689	-6.7	101	0.00
58	Fluorene	40.000	42.006	-5.0	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.130	-2.8	96	0.00
60	Diethylphthalate	40.000	42.464	-6.2	102	0.00
61	4-Chlorophenyl-phenylether	40.000	41.482	-3.7	99	0.00
62	4-Nitroaniline	40.000	44.757	-11.9	107	0.00
63	Azobenzene	40.000	42.019	-5.0	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.601	-4.0	100	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.149	-2.9	102	0.00
67	4-Bromophenyl-phenylether	40.000	40.513	-1.3	100	0.00
68	Hexachlorobenzene	40.000	40.961	-2.4	101	0.00
69	Atrazine	40.000	41.361	-3.4	104	0.00
70 C	Pentachlorophenol	40.000	37.995	5.0	96	0.00
71	Phenanthrene	40.000	41.168	-2.9	103	0.00
72	Anthracene	40.000	41.423	-3.6	103	0.00
73	Carbazole	40.000	41.923	-4.8	105	0.00
74	Di-n-butylphthalate	40.000	42.455	-6.1	107	0.00
75 C	Fluoranthene	40.000	40.414	-1.0	102	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzydine	40.000	42.515	-6.3	106	0.00
78	Pyrene	40.000	42.733	-6.8	102	0.00
79 S	Terphenyl-d14	80.000	84.787	-6.0	101	0.00
80	Butylbenzylphthalate	40.000	42.039	-5.1	102	0.00
81	Benzo(a)anthracene	40.000	41.030	-2.6	100	0.00
82	3,3'-Dichlorobenzidine	40.000	41.776	-4.4	100	0.00
83	Chrysene	40.000	41.201	-3.0	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.922	-4.8	102	0.00
85 c	Di-n-octyl phthalate	40.000	41.339	-3.3	99	0.00
86 I	Perylene-d12	20.000	20.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.830	-2.1	99	0.01
88	Benzo(b)fluoranthene	40.000	40.670	-1.7	97	0.00
89	Benzo(k)fluoranthene	40.000	40.700	-1.8	99	0.00
90 C	Benzo(a)pyrene	40.000	40.712	-1.8	97	0.00
91	Dibenzo(a,h)anthracene	40.000	41.453	-3.6	100	0.00
92	Benzo(g,h,i)perylene	40.000	40.934	-2.3	99	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072522\
Data File : BG054374.D
Acq On : 25 Jul 2022 10:57
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jul 26 00:53:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 15 15:54:11 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0