

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072618\
 Data File : BG035912.D
 Acq On : 26 Jul 2018 17:14
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 27 08:38:09 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG071218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 25 11:04:14 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.613	0.488	20.4	87	0.00
3	Pyridine	1.601	1.402	12.4	86	0.00
4	n-Nitrosodimethylamine	0.687	0.550	19.9	82	0.00
5 S	2-Fluorophenol	1.205	1.060	12.0	89	0.00
6	Aniline	2.330	2.135	8.4	93	0.00
7 S	Phenol-d6	1.797	1.718	4.4	96	0.00
8	2-Chlorophenol	1.225	1.224	0.1	101	0.00
9	Benzaldehyde	1.103	0.944	14.4	85	0.00
10 C	Phenol	1.816	1.760	3.1	97	0.00
11	bis(2-Chloroethyl)ether	1.422	1.347	5.3	96	0.00
12	1,3-Dichlorobenzene	1.480	1.391	6.0	97	0.00
13 C	1,4-Dichlorobenzene	1.503	1.413	6.0	96	0.00
14	1,2-Dichlorobenzene	1.444	1.408	2.5	100	0.00
15	Benzyl Alcohol	1.632	1.513	7.3	94	0.00
16	2,2'-oxybis(1-Chloropropane	1.192	1.064	10.7	91	0.00
17	2-Methylphenol	1.235	1.202	2.7	97	0.00
18	Hexachloroethane	0.575	0.531	7.7	96	0.00
19 P	n-Nitroso-di-n-propylamine	1.513	1.458	3.6	100	0.00
20	3+4-Methylphenols	1.713	1.699	0.8	96	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
22	Acetophenone	0.622	0.586	5.8	98	0.00
23 S	Nitrobenzene-d5	0.479	0.463	3.3	98	0.00
24	Nitrobenzene	0.481	0.454	5.6	97	0.00
25	Isophorone	0.889	0.836	6.0	96	0.00
26 C	2-Nitrophenol	0.171	0.192	-12.3	111	0.00
27	2,4-Dimethylphenol	0.289	0.287	0.7	100	0.00
28	bis(2-Chloroethoxy)methane	0.490	0.468	4.5	97	0.00
29 C	2,4-Dichlorophenol	0.330	0.349	-5.8	104	0.00
30	1,2,4-Trichlorobenzene	0.405	0.415	-2.5	106	0.00
31	Naphthalene	1.042	1.023	1.8	103	0.00
32	Benzoic acid	0.195	0.167	14.4	86	0.00
33	4-Chloroaniline	0.454	0.456	-0.4	104	0.00
34 C	Hexachlorobutadiene	0.316	0.313	0.9	102	0.00
35	Caprolactam	0.142	0.132	7.0	99	0.00
36 C	4-Chloro-3-methylphenol	0.443	0.456	-2.9	103	0.00
37	2-Methylnaphthalene	0.776	0.801	-3.2	106	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
39	1,2,4,5-Tetrachlorobenzene	0.755	0.722	4.4	109	0.00
40 P	Hexachlorocyclopentadiene	0.396	0.371	6.3	102	0.00
41 S	2,4,6-Tribromophenol	0.264	0.267	-1.1	112	0.00
42 C	2,4,6-Trichlorophenol	0.441	0.443	-0.5	106	0.00
43	2,4,5-Trichlorophenol	0.494	0.482	2.4	111	0.00
44 S	2-Fluorobiphenyl	1.493	1.426	4.5	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072618\
 Data File : BG035912.D
 Acq On : 26 Jul 2018 17:14
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 27 08:38:09 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG071218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 25 11:04:14 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.551	1.487	4.1	107	0.00
46	2-Chloronaphthalene	1.206	1.176	2.5	111	0.00
47	2-Nitroaniline	0.426	0.406	4.7	105	0.00
48	Acenaphthylene	2.020	1.958	3.1	109	0.00
49	Dimethylphthalate	1.752	1.628	7.1	107	0.00
50	2,6-Dinitrotoluene	0.357	0.367	-2.8	113	0.00
51 C	Acenaphthene	1.259	1.198	4.8	109	0.00
52	3-Nitroaniline	0.365	0.347	4.9	103	0.00
53 P	2,4-Dinitrophenol	0.158	0.147	7.0	103	0.00
54	Dibenzofuran	2.002	1.939	3.1	109	0.00
55 P	4-Nitrophenol	0.260	0.235	9.6	98	0.00
56	2,4-Dinitrotoluene	0.512	0.529	-3.3	110	0.00
57	Fluorene	1.678	1.623	3.3	111	0.00
58	2,3,4,6-Tetrachlorophenol	0.473	0.467	1.3	109	0.00
59	Diethylphthalate	1.802	1.652	8.3	104	0.00
60	4-Chlorophenyl-phenylether	0.986	0.968	1.8	112	0.00
61	4-Nitroaniline	0.382	0.361	5.5	102	0.00
62	Azobenzene	1.777	1.613	9.2	103	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
64	4,6-Dinitro-2-methylphenol	0.111	0.119	-7.2	115	0.00
65 c	n-Nitrosodiphenylamine	0.569	0.569	0.0	108	0.00
66	4-Bromophenyl-phenylether	0.232	0.241	-3.9	110	0.00
67	Hexachlorobenzene	0.239	0.243	-1.7	110	0.00
68	Atrazine	0.234	0.210	10.3	93	0.00
69 C	Pentachlorophenol	0.146	0.138	5.5	99	0.00
70	Phenanthrene	0.998	0.968	3.0	106	0.00
71	Anthracene	0.994	0.970	2.4	106	0.00
72	Carbazole	0.944	0.881	6.7	101	0.00
73	Di-n-butylphthalate	1.115	1.027	7.9	99	0.00
74 C	Fluoranthene	1.344	1.262	6.1	102	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
76	Benzidine	0.526	0.422	19.8	90	0.00
77	Pyrene	1.265	1.185	6.3	102	0.00
78 S	Terphenyl-d14	0.907	0.863	4.9	103	0.00
79	Butylbenzylphthalate	0.452	0.447	1.1	103	0.00
80	Benzo(a)anthracene	1.246	1.174	5.8	102	0.00
81	3,3'-Dichlorobenzidine	0.435	0.437	-0.5	106	0.00
82	Chrysene	1.155	1.099	4.8	103	0.00
83	Bis(2-ethylhexyl)phthalate	0.615	0.616	-0.2	104	0.00
84 c	Di-n-octyl phthalate	1.029	1.044	-1.5	105	0.00
85	Indeno(1,2,3-cd)pyrene	1.279	1.272	0.5	106	0.00
86 I	Perylene-d12	1.000	1.000	0.0	106	0.00
87	Benzo(b)fluoranthene	1.216	1.176	3.3	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072618\
Data File : BG035912.D
Acq On : 26 Jul 2018 17:14
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jul 27 08:38:09 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG071218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 25 11:04:14 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.188	1.141	4.0	104	0.00
89 C	Benzo(a)pyrene	1.131	1.102	2.6	105	0.00
90	Dibenzo(a,h)anthracene	1.110	1.081	2.6	105	0.00
91	Benzo(g,h,i)perylene	1.086	1.061	2.3	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0