


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG072715\
Data File : BG018109.D
Acq On : 27 Jul 2015 10:21
Operator : TP/UM
Sample : DFTPP20
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP20

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG072315.M
Title : SVOA CALIBRATION
Last Update : Tue Jul 28 01:54:27 2015



AutoFind: Scans 2503, 2504, 2505; Background Corrected with Scan 2498

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.0	22122	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	32.9	26920	PASS
70	69	0.00	2	0.2	51	PASS
127	198	10	80	46.9	38362	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	81829	PASS
199	198	5	9	6.9	5641	PASS
275	198	10	60	27.2	22294	PASS
365	198	1	100	4.0	3302	PASS
441	442	0.01	24	14.8	10838	PASS
442	198	50	100	89.2	73010	PASS
443	442	15	24	20.1	14663	PASS

Scan 2498 (17.364 min): BG018109.D
DFTPP20

m/z	Abundance
40.00	389.0
44.00	696.0
167.00	243.0
207.00	422.0

Instrument :
BNA_G
ClientSampleId :
DFTPP20

m/z	Abundance
37.00	160.0
38.00	253.0
39.00	1581.0
39.90	540.0
44.00	889.0
49.00	200.0
50.00	4944.0
51.00	19816.0
52.00	1254.0
55.00	228.0
56.00	522.0
57.00	1399.0
61.10	285.0
62.00	330.0
63.00	923.0
65.00	507.0
69.00	22768.0
73.00	314.0
74.00	2554.0
75.00	4132.0
76.00	1520.0
77.00	27960.0
78.00	1548.0
79.00	1778.0
80.00	1192.0
81.00	2143.0
82.00	433.0
83.00	435.0
85.00	430.0
86.00	657.0
86.90	408.0
91.00	495.0
91.90	523.0
93.00	2856.0
94.10	186.0
96.90	151.0
98.00	2569.0
99.00	1880.0
101.00	1135.0
103.00	491.0
103.80	674.0
105.00	714.0
105.90	190.0
107.00	8413.0
108.00	1293.0
109.00	534.0
110.00	15486.0
111.00	2388.0
112.00	408.0
112.90	185.0
117.00	7322.0
117.90	463.0
120.00	178.0
121.80	359.0
123.00	704.0
123.90	516.0
124.90	320.0
127.00	32056.0
128.00	2455.0
129.00	12280.0
130.00	1270.0
130.90	312.0
133.90	431.0
135.00	1157.0
136.00	448.0
137.00	890.0
138.10	220.0
141.00	1843.0
142.00	484.0
142.90	307.0
145.90	331.0
147.00	1028.0
147.90	1422.0
149.00	361.0
150.10	185.0
151.00	260.0
153.00	720.0
153.90	561.0
154.90	731.0
156.00	1389.0
156.90	285.0
157.90	447.0
158.90	220.0
159.90	606.0
161.00	820.0
161.90	238.0
165.00	525.0
166.00	625.0
167.00	2897.0
168.00	1812.0
168.90	287.0

Instrument :
BNA_G
ClientSampleId :
DFTPP20

170.90	201.0
171.90	272.0
172.90	229.0
174.00	649.0
175.00	1325.0
176.00	325.0
176.90	633.0
177.90	173.0
178.90	2361.0
180.00	1384.0
180.90	786.0
184.00	334.0
184.90	1311.0
186.00	9131.0
187.00	2017.0
187.90	418.0
189.00	666.0
190.90	444.0
192.00	792.0
193.00	825.0
194.00	356.0
196.00	1684.0
197.90	63056.0
198.90	4645.0
199.90	380.0
201.50	318.0
202.00	208.0
203.00	346.0
204.00	2499.0
205.00	4261.0
206.00	16608.0
207.00	2409.0
208.00	797.0
209.10	281.0
209.90	325.0
211.00	875.0
215.10	214.0
215.90	485.0
216.90	4605.0
217.90	491.0
221.00	3765.0
222.00	577.0
222.90	761.0
224.00	8502.0
225.00	2102.0
226.10	269.0
226.90	4342.0
228.00	560.0
228.90	754.0
229.90	167.0
231.00	429.0
233.80	285.0
234.90	291.0
236.00	183.0
236.80	198.0
237.10	188.0
238.90	181.0
240.90	274.0
242.00	574.0
242.90	673.0
244.00	6785.0
245.00	1201.0
245.90	1425.0
247.10	353.0
248.90	274.0
252.00	209.0
253.10	220.0
255.00	34136.0
256.00	5325.0
257.00	504.0
258.00	2299.0
258.90	391.0
264.90	762.0
273.00	1064.0
274.00	2360.0
275.00	15642.0
276.00	2398.0
277.00	1469.0
278.00	257.0
282.90	199.0
285.00	240.0
293.00	264.0
296.00	4312.0
297.00	639.0
301.00	173.0
303.00	521.0
314.00	234.0
314.90	527.0
315.90	251.0
321.00	179.0
323.00	1253.0
323.90	230.0
326.90	242.0
333.00	152.0
334.00	852.0

Instrument :
BNA_G
ClientSampleId :
DFTPP20

335.00	298.0
341.00	186.0
345.90	182.0
352.00	436.0
353.00	212.0
354.00	294.0
359.00	159.0
364.90	1840.0
365.90	212.0
372.00	560.0
382.90	229.0
401.90	186.0
403.00	317.0
421.00	384.0
421.90	258.0
423.00	1882.0
424.10	408.0
441.00	6183.0
442.00	37360.0
443.00	8270.0
444.00	795.0

Instrument :
BNA_G
ClientSampleId :
DFTPP20

m/z	Abundance
38.00	380.0
39.10	1955.0
40.00	531.0
44.00	631.0
48.90	223.0
50.00	7096.0
51.00	26832.0
52.00	1259.0
52.90	156.0
54.90	178.0
56.00	906.0
57.00	2232.0
61.00	269.0
62.00	529.0
63.00	1331.0
63.90	201.0
65.00	435.0
65.90	156.0
69.00	33264.0
70.00	154.0
73.00	403.0
74.00	3368.0
75.00	5524.0
76.10	1984.0
77.00	37440.0
78.00	2394.0
79.00	2453.0
80.00	2079.0
81.00	2485.0
81.90	571.0
83.00	697.0
84.90	524.0
86.00	700.0
87.00	518.0
87.70	150.0
91.00	661.0
92.10	776.0
93.00	5094.0
94.00	374.0
98.00	3536.0
99.00	2974.0
100.00	253.0
100.90	1722.0
102.10	202.0
102.90	788.0
103.90	1018.0
105.00	1125.0
106.00	435.0
107.00	12431.0
107.90	1995.0
110.00	23096.0
111.00	3348.0
112.00	462.0
116.00	981.0
117.00	10029.0
118.00	881.0
120.00	182.0
122.00	1149.0
123.00	1366.0
123.90	559.0
124.90	472.0
127.00	46904.0
128.00	3835.0
129.00	19064.0
129.90	1568.0
131.00	256.0
133.90	726.0
135.00	1925.0
135.90	648.0
137.00	928.0
138.00	292.0
139.90	310.0
140.90	2379.0
142.00	937.0
142.90	615.0
143.90	187.0
145.90	394.0
146.90	1237.0
148.00	2806.0
149.00	715.0
149.90	187.0
151.00	248.0
152.10	254.0
153.00	462.0
153.90	429.0
155.00	1204.0
156.00	1977.0
157.00	412.0
157.90	493.0
159.10	344.0
160.00	739.0

Instrument :
BNA_G
ClientSampleId :
DFTPP20

161.00	1035.0
162.00	289.0
163.90	234.0
165.00	945.0
166.00	853.0
167.00	5299.0
168.00	2202.0
169.00	611.0
170.80	293.0
171.90	562.0
173.00	659.0
174.00	882.0
175.00	1844.0
176.00	728.0
176.90	1011.0
178.00	267.0
178.90	3510.0
179.90	2110.0
181.00	1324.0
182.00	483.0
184.00	342.0
185.00	1790.0
186.00	14554.0
187.00	3738.0
187.90	341.0
188.90	1012.0
189.90	278.0
190.90	509.0
192.00	1280.0
193.00	1440.0
194.00	336.0
194.90	256.0
196.00	3408.0
197.90	97368.0
198.90	6459.0
200.00	418.0
201.50	368.0
202.90	917.0
204.00	3674.0
205.00	6299.0
206.00	26328.0
207.00	3852.0
208.00	955.0
208.90	246.0
210.20	388.0
211.00	1146.0
214.90	269.0
216.00	543.0
216.90	7593.0
217.90	851.0
221.00	5648.0
223.00	1615.0
224.00	13969.0
225.00	3711.0
226.00	400.0
227.00	6616.0
227.90	932.0
228.90	1260.0
229.90	173.0
231.00	543.0
233.10	193.0
234.00	292.0
234.90	454.0
237.00	316.0
238.80	217.0
239.90	208.0
241.00	508.0
241.90	1072.0
243.10	849.0
244.00	12463.0
245.00	1608.0
245.90	2447.0
247.00	398.0
248.00	174.0
248.90	455.0
250.90	247.0
252.00	240.0
253.00	494.0
255.00	58400.0
256.00	8599.0
256.90	783.0
258.00	3755.0
259.00	822.0
264.90	1441.0
265.90	262.0
270.90	221.0
273.00	2266.0
274.00	4752.0
275.00	25944.0
276.00	3317.0
277.00	2230.0
277.90	443.0
282.90	317.0
285.00	292.0
290.00	151.0

Instrument :
BNA_G
ClientSampleId :
DFTPP20

292.90	425.0
296.00	8589.0
297.00	954.0
302.00	220.0
303.00	861.0
303.90	275.0
314.00	613.0
315.00	1130.0
316.00	523.0
316.80	152.0
321.10	247.0
321.90	179.0
322.10	179.0
323.00	2127.0
324.10	537.0
327.00	489.0
328.00	255.0
332.90	233.0
334.00	1427.0
335.00	297.0
341.00	171.0
346.00	713.0
352.00	889.0
353.00	413.0
354.00	698.0
355.00	320.0
365.00	3917.0
366.00	526.0
370.90	168.0
372.00	1190.0
373.00	368.0
383.00	298.0
384.10	225.0
401.10	170.0
402.00	420.0
403.00	715.0
404.00	315.0
421.00	511.0
422.00	873.0
423.00	4444.0
424.00	857.0
441.00	12055.0
442.00	86120.0
443.00	17296.0
444.00	1569.0

Instrument :
BNA_G
ClientSampleId :
DFTPP20

m/z	Abundance
38.10	194.0
39.00	1160.0
40.00	652.0
44.00	511.0
50.00	5485.0
51.00	19720.0
52.00	1194.0
56.00	545.0
57.00	1375.0
61.00	341.0
62.00	333.0
63.00	1116.0
65.00	355.0
69.00	24728.0
73.00	153.0
74.00	2446.0
75.00	4331.0
76.00	1695.0
77.00	29272.0
78.00	1751.0
79.00	1680.0
80.00	1650.0
81.00	2350.0
81.90	557.0
83.00	607.0
84.80	426.0
85.90	586.0
87.00	270.0
90.90	406.0
92.00	504.0
93.00	3184.0
94.00	240.0
96.00	154.0
98.00	2593.0
98.90	2025.0
99.90	199.0
100.90	1128.0
102.00	155.0
102.90	424.0
104.00	842.0
105.00	820.0
107.00	9493.0
108.00	1481.0
110.00	18688.0
111.00	2836.0
111.90	395.0
113.00	272.0
116.00	880.0
117.00	6975.0
118.00	653.0
122.00	563.0
123.00	890.0
124.00	487.0
125.00	665.0
127.00	36128.0
128.00	3017.0
129.00	14129.0
130.00	1090.0
130.90	202.0
132.00	224.0
134.00	442.0
135.00	1174.0
136.00	423.0
137.00	639.0
140.00	281.0
141.00	2010.0
141.90	527.0
142.90	568.0
146.00	303.0
147.00	1288.0
148.00	2084.0
148.90	441.0
150.00	156.0
151.10	313.0
152.00	171.0
153.00	546.0
154.00	543.0
155.00	1036.0
156.00	1421.0
156.90	322.0
158.00	370.0
158.90	454.0
159.90	497.0
161.00	791.0
161.90	295.0
164.00	154.0
164.90	763.0
165.90	655.0
167.00	3501.0
168.00	1904.0
168.80	403.0

Instrument :
BNA_G
ClientSampleId :
DFTPP20

170.20	176.0
170.90	230.0
171.90	353.0
172.90	373.0
174.00	848.0
175.00	1893.0
175.90	402.0
177.00	681.0
177.90	306.0
178.90	2442.0
180.00	2100.0
181.00	1049.0
183.90	420.0
185.00	1478.0
186.00	11489.0
187.00	3555.0
187.90	340.0
189.00	774.0
190.80	365.0
192.00	893.0
192.90	1239.0
194.00	268.0
195.90	2245.0
197.90	85064.0
198.90	5821.0
200.00	412.0
201.50	449.0
203.00	438.0
204.00	3613.0
205.00	5394.0
206.00	22120.0
207.00	3074.0
207.90	751.0
208.90	289.0
209.90	420.0
211.00	1071.0
214.90	333.0
216.00	618.0
216.90	6089.0
218.00	887.0
220.00	177.0
220.90	4638.0
223.00	1396.0
224.00	13458.0
225.00	3580.0
226.10	415.0
227.00	5507.0
228.00	872.0
228.90	1260.0
229.80	215.0
230.90	504.0
233.90	342.0
234.90	527.0
235.80	304.0
237.00	550.0
239.20	177.0
240.00	203.0
240.90	397.0
242.00	799.0
243.00	897.0
244.00	11717.0
245.00	1116.0
245.90	1949.0
246.90	395.0
248.90	400.0
250.90	214.0
251.90	173.0
253.00	273.0
255.00	52856.0
256.00	8122.0
257.00	476.0
257.90	3131.0
258.90	529.0
260.00	202.0
265.00	1357.0
265.90	187.0
272.10	286.0
272.90	1754.0
274.00	4512.0
275.00	25296.0
276.00	3602.0
277.00	2268.0
277.90	346.0
283.00	255.0
283.90	230.0
285.00	568.0
292.10	202.0
293.00	420.0
294.00	184.0
296.00	7545.0
296.90	726.0
302.00	212.0
303.00	815.0
303.80	242.0
308.00	162.0

Instrument :
BNA_G
ClientSampleId :
DFTPP20

314.00	452.0
315.00	811.0
316.00	481.0
321.10	294.0
323.00	2533.0
324.10	429.0
326.90	455.0
328.00	297.0
333.00	219.0
334.00	1721.0
335.10	468.0
341.00	314.0
345.90	621.0
347.00	174.0
352.10	799.0
353.00	628.0
354.00	870.0
355.00	272.0
365.00	4150.0
365.90	517.0
371.00	270.0
372.00	1387.0
373.00	280.0
383.00	396.0
390.00	219.0
392.10	156.0
402.00	439.0
403.00	814.0
403.90	204.0
421.00	684.0
422.00	495.0
423.00	5099.0
424.00	986.0
441.00	14278.0
442.00	95552.0
443.00	18424.0
444.00	1639.0

Instrument :
BNA_G
ClientSampleId :
DFTPP20