


Data Path : Z:\HPCHEM1\BNA G\DATA\BG072716\  
 Data File : BG023294.D  
 Acq On : 27 Jul 2016 22:21  
 Operator : UM/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG072616.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Jul 27 17:30:33 2016



AutoFind: Scans 2578, 2579, 2580; Background Corrected with Scan 2572

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 34.0      | 52365   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.6       | 1000    | PASS             |
| 69          | 198          | 0.00         | 100          | 40.8      | 62912   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.9       | 567     | PASS             |
| 127         | 198          | 10           | 80           | 45.1      | 69429   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 154114  | PASS             |
| 199         | 198          | 5            | 9            | 7.7       | 11912   | PASS             |
| 275         | 198          | 10           | 60           | 27.3      | 42005   | PASS             |
| 365         | 198          | 1            | 100          | 4.7       | 7303    | PASS             |
| 441         | 442          | 0.01         | 24           | 17.2      | 21203   | PASS             |
| 442         | 198          | 50           | 100          | 79.9      | 123213  | PASS             |
| 443         | 442          | 15           | 24           | 19.0      | 23353   | PASS             |