


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG072717\
 Data File : BG028028.D
 Acq On : 27 Jul 2017 14:38
 Operator : SJ/JU
 Sample : DFTPP55
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP55

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG072717MA.M
 Title : SVOA CALIBRATION
 Last Update : Fri Jul 28 02:09:08 2017



AutoFind: Scans 2473, 2474, 2475; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	24.9	18488	PASS
68	69	0.00	2	1.0	181	PASS
69	198	0.00	100	25.4	18842	PASS
70	69	0.00	2	0.3	51	PASS
127	198	10	80	36.0	26736	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	74314	PASS
199	198	5	9	6.5	4867	PASS
275	198	10	60	31.6	23508	PASS
365	198	1	100	3.4	2522	PASS
441	442	0.01	100	15.3	10043	PASS
442	198	50	100	88.2	65512	PASS
443	442	15	24	19.1	12525	PASS

m/z	Abundance
38.10	205.0
39.00	1257.0
44.00	372.0
49.00	372.0
50.00	3853.0
51.00	16592.0
52.00	922.0
55.90	420.0
56.90	1315.0
59.00	177.0
61.00	346.0
61.90	421.0
63.00	861.0
65.20	233.0
69.00	16504.0
72.90	213.0
74.00	2047.0
75.00	2485.0
76.00	682.0
77.00	17584.0
78.00	1284.0
78.90	1492.0
79.90	723.0
81.00	1503.0
81.90	443.0
83.10	530.0
84.90	323.0
85.90	485.0
87.00	242.0
90.90	527.0
92.00	501.0
93.00	1942.0
94.00	247.0
97.90	1822.0
99.00	1439.0
99.90	211.0
101.00	589.0
102.80	223.0
104.10	620.0
104.90	331.0
105.90	219.0
107.00	6241.0
108.00	687.0
109.00	182.0
110.00	9509.0
110.90	1531.0
111.80	226.0
115.90	316.0
117.00	5033.0
118.00	389.0
122.00	318.0
122.90	820.0
123.90	318.0
125.00	418.0
127.00	21560.0
128.00	1620.0
129.00	8797.0
129.90	941.0
130.80	222.0
134.00	417.0
134.90	688.0
136.10	319.0
137.00	462.0
137.80	287.0
141.00	1671.0
141.90	549.0
143.00	315.0
146.00	186.0
146.90	567.0
147.90	1506.0
149.00	365.0
150.20	173.0
150.90	152.0
153.10	222.0
153.80	304.0
155.00	830.0
156.00	855.0
156.80	197.0
157.90	409.0
158.90	383.0
160.00	475.0
161.10	370.0
161.90	162.0
164.90	403.0
166.00	498.0
167.00	2426.0
168.00	1431.0
168.80	183.0
171.90	267.0
172.90	269.0
173.90	576.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

175.00	1008.0
175.90	344.0
177.00	594.0
178.00	319.0
178.90	1714.0
179.90	1454.0
180.90	524.0
184.00	364.0
184.90	896.0
186.00	7976.0
187.00	1924.0
188.00	400.0
188.90	623.0
191.00	305.0
191.90	683.0
193.00	759.0
195.20	303.0
196.00	1908.0
197.90	54824.0
198.90	3849.0
199.90	346.0
200.90	185.0
201.40	257.0
202.90	399.0
203.90	2010.0
205.00	3172.0
206.00	14396.0
207.10	2016.0
207.90	721.0
209.10	224.0
210.20	287.0
211.00	497.0
212.00	166.0
214.90	182.0
215.90	300.0
216.90	4499.0
218.00	332.0
221.00	2793.0
222.10	715.0
223.00	1345.0
224.00	7469.0
224.90	1938.0
226.10	338.0
226.90	3568.0
228.00	404.0
228.90	831.0
231.00	360.0
233.90	253.0
234.90	258.0
235.90	176.0
236.90	579.0
239.00	228.0
240.00	258.0
241.00	308.0
241.90	481.0
243.10	566.0
244.00	6821.0
245.00	605.0
245.90	1095.0
246.90	398.0
248.90	347.0
253.00	353.0
255.00	35320.0
255.90	5702.0
256.90	623.0
258.00	1865.0
258.80	374.0
264.90	548.0
265.80	181.0
273.00	829.0
273.90	2619.0
275.00	15741.0
276.00	1970.0
277.00	1320.0
277.90	315.0
279.00	207.0
292.90	341.0
293.70	160.0
296.00	4172.0
296.90	659.0
303.00	624.0
314.10	281.0
315.00	714.0
316.00	242.0
321.00	311.0
323.00	1612.0
323.90	305.0
326.90	259.0
331.80	165.0
334.00	876.0
335.10	227.0
341.00	160.0
346.00	291.0
352.00	561.0
353.00	264.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

354.10	432.0
365.00	1018.0
365.90	367.0
372.00	968.0
383.10	274.0
401.90	212.0
403.00	565.0
421.00	329.0
422.10	276.0
423.00	2139.0
424.00	795.0
441.00	6102.0
442.00	37776.0
443.00	8038.0
444.00	1037.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

m/z	Abundance
38.10	206.0
39.00	1554.0
43.90	273.0
49.00	301.0
50.00	5725.0
51.00	21464.0
52.00	1267.0
56.00	839.0
57.00	1483.0
61.00	267.0
61.90	284.0
63.00	1082.0
65.00	406.0
68.00	543.0
69.00	21120.0
72.90	518.0
74.00	1992.0
75.00	3472.0
76.00	1336.0
77.00	23920.0
78.00	1749.0
78.90	1414.0
80.00	1452.0
81.00	1395.0
82.00	310.0
82.90	536.0
85.00	416.0
85.90	569.0
87.00	346.0
88.00	172.0
91.00	296.0
92.00	547.0
92.90	2228.0
97.90	2511.0
99.00	1891.0
99.80	192.0
101.00	1142.0
102.90	626.0
104.00	834.0
105.00	759.0
107.00	8814.0
108.00	1414.0
109.00	326.0
110.00	15164.0
111.00	2513.0
112.00	556.0
113.00	174.0
115.90	764.0
117.00	6541.0
117.90	623.0
119.70	206.0
121.90	606.0
123.00	1087.0
123.90	567.0
125.00	354.0
127.00	32256.0
128.00	2823.0
129.00	12069.0
130.00	1178.0
131.00	158.0
133.90	348.0
134.90	1300.0
135.90	550.0
137.00	440.0
138.10	159.0
139.80	297.0
141.00	1569.0
141.90	579.0
142.50	267.0
143.00	300.0
145.10	175.0
146.10	270.0
146.90	975.0
147.90	1869.0
149.00	532.0
149.90	231.0
150.90	321.0
151.80	169.0
153.00	756.0
153.90	598.0
155.00	818.0
156.00	1613.0
157.00	527.0
157.90	385.0
159.00	196.0
160.00	548.0
161.00	850.0
161.80	356.0
164.90	814.0
165.90	561.0
166.90	2930.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

168.00	1605.0
169.10	383.0
171.00	257.0
171.90	301.0
173.00	410.0
174.10	988.0
175.00	1481.0
176.00	322.0
177.00	1033.0
178.00	339.0
178.90	3177.0
180.00	2053.0
181.00	1325.0
182.00	154.0
184.10	173.0
185.00	1392.0
186.00	12222.0
187.00	3567.0
187.90	372.0
188.90	681.0
191.00	499.0
192.00	1134.0
193.00	969.0
194.80	258.0
195.90	2268.0
197.90	88096.0
198.90	5719.0
199.90	448.0
201.50	336.0
202.00	163.0
203.00	581.0
204.00	3742.0
205.00	5186.0
206.00	23648.0
207.00	3407.0
207.90	952.0
209.00	360.0
209.90	271.0
211.00	1094.0
212.00	301.0
215.00	323.0
216.10	569.0
217.00	5993.0
218.00	674.0
220.20	430.0
221.00	4163.0
221.80	943.0
222.90	1283.0
224.00	13666.0
225.00	3171.0
225.80	423.0
226.90	5270.0
228.00	734.0
228.90	1407.0
229.90	247.0
231.10	503.0
231.90	185.0
233.90	244.0
234.90	368.0
236.10	229.0
236.90	499.0
238.10	291.0
239.00	396.0
240.90	343.0
241.90	738.0
242.90	997.0
244.00	11748.0
245.00	1822.0
245.90	2326.0
247.00	322.0
248.90	264.0
250.00	283.0
253.00	575.0
255.00	55464.0
256.00	8467.0
257.00	825.0
258.00	3205.0
258.90	509.0
265.00	1284.0
266.00	376.0
271.60	311.0
272.90	1725.0
274.00	5033.0
275.00	26512.0
276.00	4009.0
277.00	2080.0
277.90	377.0
278.90	172.0
280.90	268.0
283.00	216.0
283.80	271.0
285.00	386.0
290.00	277.0
293.00	744.0
294.00	221.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

294.80	227.0
296.00	7285.0
297.00	979.0
303.00	1071.0
304.00	244.0
307.80	176.0
314.00	356.0
314.90	888.0
316.00	583.0
323.10	2277.0
324.00	552.0
327.00	369.0
328.00	411.0
331.90	172.0
333.00	372.0
334.00	1427.0
335.00	471.0
339.90	157.0
341.10	248.0
345.90	386.0
352.00	755.0
353.00	605.0
354.00	653.0
355.00	173.0
365.00	3016.0
366.00	334.0
369.80	233.0
371.20	244.0
372.00	1243.0
372.80	312.0
382.90	271.0
384.10	196.0
390.00	206.0
402.00	631.0
403.00	937.0
403.90	153.0
405.10	167.0
421.00	948.0
422.00	759.0
423.00	5216.0
424.10	1127.0
425.00	213.0
441.00	11855.0
442.00	75776.0
443.00	13505.0
444.00	1525.0
445.00	166.0

Instrument :

BNA_G

ClientSampleId :

DFTPP55

m/z	Abundance
38.00	310.0
39.00	1514.0
44.00	507.0
49.00	243.0
50.00	4198.0
51.00	17408.0
52.00	740.0
55.00	201.0
55.90	989.0
56.90	1515.0
61.00	260.0
61.90	304.0
63.00	885.0
65.00	380.0
68.90	18904.0
69.80	154.0
72.90	237.0
74.00	2109.0
75.00	2527.0
76.00	1170.0
77.00	18296.0
78.00	1380.0
79.00	1595.0
80.00	999.0
80.90	1768.0
82.00	446.0
83.00	391.0
85.00	377.0
85.90	483.0
86.90	341.0
90.90	461.0
91.90	514.0
93.00	2855.0
94.00	255.0
97.90	2233.0
99.00	1337.0
100.00	305.0
101.00	1180.0
101.90	187.0
102.90	452.0
103.90	926.0
105.00	628.0
106.00	270.0
107.00	8622.0
108.00	1256.0
110.00	12307.0
110.90	2576.0
111.90	170.0
112.80	177.0
115.00	236.0
116.00	413.0
116.90	5865.0
117.90	473.0
119.80	156.0
121.90	522.0
123.00	938.0
124.00	525.0
125.00	454.0
127.00	26392.0
128.00	2448.0
129.00	11408.0
129.90	1224.0
133.90	349.0
135.00	910.0
136.00	531.0
137.00	747.0
140.10	244.0
140.90	1134.0
141.90	354.0
143.00	204.0
145.00	238.0
145.90	337.0
146.90	812.0
147.90	1030.0
149.00	314.0
152.90	472.0
154.00	642.0
154.90	1093.0
156.00	1534.0
157.00	388.0
157.90	425.0
158.90	336.0
160.00	475.0
161.00	654.0
161.70	152.0
161.90	161.0
165.00	765.0
166.00	861.0
166.90	2894.0
167.90	1527.0
168.90	313.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

171.00	279.0
171.90	404.0
172.90	269.0
174.00	686.0
175.00	1604.0
175.90	454.0
177.10	803.0
178.00	302.0
178.90	2554.0
180.00	1619.0
180.90	726.0
184.00	269.0
185.00	1533.0
186.00	10584.0
187.00	2972.0
188.00	377.0
189.00	470.0
190.80	385.0
192.00	841.0
193.00	1143.0
194.90	228.0
196.00	2035.0
197.90	80024.0
198.90	5033.0
200.00	294.0
200.60	157.0
201.50	426.0
203.00	384.0
204.00	2458.0
205.00	5294.0
206.00	21360.0
207.00	3217.0
207.90	572.0
209.10	306.0
209.90	228.0
211.00	759.0
214.90	269.0
216.10	516.0
216.90	5368.0
218.00	1034.0
219.20	232.0
221.00	3501.0
221.90	732.0
222.90	1492.0
224.00	12213.0
225.00	3542.0
225.80	590.0
227.00	5459.0
227.80	445.0
229.00	1192.0
231.00	477.0
233.90	450.0
234.90	372.0
237.00	499.0
238.90	357.0
240.10	254.0
241.00	191.0
242.00	759.0
242.90	692.0
244.00	9661.0
245.00	1610.0
246.00	2597.0
246.90	489.0
249.00	261.0
252.10	170.0
252.90	362.0
255.00	54184.0
256.00	7819.0
257.00	829.0
258.00	3197.0
258.80	376.0
264.90	1213.0
265.90	392.0
270.10	159.0
271.00	228.0
271.90	156.0
273.00	1802.0
274.00	3947.0
275.00	28272.0
276.00	3454.0
276.90	2246.0
278.00	415.0
283.00	315.0
284.90	241.0
293.00	432.0
295.00	342.0
296.00	7838.0
297.00	940.0
303.00	1153.0
304.10	260.0
314.00	420.0
314.90	754.0
316.00	509.0
321.00	422.0
322.00	189.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

323.00	2281.0
324.10	335.0
326.90	548.0
328.00	524.0
331.90	200.0
332.80	211.0
334.00	1544.0
335.00	387.0
340.90	262.0
345.90	532.0
352.00	836.0
352.90	524.0
354.00	1000.0
365.00	3534.0
365.90	406.0
371.00	233.0
372.00	1635.0
373.00	438.0
383.00	495.0
390.10	187.0
391.00	274.0
392.10	199.0
401.00	229.0
401.90	749.0
403.00	744.0
404.00	546.0
421.00	830.0
422.00	565.0
423.00	4686.0
423.90	1410.0
425.00	176.0
441.00	12172.0
442.00	82984.0
443.00	16033.0
444.00	1458.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

m/z	Abundance
43.90	568.0
207.20	362.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55