

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG072720\
 Data File : BG046086.D
 Acq On : 27 Jul 2020 13:37
 Operator : CG/JU
 Sample : PB130168BL
 Misc : PBBL-WTR-05
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK68

Quant Time: Jul 27 14:51:25 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG072320MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 27 10:02:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.96	152	104010	20.00	ng/ul	0.00
18) Naphthalene-d8	10.75	136	429337	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.59	164	299094	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.33	188	717670	20.00	ng/ul	0.00
78) Chrysene-d12	21.58	240	745272	20.00	ng/ul	0.00
86) Perylene-d12	24.69	264	755942	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.42	96	16267	7.72	ng/uL	0.00
5) Phenol-d5	7.12	99	289100	35.42	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.29	67	160391	31.72	ng/ul	0.00
9) 2-Chlorophenol-d4	7.49	132	228544	34.98	ng/ul	0.00
13) 4-Methylphenol-d8	8.66	113	237645	34.51	ng/ul	0.00
19) Nitrobenzene-d5	9.11	128	102949	33.96	ng/ul	0.00
22) 2-Nitrophenol-d4	9.84	143	109486	35.34	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.38	165	235336	33.56	ng/ul	0.00
29) 4-Chloroaniline-d4	10.88	131	295300	37.39	ng/ul	0.00
44) Dimethylphthalate-d6	13.99	166	769876	33.00	ng/ul	0.00
47) Acenaphthylene-d8	14.28	160	884646	31.96	ng/ul	0.00
52) 4-Nitrophenol-d4	14.77	143	101221	26.57	ng/ul	0.00
58) Fluorene-d10	15.58	176	694227	32.70	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.68	200	85886	23.91	ng/ul	0.00
71) Anthracene-d10	17.42	188	1056409	31.92	ng/ul	0.00
79) Pyrene-d10	19.71	212	1239560	31.45	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.48	264	1342519	32.04	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG072720\
Data File : BG046086.D
Acq On : 27 Jul 2020 13:37
Operator : CG/JU
Sample : PB130168BL
Misc : PBBL-WTR-05
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK68

Quant Time: Jul 27 14:51:25 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG072320MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jul 27 10:02:22 2020
Response via : Initial Calibration

