


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG072720\
 Data File : BG046078.D
 Acq On : 27 Jul 2020 8:17
 Operator : CG/JU
 Sample : DFTPP27
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP27

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG072320MA.M
 Title : SVOA CALIBRATION
 Last Update : Mon Jul 27 10:02:22 2020



AutoFind: Scans 2501, 2502, 2503; Background Corrected with Scan 2495

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.2	61002	PASS
68	69	0.00	2	0.7	417	PASS
69	198	0.00	100	30.8	58352	PASS
70	69	0.00	2	0.5	304	PASS
127	198	10	80	39.8	75368	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	189290	PASS
199	198	5	9	6.9	13034	PASS
275	198	10	60	30.1	57056	PASS
365	198	1	100	4.3	8114	PASS
441	198	0.01	100	14.9	28277	PASS
442	198	50	100	92.1	174250	PASS
443	442	15	24	19.9	34624	PASS

m/z	Abundance
40.00	450.0
43.10	274.0
44.00	999.0
73.00	330.0
95.90	172.0
147.00	229.0
191.00	159.0
207.10	1388.0
208.00	308.0
209.10	215.0
221.00	204.0
281.00	332.0

Instrument :
BNA_G
ClientSampleId :
DFTPP27

m/z	Abundance
38.00	650.0
39.10	2954.0
40.00	693.0
41.10	299.0
42.90	191.0
44.10	865.0
49.20	221.0
50.10	9505.0
51.10	39328.0
52.10	1625.0
55.10	206.0
56.10	1260.0
57.00	2711.0
61.10	292.0
62.10	647.0
63.10	1550.0
64.00	488.0
65.10	610.0
69.00	36472.0
70.10	270.0
73.00	534.0
74.00	4376.0
75.00	6014.0
76.00	1914.0
77.10	44408.0
78.10	3097.0
79.10	2699.0
80.10	2265.0
81.00	3080.0
82.00	1000.0
83.00	891.0
84.00	219.0
85.00	458.0
86.00	1047.0
87.00	317.0
87.90	152.0
91.00	758.0
91.90	792.0
93.10	4921.0
94.00	400.0
95.00	222.0
96.00	359.0
98.00	4016.0
99.00	3359.0
100.00	395.0
101.00	2013.0
103.00	617.0
104.00	1223.0
105.10	1073.0
107.00	13681.0
108.00	2254.0
110.00	26320.0
111.00	3687.0
112.00	509.0
113.10	229.0
116.00	866.0
117.00	11870.0
118.10	624.0
119.00	299.0
120.00	181.0
122.00	1204.0
123.10	1550.0
124.00	607.0
125.10	542.0
127.00	51384.0
128.00	4346.0
129.00	21792.0
130.00	1722.0
131.00	336.0
132.00	194.0
134.00	610.0
134.90	1341.0
136.10	655.0
137.00	846.0
138.10	236.0
140.00	388.0
141.00	2584.0
142.00	1080.0
143.00	558.0
145.00	250.0
146.10	520.0
147.00	1605.0
148.10	3474.0
149.00	823.0
149.90	270.0
151.00	333.0
151.90	239.0
153.00	998.0
154.10	1040.0
155.00	1512.0
156.10	1962.0
157.10	544.0

Instrument :
BNA_G
ClientSampleId :
DFTPP27

157.90	526.0
159.00	692.0
159.90	1002.0
161.00	1233.0
162.10	592.0
164.10	161.0
165.00	1331.0
166.00	866.0
167.00	6164.0
168.00	3185.0
168.90	405.0
170.00	247.0
171.00	256.0
172.00	647.0
173.00	926.0
174.00	1580.0
175.10	2952.0
176.10	1047.0
177.00	1115.0
178.10	481.0
179.00	5715.0
180.00	4008.0
181.00	1744.0
182.10	170.0
182.90	217.0
184.00	432.0
185.00	2637.0
186.00	19936.0
187.00	5673.0
188.10	557.0
189.00	1432.0
189.90	191.0
191.00	636.0
192.00	2142.0
193.00	2061.0
194.00	437.0
194.90	505.0
196.00	4095.0
198.00	142144.0
199.00	8506.0
200.10	675.0
200.70	160.0
201.50	587.0
203.10	993.0
204.00	5111.0
205.10	9277.0
206.00	41120.0
207.10	6882.0
208.00	1891.0
208.90	701.0
211.00	1252.0
214.90	459.0
216.00	825.0
217.00	10673.0
218.00	1591.0
219.00	188.0
221.00	5848.0
221.90	1397.0
223.00	2548.0
224.10	23184.0
225.00	6353.0
226.10	759.0
227.00	9699.0
228.00	1255.0
229.00	2101.0
230.00	351.0
231.00	846.0
233.10	374.0
233.90	673.0
235.00	730.0
236.00	482.0
237.00	1062.0
239.00	532.0
240.00	259.0
241.00	459.0
242.00	1107.0
243.00	1381.0
244.10	18440.0
245.10	2507.0
246.00	3506.0
247.00	860.0
248.10	236.0
249.00	643.0
250.90	176.0
252.10	207.0
253.00	533.0
255.00	103960.0
256.00	15041.0
257.00	1364.0
258.00	6430.0
259.10	749.0
259.90	245.0
261.10	179.0
264.10	193.0
265.00	2355.0
265.90	436.0

Instrument :
BNA_G
ClientSampleId :
DFTPP27

270.90	238.0
272.10	267.0
273.00	3470.0
274.00	8394.0
275.00	47552.0
276.00	5500.0
277.00	3878.0
278.00	575.0
281.10	343.0
282.00	275.0
283.00	642.0
284.10	357.0
285.00	602.0
288.90	161.0
290.00	221.0
292.00	262.0
293.00	939.0
294.00	321.0
295.00	275.0
296.00	13671.0
297.00	2072.0
298.00	174.0
301.00	308.0
302.20	374.0
303.00	1733.0
304.10	418.0
308.00	365.0
309.10	150.0
309.90	249.0
313.90	628.0
315.00	1438.0
316.00	880.0
317.10	209.0
321.10	508.0
321.90	251.0
323.10	5218.0
324.00	994.0
326.90	880.0
328.00	510.0
332.00	388.0
333.00	393.0
334.00	2975.0
335.10	754.0
339.80	158.0
340.90	518.0
341.90	256.0
346.00	1049.0
346.90	190.0
352.10	1542.0
353.10	1190.0
354.00	1997.0
355.10	352.0
359.20	174.0
365.00	6956.0
366.00	999.0
371.10	459.0
372.00	2541.0
372.90	574.0
383.00	1039.0
383.90	290.0
390.20	555.0
390.90	279.0
402.00	1309.0
403.00	1645.0
404.00	588.0
421.00	1521.0
422.00	1608.0
423.00	9651.0
424.00	2393.0
425.00	464.0
441.00	26808.0
442.00	166080.0
443.00	32032.0
444.00	2885.0

Instrument :
BNA_G
ClientSampleId :
DFTPP27

m/z	Abundance
37.10	324.0
38.10	756.0
39.10	4539.0
40.00	743.0
41.10	399.0
43.10	152.0
44.00	979.0
45.00	239.0
49.20	617.0
50.10	17288.0
51.10	68880.0
52.10	3179.0
53.00	306.0
55.00	426.0
56.00	2254.0
57.00	5181.0
58.10	220.0
59.90	158.0
61.10	1084.0
62.10	1065.0
63.00	2705.0
64.00	432.0
65.10	1514.0
66.00	175.0
68.10	1251.0
69.00	65936.0
70.10	326.0
73.10	660.0
74.10	6782.0
75.10	10927.0
76.10	3593.0
77.10	76344.0
78.10	5417.0
79.00	4970.0
80.00	3797.0
81.00	5721.0
82.00	1764.0
83.00	1104.0
85.00	987.0
86.00	1578.0
87.00	772.0
88.00	219.0
89.10	202.0
91.00	1025.0
92.10	1036.0
93.00	8377.0
94.00	506.0
95.00	217.0
96.00	603.0
98.00	6403.0
99.00	5545.0
100.00	437.0
101.10	2960.0
101.80	189.0
103.00	1134.0
104.00	2245.0
105.00	2052.0
106.00	767.0
107.00	25448.0
108.00	3748.0
110.00	46256.0
111.00	6625.0
112.00	1043.0
113.10	227.0
116.00	1580.0
117.00	19672.0
118.00	987.0
119.00	303.0
120.10	521.0
120.90	166.0
122.00	1446.0
123.00	3539.0
124.00	961.0
125.10	1356.0
127.00	85168.0
128.10	6646.0
129.00	34072.0
130.10	2659.0
131.00	594.0
131.80	241.0
132.20	324.0
133.10	237.0
134.10	969.0
135.00	2661.0
136.00	1188.0
137.10	1821.0
137.90	414.0
138.90	220.0
140.00	474.0
141.00	4633.0
142.10	1457.0
143.00	895.0

Instrument :
BNA_G
ClientSampleId :
DFTPP27

144.00	268.0
144.90	288.0
146.00	870.0
147.00	2465.0
148.00	5727.0
149.00	1274.0
150.00	277.0
151.20	603.0
151.90	403.0
153.00	1393.0
154.00	1164.0
155.00	2485.0
156.00	4284.0
157.10	898.0
158.00	1432.0
159.00	840.0
160.00	1614.0
161.00	2581.0
161.90	869.0
163.10	199.0
164.00	264.0
165.00	2094.0
166.00	1624.0
167.00	10894.0
168.00	4931.0
169.00	1040.0
170.00	251.0
171.10	389.0
172.00	877.0
173.00	1596.0
174.00	1987.0
175.00	3802.0
176.00	1355.0
177.10	1891.0
178.00	682.0
179.00	8443.0
180.00	5863.0
181.00	2685.0
182.00	531.0
182.80	320.0
184.00	636.0
185.00	3979.0
186.00	29800.0
187.00	7973.0
188.10	1037.0
189.00	1909.0
190.10	390.0
191.00	1021.0
192.00	2576.0
193.00	3024.0
194.00	783.0
194.90	454.0
196.00	6347.0
198.00	217472.0
199.00	15312.0
200.00	1275.0
201.40	885.0
203.00	1654.0
204.00	8505.0
205.10	14787.0
206.10	63040.0
207.10	9377.0
208.00	2299.0
209.10	1013.0
210.30	995.0
211.00	2028.0
211.90	321.0
214.10	162.0
215.00	641.0
216.00	1799.0
217.00	17352.0
218.00	2119.0
218.90	172.0
221.00	9848.0
222.00	1801.0
223.00	3630.0
224.10	34888.0
225.10	8915.0
226.10	1071.0
227.00	14894.0
228.00	2118.0
229.00	2743.0
230.10	539.0
231.10	1270.0
232.10	243.0
233.00	346.0
234.00	1106.0
235.00	987.0
236.00	1067.0
237.00	1478.0
238.10	287.0
239.00	627.0
240.10	365.0
241.00	1013.0
242.00	1914.0
243.10	2203.0

Instrument :
BNA_G
ClientSampleId :
DFTPP27

244.10	27504.0
245.00	3709.0
246.00	5730.0
247.00	1130.0
248.10	264.0
249.00	1158.0
249.90	201.0
251.00	238.0
252.00	359.0
253.00	605.0
255.00	149952.0
256.00	22632.0
257.00	1740.0
258.00	9167.0
259.00	1207.0
260.10	237.0
261.10	177.0
265.00	3679.0
266.00	426.0
270.10	284.0
271.00	538.0
272.00	635.0
273.00	4335.0
274.00	11872.0
275.00	66800.0
276.00	9324.0
277.00	5830.0
278.00	925.0
278.90	249.0
281.00	413.0
282.10	213.0
283.00	694.0
284.00	626.0
285.00	1016.0
286.00	207.0
289.00	282.0
290.00	172.0
291.10	154.0
292.00	429.0
293.00	1381.0
294.00	408.0
296.00	20384.0
297.10	2702.0
297.90	252.0
301.00	339.0
301.90	365.0
303.00	2574.0
304.00	480.0
308.00	338.0
310.10	230.0
313.10	187.0
314.00	939.0
315.00	2311.0
316.00	1182.0
317.00	234.0
321.10	663.0
322.00	289.0
323.10	6194.0
324.00	1198.0
327.00	1507.0
328.10	668.0
329.10	210.0
332.00	634.0
333.10	638.0
334.00	3893.0
335.10	1097.0
341.10	849.0
346.00	1333.0
347.10	175.0
352.00	1968.0
353.00	1222.0
354.00	1878.0
355.10	323.0
358.80	169.0
365.00	9739.0
366.00	1722.0
370.00	159.0
371.00	483.0
372.00	3519.0
373.00	898.0
383.00	936.0
384.00	455.0
390.20	376.0
390.90	559.0
391.80	385.0
401.00	271.0
402.00	1400.0
403.00	2110.0
404.00	659.0
421.00	1714.0
422.10	1509.0
423.00	11563.0
424.10	2950.0
425.00	435.0
441.00	33184.0
442.00	204224.0

Instrument :
BNA_G
ClientSampleId :
DFTPP27

443.00	41296.0
444.00	3807.0
445.20	262.0

Instrument :
BNA_G
ClientSampleId :
DFTPP27

m/z	Abundance
37.00	344.0
38.10	964.0
39.10	5488.0
40.00	648.0
41.20	264.0
43.00	251.0
44.00	738.0
49.00	496.0
50.10	18544.0
51.10	74800.0
52.10	3911.0
53.10	277.0
55.00	314.0
56.00	2471.0
57.00	5209.0
58.10	385.0
59.00	209.0
61.10	999.0
62.10	1163.0
63.10	3003.0
64.00	555.0
65.10	1630.0
69.00	72648.0
70.10	317.0
73.00	1124.0
74.00	7225.0
75.10	12104.0
76.10	4456.0
77.10	83816.0
78.10	5713.0
79.00	5333.0
80.00	4270.0
81.00	5523.0
82.00	1710.0
83.00	1416.0
84.10	211.0
85.00	1156.0
86.00	1551.0
87.10	499.0
88.10	258.0
89.00	233.0
91.00	1167.0
92.10	1292.0
93.00	9883.0
94.00	679.0
94.90	196.0
96.00	686.0
98.00	6792.0
99.00	6080.0
100.00	580.0
101.00	3143.0
103.00	1023.0
104.00	2224.0
105.00	1675.0
106.00	592.0
107.00	24896.0
108.00	3783.0
110.00	47840.0
111.00	7654.0
112.00	590.0
113.00	279.0
116.00	1558.0
117.00	19912.0
118.10	1838.0
119.00	200.0
120.10	402.0
122.00	1639.0
123.10	3149.0
124.00	1215.0
125.00	1302.0
127.00	89552.0
128.00	6772.0
129.00	34800.0
130.00	3205.0
131.00	474.0
132.10	322.0
133.00	333.0
134.00	857.0
135.00	2989.0
136.00	1277.0
137.00	1791.0
138.10	187.0
138.90	184.0
140.00	432.0
141.00	4625.0
142.00	1644.0
143.00	1052.0
144.10	398.0
145.10	357.0
146.00	1249.0
147.00	2801.0
148.00	4886.0

Instrument :
BNA_G
ClientSampleId :
DFTPP27

149.00	1370.0
149.90	359.0
151.00	767.0
152.10	499.0
153.00	1761.0
154.00	1305.0
155.10	2586.0
156.00	4203.0
157.00	641.0
157.90	1065.0
158.90	724.0
160.10	1436.0
161.00	2550.0
162.10	621.0
163.10	271.0
164.10	280.0
165.00	2160.0
166.10	1783.0
167.00	9017.0
168.00	5336.0
169.00	1101.0
169.90	428.0
170.80	400.0
172.00	1014.0
173.00	1076.0
174.00	2202.0
175.00	3808.0
176.00	989.0
177.10	1790.0
178.10	632.0
179.00	7853.0
180.00	5635.0
181.00	2186.0
182.20	410.0
183.00	360.0
184.00	632.0
185.00	4031.0
186.00	29296.0
187.00	8353.0
188.00	891.0
189.00	1796.0
189.90	252.0
191.00	1097.0
192.00	2595.0
193.00	3250.0
194.00	576.0
194.90	473.0
196.00	5934.0
198.00	208256.0
199.00	15286.0
200.10	1060.0
201.50	879.0
203.10	1701.0
204.10	7282.0
205.00	13801.0
206.00	58040.0
207.10	9209.0
208.00	2280.0
209.00	885.0
210.20	1298.0
211.10	1772.0
212.90	158.0
215.00	746.0
216.10	1399.0
217.00	14663.0
218.00	1801.0
219.10	304.0
221.00	9098.0
221.90	2026.0
223.00	3128.0
224.10	31384.0
225.00	8409.0
225.90	923.0
227.00	14451.0
228.00	2083.0
229.00	2826.0
230.00	403.0
231.00	1262.0
232.10	249.0
232.90	191.0
234.00	1203.0
235.00	1079.0
236.00	847.0
237.10	1310.0
238.00	235.0
239.00	594.0
239.90	354.0
240.90	930.0
242.00	1847.0
243.00	1638.0
244.00	23840.0
245.00	3141.0
246.00	5372.0
247.00	896.0
248.00	305.0
248.90	769.0

Instrument :
BNA_G
ClientSampleId :
DFTPP27

249.90	234.0
250.90	307.0
252.20	281.0
252.90	549.0
255.00	134784.0
256.00	20456.0
257.00	1284.0
258.00	8174.0
259.00	1050.0
260.10	218.0
262.90	173.0
263.90	184.0
265.00	3252.0
266.00	335.0
269.90	190.0
271.00	152.0
272.10	367.0
273.00	3930.0
274.00	10588.0
275.00	56816.0
276.00	7774.0
277.00	5585.0
278.00	972.0
278.90	185.0
281.00	432.0
283.00	580.0
284.00	462.0
285.00	868.0
288.80	175.0
290.00	360.0
291.10	198.0
292.00	231.0
293.00	1201.0
293.90	438.0
296.00	17336.0
297.00	2246.0
301.00	247.0
301.90	382.0
303.00	2078.0
304.10	609.0
308.10	306.0
309.10	218.0
310.10	358.0
314.00	1006.0
315.00	2031.0
316.00	1544.0
317.10	409.0
319.00	205.0
320.90	652.0
321.90	271.0
323.10	5163.0
324.00	1147.0
325.10	232.0
325.90	150.0
326.90	1073.0
328.00	475.0
332.00	425.0
333.10	494.0
334.10	3201.0
335.00	869.0
340.00	183.0
341.00	684.0
342.20	199.0
346.00	1217.0
347.00	258.0
351.20	180.0
352.00	1571.0
353.00	1133.0
354.00	1474.0
355.00	374.0
359.10	167.0
365.00	7647.0
365.90	846.0
371.10	375.0
372.10	3397.0
373.10	540.0
382.90	571.0
384.00	306.0
385.00	190.0
389.90	289.0
391.00	184.0
391.90	263.0
401.00	263.0
402.00	954.0
403.00	1660.0
403.90	626.0
405.00	157.0
421.00	1296.0
422.10	1430.0
423.00	8713.0
424.00	2581.0
425.10	305.0
441.10	24840.0
442.00	152448.0
443.00	30544.0
444.00	3232.0

Instrument :
BNA_G
ClientSampleId :
DFTPP27

445.20

291.0

Instrument :
BNA_G
ClientSampleId :
DFTPP27