

Data Path : Z:\HPCHEM1\BNA G\DATA\BG072816\
 Data File : BG023319.D
 Acq On : 28 Jul 2016 23:06
 Operator : UM/SJ
 Sample : H4175-02RE
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 EFF-WWRE

Quant Time: Jul 29 01:36:58 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG072616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 27 17:30:33 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	214025	20.00	ng	0.00
21) Naphthalene-d8	10.97	136	913735	20.00	ng	0.00
38) Acenaphthene-d10	14.81	164	648156	20.00	ng	0.00
63) Phenanthrene-d10	17.58	188	1123653	20.00	ng	0.00
75) Chrysene-d12	21.91	240	1104096	20.00	ng	0.02
86) Perylene-d12	25.30	264	1152951	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.68	112	35785	2.85	ng	0.00
7) Phenol-d6	7.34	99	140076	7.28	ng	0.05
23) Nitrobenzene-d5	9.32	82	1247509	66.77	ng	0.00
41) 2,4,6-Tribromophenol	16.31	330	15199	2.27	ng	0.00
44) 2-Fluorobiphenyl	13.42	172	2375777	68.60	ng	0.00
78) Terphenyl-d14	20.20	244	2249571	66.23	ng	0.02
Target Compounds						
10) Phenol	7.37	94	352617	17.36	ng	Qvalue 82
15) Benzyl Alcohol	8.42	79	2203852	170.13	ng	# 88
17) 2-Methylphenol	8.59	107	171860	12.38	ng	# 88
20) 3+4-Methylphenols	8.92	107	268098	14.20	ng	82
22) Acetophenone	8.98	105	2242745	94.01	ng	# 93
32) Benzoic acid	10.25	122	106860	16.01	ng	# 83
45) 1,1'-Biphenyl	13.63	154	305471	6.76	ng	99
49) Dimethylphthalate	14.25	163	431255	9.02	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG072816\
Data File : BG023319.D
Acq On : 28 Jul 2016 23:06
Operator : UM/SJ
Sample : H4175-02RE
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
EFF-WWRE

Quant Time: Jul 29 01:36:58 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG072616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 27 17:30:33 2016
Response via : Initial Calibration

