

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072822\
 Data File : BG054432.D
 Acq On : 28 Jul 2022 15:03
 Operator : CG/JU
 Sample : N3893-11
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 RVE-11

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 07/29/2022
 Supervised By : Jagrut Upadhyay 07/29/2022

Quant Time: Jul 29 04:08:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 28 04:30:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.985	152	17732	20.000	ng	0.00
21) Naphthalene-d8	10.800	136	68298	20.000	ng	# 0.00
39) Acenaphthene-d10	14.625	164	53805	20.000	ng	0.00
64) Phenanthrene-d10	17.369	188	130287	20.000	ng	# 0.00
76) Chrysene-d12	21.634	240	144941	20.000	ng	# 0.00
86) Perylene-d12	24.836	264	155515	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.565	112	91289	107.490	ng	0.00
7) Phenol-d6	7.169	99	126476	108.682	ng	0.00
23) Nitrobenzene-d5	9.149	82	85304	68.017	ng	0.00
42) 2,4,6-Tribromophenol	16.117	330	116385	103.022	ng	0.00
45) 2-Fluorobiphenyl	13.244	172	265281	69.553	ng	0.00
79) Terphenyl-d14	19.977	244	542420	74.241	ng	0.00
Target Compounds						Qvalue
75) Fluoranthene	19.419	202	24955	2.822	ng	97
78) Pyrene	19.783	202	44940	5.404	ng	99
83) Chrysene	21.681	228	29870	3.436	ng	# 92
88) Benzo(b)fluoranthene	23.826	252	39490	4.294	ng	98
90) Benzo(a)pyrene	24.683	252	34611	4.407	ng	95
92) Benzo(g,h,i)perylene	29.648	276	32540m	3.389	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG072822\
Data File : BG054432.D
Acq On : 28 Jul 2022 15:03
Operator : CG/JU
Sample : N3893-11
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
RVE-11

Quant Time: Jul 29 04:08:48 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 28 04:30:50 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 07/29/2022
Supervised By : Jagrut Upadhyay 07/29/2022

