


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG072915\
 Data File : BG018165.D
 Acq On : 29 Jul 2015 14:14
 Operator : TP/UM
 Sample : DFTPP21
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP21

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG072915.M
 Title : SVOA CALIBRATION
 Last Update : Thu Jul 30 03:22:05 2015



AutoFind: Scans 2502, 2503, 2504; Background Corrected with Scan 2497

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	23.4	12445	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	30.7	16311	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	44.9	23834	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	53109	PASS
199	198	5	9	6.8	3588	PASS
275	198	10	60	27.9	14826	PASS
365	198	1	100	4.2	2207	PASS
441	442	0.01	24	14.9	7552	PASS
442	198	50	100	95.2	50560	PASS
443	442	15	24	20.3	10270	PASS

Scan 2497 (17.358 min): BG018165.D
DFTPP21

m/z	Abundance
39.90	172.0
44.00	336.0
55.10	191.0
73.00	166.0
166.90	251.0
207.00	433.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

m/z	Abundance
38.00	177.0
39.00	680.0
39.90	247.0
41.00	230.0
44.00	262.0
49.00	159.0
50.00	2831.0
51.00	11125.0
52.00	581.0
55.00	168.0
55.90	368.0
57.00	1047.0
62.00	253.0
63.00	433.0
65.00	307.0
69.00	13657.0
73.00	337.0
74.00	1549.0
75.00	2215.0
76.00	1009.0
77.00	15423.0
78.00	1186.0
79.00	970.0
79.90	805.0
81.00	994.0
82.00	367.0
83.00	467.0
85.00	337.0
85.90	355.0
91.00	345.0
92.00	368.0
92.90	1834.0
93.90	158.0
98.00	1452.0
99.00	1006.0
100.90	729.0
102.90	221.0
103.90	399.0
105.00	401.0
105.90	184.0
107.00	5391.0
108.00	894.0
110.00	8927.0
111.00	1664.0
116.10	466.0
117.00	4410.0
118.00	275.0
121.90	360.0
123.00	667.0
123.90	324.0
125.00	272.0
127.00	18496.0
128.00	1515.0
129.00	7412.0
130.00	662.0
130.90	235.0
134.00	264.0
135.00	606.0
136.00	190.0
137.00	370.0
141.00	1108.0
142.00	249.0
143.00	333.0
146.00	295.0
146.90	667.0
148.00	1247.0
149.00	261.0
151.00	173.0
151.20	172.0
152.90	331.0
154.00	346.0
155.00	437.0
155.90	753.0
157.80	200.0
160.00	356.0
160.90	424.0
164.90	400.0
166.00	375.0
167.00	1820.0
168.00	1009.0
168.90	166.0
172.00	264.0
173.00	155.0
174.00	423.0
175.00	713.0
175.90	201.0
176.90	369.0
178.00	195.0
178.90	1576.0
180.00	850.0
180.90	453.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

185.00	885.0
186.00	5540.0
187.00	1624.0
187.90	211.0
189.00	389.0
191.00	151.0
191.90	567.0
193.00	473.0
196.00	1042.0
197.90	38416.0
198.90	2590.0
199.90	354.0
201.40	219.0
203.00	307.0
203.90	1372.0
205.00	2751.0
206.00	9173.0
207.00	1617.0
208.00	324.0
208.90	154.0
209.90	178.0
211.00	439.0
216.00	294.0
216.90	2752.0
217.90	331.0
221.00	2071.0
223.00	559.0
224.00	4963.0
225.00	1593.0
226.90	2412.0
227.90	259.0
228.90	614.0
231.00	393.0
235.00	222.0
236.90	349.0
241.00	150.0
242.00	336.0
242.90	406.0
244.00	4087.0
245.00	556.0
246.00	831.0
246.90	236.0
248.90	226.0
255.00	20888.0
256.00	3111.0
257.10	275.0
257.90	1335.0
259.00	150.0
265.00	711.0
272.90	547.0
274.00	1869.0
275.00	9351.0
276.00	1370.0
277.00	891.0
293.00	208.0
296.00	2541.0
297.00	445.0
303.00	311.0
314.90	243.0
323.00	803.0
324.00	258.0
334.00	379.0
340.90	158.0
351.90	251.0
354.00	245.0
364.90	1234.0
366.00	236.0
372.00	518.0
403.00	185.0
420.90	236.0
422.00	188.0
423.00	1572.0
424.00	331.0
441.00	3573.0
442.00	26016.0
443.00	5465.0
444.00	506.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

m/z	Abundance
38.00	186.0
39.00	1295.0
40.00	260.0
44.00	389.0
49.10	176.0
50.00	4121.0
51.00	14845.0
52.00	891.0
55.00	255.0
56.00	508.0
57.00	1382.0
61.00	289.0
62.00	239.0
63.00	768.0
65.00	406.0
69.00	19440.0
73.10	205.0
74.00	2253.0
75.00	3102.0
76.00	1119.0
77.00	22616.0
78.00	1306.0
79.00	1379.0
80.00	1011.0
81.00	1456.0
81.90	388.0
82.90	454.0
83.90	157.0
84.90	344.0
85.90	485.0
87.00	159.0
91.00	452.0
92.00	481.0
93.00	2772.0
97.10	202.0
98.00	2369.0
99.00	1597.0
101.00	1113.0
103.00	367.0
103.90	702.0
104.90	767.0
107.00	8053.0
108.00	1214.0
110.00	14470.0
111.00	2299.0
111.90	220.0
116.00	565.0
117.00	6275.0
118.00	455.0
122.00	713.0
122.90	1012.0
124.00	395.0
124.90	259.0
127.00	29824.0
128.00	2591.0
129.00	12285.0
129.90	994.0
130.90	212.0
134.00	319.0
135.00	1160.0
136.00	513.0
137.00	498.0
137.90	163.0
141.00	1453.0
141.90	499.0
142.90	380.0
143.90	181.0
145.90	470.0
146.90	1109.0
147.90	1851.0
149.00	367.0
151.00	227.0
152.00	159.0
153.00	523.0
153.90	250.0
155.00	881.0
156.00	1423.0
157.00	189.0
157.90	159.0
159.00	275.0
159.90	640.0
161.00	759.0
162.00	226.0
165.00	617.0
165.90	618.0
167.00	3023.0
167.90	1400.0
169.00	223.0
172.00	385.0
173.00	470.0
174.00	687.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

175.00	1379.0
176.00	272.0
177.00	537.0
178.00	177.0
178.90	1995.0
180.00	1511.0
180.90	737.0
185.00	1297.0
186.00	8815.0
187.00	2670.0
188.00	255.0
189.00	795.0
190.90	335.0
191.90	953.0
192.90	737.0
194.00	250.0
196.00	1784.0
196.60	453.0
197.90	64624.0
198.90	4528.0
200.00	360.0
201.40	248.0
202.90	474.0
204.00	2782.0
205.00	3998.0
206.00	16744.0
207.00	2655.0
207.90	675.0
208.90	243.0
209.90	208.0
210.90	868.0
214.90	276.0
216.00	401.0
216.90	4628.0
217.90	415.0
221.00	4326.0
221.90	548.0
223.00	985.0
224.00	9676.0
225.00	2372.0
226.00	202.0
227.00	4011.0
227.90	683.0
229.00	1052.0
230.00	167.0
231.00	300.0
234.00	457.0
234.90	351.0
235.90	188.0
237.00	367.0
238.90	212.0
240.90	284.0
241.90	486.0
243.00	676.0
244.00	7556.0
245.00	1014.0
245.90	1821.0
247.00	354.0
248.90	316.0
252.90	173.0
255.00	38024.0
256.00	6020.0
257.00	571.0
258.00	2507.0
258.90	546.0
264.90	972.0
272.00	168.0
273.00	1252.0
274.00	3170.0
275.00	18144.0
276.00	2527.0
277.00	1608.0
283.00	372.0
285.00	266.0
288.90	154.0
293.00	385.0
296.00	5197.0
297.00	843.0
303.00	594.0
304.00	167.0
314.00	312.0
314.90	622.0
316.00	436.0
321.00	179.0
323.00	1614.0
323.90	253.0
327.00	305.0
327.90	301.0
332.10	199.0
333.00	181.0
334.00	1130.0
335.00	301.0
341.00	316.0
345.90	425.0
351.90	497.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

353.00	276.0
354.00	454.0
365.00	2841.0
365.90	284.0
370.90	196.0
372.00	735.0
373.10	244.0
383.00	294.0
391.10	159.0
402.00	421.0
403.00	483.0
404.00	195.0
421.00	472.0
421.90	277.0
423.00	3259.0
424.00	872.0
441.00	8638.0
442.00	57264.0
443.00	11688.0
444.00	1015.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

m/z	Abundance
39.00	1035.0
40.00	201.0
43.90	284.0
50.00	3219.0
51.00	11366.0
52.00	463.0
54.90	154.0
55.90	467.0
57.00	933.0
61.00	244.0
62.00	203.0
63.00	655.0
65.10	237.0
69.00	15838.0
72.90	429.0
74.00	1893.0
75.00	2546.0
76.00	868.0
77.00	17680.0
78.00	1134.0
79.00	1287.0
79.90	1019.0
81.00	1417.0
82.00	342.0
83.00	362.0
85.00	333.0
86.00	530.0
86.90	151.0
91.00	385.0
91.90	444.0
93.00	2065.0
94.00	194.0
97.00	189.0
98.00	1702.0
99.00	1270.0
100.10	230.0
101.00	794.0
103.00	290.0
103.90	452.0
105.00	727.0
106.00	225.0
107.00	6208.0
108.00	1129.0
110.00	10640.0
111.00	1798.0
111.90	164.0
116.00	482.0
117.00	5055.0
117.90	388.0
121.90	435.0
122.90	699.0
124.00	340.0
125.00	301.0
127.00	23184.0
128.00	1958.0
129.00	10304.0
129.90	840.0
133.00	184.0
133.90	301.0
135.00	849.0
135.90	363.0
136.90	454.0
138.00	180.0
140.90	1195.0
142.00	415.0
143.00	245.0
145.90	162.0
147.00	645.0
147.90	1279.0
149.00	317.0
151.00	182.0
152.90	360.0
153.90	329.0
154.90	618.0
156.00	1098.0
158.00	283.0
159.00	171.0
159.90	390.0
160.90	691.0
161.90	173.0
164.90	568.0
166.00	461.0
167.00	2473.0
168.00	897.0
168.90	306.0
171.00	230.0
172.00	345.0
173.00	337.0
173.90	693.0
175.00	1148.0
175.90	348.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

177.00	579.0
178.00	217.0
178.90	1962.0
180.00	1310.0
180.90	663.0
182.00	174.0
183.90	165.0
184.90	1252.0
186.00	8252.0
187.00	2056.0
187.90	158.0
188.90	546.0
191.00	242.0
192.00	912.0
192.90	635.0
194.90	190.0
195.90	1472.0
197.90	56288.0
199.00	3646.0
200.00	336.0
201.40	341.0
203.00	390.0
204.00	2355.0
205.00	3685.0
206.00	14985.0
207.00	2281.0
208.00	520.0
209.90	414.0
211.00	629.0
214.90	196.0
216.00	471.0
216.90	4780.0
217.90	487.0
219.90	173.0
221.00	3294.0
222.00	514.0
222.90	1123.0
224.00	8840.0
225.00	2214.0
225.90	316.0
227.00	3558.0
228.00	457.0
228.90	812.0
230.90	294.0
234.00	295.0
235.00	336.0
237.00	252.0
238.90	190.0
241.00	207.0
242.00	518.0
243.00	508.0
244.00	7236.0
245.00	1125.0
246.00	1181.0
246.90	367.0
248.90	255.0
253.00	218.0
255.00	36384.0
256.00	5229.0
257.00	385.0
258.00	2419.0
258.90	372.0
265.00	994.0
273.00	1116.0
274.00	3189.0
275.00	16984.0
276.00	2016.0
276.90	1603.0
278.00	354.0
283.00	153.0
283.90	219.0
285.00	350.0
293.00	203.0
295.00	157.0
296.00	5327.0
297.00	757.0
303.00	774.0
304.00	175.0
314.00	410.0
315.00	671.0
316.00	324.0
321.10	175.0
323.00	1592.0
324.00	289.0
327.00	416.0
327.90	151.0
332.90	233.0
334.00	1203.0
334.90	292.0
340.90	289.0
345.90	476.0
352.00	522.0
353.00	344.0
354.00	504.0
364.90	2548.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

365.90	451.0
370.90	151.0
372.00	920.0
373.00	281.0
382.90	189.0
402.00	318.0
402.90	390.0
420.90	632.0
422.00	405.0
423.00	3657.0
424.00	739.0
441.00	10447.0
442.00	68400.0
443.00	13659.0
444.00	1318.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21