


Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG073019\
 Data File : BG041794.D
 Acq On : 30 Jul 2019 8:13
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG072719.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jul 31 08:57:26 2019



AutoFind: Scans 2519, 2520, 2521; Background Corrected with Scan 2513

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.1	33016	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	29.6	31405	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	40.3	42864	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	106245	PASS
199	198	5	9	7.2	7651	PASS
275	198	10	60	27.2	28914	PASS
365	198	1	100	3.3	3547	PASS
441	198	0.01	100	10.1	10737	PASS
442	198	50	100	64.1	68058	PASS
443	442	15	24	19.5	13296	PASS