

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG073124\
 Data File : BG062222.D
 Acq On : 31 Jul 2024 17:42
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020587

Quant Time: Aug 01 04:04:39 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG073124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 01 03:46:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.982	152	36725	20.000	ng/u1	0.00
20) Naphthalene-d8	10.778	136	175585	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.597	164	134801	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.340	188	309816	20.000	ng/u1	0.00
79) Chrysene-d12	21.587	240	317183	20.000	ng/u1	0.00
88) Perylene-d12	24.701	264	344279	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.441	96	6451	7.476	ng/uL	0.00
4) Pyridine-d5	3.847	84	51899	19.694	ng/u1	0.00
7) Phenol-d5	7.130	99	67671	19.254	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.307	67	42889	19.787	ng/u1	0.00
11) 2-Chlorophenol-d4	7.512	132	49010	20.732	ng/u1	0.00
15) 4-Methylphenol-d8	8.669	113	57087	19.516	ng/u1	0.00
21) Nitrobenzene-d5	9.134	128	19795	27.181	ng/u1	0.00
24) 2-Nitrophenol-d4	9.850	143	18426	27.007	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.391	165	57095	22.349	ng/u1	0.00
31) 4-Chloroaniline-d4	10.902	131	85300	19.735	ng/u1	0.00
46) Dimethylphthalate-d6	14.009	166	205313	20.210	ng/u1	0.00
49) Acenaphthylene-d8	14.291	160	219903	19.608	ng/u1	0.00
54) 4-Nitrophenol-d4	14.767	143	28338	23.242	ng/u1	0.00
60) Fluorene-d10	15.589	176	171539	19.004	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.695	200	14981	20.399	ng/u1	0.00
73) Anthracene-d10	17.440	188	283480	19.855	ng/u1	0.00
81) Pyrene-d10	19.719	212	360276	20.296	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.489	264	338928	20.137	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.476	88	8212	7.715	ng/uL	96
5) Pyridine	3.864	79	56400	20.071	ng/u1	93
6) Benzaldehyde	7.119	77	45484	23.976	ng/u1	96
8) Phenol	7.154	94	70152	19.351	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.401	93	53927	19.716	ng/u1	97
12) 2-Chlorophenol	7.542	128	50454	20.546	ng/u1	99
13) 2-Methylphenol	8.411	108	54707	19.548	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.505	45	114529	19.677	ng/u1	98
16) Acetophenone	8.799	105	94838	19.963	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.787	70	52837	20.189	ng/u1#	96
18) 4-Methylphenol	8.740	108	60941	19.769	ng/u1	93
19) Hexachloroethane	9.063	117	20028	20.488	ng/u1	92
22) Nitrobenzene	9.175	77	58817	26.960	ng/u1	96
23) Isophorone	9.697	82	149658	19.304	ng/u1	98
25) 2-Nitrophenol	9.885	139	22481	28.445	ng/u1	95
26) 2,4-Dimethylphenol	9.944	107	67691	19.654	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.185	93	82218	19.596	ng/u1	99
29) 2,4-Dichlorophenol	10.414	162	54532	21.389	ng/u1	98
30) Naphthalene	10.825	128	192124	19.689	ng/u1	99
32) 4-Chloroaniline	10.925	127	82434	19.850	ng/u1	95
33) Hexachlorobutadiene	11.119	225	39012	19.329	ng/u1	96
34) Caprolactam	11.671	113	8623	7.939	ng/u1	92
35) 4-Chloro-3-methylphenol	12.035	107	67297	20.301	ng/u1	99
36) 2-Methylnaphthalene	12.429	142	134264	19.468	ng/u1	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG073124\
 Data File : BG062222.D
 Acq On : 31 Jul 2024 17:42
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 10 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SSTD020587


Quant Time: Aug 01 04:04:39 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG073124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 01 03:46:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.646	142	139379	19.608	ng/ul#	96
39) 1,2,4,5-Tetrachloroben...	12.793	216	81272	20.372	ng/ul	99
40) Hexachlorocyclopentadiene	12.782	237	42389	20.461	ng/ul	95
41) 2,4,6-Trichlorophenol	13.028	196	46498	23.587	ng/ul	99
42) 2,4,5-Trichlorophenol	13.093	196	53357	23.935	ng/ul	92
43) 1,1'-Biphenyl	13.439	154	193640	19.913	ng/ul	97
44) 2-Chloronaphthalene	13.475	162	149340	19.582	ng/ul	98
45) 2-Nitroaniline	13.669	65	37399	27.737	ng/ul	93
47) Dimethylphthalate	14.056	163	209968	19.918	ng/ul	98
48) 2,6-Dinitrotoluene	14.168	165	27122	27.312	ng/ul	94
50) Acenaphthylene	14.321	152	246955	19.518	ng/ul	99
51) 3-Nitroaniline	14.485	138	29490	23.823	ng/ul#	91
52) Acenaphthene	14.661	153	173650	19.419	ng/ul	98
53) 2,4-Dinitrophenol	14.691	184	8942	17.187	ng/ul	96
55) 4-Nitrophenol	14.785	109	26664	22.613	ng/ul	88
56) Dibenzofuran	14.996	168	239963	19.366	ng/ul	98
57) 2,4-Dinitrotoluene	14.949	165	38382	28.291	ng/ul	92
58) 2,3,4,6-Tetrachlorophenol	15.219	232	48700	22.952	ng/ul	99
59) Diethylphthalate	15.425	149	210044	19.435	ng/ul	99
61) Fluorene	15.648	166	186248	18.669	ng/ul	95
62) 4-Chlorophenyl-phenyle...	15.648	204	94769	18.768	ng/ul	97
63) 4-Nitroaniline	15.648	138	32746	23.795	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.713	198	16979	20.644	ng/ul	96
67) N-Nitrosodiphenylamine	15.854	169	171037	19.839	ng/ul	97
68) 4-Bromophenyl-phenylether	16.535	248	64635	19.479	ng/ul	99
69) Hexachlorobenzene	16.647	284	69922	19.493	ng/ul	96
70) Atrazine	16.805	200	69990	21.096	ng/ul	98
71) Pentachlorophenol	16.982	266	38350	16.821	ng/ul	94
72) Phenanthrene	17.381	178	330377	19.765	ng/ul	99
74) Anthracene	17.475	178	341608	19.743	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.398	216	80387	21.036	ng/ul	97
76) Pentachlorobenzene	14.920	250	83830	20.744	ng/ul	98
77) Carbazole	17.734	167	296417	20.874	ng/ul	99
78) Di-n-butylphthalate	18.321	149	372122	21.131	ng/ul	100
80) Fluoranthene	19.390	202	417046	20.060	ng/ul	99
82) Pyrene	19.749	202	433302	19.994	ng/ul	100
83) Butylbenzylphthalate	20.653	149	156820	21.564	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.487	252	150488	21.821	ng/ul	95
85) Benzo(a)anthracene	21.570	228	444932	19.741	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.517	149	240818	21.184	ng/ul	97
87) Chrysene	21.634	228	417417	19.817	ng/ul	99
89) Di-n-octyl phthalate	22.715	149	402983	20.848	ng/ul	100
90) Benzo(b)fluoranthene	23.720	252	419471	19.585	ng/ul	98
91) Benzo(k)fluoranthene	23.784	252	426089	20.011	ng/ul	99
93) Benzo(a)pyrene	24.554	252	404580	19.973	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	28.219	276	496300	20.838	ng/ul	99
95) Dibenzo(a,h)anthracene	28.296	278	415835	20.824	ng/ul	99
96) Benzo(g,h,i)perylene	29.312	276	396571	20.743	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Aug 01 04:04:39 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG073124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Aug 01 03:46:07 2024
Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 7.715 ng/uL
RT: 3.476 min Scan# 31
Delta R.T. 0.004 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Instrument :
BNA_G
ClientSampleId :
SSTD020587







Tgt Ion: 88 Resp: 8212
Ion Ratio Lower Upper
88 100
43 38.2 26.4 39.6
58 85.2 69.7 104.5



#5
Pyridine
Concen: 20.071 ng/uL
RT: 3.864 min Scan# 98
Delta R.T. 0.004 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion: 79 Resp: 56400
Ion Ratio Lower Upper
79 100
52 98.2 85.8 128.8
51 46.2 39.6 59.4









#6
Benzaldehyde
Concen: 23.976 ng/ul
RT: 7.119 min Scan# 611
Delta R.T. 0.004 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion: 77 Resp: 45484
Ion Ratio Lower Upper
77 100
105 94.2 75.0 112.6
106 93.3 68.2 102.4



#8
Phenol
Concen: 19.351 ng/ul
RT: 7.154 min Scan# 658
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion: 94 Resp: 70152
Ion Ratio Lower Upper
94 100
65 28.3 24.5 36.7
66 38.3 30.8 46.2









#10
 Bis(2-Chloroethyl)ether
 Concen: 19.716 ng/ul
 RT: 7.401 min Scan# 700
 Delta R.T. -0.002 min
 Lab File: BG062222.D
 Acq: 31 Jul 2024 17:42




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020587


Tgt Ion: 93	Resp: 53927
Ion Ratio	Lower Upper
93	100
63	87.5 68.0 102.0
95	30.9 26.4 39.6



#12
 2-Chlorophenol
 Concen: 20.546 ng/ul
 RT: 7.542 min Scan# 724
 Delta R.T. -0.002 min
 Lab File: BG062222.D
 Acq: 31 Jul 2024 17:42

Tgt Ion: 128	Resp: 50454
Ion Ratio	Lower Upper
128	100
64	51.1 40.6 61.0
130	30.8 25.5 38.3





#13

2-Methylphenol

Concen: 19.548 ng/ul

RT: 8.411 min Scan# 81

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42

Instrument :

BNA_G

ClientSampleId :





SSTD020587

Tgt Ion:108 Resp: 54707

Ion Ratio Lower Upper

108 100

107 88.7 68.2 102.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.677 ng/ul

RT: 8.505 min Scan# 888

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42




Tgt Ion: 45 Resp: 114529


Ion Ratio Lower Upper

45 100

77 10.6 7.8 11.6

79 7.1 5.1 7.7









#16
Acetophenone
Concen: 19.963 ng/ul
RT: 8.799 min Scan# 91
Delta R.T. 0.004 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:105 Resp: 94838
Ion Ratio Lower Upper
105 100
77 80.0 64.8 97.2
51 35.0 28.6 42.8



#17
N-Nitroso-di-n-propylamine
Concen: 20.189 ng/ul
RT: 8.787 min Scan# 936
Delta R.T. 0.004 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion: 70 Resp: 52837
Ion Ratio Lower Upper
70 100
42 75.3 62.9 94.3
101 9.7 5.9 8.9#
130 19.4 14.2 21.2









#18
4-Methylphenol
Concen: 19.769 ng/ul
RT: 8.740 min Scan# 91
Delta R.T. 0.004 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:108 Resp: 60941
Ion Ratio Lower Upper
108 100
107 102.8 88.3 132.5



#19
Hexachloroethane
Concen: 20.488 ng/ul
RT: 9.063 min Scan# 983
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:117 Resp: 20028
Ion Ratio Lower Upper
117 100
201 109.1 94.3 141.5
199 69.2 61.4 92.2





#22

Nitrobenzene

Concen: 26.960 ng/ul

RT: 9.175 min Scan# 1002

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42

Instrument :

BNA_G

ClientSampleId :

SSTD020587





Tgt Ion: 77 Resp: 58817

Ion Ratio Lower Upper

77 100

123 42.0 31.1 46.7

65 13.4 11.0 16.4



#23

Isophorone

Concen: 19.304 ng/ul

RT: 9.697 min Scan# 1091

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42




Tgt Ion: 82 Resp: 149658


Ion Ratio Lower Upper

82 100

95 7.1 5.4 8.0

138 15.1 13.1 19.7









#25
2-Nitrophenol
Concen: 28.445 ng/ul
RT: 9.885 min Scan# 1123
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:139 Resp: 22481
Ion Ratio Lower Upper
139 100
65 55.0 46.2 69.2
109 33.7 30.2 45.2



#26
2,4-Dimethylphenol
Concen: 19.654 ng/ul
RT: 9.944 min Scan# 1133
Delta R.T. 0.004 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:107 Resp: 67691
Ion Ratio Lower Upper
107 100
121 50.3 40.2 60.2
122 87.7 68.7 103.1









#27
Bis(2-Chloroethoxy)methane
Concen: 19.596 ng/ul
RT: 10.185 min Scan# 1174
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion: 93 Resp: 82218
Ion Ratio Lower Upper
93 100
95 31.8 25.8 38.6
123 10.7 9.0 13.4



#29
2,4-Dichlorophenol
Concen: 21.389 ng/ul
RT: 10.414 min Scan# 1213
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:162 Resp: 54532
Ion Ratio Lower Upper
162 100
164 64.4 52.2 78.4
98 38.7 28.8 43.2









#30
Naphthalene
Concen: 19.689 ng/ul
RT: 10.825 min Scan# 11
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:128 Resp: 192124
Ion Ratio Lower Upper
128 100
129 11.2 9.1 13.7
127 13.1 11.0 16.4



#32
4-Chloroaniline
Concen: 19.850 ng/ul
RT: 10.925 min Scan# 1300
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:127 Resp: 82434
Ion Ratio Lower Upper
127 100
129 33.5 24.8 37.2









#33
Hexachlorobutadiene
Concen: 19.329 ng/ul
RT: 11.119 min Scan# 11
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:225 Resp: 39012
Ion Ratio Lower Upper
225 100
223 66.5 48.6 72.8
227 65.5 52.6 78.8



#34
Caprolactam
Concen: 7.939 ng/ul
RT: 11.671 min Scan# 1427
Delta R.T. 0.004 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:113 Resp: 8623
Ion Ratio Lower Upper
113 100
55 228.6 177.0 265.4
56 162.0 117.6 176.4









#35
 4-Chloro-3-methylphenol
 Concen: 20.301 ng/ul
 RT: 12.035 min Scan# 1489
 Delta R.T. -0.002 min
 Lab File: BG062222.D
 Acq: 31 Jul 2024 17:42

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020587

Tgt Ion:107 Resp: 67297




Ion	Ratio	Lower	Upper
107	100		
144	24.0	19.0	28.4
142	71.7	57.8	86.8




#36
 2-Methylnaphthalene
 Concen: 19.468 ng/ul
 RT: 12.429 min Scan# 1556
 Delta R.T. -0.002 min
 Lab File: BG062222.D
 Acq: 31 Jul 2024 17:42

Tgt Ion:142 Resp: 134264

Ion	Ratio	Lower	Upper
142	100		
141	88.7	73.7	110.5





#37

1-Methylnaphthalene

Concen: 19.608 ng/ul

RT: 12.646 min Scan# 1

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42

Instrument :

BNA_G

ClientSampleId :

SSTD020587





Tgt Ion:142 Resp: 139379

Ion Ratio Lower Upper

142 100

141 96.8 74.0 111.0

116 4.4 2.7 4.1#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.372 ng/ul

RT: 12.793 min Scan# 1618

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42

Tgt Ion:216 Resp: 81272




Ion Ratio Lower Upper


216 100

214 79.6 64.0 96.0

179 19.7 17.4 26.2

108 16.2 12.8 19.2





#40

Hexachlorocyclopentadiene

Concen: 20.461 ng/ul

RT: 12.782 min Scan# 1

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42

Instrument :

BNA_G

ClientSampleId :

SSTD020587





Tgt Ion:237 Resp: 42389

Ion Ratio Lower Upper

237 100

235 65.7 48.9 73.3

272 12.5 10.4 15.6



#41

2,4,6-Trichlorophenol

Concen: 23.587 ng/ul

RT: 13.028 min Scan# 1658

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42




Tgt Ion:196 Resp: 46498


Ion Ratio Lower Upper

196 100

198 100.5 81.0 121.6

200 31.0 25.2 37.8









#42
2,4,5-Trichlorophenol
Concen: 23.935 ng/ul
RT: 13.093 min Scan# 1
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:196 Resp: 53357
Ion Ratio Lower Upper
196 100
198 92.3 81.7 122.5
200 31.1 25.8 38.6



#43
1,1'-Biphenyl
Concen: 19.913 ng/ul
RT: 13.439 min Scan# 1728
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:154 Resp: 193640
Ion Ratio Lower Upper
154 100
153 41.1 31.3 46.9
76 12.8 10.6 15.8





#44

2-Chloronaphthalene

Concen: 19.582 ng/ul

RT: 13.475 min Scan# 1

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42

Instrument :

BNA_G

ClientSampleId :

SSTD020587





Tgt Ion:162 Resp: 149340

Ion Ratio Lower Upper

162 100

164 32.9 25.2 37.8

127 39.5 30.9 46.3



#45

2-Nitroaniline

Concen: 27.737 ng/ul

RT: 13.669 min Scan# 1767

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42




Tgt Ion: 65 Resp: 37399


Ion Ratio Lower Upper

65 100

92 63.7 47.2 70.8

138 92.0 68.6 103.0









#47
Dimethylphthalate
Concen: 19.918 ng/ul
RT: 14.056 min Scan# 1833
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:163 Resp: 209968
Ion Ratio Lower Upper
163 100
194 3.6 3.0 4.6
164 10.4 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 27.312 ng/ul
RT: 14.168 min Scan# 1852
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:165 Resp: 27122
Ion Ratio Lower Upper
165 100
89 61.9 55.0 82.4
121 24.3 19.4 29.2









#50
Acenaphthylene
Concen: 19.518 ng/ul
RT: 14.321 min Scan# 11
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:152 Resp: 246955
Ion Ratio Lower Upper
152 100
151 20.4 16.2 24.4
153 14.1 11.0 16.4



#51
3-Nitroaniline
Concen: 23.823 ng/ul
RT: 14.485 min Scan# 1906
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:138 Resp: 29490
Ion Ratio Lower Upper
138 100
108 22.0 26.0 39.0#
92 137.3 103.4 155.2









#52
Acenaphthene
Concen: 19.419 ng/ul
RT: 14.661 min Scan# 1936
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:153 Resp: 173650
Ion Ratio Lower Upper
153 100
152 45.3 35.0 52.4
154 84.5 66.9 100.3



#53
2,4-Dinitrophenol
Concen: 17.187 ng/ul
RT: 14.691 min Scan# 1941
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:184 Resp: 8942
Ion Ratio Lower Upper
184 100
63 65.9 55.3 82.9
154 95.2 72.9 109.3









#55
4-Nitrophenol
Concen: 22.613 ng/ul
RT: 14.785 min Scan# 1957
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:109 Resp: 26664
Ion Ratio Lower Upper
109 100
139 122.3 91.2 136.8
65 139.2 97.0 145.4



#56
Dibenzofuran
Concen: 19.366 ng/ul
RT: 14.996 min Scan# 1993
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:168 Resp: 239963
Ion Ratio Lower Upper
168 100
139 36.3 29.8 44.8









#57
2,4-Dinitrotoluene
Concen: 28.291 ng/ul
RT: 14.949 min Scan# 1985
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:165 Resp: 38382
Ion Ratio Lower Upper
165 100
63 48.1 43.4 65.0
182 2.5 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 22.952 ng/ul
RT: 15.219 min Scan# 2031
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:232 Resp: 48700
Ion Ratio Lower Upper
232 100
131 37.3 29.9 44.9
130 2.1 1.8 2.6
166 27.5 22.8 34.2









#59
Diethylphthalate
Concen: 19.435 ng/ul
RT: 15.425 min Scan# 2066
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:149 Resp: 210044
Ion Ratio Lower Upper
149 100
177 22.7 18.2 27.4
150 12.8 9.7 14.5



#61
Fluorene
Concen: 18.669 ng/ul
RT: 15.648 min Scan# 2104
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:166 Resp: 186248
Ion Ratio Lower Upper
166 100
165 104.5 79.4 119.2
167 13.8 10.1 15.1









#62
4-Chlorophenyl-phenylether
Concen: 18.768 ng/ul
RT: 15.648 min Scan# 2104
Delta R.T. 0.004 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:204 Resp: 94769
Ion Ratio Lower Upper
204 100
206 35.8 26.2 39.2
141 56.7 46.4 69.6



#63
4-Nitroaniline
Concen: 23.795 ng/ul
RT: 15.648 min Scan# 2104
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:138 Resp: 32746
Ion Ratio Lower Upper
138 100
92 58.7 48.6 73.0
108 91.8 78.6 118.0









#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.644 ng/ul
 RT: 15.713 min Scan# 2115
 Delta R.T. 0.004 min
 Lab File: BG062222.D
 Acq: 31 Jul 2024 17:42




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020587


Tgt Ion:198 Resp: 16979
 Ion Ratio Lower Upper
 198 100
 182 4.6 3.1 4.7
 77 21.8 19.0 28.6



#67
 N-Nitrosodiphenylamine
 Concen: 19.839 ng/ul
 RT: 15.854 min Scan# 2139
 Delta R.T. -0.002 min
 Lab File: BG062222.D
 Acq: 31 Jul 2024 17:42

Tgt Ion:169 Resp: 171037
 Ion Ratio Lower Upper
 169 100
 168 69.0 57.4 86.0
 167 34.7 28.5 42.7









#68
4-Bromophenyl-phenylether
Concen: 19.479 ng/ul
RT: 16.535 min Scan# 21
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:248 Resp: 64635
Ion Ratio Lower Upper
248 100
250 99.9 80.0 120.0
141 73.1 56.3 84.5



#69
Hexachlorobenzene
Concen: 19.493 ng/ul
RT: 16.647 min Scan# 2274
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:284 Resp: 69922
Ion Ratio Lower Upper
284 100
142 30.5 25.4 38.2
249 30.7 26.9 40.3





#70

Atrazine

Concen: 21.096 ng/ul

RT: 16.805 min Scan# 21

Delta R.T. 0.004 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42

Instrument :

BNA_G

ClientSampleId :

SSTD020587





Tgt Ion:200 Resp: 69990

Ion Ratio Lower Upper

200 100

173 24.8 20.6 30.8

215 48.3 37.8 56.6



#71

Pentachlorophenol

Concen: 16.821 ng/ul

RT: 16.982 min Scan# 2331

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42




Tgt Ion:266 Resp: 38350


Ion Ratio Lower Upper

266 100

264 58.9 52.7 79.1

268 61.2 47.1 70.7









#72
Phenanthrene
Concen: 19.765 ng/ul
RT: 17.381 min Scan# 21
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:178 Resp: 330377
Ion Ratio Lower Upper
178 100
179 15.1 12.2 18.2
176 19.4 15.3 22.9



#74
Anthracene
Concen: 19.743 ng/ul
RT: 17.475 min Scan# 2415
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:178 Resp: 341608
Ion Ratio Lower Upper
178 100
179 14.4 12.6 18.8
176 18.3 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.036 ng/ul

RT: 13.398 min Scan# 1

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42

Instrument :

BNA_G

ClientSampleId :

SSTD020587





Tgt Ion:216 Resp: 80387

Ion Ratio Lower Upper

216 100

214 80.5 62.8 94.2

218 48.2 36.7 55.1



#76

Pentachlorobenzene

Concen: 20.744 ng/ul

RT: 14.920 min Scan# 1980

Delta R.T. 0.004 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42




Tgt Ion:250 Resp: 83830


Ion Ratio Lower Upper

250 100

248 60.8 49.8 74.8

252 60.4 49.7 74.5





#77

Carbazole

Concen: 20.874 ng/ul

RT: 17.734 min Scan# 2459

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42

Instrument :

BNA_G

ClientSampleId :

SSTD020587





Tgt Ion:167 Resp: 296417

Ion Ratio Lower Upper

167 100

166 22.5 17.7 26.5

139 12.5 10.4 15.6



#78

Di-n-butylphthalate

Concen: 21.131 ng/ul

RT: 18.321 min Scan# 2559

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42




Tgt Ion:149 Resp: 372122


Ion Ratio Lower Upper

149 100

150 8.9 7.1 10.7

104 4.9 4.2 6.2









#80
Fluoranthene
Concen: 20.060 ng/ul
RT: 19.390 min Scan# 21
Delta R.T. 0.004 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:202 Resp: 417046
Ion Ratio Lower Upper
202 100
101 9.7 8.2 12.4
100 8.2 6.9 10.3



#82
Pyrene
Concen: 19.994 ng/ul
RT: 19.749 min Scan# 2802
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:202 Resp: 433302
Ion Ratio Lower Upper
202 100
101 11.5 9.0 13.6
100 10.0 8.0 12.0





#83

Butylbenzylphthalate

Concen: 21.564 ng/ul

RT: 20.653 min Scan# 2956

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42

Instrument :

BNA_G

ClientSampleId :

SSTD020587





Tgt Ion:149 Resp: 156820

Ion Ratio Lower Upper

149 100

91 73.9 60.0 90.0

206 21.9 17.2 25.8



#84

3,3'-Dichlorobenzidine

Concen: 21.821 ng/ul

RT: 21.487 min Scan# 3098

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42




Tgt Ion:252 Resp: 150488


Ion Ratio Lower Upper

252 100

254 66.2 49.7 74.5

126 9.5 6.9 10.3









#85
Benzo(a)anthracene
Concen: 19.741 ng/ul
RT: 21.570 min Scan# 3112
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:228 Resp: 444932
Ion Ratio Lower Upper
228 100
229 19.6 15.9 23.9
226 27.2 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.184 ng/ul
RT: 21.517 min Scan# 3103
Delta R.T. -0.008 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:149 Resp: 240818
Ion Ratio Lower Upper
149 100
167 25.2 21.4 32.0
279 3.3 2.8 4.2





#87

Chrysene

Concen: 19.817 ng/ul

RT: 21.634 min Scan# 3123

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42

Instrument :

BNA_G

ClientSampleId :

SSTD020587





Tgt Ion:228 Resp: 417417

Ion Ratio Lower Upper

228 100

226 30.2 24.2 36.4

229 18.7 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 20.848 ng/ul




RT: 22.715 min Scan# 3307


Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42

Tgt Ion:149 Resp: 402983





#90

Benzo(b)fluoranthene

Concen: 19.585 ng/ul

RT: 23.720 min Scan# 3478

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42

Instrument :

BNA_G

ClientSampleId :

SSTD020587





Tgt Ion:252 Resp: 419471

Ion Ratio Lower Upper

252 100

253 21.8 17.0 25.6

125 9.4 6.7 10.1



#91

Benzo(k)fluoranthene

Concen: 20.011 ng/ul

RT: 23.784 min Scan# 3489

Delta R.T. -0.002 min

Lab File: BG062222.D

Acq: 31 Jul 2024 17:42




Tgt Ion:252 Resp: 426089


Ion Ratio Lower Upper

252 100

253 21.6 17.5 26.3

125 8.8 7.3 10.9









#93
Benzo(a)pyrene
Concen: 19.973 ng/ul
RT: 24.554 min Scan# 3620
Delta R.T. -0.008 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42




Instrument :
BNA_G
ClientSampleId :
SSTD020587


Tgt Ion:252 Resp: 404580
Ion Ratio Lower Upper
252 100
253 21.2 17.4 26.0
125 9.6 7.5 11.3



#94
Indeno(1,2,3-cd)pyrene
Concen: 20.838 ng/ul
RT: 28.219 min Scan# 4244
Delta R.T. -0.002 min
Lab File: BG062222.D
Acq: 31 Jul 2024 17:42

Tgt Ion:276 Resp: 496300
Ion Ratio Lower Upper
276 100
138 16.8 13.5 20.3
277 24.0 19.8 29.6









#95
 Dibenzo(a,h)anthracene
 Concen: 20.824 ng/ul
 RT: 28.296 min Scan# 41
 Delta R.T. -0.002 min
 Lab File: BG062222.D
 Acq: 31 Jul 2024 17:42

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020587

Tgt Ion:278 Resp: 415835
 Ion Ratio Lower Upper
 278 100
 139 13.2 10.2 15.2
 279 23.5 18.8 28.2



#96
 Benzo(g,h,i)perylene
 Concen: 20.743 ng/ul
 RT: 29.312 min Scan# 4430
 Delta R.T. -0.008 min
 Lab File: BG062222.D
 Acq: 31 Jul 2024 17:42

Tgt Ion:276 Resp: 396571
 Ion Ratio Lower Upper
 276 100
 138 17.1 15.4 23.2
 277 23.6 19.4 29.0

