



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG073124\
Data File : BG062213.D
Acq On : 31 Jul 2024 8:27
Operator : MA/JU
Sample : DFTPP405
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP405

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG073124.MA.M
Title : SVOA CALIBRATION
Last Update : Thu Aug 01 03:46:07 2024



AutoFind: Scans 2470, 2471, 2472; Background Corrected with Scan 2464

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	41.1	33781	PASS
68	69	0.00	2	0.9	234	PASS
69	198	0.00	100	32.3	26541	PASS
70	69	0.00	2	0.8	215	PASS
127	198	10	80	39.9	32805	PASS
197	198	0.00	2	0.2	172	PASS
198	198	100	100	100.0	82229	PASS
199	198	5	9	6.5	5325	PASS
275	198	10	60	26.4	21669	PASS
365	198	1	100	3.7	3030	PASS
441	198	0.01	100	10.0	8241	PASS
442	198	50	100	58.3	47936	PASS
443	442	15	24	18.8	9007	PASS

Scan 2464 (17.763 min): BG062213.D\data.ms
DFTPP405

m/z	Abundance
44.15	1485.0
207.15	494.0

Instrument :

BNA_G

ClientSampleId :

DFTPP405

m/z	Abundance
37.15	152.0
38.15	474.0
39.15	2979.0
40.05	303.0
41.05	223.0
44.05	1815.0
49.15	272.0
50.15	10286.0
51.15	38656.0
52.15	2118.0
53.05	291.0
55.05	395.0
56.05	1219.0
57.05	2700.0
60.95	476.0
62.15	477.0
63.15	1639.0
65.05	603.0
69.05	30320.0
70.05	175.0
73.15	265.0
74.15	3250.0
75.15	4792.0
77.15	36720.0
78.15	2538.0
79.05	2192.0
80.05	1645.0
81.05	2975.0
82.15	584.0
83.05	444.0
85.15	389.0
86.05	717.0
87.05	314.0
91.15	523.0
92.15	451.0
93.05	3919.0
94.05	358.0
95.25	275.0
96.15	255.0
97.05	220.0
98.05	3100.0
99.05	2232.0
100.05	184.0
101.05	1597.0
103.15	412.0
104.05	928.0
105.05	913.0
107.15	11380.0
108.15	1624.0
110.05	19360.0
111.05	3225.0
112.05	436.0
116.05	605.0
117.05	9207.0
118.05	603.0
119.95	167.0
122.15	589.0
123.05	1352.0
124.05	706.0
125.05	416.0
127.05	35456.0
128.05	2899.0
129.05	13492.0
130.05	1153.0
131.05	296.0
134.05	378.0
135.15	1502.0
135.95	405.0
137.05	637.0
141.05	2029.0
142.15	820.0
143.15	309.0
145.05	195.0
146.05	464.0
147.05	905.0
148.05	2453.0
149.15	547.0
150.15	177.0
151.05	151.0
152.05	178.0
153.15	680.0
154.05	541.0
155.05	1163.0
156.15	1531.0
157.15	449.0
158.05	468.0
158.95	376.0
160.05	621.0
160.95	871.0
162.05	371.0
165.05	855.0
166.15	586.0

Instrument :
BNA_G
ClientSampleId :
DFTPP405

167.05	3456.0
168.15	1844.0
169.05	270.0
170.15	155.0
171.05	173.0
171.95	362.0
173.05	604.0
174.05	954.0
175.15	1724.0
176.05	492.0
177.05	874.0
178.05	413.0
179.05	3227.0
180.15	1732.0
181.15	972.0
182.05	153.0
183.05	233.0
184.05	322.0
185.05	1677.0
186.15	11650.0
187.05	2956.0
188.05	255.0
189.05	709.0
191.05	287.0
192.15	952.0
193.15	1118.0
193.95	289.0
196.05	2341.0
198.05	81912.0
199.05	5644.0
199.95	463.0
201.55	341.0
203.15	875.0
204.05	3305.0
205.05	5686.0
206.15	22504.0
207.15	3318.0
208.05	851.0
209.05	297.0
211.05	1090.0
215.05	210.0
215.95	553.0
217.05	6774.0
218.05	849.0
220.05	240.0
221.05	3261.0
223.05	1518.0
224.05	13420.0
225.05	3375.0
226.05	295.0
227.05	6754.0
227.95	933.0
229.05	1128.0
230.05	324.0
231.05	423.0
233.05	169.0
234.05	338.0
234.95	370.0
236.15	292.0
237.05	425.0
239.05	243.0
239.95	208.0
240.15	211.0
241.05	251.0
242.05	550.0
243.25	650.0
244.05	9199.0
245.05	1304.0
246.05	2346.0
246.95	522.0
248.95	369.0
253.05	268.0
255.05	51880.0
256.05	7820.0
257.05	772.0
258.05	3042.0
259.05	488.0
265.05	1524.0
271.05	216.0
272.05	195.0
273.05	1715.0
274.05	4046.0
275.05	20368.0
276.05	2975.0
277.05	2218.0
277.95	311.0
282.15	150.0
283.15	300.0
284.15	181.0
285.15	452.0
289.05	159.0
292.95	384.0
293.95	164.0
296.05	6732.0
297.05	900.0
303.15	655.0

Instrument :
BNA_G
ClientSampleId :
DFTPP405

304.05	179.0
313.95	294.0
314.95	941.0
316.05	497.0
320.95	202.0
323.05	1891.0
324.05	381.0
326.95	482.0
327.95	212.0
331.95	176.0
333.25	171.0
334.05	1088.0
335.05	277.0
341.05	263.0
346.05	543.0
352.05	601.0
353.05	518.0
353.95	466.0
364.95	2351.0
365.95	437.0
372.05	706.0
372.95	221.0
382.95	188.0
391.05	173.0
401.85	373.0
403.05	360.0
404.05	222.0
421.05	472.0
422.05	316.0
423.05	2440.0
423.95	439.0
441.05	6495.0
442.05	40448.0
443.05	7930.0
444.05	718.0

Instrument :
BNA_G
ClientSampleId :
DFTPP405

m/z	Abundance
38.05	494.0
39.15	2732.0
40.15	316.0
41.05	250.0
44.15	1404.0
50.15	9162.0
51.15	39440.0
52.15	1668.0
55.15	348.0
56.15	1326.0
57.15	2313.0
61.15	434.0
62.15	698.0
63.15	1507.0
64.05	231.0
65.05	855.0
68.15	701.0
69.05	29680.0
70.05	311.0
73.15	258.0
74.15	3703.0
75.15	5244.0
76.15	2048.0
77.15	37936.0
78.15	2759.0
79.05	2639.0
80.05	1585.0
81.15	2906.0
82.05	762.0
83.05	476.0
85.05	496.0
86.05	692.0
87.15	306.0
88.05	153.0
91.05	659.0
91.95	582.0
93.05	4842.0
93.95	418.0
95.05	271.0
96.25	194.0
98.15	3184.0
99.05	2449.0
100.15	193.0
101.05	1462.0
102.95	359.0
104.05	1023.0
105.05	913.0
107.05	11504.0
108.15	1720.0
110.05	19880.0
111.05	3219.0
112.05	381.0
113.05	191.0
116.05	694.0
117.05	9975.0
118.05	728.0
120.05	222.0
122.05	684.0
123.05	1146.0
124.05	582.0
125.05	559.0
127.05	38912.0
128.05	2811.0
129.05	16122.0
130.15	1453.0
131.15	313.0
132.95	163.0
133.95	433.0
135.05	1394.0
136.15	462.0
137.05	777.0
138.15	165.0
140.05	185.0
141.05	1926.0
141.95	607.0
142.95	422.0
144.15	277.0
145.95	470.0
147.15	1199.0
148.15	2531.0
148.95	462.0
151.05	319.0
153.05	586.0
153.95	606.0
155.15	1499.0
156.15	1796.0
157.05	291.0
157.85	375.0
159.05	437.0
160.05	584.0
161.15	1133.0
162.05	395.0

Instrument :
BNA_G
ClientSampleId :
DFTPP405

165.05	727.0
165.95	423.0
167.05	4296.0
168.15	2310.0
169.15	370.0
170.05	178.0
171.05	337.0
172.05	457.0
173.05	495.0
174.05	970.0
175.05	1946.0
176.05	400.0
177.05	975.0
178.05	303.0
179.05	3824.0
180.05	2120.0
181.15	1253.0
182.15	193.0
184.15	413.0
185.05	1868.0
186.05	14283.0
187.15	3524.0
188.05	283.0
189.05	763.0
190.05	193.0
190.95	525.0
192.05	1388.0
193.05	1267.0
194.15	458.0
195.05	324.0
196.05	2945.0
198.05	97192.0
199.05	6299.0
200.05	562.0
201.75	381.0
203.15	748.0
204.05	3788.0
205.05	6491.0
206.05	26880.0
207.05	3706.0
208.05	909.0
209.15	352.0
211.15	1259.0
215.05	459.0
216.05	624.0
217.05	7233.0
218.15	822.0
221.15	2887.0
221.85	1055.0
223.05	1860.0
224.05	15862.0
225.05	4420.0
226.05	506.0
227.05	7204.0
228.15	946.0
229.05	1303.0
229.95	204.0
230.95	734.0
233.15	174.0
234.05	464.0
235.05	428.0
236.05	303.0
237.05	401.0
239.05	265.0
239.95	236.0
241.05	409.0
242.15	736.0
243.05	957.0
244.05	11659.0
245.05	1527.0
246.05	2283.0
247.05	597.0
248.95	449.0
252.05	182.0
253.05	309.0
255.05	62576.0
256.05	9524.0
256.95	662.0
258.05	4226.0
258.95	593.0
265.05	1748.0
265.95	227.0
271.05	172.0
272.05	170.0
273.05	1920.0
274.05	4673.0
275.05	25672.0
276.05	3224.0
276.95	2575.0
278.05	451.0
283.05	170.0
284.05	164.0
285.05	374.0
293.05	604.0
294.25	186.0
296.05	8737.0

Instrument :
BNA_G
ClientSampleId :
DFTPP405

296.95	1026.0
302.15	167.0
303.15	1069.0
304.05	347.0
308.05	181.0
309.95	153.0
314.05	388.0
314.95	947.0
315.95	442.0
321.05	158.0
323.15	2193.0
324.15	613.0
327.05	584.0
328.15	251.0
331.95	292.0
332.95	236.0
334.05	1700.0
335.05	361.0
340.95	258.0
345.95	496.0
352.05	770.0
353.05	308.0
354.05	752.0
365.05	3680.0
366.15	626.0
371.05	218.0
372.05	1340.0
372.95	319.0
383.05	547.0
401.85	343.0
402.05	343.0
402.95	652.0
404.15	208.0
421.15	624.0
421.95	459.0
423.05	2722.0
423.95	670.0
441.05	9988.0
442.05	56368.0
443.05	10706.0
443.95	983.0

Instrument :
BNA_G
ClientSampleId :
DFTPP405

m/z	Abundance
38.05	262.0
39.05	1748.0
40.15	205.0
41.15	154.0
44.15	1352.0
49.15	233.0
50.15	6186.0
51.15	23248.0
52.15	1488.0
55.15	436.0
56.15	817.0
57.15	1469.0
61.15	286.0
62.15	257.0
63.15	1100.0
65.15	514.0
67.05	152.0
69.05	19624.0
70.05	158.0
73.05	202.0
74.15	2056.0
75.05	2968.0
76.15	1202.0
77.15	22552.0
78.15	1762.0
79.05	1479.0
80.05	1184.0
81.05	2179.0
82.05	387.0
83.05	545.0
85.05	369.0
86.05	493.0
86.95	245.0
91.15	425.0
91.95	422.0
93.05	2866.0
93.95	277.0
98.05	1963.0
99.15	1582.0
100.05	239.0
101.05	951.0
103.15	284.0
104.15	884.0
105.05	451.0
106.15	161.0
107.05	7850.0
108.05	1387.0
110.05	14074.0
111.05	2030.0
112.05	247.0
116.15	472.0
117.15	6911.0
117.95	415.0
120.95	150.0
121.95	540.0
123.05	750.0
124.05	454.0
125.15	313.0
127.05	24048.0
128.05	2177.0
129.05	10051.0
129.95	848.0
131.05	238.0
134.15	226.0
135.05	981.0
136.05	332.0
137.05	381.0
141.15	1411.0
142.05	492.0
142.95	313.0
145.95	261.0
147.05	851.0
148.05	1611.0
149.15	397.0
150.05	194.0
150.95	209.0
153.05	521.0
154.15	364.0
155.05	798.0
156.05	1035.0
157.05	358.0
157.95	176.0
158.95	242.0
160.05	451.0
161.05	676.0
161.85	173.0
162.15	169.0
165.05	668.0
166.05	530.0
167.15	2825.0
168.05	1512.0
168.95	315.0

Instrument :
BNA_G
ClientSampleId :
DFTPP405

170.05	166.0
170.95	225.0
171.95	389.0
173.05	470.0
174.05	665.0
175.05	1357.0
176.15	254.0
177.05	493.0
178.05	293.0
179.05	2838.0
180.05	1598.0
181.05	966.0
183.85	215.0
185.05	1199.0
186.05	9431.0
187.05	2947.0
188.05	249.0
189.05	741.0
190.15	209.0
191.05	283.0
192.05	861.0
193.15	1166.0
194.05	186.0
196.05	1980.0
196.75	516.0
198.05	67584.0
199.05	4033.0
199.95	214.0
201.35	269.0
201.75	289.0
203.05	539.0
204.05	2972.0
205.05	4735.0
206.05	19320.0
207.05	2481.0
208.05	799.0
209.05	302.0
210.35	455.0
211.05	899.0
214.95	276.0
216.05	597.0
217.05	4901.0
218.05	660.0
220.05	171.0
221.05	2864.0
221.95	781.0
223.05	1432.0
224.05	10930.0
225.05	2607.0
226.15	288.0
227.05	5422.0
227.95	760.0
229.05	1074.0
229.95	202.0
230.95	417.0
234.05	416.0
234.95	369.0
236.15	312.0
237.05	436.0
239.05	357.0
239.95	161.0
241.05	523.0
242.05	678.0
242.95	657.0
244.05	8681.0
245.05	946.0
246.05	1986.0
246.95	236.0
248.95	367.0
252.15	161.0
253.15	167.0
255.05	46872.0
256.05	6682.0
256.95	489.0
258.05	2962.0
259.05	484.0
265.05	1335.0
266.05	185.0
272.15	198.0
273.05	1340.0
274.05	4044.0
275.05	18968.0
276.05	2598.0
277.05	1782.0
277.85	198.0
283.05	312.0
284.95	239.0
292.95	406.0
296.05	6463.0
297.05	728.0
302.05	210.0
303.05	595.0
304.05	199.0
313.95	278.0
314.95	924.0
316.05	337.0

Instrument :
BNA_G
ClientSampleId :
DFTPP405

320.95	181.0
323.05	1731.0
323.85	362.0
326.95	454.0
327.95	298.0
332.95	152.0
334.05	1337.0
335.05	293.0
341.05	252.0
346.05	577.0
351.95	511.0
352.25	469.0
353.05	392.0
354.05	549.0
364.95	3059.0
365.85	451.0
371.15	249.0
372.05	915.0
373.05	372.0
382.95	284.0
390.85	174.0
401.95	457.0
402.95	543.0
404.05	167.0
421.05	330.0
422.05	396.0
423.05	2409.0
424.05	492.0
424.95	164.0
441.05	8239.0
442.05	46992.0
443.05	8386.0
444.05	951.0

Instrument :
BNA_G
ClientSampleId :
DFTPP405