

Data Path : Z:\HPCHEM1\BNA_G\Data\BG080217\
 Data File : BG028144.D
 Acq On : 3 Aug 2017 16:48
 Operator : SJ/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02096

Quant Time: Aug 03 20:03:56 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 03 19:32:32 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	130	0.00
2	1,4-Dioxane	0.451	0.464	-2.9	137	0.00
3 S	1,4-Dioxane-d8	0.429	0.368	14.2	123	0.01
4	Benzaldehyde	1.269	1.297	-2.2	124	0.00
5 S	Phenol-d5	2.197	2.158	1.8	134	0.00
6	Phenol	2.183	2.168	0.7	134	0.01
7 S	Bis-(2-Chloroethyl)ether-d8	1.391	1.357	2.4	126	0.00
8	Bis(2-Chloroethyl)ether	1.487	1.414	4.9	125	0.00
9 S	2-Chlorophenol-d4	1.386	1.443	-4.1	132	0.00
10	2-Chlorophenol	1.350	1.379	-2.1	133	0.00
11	2-Methylphenol	1.541	1.481	3.9	126	0.00
12	2,2'-oxybis(1-Chloropropane	3.610	3.443	4.6	121	0.00
13 S	4-Methylphenol-d8	1.836	1.808	1.5	128	0.00
14	Acetophenone	2.618	2.597	0.8	126	0.00
15 P	N-Nitroso-di-n-propylamine	1.488	1.514	-1.7	126	0.00
16	4-Methylphenol	1.740	1.772	-1.8	132	0.00
17	Hexachloroethane	0.560	0.549	2.0	126	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	134	0.00
19 S	Nitrobenzene-d5	0.156	0.155	0.6	135	0.00
20	Nitrobenzene	0.453	0.437	3.5	132	0.00
21	Isophorone	0.860	0.833	3.1	131	0.00
22 S	2-Nitrophenol-d4	0.171	0.185	-8.2	146	0.00
23 C	2-Nitrophenol	0.173	0.176	-1.7	136	0.00
24	2,4-Dimethylphenol	0.430	0.419	2.6	130	0.00
25	Bis(2-Chloroethoxy)methane	0.469	0.443	5.5	126	0.00
26 S	2,4-Dichlorophenol-d3	0.355	0.362	-2.0	140	0.00
27 C	2,4-Dichlorophenol	0.323	0.332	-2.8	142	0.00
28	Naphthalene	0.991	0.969	2.2	132	0.00
29 S	4-Chloroaniline-d4	0.395	0.442	-11.9	135	0.00
30	4-Chloroaniline	0.383	0.417	-8.9	132	0.00
31 C	Hexachlorobutadiene	0.238	0.237	0.4	138	0.00
32	Caprolactam	0.134	0.136	-1.5	139	0.00
33 C	4-Chloro-3-methylphenol	0.412	0.442	-7.3	143	0.00
34	2-Methylnaphthalene	0.796	0.791	0.6	135	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	134	0.00
36	1,2,4,5-Tetrachlorobenzene	0.645	0.636	1.4	139	0.00
37	Hexachlorocyclopentadiene	0.356	0.284	20.2#	122	0.00
38 C	2,4,6-Trichlorophenol	0.402	0.449	-11.7	152#	0.00
39	2,4,5-Trichlorophenol	0.438	0.468	-6.8	146	0.00
40	1,1'-Biphenyl	1.461	1.475	-1.0	139	0.00
41	2-Chloronaphthalene	1.154	1.167	-1.1	138	0.00
42	2-Nitroaniline	0.484	0.498	-2.9	133	0.00
43 S	Dimethylphthalate-d6	1.779	1.685	5.3	126	0.00
44	Dimethylphthalate	1.638	1.565	4.5	130	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG080217\
 Data File : BG028144.D
 Acq On : 3 Aug 2017 16:48
 Operator : SJ/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02096

Quant Time: Aug 03 20:03:56 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 03 19:32:32 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.338	0.343	-1.5	133	0.00
46 S	Acenaphthylene-d8	2.006	2.018	-0.6	137	0.00
47	Acenaphthylene	1.917	1.910	0.4	135	0.00
48	3-Nitroaniline	0.305	0.313	-2.6	134	0.00
49 C	Acenaphthene	1.293	1.286	0.5	133	0.00
50	2,4-Dinitrophenol	0.191	0.190	0.5	158#	0.00
51 S	4-Nitrophenol-d4	0.283	0.252	11.0	123	0.01
52	4-Nitrophenol	0.293	0.260	11.3	119	0.00
53	Dibenzofuran	1.886	1.891	-0.3	136	0.00
54	2,4-Dinitrotoluene	0.512	0.529	-3.3	138	0.00
55	2,3,4,6-Tetrachlorophenol	0.408	0.432	-5.9	144	0.00
56	Diethylphthalate	1.921	1.780	7.3	127	0.00
57 S	Fluorene-d10	1.596	1.580	1.0	135	0.00
58	Fluorene	1.669	1.639	1.8	135	0.00
59	4-Chlorophenyl-phenylether	0.896	0.857	4.4	133	0.00
60	4-Nitroaniline	0.361	0.331	8.3	127	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	124	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.130	0.131	-0.8	140	0.00
63	4,6-Dinitro-2-methylphenol	0.132	0.131	0.8	137	0.00
64	N-Nitrosodiphenylamine	0.526	0.551	-4.8	131	0.00
65	4-Bromophenyl-phenylether	0.215	0.227	-5.6	132	0.00
66	Hexachlorobenzene	0.229	0.240	-4.8	134	0.00
67	Atrazine	0.229	0.225	1.7	123	0.00
68 C	Pentachlorophenol	0.116	0.115	0.9	145	0.00
69	Phenanthrene	1.007	1.000	0.7	124	0.00
70 S	Anthracene-d10	0.959	0.951	0.8	127	0.00
71	Anthracene	1.029	1.038	-0.9	128	0.00
72	Carbazole	0.895	0.858	4.1	122	0.00
73	Di-n-butylphthalate	1.114	1.100	1.3	122	0.00
74 C	Fluoranthene	1.223	1.131	7.5	115	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	116	0.00
76 S	Pyrene-d10	1.026	1.016	1.0	113	0.00
77	Pyrene	1.193	1.183	0.8	114	0.00
78	Butylbenzylphthalate	0.447	0.497	-11.2	129	0.00
79	3,3'-Dichlorobenzidine	0.386	0.407	-5.4	119	0.00
80	Benzo(a)anthracene	1.156	1.185	-2.5	120	0.00
81	Bis(2-ethylhexyl)phthalate	0.636	0.700	-10.1	130	0.00
82	Chrysene	1.041	1.076	-3.4	120	0.00
83 I	Perylene-d12	1.000	1.000	0.0	119	0.00
84	Di-n-octyl phthalate	1.199	1.254	-4.6	129	0.00
85	Benzo(b)fluoranthene	1.197	1.191	0.5	122	0.00
86	Benzo(k)fluoranthene	1.169	1.178	-0.8	118	0.01
87 S	Benzo(a)pyrene-d12	0.936	0.939	-0.3	121	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG080217\
Data File : BG028144.D
Acq On : 3 Aug 2017 16:48
Operator : SJ/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02096

Quant Time: Aug 03 20:03:56 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Aug 03 19:32:32 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.159	1.171	-1.0	122	0.00
89	Indeno(1,2,3-cd)pyrene	1.357	1.381	-1.8	126	0.00
90	Dibenzo(a,h)anthracene	1.137	1.153	-1.4	125	0.01
91	Benzo(g,h,i)perylene	1.116	1.116	0.0	122	0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0