


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG080217\
Data File : BG028101.D
Acq On : 2 Aug 2017 10:04
Operator : SJ/JU
Sample : DFTPP58
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP58

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
Title : SVOA CALIBRATION
Last Update : Thu Aug 03 20:26:11 2017



AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.5	31448	PASS
68	69	0.00	2	0.9	241	PASS
69	198	0.00	100	39.3	26602	PASS
70	69	0.00	2	0.5	133	PASS
127	198	10	80	44.7	30250	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	67674	PASS
199	198	5	9	6.8	4593	PASS
275	198	10	60	30.0	20288	PASS
365	198	1	100	4.2	2871	PASS
441	442	0.01	100	16.3	8950	PASS
442	198	50	100	81.0	54829	PASS
443	442	15	24	18.7	10279	PASS

m/z	Abundance
38.10	354.0
39.10	1956.0
39.90	233.0
41.20	154.0
44.10	691.0
49.10	300.0
50.10	7994.0
51.10	33648.0
52.10	2430.0
55.20	151.0
56.10	1349.0
57.10	2763.0
61.10	277.0
62.00	336.0
63.10	1521.0
65.10	821.0
68.00	725.0
69.00	27016.0
70.00	207.0
73.10	243.0
74.10	3148.0
75.00	4333.0
76.10	1885.0
77.10	31888.0
78.10	1807.0
79.00	2036.0
80.10	1727.0
81.00	2076.0
82.10	426.0
83.00	873.0
85.10	405.0
85.90	874.0
86.90	368.0
88.10	169.0
91.00	508.0
92.10	916.0
93.00	4047.0
94.10	218.0
98.00	2109.0
99.10	2488.0
100.10	201.0
101.00	1169.0
103.10	342.0
104.00	702.0
105.00	786.0
107.00	9862.0
108.00	1440.0
108.90	349.0
110.00	15320.0
111.00	2368.0
111.80	207.0
112.90	319.0
115.90	606.0
117.00	6232.0
118.00	697.0
122.00	561.0
123.00	828.0
124.00	474.0
125.00	304.0
127.00	29800.0
128.00	2674.0
129.00	12206.0
130.00	1004.0
131.10	382.0
133.00	169.0
133.90	333.0
135.00	1187.0
136.00	544.0
137.00	203.0
141.00	1596.0
142.00	569.0
143.00	243.0
146.10	553.0
146.90	634.0
147.90	2070.0
149.10	531.0
151.10	286.0
151.70	263.0
152.90	566.0
153.90	537.0
155.00	1169.0
156.00	1241.0
157.00	296.0
157.90	505.0
158.90	248.0
159.90	418.0
161.00	659.0
162.00	178.0
163.80	189.0
165.00	944.0
165.90	589.0

Instrument :
BNA_G
ClientSampleId :
DFTPP58

167.00	3870.0
167.90	1917.0
169.20	285.0
171.00	245.0
172.00	342.0
173.00	509.0
174.00	650.0
175.00	1250.0
175.90	453.0
177.00	600.0
177.90	309.0
179.00	2713.0
180.00	1521.0
181.00	1099.0
184.00	392.0
185.00	1532.0
186.00	10273.0
187.00	3037.0
188.00	513.0
189.00	758.0
191.00	366.0
192.00	1045.0
193.00	884.0
193.90	396.0
196.00	2119.0
198.00	62248.0
198.90	4412.0
199.90	320.0
201.30	200.0
203.00	462.0
204.00	2769.0
205.00	4624.0
206.00	19944.0
207.00	2571.0
208.00	638.0
208.90	351.0
210.00	283.0
211.00	745.0
212.00	193.0
215.00	153.0
216.00	479.0
217.00	4490.0
218.00	931.0
220.90	1710.0
221.80	792.0
222.30	407.0
223.00	1543.0
224.00	10454.0
224.90	2517.0
225.90	197.0
227.00	4818.0
227.90	626.0
229.00	855.0
231.00	296.0
235.00	494.0
237.00	386.0
237.90	163.0
238.90	170.0
239.80	239.0
241.00	207.0
242.10	565.0
243.00	569.0
244.00	7569.0
245.00	1304.0
246.00	1719.0
247.00	235.0
249.00	273.0
252.30	154.0
252.90	295.0
255.00	42584.0
256.00	5693.0
257.00	639.0
258.00	2346.0
259.00	284.0
264.90	947.0
265.90	283.0
271.10	185.0
273.00	1317.0
274.00	3187.0
275.00	17664.0
276.00	1993.0
277.00	1613.0
278.00	307.0
282.90	292.0
283.90	151.0
285.10	377.0
293.00	217.0
294.00	207.0
296.00	4757.0
297.00	1116.0
301.00	216.0
303.10	513.0
308.10	175.0
314.10	311.0
315.00	565.0

Instrument :
BNA_G
ClientSampleId :
DFTPP58

316.10	244.0
321.00	155.0
323.00	1250.0
324.10	469.0
327.00	390.0
328.00	370.0
334.00	962.0
335.10	179.0
345.90	244.0
352.00	618.0
353.00	306.0
353.90	326.0
364.90	1979.0
365.90	248.0
372.00	822.0
390.10	157.0
402.20	261.0
403.00	625.0
404.00	248.0
405.00	170.0
420.90	365.0
422.10	315.0
423.00	2227.0
424.00	363.0
441.00	6750.0
442.00	38880.0
443.00	7040.0
444.00	792.0
445.10	226.0

Instrument :
BNA_G
ClientSampleId :
DFTPP58

m/z	Abundance
38.10	245.0
39.10	2716.0
40.20	170.0
40.90	201.0
44.00	244.0
49.10	267.0
50.10	8685.0
51.10	36512.0
52.10	2039.0
56.00	1180.0
57.10	3001.0
58.10	231.0
61.00	725.0
62.10	643.0
63.10	1333.0
64.10	240.0
65.10	1007.0
69.00	31416.0
69.90	192.0
73.10	386.0
74.10	3432.0
75.10	5455.0
76.10	1980.0
77.10	35984.0
78.10	2054.0
79.00	2226.0
80.00	2074.0
81.00	2324.0
82.00	637.0
83.00	499.0
85.00	447.0
86.10	983.0
87.20	321.0
90.90	445.0
92.00	967.0
93.00	4007.0
94.10	253.0
96.10	239.0
98.10	2657.0
99.00	3168.0
99.90	282.0
101.10	841.0
103.00	492.0
104.00	673.0
105.00	818.0
107.00	12023.0
108.00	1569.0
110.00	17848.0
111.00	2685.0
112.10	483.0
113.10	181.0
116.10	656.0
117.00	7882.0
117.90	452.0
120.20	236.0
122.10	555.0
123.00	1078.0
124.00	508.0
125.10	420.0
127.00	35888.0
128.10	3141.0
129.00	17280.0
130.00	1192.0
131.00	227.0
133.80	234.0
135.00	1343.0
136.00	523.0
137.10	572.0
138.10	249.0
140.00	285.0
141.00	1917.0
142.00	608.0
143.00	415.0
144.00	254.0
146.00	359.0
147.00	917.0
148.00	1892.0
149.00	512.0
151.00	355.0
152.90	705.0
154.10	351.0
155.00	1041.0
156.00	1793.0
157.00	329.0
158.00	327.0
159.10	251.0
160.00	567.0
161.00	949.0
161.90	204.0
164.90	774.0
166.00	681.0

Instrument :
BNA_G
ClientSampleId :
DFTPP58

167.00	3942.0
168.00	1871.0
169.10	404.0
171.90	202.0
173.10	343.0
174.00	1191.0
174.90	1895.0
176.00	339.0
177.10	704.0
177.90	421.0
179.00	3396.0
180.00	2729.0
180.90	1024.0
182.10	182.0
183.90	170.0
185.00	1949.0
186.00	12238.0
187.00	4374.0
188.00	628.0
188.90	595.0
190.00	307.0
190.80	404.0
191.90	971.0
193.00	1586.0
193.90	402.0
196.00	2094.0
198.00	79784.0
199.00	5557.0
200.00	375.0
201.40	481.0
202.10	190.0
202.90	932.0
204.00	3720.0
205.00	6734.0
206.00	25216.0
207.00	2756.0
208.00	1558.0
209.10	303.0
210.10	336.0
211.10	955.0
213.00	239.0
214.90	483.0
216.00	405.0
217.00	6790.0
217.90	830.0
221.00	3238.0
221.90	1041.0
223.10	1249.0
224.00	13269.0
225.00	3316.0
225.90	207.0
227.00	6948.0
228.00	1250.0
229.00	1529.0
230.00	215.0
231.10	595.0
232.10	249.0
232.90	235.0
234.00	399.0
235.10	315.0
236.00	303.0
237.00	560.0
238.10	205.0
238.90	341.0
240.00	381.0
240.90	426.0
242.00	532.0
243.00	876.0
244.00	9622.0
245.00	1112.0
246.00	2876.0
246.80	217.0
248.00	223.0
249.00	259.0
252.20	176.0
252.80	494.0
255.00	58640.0
256.00	8582.0
257.00	526.0
257.90	3369.0
259.00	633.0
263.90	150.0
265.00	1138.0
265.90	174.0
269.20	183.0
269.90	259.0
271.00	168.0
271.80	278.0
273.00	1338.0
274.00	4157.0
275.00	23576.0
276.00	3622.0
277.00	2142.0
278.00	525.0
283.00	227.0

Instrument :
BNA_G
ClientSampleId :
DFTPP58

284.10	174.0
285.00	466.0
291.80	210.0
293.00	452.0
293.90	223.0
294.90	157.0
296.00	6373.0
297.00	1127.0
303.00	966.0
303.90	229.0
309.80	225.0
314.00	457.0
315.00	1023.0
316.00	550.0
320.90	181.0
323.00	1966.0
324.00	261.0
327.00	578.0
328.10	187.0
332.00	276.0
333.10	354.0
334.00	1359.0
335.00	351.0
340.90	387.0
346.00	306.0
352.00	779.0
352.90	546.0
354.00	482.0
355.00	153.0
365.00	3226.0
365.90	417.0
371.00	314.0
372.00	1048.0
373.10	243.0
374.10	176.0
382.90	311.0
391.00	244.0
402.00	409.0
403.00	608.0
403.90	255.0
420.90	570.0
422.00	514.0
423.00	3115.0
424.00	990.0
424.90	213.0
441.00	10163.0
442.00	64824.0
443.00	12631.0
444.00	973.0

Instrument :
BNA_G
ClientSampleId :
DFTPP58

m/z	Abundance
38.10	341.0
39.10	1415.0
44.10	719.0
50.10	5385.0
51.10	24184.0
52.10	1106.0
55.10	233.0
56.00	1006.0
57.00	1367.0
61.00	286.0
63.10	849.0
65.10	470.0
69.00	21376.0
74.10	2405.0
75.00	3686.0
76.10	1089.0
77.10	25896.0
78.10	1617.0
79.10	1755.0
80.00	1645.0
81.10	1810.0
82.00	470.0
83.00	255.0
84.00	181.0
84.90	265.0
86.00	545.0
86.90	271.0
87.50	198.0
91.00	489.0
92.10	400.0
93.00	2750.0
98.00	1684.0
99.00	1375.0
100.00	199.0
101.00	1087.0
103.10	255.0
104.00	650.0
105.10	671.0
106.20	399.0
107.00	7794.0
108.00	1222.0
109.10	685.0
110.00	12951.0
111.00	2444.0
112.00	205.0
112.70	164.0
116.10	488.0
117.00	6028.0
118.00	282.0
119.90	183.0
122.00	404.0
123.00	510.0
124.10	373.0
124.80	258.0
125.20	266.0
127.00	25064.0
128.00	1741.0
129.00	10263.0
130.00	996.0
131.00	199.0
133.90	438.0
135.00	1035.0
136.00	499.0
137.00	345.0
141.00	1394.0
141.90	417.0
142.50	229.0
143.10	400.0
145.90	267.0
147.10	1130.0
148.00	1565.0
149.10	345.0
151.10	170.0
152.10	203.0
153.10	563.0
153.90	171.0
155.00	711.0
156.00	1308.0
157.20	281.0
157.80	372.0
159.00	383.0
160.10	522.0
161.00	415.0
165.00	867.0
166.00	549.0
167.10	2737.0
168.00	1668.0
169.00	181.0
171.90	236.0
172.90	326.0
174.10	597.0

Instrument :
BNA_G
ClientSampleId :
DFTPP58

175.00	1542.0
176.10	239.0
177.00	570.0
178.00	234.0
178.90	2600.0
180.00	1711.0
180.90	592.0
184.10	332.0
185.00	1279.0
186.00	9526.0
187.00	3164.0
188.00	309.0
189.00	926.0
190.00	156.0
190.90	290.0
192.00	767.0
192.90	942.0
194.00	191.0
195.20	381.0
196.00	1467.0
198.00	60992.0
199.00	3811.0
199.90	311.0
201.10	227.0
201.50	166.0
201.70	162.0
202.90	608.0
204.00	3413.0
205.00	4820.0
206.00	18928.0
207.10	1973.0
208.00	561.0
209.20	166.0
209.90	153.0
210.50	305.0
211.00	529.0
212.00	151.0
214.80	186.0
215.90	448.0
217.00	5270.0
218.00	482.0
221.00	1633.0
221.80	655.0
222.90	1327.0
224.00	10877.0
225.00	2599.0
225.90	370.0
227.00	4592.0
227.90	656.0
229.00	1068.0
231.00	558.0
232.90	301.0
233.90	524.0
234.80	172.0
237.00	499.0
239.00	279.0
239.90	159.0
240.80	165.0
242.00	677.0
243.10	492.0
244.00	7386.0
244.90	847.0
246.00	2006.0
247.10	327.0
247.90	176.0
251.90	190.0
253.00	415.0
255.00	46080.0
256.00	6414.0
257.00	393.0
258.00	3000.0
258.90	372.0
265.00	904.0
265.80	370.0
271.00	160.0
271.80	151.0
273.00	1370.0
274.00	3077.0
275.00	19624.0
276.00	2501.0
277.00	1946.0
277.90	285.0
283.00	248.0
284.10	228.0
285.10	221.0
293.00	431.0
293.90	167.0
295.10	302.0
296.00	6035.0
296.90	898.0
302.20	187.0
303.00	702.0
304.00	217.0
314.00	345.0
315.00	726.0

Instrument :
BNA_G
ClientSampleId :
DFTPP58

318.00	168.0
321.00	365.0
323.00	1918.0
324.00	496.0
326.90	368.0
328.00	208.0
333.10	162.0
334.00	1548.0
334.90	288.0
341.10	342.0
346.00	389.0
346.90	208.0
352.00	640.0
353.00	802.0
354.00	350.0
365.00	3410.0
365.80	557.0
372.00	929.0
373.00	277.0
382.90	393.0
390.00	170.0
401.90	424.0
403.00	678.0
403.90	271.0
421.00	421.0
422.00	522.0
423.00	2901.0
424.00	559.0
424.80	208.0
441.00	9939.0
442.00	60784.0
443.00	11168.0
443.80	740.0

Instrument :
BNA_G
ClientSampleId :
DFTPP58

Scan 2467 (18.082 min): BG028101.D
DFTPP58

m/z	Abundance
39.90	189.0
44.10	913.0
206.90	156.0

Instrument :
BNA_G
ClientSampleId :
DFTPP58