

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080221\
 Data File : BG049602.D
 Acq On : 2 Aug 2021 12:58
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 02 14:49:52 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 21 16:57:05 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	94	0.00
2	1,4-Dioxane	40.000	37.314	6.7	90	-0.02
3	Pyridine	40.000	41.536	-3.8	91	-0.02
4	n-Nitrosodimethylamine	40.000	47.512	-18.8	107	-0.02
5 S	2-Fluorophenol	80.000	78.700	1.6	89	-0.02
6	Aniline	40.000	40.692	-1.7	95	-0.01
7 S	Phenol-d6	80.000	82.361	-3.0	94	-0.01
8	2-Chlorophenol	40.000	41.999	-5.0	95	0.00
9	Benzaldehyde	40.000	37.962	5.1	87	-0.01
10 C	Phenol	40.000	41.678	-4.2	96	0.00
11	bis(2-Chloroethyl)ether	40.000	42.245	-5.6	100	-0.01
12	1,3-Dichlorobenzene	40.000	39.937	0.2	93	-0.01
13 C	1,4-Dichlorobenzene	40.000	40.135	-0.3	93	-0.01
14	1,2-Dichlorobenzene	40.000	39.795	0.5	93	0.00
15	Benzyl Alcohol	40.000	42.932	-7.3	99	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.257	-0.6	95	-0.01
17	2-Methylphenol	40.000	41.839	-4.6	96	-0.01
18	Hexachloroethane	40.000	40.144	-0.4	95	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	43.919	-9.8	99	0.00
20	3+4-Methylphenols	40.000	43.395	-8.5	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	104	0.00
22	Acetophenone	40.000	39.364	1.6	102	0.00
23 S	Nitrobenzene-d5	80.000	78.701	1.6	102	-0.01
24	Nitrobenzene	40.000	40.674	-1.7	104	0.00
25	Isophorone	40.000	40.579	-1.4	105	0.00
26 C	2-Nitrophenol	40.000	40.841	-2.1	102	-0.01
27	2,4-Dimethylphenol	40.000	38.860	2.9	99	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.120	4.7	98	0.00
29 C	2,4-Dichlorophenol	40.000	41.374	-3.4	105	0.00
30	1,2,4-Trichlorobenzene	40.000	38.299	4.3	100	0.00
31	Naphthalene	40.000	38.578	3.6	99	0.00
32	Benzoic acid	40.000	34.112	14.7	92	0.01
33	4-Chloroaniline	40.000	41.639	-4.1	105	-0.01
34 C	Hexachlorobutadiene	40.000	39.245	1.9	102	-0.01
35	Caprolactam	40.000	41.305	-3.3	105	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.718	-1.8	105	-0.01
37	2-Methylnaphthalene	40.000	40.086	-0.2	103	0.00
38	1-Methylnaphthalene	40.000	40.191	-0.5	104	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	107	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.905	2.7	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.356	-3.4	114	0.00
42 S	2,4,6-Tribromophenol	80.000	88.756	-10.9	120	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.992	-2.5	112	0.00
44	2,4,5-Trichlorophenol	40.000	40.659	-1.6	109	0.00
45 S	2-Fluorobiphenyl	80.000	78.705	1.6	108	0.00
46	1,1'-Biphenyl	40.000	38.327	4.2	107	0.00
47	2-Chloronaphthalene	40.000	39.980	0.1	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080221\
 Data File : BG049602.D
 Acq On : 2 Aug 2021 12:58
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 02 14:49:52 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 21 16:57:05 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.437	-6.1	114	0.00
49	Acenaphthylene	40.000	39.378	1.6	107	0.00
50	Dimethylphthalate	40.000	41.039	-2.6	113	0.00
51	2,6-Dinitrotoluene	40.000	41.525	-3.8	116	0.00
52 C	Acenaphthene	40.000	38.439	3.9	105	0.00
53	3-Nitroaniline	40.000	39.173	2.1	104	0.00
54 P	2,4-Dinitrophenol	40.000	43.003	-7.5	128	0.00
55	Dibenzofuran	40.000	39.191	2.0	109	0.00
56 P	4-Nitrophenol	40.000	37.951	5.1	108	0.00
57	2,4-Dinitrotoluene	40.000	43.411	-8.5	116	0.00
58	Fluorene	40.000	39.349	1.6	109	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.856	-4.6	112	0.00
60	Diethylphthalate	40.000	40.646	-1.6	111	0.00
61	4-Chlorophenyl-phenylether	40.000	41.128	-2.8	111	0.00
62	4-Nitroaniline	40.000	38.713	3.2	104	0.00
63	Azobenzene	40.000	39.909	0.2	110	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	113	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.017	-15.0	128	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.641	5.9	110	0.00
67	4-Bromophenyl-phenylether	40.000	39.821	0.4	113	0.00
68	Hexachlorobenzene	40.000	39.049	2.4	115	0.00
69	Atrazine	40.000	38.698	3.3	114	0.00
70 C	Pentachlorophenol	40.000	43.686	-9.2	132	0.00
71	Phenanthrene	40.000	37.694	5.8	109	0.00
72	Anthracene	40.000	37.613	6.0	108	0.00
73	Carbazole	40.000	37.098	7.3	106	0.00
74	Di-n-butylphthalate	40.000	37.649	5.9	108	0.00
75 C	Fluoranthene	40.000	37.279	6.8	107	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	103	0.00
77	Benzydine	40.000	38.417	4.0	92	0.00
78	Pyrene	40.000	40.843	-2.1	108	0.00
79 S	Terphenyl-d14	80.000	84.133	-5.2	109	0.00
80	Butylbenzylphthalate	40.000	42.113	-5.3	105	0.00
81	Benzo(a)anthracene	40.000	39.998	0.0	104	0.00
82	3,3'-Dichlorobenzidine	40.000	40.356	-0.9	105	0.00
83	Chrysene	40.000	39.796	0.5	105	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.015	-2.5	106	0.00
85 c	Di-n-octyl phthalate	40.000	40.342	-0.9	104	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.516	-1.3	106	0.00
88	Benzo(b)fluoranthene	40.000	40.561	-1.4	104	0.00
89	Benzo(k)fluoranthene	40.000	39.709	0.7	104	0.00
90 C	Benzo(a)pyrene	40.000	40.150	-0.4	104	-0.01
91	Dibenzo(a,h)anthracene	40.000	41.313	-3.3	108	0.00
92	Benzo(g,h,i)perylene	40.000	40.113	-0.3	105	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080221\
Data File : BG049602.D
Acq On : 2 Aug 2021 12:58
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Aug 02 14:49:52 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG072121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 21 16:57:05 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0